

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: <i>Overview Hospital</i> <i>Noblesville, IN</i> REPORT NUMBER(S) <i>2008-001</i>	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrentonville Road Suite 210 Lisle, Illinois 60532-4351
--	--

3. DOCKET NUMBER(S) <i>030-10729</i>	4. LICENSEE NUMBER(S) <i>13-16286-01</i>	5. DATE(S) OF INSPECTION <i>July 24, 2008</i>
--	--	---

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
 (Violations and Corrective Actions)



Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Deborah A. Piskura	<i>Deborah A. Piskura</i>	<i>7/24/08</i>

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE Riverview Hospital REPORT NUMBER(S) 2008-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-10729	4. LICENSE NUMBER(S) 13-16286-01	5. DATE(S) OF INSPECTION July 24, 2008	
6. INSPECTION PROCEDURES USED 87130 & 87131		7. INSPECTION FOCUS AREAS 03.01 – 03.08	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02120	2. PRIORITY 3	3. LICENSEE CONTACT Phillip R. Partlan, RT	4. TELEPHONE NUMBER 317-773-0760
-----------------------------	------------------	---	-------------------------------------

<input checked="checked" type="checkbox"/> Main Office Inspection	Next Inspection Date: July 2011
<input type="checkbox"/> Field Office	
<input type="checkbox"/> Temporary Job Site Inspection	

PROGRAM SCOPE

This licensee was a 150-bed hospital, authorized to use licensed material permitted by Sections 35.100, 35.200, and 35.300. The nuclear medicine department was staffed with two full-time technologists and two agency technologists who performed approximately 80-100+ diagnostic nuclear medicine procedures per month. The licensee received unit doses from a licensed radiopharmacy. The hospital performed a full spectrum of nuclear diagnostic imaging studies. Typically, in a year the hospital administered 8-10 iodine-131 thyroid carcinoma therapies, 20-25 hyperthyroidism treatments, and 3-5 whole body CA follow up studies. The hospital obtained its I-131 in capsule form only. The licensee retained the services of a consulting physicist to audit the radiation safety program on a quarterly basis (last audit 6/18/2008 with no findings or violations identified).

This inspection consisted of interviews with licensee personnel, a review of selected records, tours of the nuclear medicine department, and independent measurements. The inspector observed licensee nuclear medicine personnel prepare, assay and administer several unit doses for various imaging procedures. The inspection included observations of dose calibrator QA checks, package receipts and surveys, and area surveys.