

Associates in Cardiovascular Disease
211 Mountain Avenue
Springfield, NJ 07081
973-467-0006

Licensing Assistant Section
Nuclear Materials Safety Branch
U.S. Nuclear Regulatory Commission, Region I
475 Allendale Road
King of Prussia, PA 19406-1415

Br. 2

Re: Notification/ Decommissioning NRC License # 29-28773-12
Control #142199

01 RLT 03032847

June 30, 2008


Dear Sir/ Madam:

We have completed construction and have relocated our nuclear medicine department to another area of the building, but still located at the same physical address.

Enclosed is a decommissioning report for the old department, the new department is fully operational. All license conditions were met during this transition. The results of the survey are attached no removable contamination was detected.

Please contact Elaine Rovazzi, M.S., DABR, Physicist @ 973-322-5118 if you need additional information. Thank you in advance for your assistance.

Sincerely,



Michael Lux, M.D./ RSO
Management, Associates in Cardiovascular Disease

enc. Decommissioning report
Survey and Wipe test results

2008 AUG -6 AM 10:13

RECEIVED
INSTRUMENT

(Ref. 142199)

142692
NRCSS/RGNI MATER. ALS-002

Michael Lux, M.D.
Associates in Cardiovascular Disease
211 Mountain Avenue
Springfield, NJ 07081

NJBRP License #20402
NRC License #29-28773-01

DECOMMISSIONING SURVEY NUCLEAR MEDICINE
June 30, 2008

Area Survey

Locations surveyed: 1. Hot Lab/ Scanning Area/Stress lab

Instrument: Ludlum Model 14-C Serial # 958601
Pancake Probe Model 44-9 /093288
Calibration Date - 3/18/08 (Check Source: 0.6 mR/hr)
Check Source Reading 0.6 mR/hr

Background: Less than 0.02 mR/hr

Actual Readings: All areas were less than 0.02 mR/hr

Area Wipe Test

Locations surveyed: 1. Hot Lab/ Scanning Room/Stress Lab

Instrument: Ludlum 2200 NaI Well Counter Model
Serial # 2621/48277
Efficiency for ¹³⁷Cs: 27.5 %
MDA for 1 min: 0.000051 uCi

Background: 865 cpm

Results: 1. Hot Lab - 895 cpm (90 dpm)
2. Scanning Room - 883 cpm (54 dpm)
3. Stress Room - 829 cpm (0 dpm)
ALL AREAS WITHIN STATISTICAL BACKGROUND

Decommissioning Results: **No radiation detected in any area**

Clairn Rovazzi, M.S.

Elaine Rovazzi, M.S., D.A.B.R.
Board Certified:
American Board of Radiology

June 30, 2008
Date

This is to acknowledge the receipt of your letter/application dated 8/6/08 ^{received}, and to inform you that the initial processing which includes an administrative review has been performed.

Amendment (19-28773-01) There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 142692.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.