Radiological Decontamination and Decommissioning (D&D)

Final Status Survey Report

JUNE 16, 2000



NMRC Diving Buildings Bethesda, Maryland 20889-5607

Prepared by Radiation Safety Office Naval Medical Research Center Bethesda, Maryland

TABLE OF CONTENTS

I.	Dedication and Acknowledgements 5			
II.	Overview			
III.	Abbreviations and Definitions			
IV.	Executive Summary			
V.	Reason for Decommissioning			
VI.	Guidance for Decommissioning Efforts			
VII.	Conceptual Model and Site Information			
VIII.	Decommissioning Activities			

	K.	<pre>Sampling and Preparation for Laboratory Measurements</pre>
IX.	Deconta	amination Procedures
х.	Final S A. B. C. D. E. F.	Status Survey<
XI.	Referen	nces
XII.	Append	ices
	Α.	Conceptual Model and Floor Diagrams
	в.	Photographs

с.	<pre>Standard Operating Procedures and Quality Assurance Program</pre>
D.	<pre>Quality Assurance Information and Data85 D-1. Training Information and Certificates .86 D-2. Service and Calibration Certificates for Instruments, Equipment and Check Sources</pre>
E.	<pre>Survey Unit Release Reports (Narratives, Data, Floor Plans, and Grid Maps)</pre>

I. DEDICATION AND ACKNOWLEDGEMENTS

DEDICATION

This work is dedicated in memory of my father, Mr. John S. Gaiter, Sr. (May 1, 1932 - August 19, 1999) and my youngest sister, Bergenia R. Gaiter (June 13, 1967 - October 11, 1999).

ACKNOWLEDGMENTS

Without the noteworthy contributions, support and assistance of numerous persons, this work would not have been possible.

NMRC Radiological Decontamination and Decommissioning (D&D) Team

- Schleurious L. Gaiter; LCDR, MSC, USN: Radiation Safety Officer, Decommissioning Officer, and Quality Control Officer
- Saifeddin Ahmad; Assistant Radiation Safety Officer, Team Leader, Lead Health Physics Technician, Chief of Office and Laboratory Operations
- Willie Deadwyler; Team Leader, Health Physics Technician, Chief of Field Operations
- Carolyn Hall; Technical Writer, Chief Editor, Publications Officer
- Glen Garth; Health Physics Assistant, Data Processing Specialist
- Rodney McGhee; Health Physics Assistant, Data Processing Specialist
- HM2 Michelle McGhee; USN: Safety Assistant

Command and Navy Personnel

- Dr. Charles Odeyale; Ph.D., NMRC Environmental and Chemical Hygiene Officer
- HM1 James Bruce; USN, Hazardous Materials Technician
- Daniel Fischer; NMRC Safety Assistant
- Stephen Krauz; LT, MSC, USN, Supply Officer
- Veronica Serrano; NMRC Purchase and Contract Specialist
- Raymond Albright; NMRC Purchase and Contract Specialist
- Donald Chandler; NMRC BRAC Officer
- Sandra Hawkins; NMRC Comptroller
- Regina Hunt; NMRC Historian
- Michael Bush; LT, MSC, USN, Navy Environmental Health Center
- Paul Tveten; Navy Environmental Health Center
- Sandy J. Wolff; LCDR, MSC, USN, Navy Environmental Health Center

II. OVERVIEW

This document, the Naval Medical Research Center (NMRC) Diving Buildings Radiological Decontamination and Decommissioning (D&D) Final Status Survey Report, was prepared in support of NMRC's decommissioning efforts. The NMRC Diving Buildings consist of buildings 28, 53, 59, 69 and 79. This report includes all survey results, reference background readings, check-source evaluations, calculations, procedures, problems/deviations from plans, recommendations for survey-site status, and other pertinent information.

This report describes the actions and levels of effort that NMRC undertook to demonstrate to regulatory authorities that the vacated NMRC Diving Buildings meet the release criteria for unrestricted future use. To achieve this objective, guidance provided in the Multi-Agency Survey and Site Investigation Manual (MARSSIM, ref 11.1) and other regulatory documents were followed and parameters chosen to meet or exceed the release criteria.

The NMRC Historical Site Assessment (HSA, ref 11.2) served as the primary document for information regarding potential contaminants, potential contaminated areas, and potential contaminated media.

The NMRC Diving Buildings are located at 8901 Wisconsin Avenue, Bethesda, Maryland 20889-5607 on the NNMC Bethesda campus. Table A provides a list of tables in Sections I through XI.

Table 1 provides a list of radionuclides that were used in the NMRC Diving Buildings. This list includes isotopes that are not of concern for this decommissioning work because of the elapsed time since their last use in the buildings.

Table 2 provides a summary of the buildings that comprise the NMRC Diving Buildings. Medical studies and biomedical research were conducted which involved the use of chemicals and other hazardous materials, biological materials, animals and animal products, and radiological materials. All the surrounding properties are located upon the National Naval Medical Center (NNMC) campus (ref 11.3, ESA 1997).

Table A.	List of Tables in Sections I through XI	
Table Number	Title	Page Number
1	Radioisotopes Used in NMRC Diving Buildings	9
2	NMRC Diving Buildings Information	10
3	NMRC Diving Buildings Impacted Class 3 Areas and Non-impacted Areas	26
4	NMRC Diving Buildings Radiological D&D Conceptual Model Information	29
5	Summary of 10 CFR Part 20 Subpart E	33
6	Acceptable License Termination Screening Values of Common Radionuclides for Building Surface Contamination (ref 11.13)	35
7	Suggested Areas for Survey Units (MARSSIM, Roadmap Table 1)	37
8	Typical Measurement Sensitivities for Laboratory Radiometric Procedures (MARSSIM, Chapter 7)	40
9	Interpretation of Sample Measurements When No Reference Area is Used (MARSSIM, Table 2.5)	46
10	Determination of the Relative Shift, Δ/σ , Using Swipe Data from a Contaminated Area (MARSSIM, Section 5.5.2.3)	48
11	Recommended Survey Coverage for Structures and Land Areas (MARSSIM, Table 5.9)	49
12	Data Conversion for Direct Readings of Readings of Surface Activity Using a Gas Proportional Detector (MARSSIM, Section 6.6.1)	52

Table 1. Radioisotopes Used At NMRC Diving Buildings						
SYMBOL	RADIOISOTOPE	YEAR OF LAST USE	HALF LIFE	MAJOR RADIATIONS	OF DECOM- MISSIONING CONCERN DURING THESE D&D EFFORTS?	
Ba133	Barium-133 (rooms 112 and 122)	1992	7.2 Y	Gamma	Ŷ	
C14	Carbon-14	1999	5730 Y	Beta	Y	
Co57	Cobalt-57	1992	270.9 D	Gamma	Y	
Cs137	Cesium-137 (rooms 112 and 122)	1992	30 Y	Beta, Gamma	Ŷ	
НЗ	Hydrogen-3	1999	12.3 Y	Beta	Y	
S35	Sulfur-35	1998	87.4 D	Beta	Ŷ	
Ar41	Argon-41	1992	1.83 H	Beta, Gamma	N	
Ca45	Calcium-45	1997	165 D	Beta	N	
Ce141	Cerium-141	1993	32.5 D	Beta, Gamma	N	
Cr51	Chromium-51	1992	27.7 D	Gamma	N	
I125	Iodine-125	1994	60.2 D	Gamma	N	
I131	Iodine-131	1991	8.1 D	Beta, Gamma	N	
In111	Indium-111	1990	2.8 D	Beta, Gamma	N	
Kr79	Krypton-79	1992	34.9 D	Gamma	N	
Kr85m	Krypton-85m	1992	4.4 H	Beta, Gamma	N	
Nb95	Niobium-95	1993	35 D	Beta, Gamma	N	
P32	Phosphorus-32	1999	14.3 D	Beta	N	
P33	Phosphorus-33	1998	24.4 D	Beta	N	
Ru103	Ruthenium-103	1993	39.5 D	Gamma	N	
Sc46	Scandium-46	1986	83.8 D	Beta, Gamma	N	
Sn113	Tin-113	1990	115 D	Gamma	N	
Tc99m	Technetium-99m	1992	6.05 н	Gamma	N	
Xe127	Xenon-127	1992	36.4 D	Gamma	N	
Xe133	Xenon-133	1992	5.3 D	Beta, Gamma	N	

NMRC	Di	ving	Bu	ildings	Radiolo	gical	(D&D)
Fin	al	Stat	us	Survey	Report,	June	2000	

Table 2. NMRC Diving Buildings Information						
Building Number	Building Uses	Floor Area (square	Occupancy or Years of RAM Use		Isotopes of Concern	
		feet)	Start	Stop		
28	Offices, labs, storerooms, machine rooms	5,856	1952	1999	None	
53	Diving facility, offices, labs, store rooms	22,136	1976	1999	H3, C14, Co57, Ba133, Cs137	
59	Diving facility	2,695	1989	1999	None	
69	Diving facility	805	1992	1999	None	
79	Equipment facility	1,195	1992	1999	None	
Total		32,687				

NMRC's history of military medical research during 1976 to 1999 at the Diving Buildings site included the use of unsealed radioactive materials. The use of unsealed radioactive materials in laboratories, rooms, and areas dictated that some level of decommissioning was required for each area of radioactive material use to ensure and document compliance with the regulatory release criteria for unrestricted use.

The radioactive contaminants of concern were Hydrogen-3, Carbon-14, Cobalt-57, Barium-133, and Cesium-137.

The examination for potential radioactive contaminants included surveys, direct measurements, sampling and analysis, and scanning with appropriate instruments and equipment. The examinations included:

- (1) all accessible floor space,
- (2) the lower half of all walls extending up approximately two meters from the floor, and
- (3) other selected locations.

The survey units were classified as either Impacted Class 3 or Non-impacted. There were no survey sites classified as Impacted Class 1 or Class 2, which are the two remaining classifications requiring the most extensive decommissioning effort. Selected common areas, restrooms, and non-impacted areas were evaluated for potential radioactive contamination. After equipment, materials and wastes were removed from radioactive-materials-use areas, residual contamination was confined to small areas within a small percentage of the survey units.

The 5-month decommissioning period began in February 2000 and was concluded in June 2000. The cost of performing the necessary decommissioning efforts included instrument and equipment costs, materials, supplies and contracted labor.

III. Abbreviations and Definitions

A. ACRONYMS AND ABBREVIATIONS

μCi	Micro-Curie
µR/hr	Micro-Roentgens per hour (exposure rate)
Annex	NMRC Rockville Annex Laboratories, Washington
	Avenue, Rockville, MD
BRAC	Base Realignment and Closure
BUMED	Bureau of Medicine and Surgery
CFR	Code of Federal Regulations
cpm	Counts per minute
D&D	Decontamination and decommissioning
DOD	Department of Defense
DOT	Department of Transportation
dpm	Disintegrations per minute
Gy	Gray, SI unit of absorbed dose, 1 Gy = 100
	rads
Irradiator	Sealed-source, Cesium-137 gamma irradiator
MDA	Minimum detectable activity
NEHC	Navy Environmental Health Center
NMRC	Naval Medical Research Center, formerly NMRI
NMRI	Naval Medical Research Institute
NNMC	National Naval Medical Center
NRC	U. S. Nuclear Regulatory Commission
NRMP	NMRC's Navy Radioactive Materials Permit
NUREG	Nuclear Regulatory Guide
R, R/hr	Roentgen (exposure), Roentgens per hour
Rad, Rad/hr	Rad (dose), Rads per hour (dose rate)
RAM	Radioactive materials
Rem, Rem/hr	Rem (dose equivalent), Rem per hour
SOP	Standard operating procedure
Sv	Sievert, SI unit of dose equivalent, 1 Sv =
	100 rems
US	United States
USN	United States Navy
WSSC	Washington Suburban Sanitary Commission

B. KEY TERMS AND DEFINITIONS (reference 11.1, MARSSIM)

a, Alpha	The specified maximum probability of Type
	I error; i.e., the maximum probability of
	rejecting the null hypothesis when it is
	true. Alpha is referred to as the size of
	the test. Alpha reflects the amount of
(evidence the decision maker would like to
	see before abandoning the null hypothesis.
Area	A term referring to any portion of a site.
	up to and including the entire site.
Background	Radiation from cosmic sources, naturally
radiation	occurring radioactive material, radon, and
	global fallout as it exists in the
	environment from testing of nuclear
	explosive devices or from puckear
	accidents Background radiation does not
	include radiation from source hyproduct
1	or special nuclear materials regulated by
	the fodoral or state agonau
Pagguaral (Pg)	The unit of redicectivity equivalent to
pecquerer (pq)	and suclear transformation non second
	The probability of The Transformation per second.
β, Beta	The probability of a type II error; I.e.,
	the probability of accepting the null
	hypothesis when it is false. The
	compliment of beta $(1-\beta)$ is referred to as
	the power of test.
Beta emitter	A radioactive material emitting beta
)	particles. Beta particles are electrons
	emitted from the nucleus during
	radioactive decay.
Byproduct	Licensed or radioactive material regulated
material	by the NRC.
Class 1 areas	Impacted areas with the highest potential
J	for contamination or insufficient evidence
	to support reclassification as Class 2 or
	3.
Class 2 areas	Impacted areas with low potential for
	delivering a dose above the release
	criterion and little or no potential for
	small areas of elevated activity.
	-

Class 3 areas	Impacted areas with little or no potential for delivering a dose above the release criterion and little or no potential for small areas of elevated activity.
Classification	The act or result of separating areas or survey units into one of the three designated classes: Class 1 area, Class 2 area, or Class 3 area.
Cleanup standard	A numerical limit set by a regulatory agency as a requirement for releasing a site after cleanup.
Contamination	The presence of residual radioactivity in excess of levels which are acceptable for release of a site or facility for unrestricted use.
Curie	A unit of radioactivity equal to 37 billion becquerels.
Derived Concentration Guideline Level (DCGL)	A derived, radionuclide-specific activity concentration within a survey unit corresponding to the release criterion. The DCGL is based on the spatial (uniform) distribution of the contaminant and hence is derived differently for the non- parametric statistical test (DCGL _W) and the elevated measurement comparison (DCGL _{EMC}). DCGLs are derived from activity/dose relationships through various exposure pathway scenarios.
Decommission	To remove (as a facility) safely from service and reduce residual radioactivity to a level that permits release of the property for unrestricted use and termination of the Nuclear Regulatory Commission license or Navy radioactive material permit.
Decommissioning	The process of removing a facility or site from operation, followed by decontamination, and license termination (or termination of authorization for operation) if appropriate.
Decontamination	The removal of radiological contaminants from persons, objects, or areas to within regulatory levels.

Direct	Radioactivity measurement obtained by
measurement	placing the detector near the surface or
	media being surveyed. An indication of
	the resulting radioactivity level is read
	out directly.
Final status	Measurements and sampling to describe the
survey	radiological conditions of a site,
-	following completion of decontamination
	activities (if warranted) in preparation
	for release.
Gamma (v)	Penetrating high-energy, short-wavelength
radiation	electromagnetic radiation (similar to X-
	rays) emitted during radioactive decay.
	Gamma rays are very penetrating and
	require dense materials (such as lead or
	steel) for shielding
Impacted areas	Any area that is not classified as non-
impaceed areas	impacted Areas with a possibility of
	containing residual radioactivity in
	excess of natural background or fallout
	levels
Investigation	A derived media-specific, radionuclide-
lovel	specific concentration or activity level
10001	of radioactivity that is based on the
	regulatory release criteria and triggers a
	response that further investigation or
	action is necessary if exceeded
MADCCTM	Multi-Daopau Padiation Survey and Site
MARSSIN	Trucchigation Manual (DOE DOD ED) NPC);
	nvestigation Manual (bob, bob, EfA, NRC),
	provides for conducting decommissioning
	activities to satisfy regulatory release
Non-impacted	Areas where there is no reasonable
Non-impacted	Areas where there is no reasonable
areas	possibility (extremely low probability) of
New newspatia	Testqual contamination.
Non-parametric	A test based on relatively lew assumptions
test	about the underlying probability
	aistributions of the measurements. As a
	consequence, non-parametric tests are
	generally valid for a fairly broad class
	of distributions. The Wilcoxon Rank Sum
	test and the Sign test are examples of
	non-parametric tests,

Outlier	Measurements that are unusually large or small relative to the rest and therefore are suspect of misrepresenting the population from which they were collected.
Radioactive material (RAM)	Solid, liquid or gaseous material that contains radionuclides regulated by the NRC.
Radioactive waste	Solid or liquid trash or excess material that contains radionuclides regulated by the NRC.
Release criterion	A regulatory limit expressed in terms of dose or risk; compliance demonstration is simply a decision as to whether or not a survey unit meets the release criterion.
Residual radioactivity	Radioactivity in structures, materials, soil, groundwater, and other media at a site resulting from activities under the cognizant organization's control.
Scanning	An evaluation technique performed by moving a detection device over a surface at a specified speed and distance above the surface to detect radiation.
Sign test	A non-parametric statistical test used to demonstrate compliance with the release criterion when the radionuclide of interest is not present in background and the distribution of the data is not symmetric.
Site	Any installation, facility, or discrete, physically separate parcel of land, or any building or structure or portion thereof, that is being considered for survey and investigation.
Survey	A systematic evaluation and documentation of radiological measurements with a correctly calibrated instrument or instruments that meet the sensitivity required by the objective of the evaluation.

Survey unit	A geographical area consisting of a room, hallway, structures or land areas of specified size and shape for which a separate decision will be made whether the unit attains the cleanup standard. Survey units are established to facilitate the	
	analysis of the survey data.	
Wilcoxon test	A non-parametric statistical test used to demonstrate compliance with the release criterion when the radionuclide of interest is present in background and the distribution of the data is not symmetric.	

IV. EXECUTIVE SUMMARY

A. PURPOSE

A final status decommissioning survey was conducted to determine the presence and extent of any radiological health hazards in the Diving Buildings used by NMRC at the National Naval Medical Center (NNMC) Bethesda campus. Appropriate actions were taken to verify that any residual radioactivity in the vacated NMRC buildings was in compliance with regulatory guidelines for the decontamination of facilities prior to release for unrestricted use.

B. PLAN

- Develop and implement an approved decontamination and decommissioning plan.
- Stop all work involving the use of radioactive materials.
- Dispose, remove or transfer all licensed, radioactive materials and equipment to authorized recipients.
- Minimize hazardous and low-level radioactive waste generated.
- Decontaminate known areas of contamination to background levels.
- Meet regulatory radiological, environmental, and safety regulations.
- Control costs and complete closure within the allocated timeframe.
- Prepare final status survey reports.

A radiological survey plan was developed to survey over 29,000 square feet (of floor and wall surface area) with portable instrumentation and to take and analyze over 600 swipe samples and more than 100 direct readings throughout the NMRC Diving Buildings. The work history and probability of residual radioactive material were used to identify areas selected for surveying and sampling. The closure operations used basic engineering principles and common industrial practices.

Those few areas where low-level radioactivity in excess of permissible limits was detected were remediated and released for unrestricted use. After completing final status survey efforts for contaminated floor and workbench areas, it was determined that no further decontamination efforts were necessary and that the residual-radioactivity level was well below the NRC screening values for residual contamination.

C. CONCLUSIONS

The plan described in item B above has been fully implemented, completed, and documented.

The decontamination and decommissioning efforts involved the satisfactory performance of final status surveys for 100% (12 of 12) of the Impacted Class 3 units and 44% (34 of 78) of the Non-impacted units.

Operations involving the use of licensed, radioactive materials at the NMRC Diving Buildings have had no adverse effect on the leased commercial building. All radioactive materials as a result of NMRC operations have been removed from the building. Regulatory agency criteria for release of the building for unrestricted use have been fulfilled.

V. REASON FOR DECOMMISSIONING

Base Realignment and Closure (BRAC) legislation passed and signed into law in 1995 mandated that the Naval Medical Research Center (NMRC) relocate some of its programs and research efforts and cease operations at the NNMC campus in Bethesda, Maryland and at the NMRC Rockville Annex in 1999.

NMRC, the Navy's largest medical research facility, opened its doors in Bethesda on October 27, 1942 to conduct research, development, tests, and evaluations to enhance the health, safety, and readiness of Navy and Marine Corps personnel in the effective performance of peacetime and contingency missions. NMRC had specific environmental obligations to fulfill before returning the NMRC Diving Buildings to the NNMC. These obligations involved the completion of biological, chemical, environmental and radiological decommissioning efforts.

VI. GUIDANCE FOR DECOMMISSIONING EFFORTS

Regulatory guidance for conducting this radiological decommissioning activity was obtained from various sources. The Environmental Protection Agency (EPA), the Nuclear Regulatory Commission (NRC), and the Department of Energy (DOE) are responsible for the release of sites following cleanup. These responsibilities apply to facilities under the control of Federal agencies, such as the DOE and the Department of Defense (DOD), and to sites licensed by the NRC and its Agreement States. Some states have responsibilities for similar sites under their control.

The NRC requires the radiological decommissioning of sites, buildings and outdoor areas where licensed activities have ceased radiological operations, even while licensed activities continue to be conducted at other site locations. Use of licensed materials at the NMRC Diving Buildings at Bethesda, Maryland ceased in December 1999. These sites required decommissioning and return to the NNMC for unrestricted use.

The detailed guidance provided in the Multi-Agency Survey and Site Investigation Manual (MARSSIM, ref 11.1) was used for planning, implementing, and evaluating environmental and facility radiological surveys that NMRC conducted to demonstrate compliance with the current dose-based regulation (10CFR20.1402, ref 11.4). Other regulatory documents (referenced in Section XI.) were used to provide estimates for derived concentration guideline values, dose modeling, and other necessary information and techniques.

The MARSSIM provides a nationally consistent consensus approach to conducting radiation surveys and investigations at potentially contaminated sites. The decommissioning that follows remediation will normally require a demonstration to the responsible federal or state agency that the cleanup effort was successful and that the release criterion was met. In MARSSIM, this demonstration is given the name "final status survey." The MARSSIM assists site personnel and others in performing or assessing such a demonstration.

Unique site-specific cases did not arise that required a modified approach beyond what is presently described in MARSSIM.

VII. CONCEPTUAL MODEL AND SITE INFORMATION

A. POTENTIAL RADIOACTIVE CONTAMINANTS

The potential radioactive contaminants included Hydrogen-3 (tritium), Carbon-14, Cobalt-57, Barium-133 and Cesium-137. The Barium-133 and Cesium-137 sources were small sealed sources of millicurie activity that were used only in rooms 112 and 122 of Building 53. Table 1 summarizes the radioisotopes that were used in the Diving Buildings, more specifically, Building 53. Radioactive materials were obtained and used in the form of gaseous and liquid solutions. There was no likely potential for contamination as a result of airborne release of radioactive materials due to the nature of operations and the elapsed period since the gaseous materials were used. Therefore, ceiling surfaces were not evaluated or surveyed for residual contamination. Low-level radioactive liquid and solid wastes were generated and properly disposed. Liquid (aqueous) and solid wastes were collected by the user and turned into the NMRC Radiation Safety Office for proper disposal. Radioactive (solid and liquid) wastes were removed from NMRC Diving Buildings for disposal.

NMRC researchers and technicians who used radioactive materials were required to maintain separate waste containers for each isotope used. Further segregation occurred for liquids, solids, carcasses, scintillation vials, and sharps. All radioactive material (RAM) wastes were turned in to the Radiation Safety Office for tracking, storage and subsequent disposal and/or transfer. Logbooks were maintained. Burial of waste was prohibited. Solid waste disposal via decay-instorage occurred for shorter half-life RAM (< 65 days). Long half-life RAM wastes were transferred to the National Naval Medical Center (NNMC) at Bethesda, Maryland, for transfer to a commercial contractor.

The review of NMRC records included summaries of NMRC's annual receipts (inventories) of radioactive materials, annual solid radioactive waste disposal logbook entries, and annual sanitary sewer disposal of

liquid radioactive materials logbook entries. Table 2 provides a list of radioisotopes that were used at NMRC Diving Buildings but are not of D&D contamination concern because of the elapsed time since their last use.

B. POTENTIAL CONTAMINATED AREAS

All available NMRC daily, weekly, monthly, and quarterly wipe test and survey monitoring results (1980-1999) were reviewed for evidence of residual contamination in the NMRC Diving Buildings. Incident and spills records were also reviewed. When contamination was detected during the course of decommissioning the NMRC buildings, characterization surveys, remedial actions, and remediation action surveys were completed and documented in the final survey summaries.

Description of spill incident, leakage of Xenon-127 gas in hood, 16 August to 25 October 1985: The laboratory in Room 112, Building 53, NMRC, had recently obtained a radiation dose calibrator to determine the actual doses of radioactive Xenon gas utilized during experiments. On 16 August 1985, a vial of Xenon-127 gas was obtained from Oak Ridge. The gas was presumed to contain 208 mCi of gas based on information obtained from Oak Ridge.

The gas was transferred without incident into an RGD 700 container that had previously held a vacuum for periods up to a week of leak testing. Over the following nine weeks, the gas was withdrawn into a sample syringe for equipment calibration and then returned into the container; no samples were used in actual experiments. No evidence of leakage was noted on the Xenon-133 monitor, which was continuously sampling the air in the hood of Room 112 during this time.

The monitor was periodically checked and found to be very sensitive to minute quantities of Xenon-127 and Xenon-133. On 25 October 1985, the amount of Xenon-

127 gas in the RGD container was estimated to be 56 mCi. Two samples of gas were assayed from the container and had an average activity of 3.55 mCi/5 cc in the radiation dose calibrator. Back calculation revealed that there was 0.2 mCi of gas in the container on 16 August 1985; this implied a loss between 100 mCi (if lost on 16 August) and 28 mCi (if lost on 25 October) of gas from the container. Assuming a worst case of a gradual 100 mCi loss over 9.4 weeks, this would have amounted to approximately 1% MPC (maximum permissible concentration) for a restricted area and 12% MPC for unrestricted areas.

Further testing revealed a light leak at the syringe/container junction during gas transfer. This leak was too small to account for the amount of gas lost so an episode in which the RGD 700 container was left with the valves cracked open was presumed to explain the amount of gas lost. Procedures were revised to prevent recurrence of this incident. No decontamination efforts were required. No residual contamination exists in 2000 as a result of this 1985 Xenon-127 leakage.

Table 3 provides a list of all NMRC Diving-Building survey sites. These sites included offices, lunchrooms, lavatories, storage areas, common areas, laboratories, equipment rooms, and radiological waste storage rooms.

C. IMPACTED AREAS - KNOWN AND POTENTIAL

Impacted areas have some potential for residual contamination. Impacted areas are further divided into three classifications. **Table 3** provides a list of Impacted Class 3 Areas and Non-impacted Areas. No NMRC Diving-building areas were classified as Impacted Class 1 or 2.

1. CLASS 1 AREAS

Class 1 areas are impacted areas that, prior to remediation, are expected to have concentrations of residual radioactivity that exceed the DCGL (DCGL is defined in MARSSIM as derived concentration guideline levels or action limits). None of the survey sites were classified as Class 1.

2. CLASS 2 AREAS

Class 2 areas are impacted areas that, prior to remediation, are not likely to have concentrations of residual activity that exceed the DCGL. None of the survey sites were classified as Class 2.

3. CLASS 3 AREAS

Class 3 areas are impacted areas that have a low probability of containing residual radioactivity. **Table 3** lists NMRC Impacted Class 3 Areas. Of the 90 NMRC Diving Buildings survey units, 12 were classified as Impacted Class 3. The Impacted Class 3 sites consisted of rooms where radioactive materials were used or stored.

4. NON-IMPACTED AREAS

Non-impacted areas have no reasonable potential for residual contamination. Table 3 lists NMRC Non-impacted Areas. These non-impacted areas were offices, common areas, and other non-radioactivematerial-use areas.

BUILDING	ROOM	DESCRIPTION	ISOTOPES OF CONCERN	D&D
	OR			CLASSIFICATION
	AREA			
53	101	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	102	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	111	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	112	Office/Lab	H3, C14, Co57,	Impacted,
			Ba133, Cs137	Class 3
53	112A	Laboratory	НЗ, С14, Со57,	Impacted,
			Ba133, Cs137	Class 3
53	113	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	117	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	118	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	119	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	120	Office	H3, C14, Co57	Impacted,
				Class 3
53	121	Laboratory	H3, C14, Co57	Impacted,
				Class 3
53	122	Laboratory	H3, C14, Co57,	Impacted,
			Ba133, Cs137	Class 3
28	100	Supply		Non-impacted
28	101	Mechanical		Non-impacted
28	102	Storage		Non-impacted
28	103	Storage		Non-impacted
28	106	Office		Non-impacted
28	106 A	Laboratory		Non-impacted
28	107	Office		Non-impacted
28	108	Office		Non-impacted
28	200	Office		Non-impacted
28	201	Office		Non-impacted
28	201 A	Office		Non-impacted
28	202 A	Office		Non-impacted
28	202 B	Storage		Non-impacted
28	203	Supply		Non-impacted
28	2.04	Shop		Non-impacted
28	205	Entryway		Non-impacted
53	10	Shop		Non-impacted
53	100	Foyer		Non-impacted
53	100 B	Office		Non-impacted

Table 3. NMRC Diving-Buildings Impacted Class 3 Areas and Non-impacted Areas

53	103	Laboratory	Non-impacted
53	104	Office	Non-impacted
53	105	Office	Non-impacted
53	108	Conference	Non-impacted
		room	
53	109	Office	Non-impacted
53	11	Shop	Non-impacted
53	110	Office	Non-impacted
53	114	Office	Non-impacted
53	115	Gear Lab	Non-impacted
53	116	Office	Non-impacted
53	120 B	Office	Non-impacted
53	120 C	Office	Non-impacted
53	120 D	Office	Non-impacted
53	123	Office	Non-impacted
53	124	Office	Non-impacted
53	124 A	Office	Non-impacted
53	125	Office	Non-impacted
53	126	Office	Non-impacted
53	126 A	Office	Non-impacted
53	126 B	Office	Non-impacted
53	127	Office	Non-impacted
53	128	Telephone	Non-impacted
53	13	Office	Non-impacted
53	14	Storage	Non-impacted
53	15	Storage	Non-impacted
53	22	Galley	Non-impacted
53	23	Compressors	Non-impacted
53	24	NRCC	Non-impacted
53	25	Office	Non-impacted
53	25 A	Office	Non-impacted
53	25 B	Office	Non-impacted
53	26	Office	Non-impacted
53	26 A	Sickbay	Non-impacted
53	26 B	Office	Non-impacted
53	28	Sickbay	Non-impacted
53	29	Office	Non-impacted
53	30	Office	Non-impacted
53	31	Instrument	Non-impacted
		room	
53	32	Mechanical	Non-impacted
53	33	Mechanical	Non-impacted
53	S-1	Entryway	Non-impacted
59	100	Research	Non-impacted
		tank	
59	101	Laboratory	Non-impacted
59	102	Laboratory	Non-impacted
59	103	Electrical	Non-impacted
		room	
59	104	Closet	Non-impacted

59	106	Laboratory	Non-impacted
59	11	Scuba Locker	Non-impacted
69	100	Shop/mech. room	Non-impacted
69	102	Laboratory	Non-impacted
69	103	Laboratory	Non-impacted
69	104	Control room	Non-impacted
79	100	Storage	Non-impacted
79	101	Mechanical shop	Non-impacted
79	103	Office	Non-impacted
79	104	Laboratory	Non-impacted
79	107	Laboratory	Non-impacted
79	120A	Office	Non-impacted
79	S-1	Entryway	Non-impacted

D. POTENTIAL CONTAMINATED MEDIA

The potential contaminated media were building surfaces such as walls, floors, workbenches, sinks and shelves.

E. CONCEPTUAL MODEL FOR DECOMMISSIONING

For the convenience of scheduling necessary radiological activities and to maximize the use of available resources, NMRC's Decontamination and Decommissioning (D&D) Plan was divided into four distinct phases. The NMRC Annex was the first phase to undergo radiological D&D efforts. The NMRC Diving Buildings comprised the second D&D phase. The other two phases included the NMRC main buildings and NMRC Building 150. **Table 4** summarizes the NMRC Diving Buildings Radiological D&D Conceptual Model Information. To accommodate radiological monitoring and surveys, some of the survey units were subdivided.

TABLE 4. NMRC Diving-Buildings Radiological D&D					
Conceptual Model Information					
Building	Number o	f Survey	Floor	Floor Area	
Number	Units (Rooms)		(square feet)		
	Non-	Impacted	Non-	Impacted	
	impacted	Class 3	impacted	Class 3	
	Areas	Areas	Areas	Areas	
28	16	0	3,415	0	
53	44	12	14,026	7,673	
59	7	0	2,695	0	
69	4	0	805	0	
79	7	0	1,195	0	
Total	78	12	22,136	7,673	

The examination for potential radioactive contaminants included direct measurements using a gas proportional counter, sampling and analysis using a liquid scintillation counter, scanning with appropriate instruments and equipment, and data analyses. For Impacted Class 3 areas, the survey unit examinations included:

- (1) all accessible floor space,
- (2) the lower half of all walls extending up approximately two meters from the floor, and
- (3) other selected locations.

After equipment, materials and wastes were removed from radioactive-materials-use areas, it was found that residual contamination was confined to small areas within a small percentage of the survey units.

VIII. DECOMMISSIONING ACTIVITIES

A. OBJECTIVE

The objective of this decommissioning effort was to demonstrate that the sites are suitable for unrestricted release in accordance with the criteria for decommissioning in Subpart E, "Radiological Criteria for License Termination," of 10 CFR Part 20, "Standards of Protection Against Radiation" (ref 11.4).

To achieve this objective, guidance provided in MARSSIM and other regulatory documents was followed. Parameters were chosen to ensure that the following objectives were met:

- Identification of the premises and radiological operations;
- (2) Demonstration that reasonable efforts were planned to reduce/eliminate residual contamination to as low as practicable;
- (3) Documentation of decontamination and decommissioning plan and survey standard operating procedures;
- (4) Preparation of the Final Status Survey Report.

B. COST ESTIMATES FOR DECOMMISSIONING ACTIVITIES

Cost estimates were prepared to approximate the cost for various decommissioning activities, including labor, materials, equipment, services, shipping expenses, and other related expenses. Estimates from commercial firms were received regarding some of NMRC's planned decommissioning efforts. In addition, estimates for equipment purchases and disposals were obtained. One vendor quoted a labor rate of \$55.00 per hour for technician support working under NMRC supervision. The manpower, equipment and other essential resources for the radiological D&D operations for the NMRC Diving Buildings cost approximately \$25,000.

C. RELEASE LIMITS AND DOSE PATHWAY MODELING

The release limits, which include fixed and loose surface contamination, were based upon the NRC guideline screening values. Because NMRC established action limits at a fraction of NRC screening values, dose modeling was not required. Site-specific parameters were not required and were not developed. Dose/pathway modeling was not required because detectable radioactivity was remediated to background levels by decontamination efforts or removal of the contaminated media.

D. CRITERIA FOR LICENSE TERMINATION (NUREG-1549)

The criteria for releasing a site for unrestricted use is described below; the criteria for releasing a site for unrestricted and restricted use are summarized in Table 5.

<u>Subpart E, 10 CFR 20, paragraph 20.1402 - Criteria for</u> <u>unrestricted use</u> - a site is considered acceptable for unrestricted use if the residual radioactivity that is distinguishable from background radiation results in a Total Effective Dose Equivalent (TEDE) to an average member of the critical group that does not exceed 25 mrem/yr, including that from groundwater sources of drinking water, and the residual radioactivity has been reduced to levels that are as low as reasonably achievable (ALARA).

Table 5. Summary of 10 CFR Part 20 Subpart E				
	Unrestricted Use	Restricted Release		
Dose	25 mrem TEDE per	25 mrem TEDE	100 mrem or 500	
Criterion	year peak annual	per year peak	mrem TEDE per	
	dose to the	annual dose to	year peak	
	average member of	the average	annual dose to	
	the critical	member of the	the average	
	group	critical group	member of the	
		while controls	critical group	
		are in place	upon failure of	
			controls	
Time Frame	1000 years	1000 years	1000 years	
Other	ALARA	ALARA,	ALARA,	
Requirements		financial	financial	
		assurance,	assurance,	
		public	public	
		participation	participation	

E. DERIVED CONCENTRATION GUIDELINE LEVELS (DCGLS)

A concentration of residual radioactivity distinguishable from background that, if distributed uniformly throughout a survey unit, would result in a total effective dose equivalent (TEDE) of 25 millirems per year to an average member of the critical group is called the derived concentration guideline level(DCGL).

Residual levels of radioactive material that correspond to allowable radiation dose standards are calculated (derived) by analysis of various pathways and scenarios (direct radiation, inhalation, ingestion, etc.) through which exposures occur. These derived levels, DCGLs, usually refer to average levels of radiation or radioactivity above appropriate background levels.

For NMRC, as would be the case for the majority of MARSSIM users, the applicable DCGLs were obtained using regulatory agency guidance (screening values)

based on default values. Other users may elect to perform site-specific pathway modeling to determine DCGLs. In both cases, the DCGL is based on the spatial distribution of the contaminant, and each derivation can produce different values depending on the specific radionuclide distribution and pathway modeling. Because residual activity levels, when found, were reduced to background levels, DCGLs or screening values were compared directly to survey data to demonstrate compliance. It is understood that this simple approach would not have been possible if more than one radionuclide was present (in the contaminated area). Residual radioactivity limits for the radionuclides present at the site were predetermined. NMRC action limits were set at a small fraction of D&D screening values and were set such that the NRC release criteria would be satisfied.

Table 6 provides the NRC D&D screening values for common radionuclides. The screening values provided in Table 6 will serve as basis for establishing the action limits or DCGLs for NMRC's D&D final status survey data analysis. NMRC's action limit will be set at a small fraction of applicable NRC screening values, 700 dpm, for H-3 and C-14 swipe analyses and be set at a level of two to three times background levels for all other measurements (e.g., beta-gamma scans and fixed-contamination evaluations). Procedures provided in the MARSSIM and 10 CFR 20, Appendix B, would have been used if multiple radionuclides were present. Also provided in Table 6 are NMRC action limits for other beta emitters and gamma emitters.

Table 6. Acceptable License Termination Screening Values of Common Radionuclides for Building Surface Contamination (ref 11.13)

Radionuclide	Symbol	Acceptable	NMRC
		screening	Action
		levels ¹ for	Limits
		unrestricted	
		release (dpm/100	(dpm/100
		cm^2) ²	cm ²)
Hydrogen-3 (Tritium)	³ H	1.2E+08	700
Carbon-14	14C	3.7E+06	700
Sodium-22	²² Na	9.5E+03	
Sulfur-35	³⁵ S	1.3E+07	700
Chlorine-36	³⁶ C1	5.0E+05	
Manganese-54	⁵⁴ Mn	3.2E+04	
Iron-55	⁵⁵ Fe	4.5E+06	
Cobalt-60	⁶⁰ Co	7.1E+03	700
Nickel-63	⁶³ Ni	1.8E+06	
Strontium-90	⁹⁰ Sr	8.7E+03	
Technitium-99	⁹⁹ TC	1.3E+06	
Iodine-129	129 I	3.5E+04	
Cesium-137	¹³⁷ Cs	2.8E+04	
Iridium-192	¹⁹² Ir	7.4E+04	
Other beta emitters	³² P, ³³ P, ⁴⁵ Ca		700
Other gamma emitters	⁵¹ Cr, ¹³³ Ba, ¹³⁷ Cs		700

¹ Screening levels are based on the assumption that the fraction of removable surface contamination is equal to 0.1. For cases when the fraction of removable contamination is undetermined or higher than 0.1, users may assume, for screening purposes, that 100% of surface contamination is removable, and therefore the screening values should be decreased by a factor of 10. Alternatively, users having site-specific data on the fraction of removable contamination (e.g., within the 10% to 100% range) may calculate the site-specific screening levels using DandD Version 1.

³ Units are disintegrations per minute per 100 square centimeters (dpm/100 cm²). One dpm is equivalent to 0.0167 becquerel (Bq). The screening values represent surface concentrations of individual radionuclides that would be deemed in compliance with the 0.25 Sv/year (25 mrem/year) unrestricted dose limit in 10 CFR 20.1402. For radionuclides in a mixture, the "sum of fractions" rule applies; see 10 CFR Part 20, Appendix B, Note 4. Refer to NRC Draft Guidance DG-4006 for further information on application of the values in this table.

F. PRE-SURVEY SAFETY ACTIONS

Pre-survey walk-through inspections were conducted to identify and remedy safety concerns before the designated sites were surveyed.

G. BACKGROUND/BASELINE LEVELS

The primary contaminants of interest (H-3, C-14, Co-57, Ba-133, and Cs-137) did not occur in background at the survey sites, or the background levels were small fractions of the radionuclide-specific DCGL_w values. Therefore, the survey unit radiological conditions, in dpm, were compared directly to the NMRC action limit (a small fraction of the DCGL). Reference area background surveys were conducted on each day that impacted areas were surveyed. The reference area background survey data is included in this final status survey report.

From the vicinity of a survey site or group of survey sites, background survey-meter readings were recorded, and background swipe samples were collected, analyzed and documented.

H. SURVEY UNITS

A survey unit was defined as a physical area consisting of a specified size and shape for which a separate decision will be made as to whether or not that area exceeds the release criterion. The survey unit is the primary entity for demonstrating compliance with the release criteria.

Per MARSSIM, survey units should be limited in size based on classification, exposure pathway modeling assumptions, and site-specific conditions. The suggested areas for survey units are provided in Table 7 as follows:
Table	 Suggested Are (MARSSIM, Road 	eas for Survey Units Admap Table 1)
	CLASSIFICATION	SUGGESTED AREA
Class :	Structures	Up to 100 m ² floor area
	Land Areas	Up to 2,000 m ²
Class :	2 Structures	100 to 1,000 m ²
	Land Areas	2,000 to 10,000 m ²
Class :	3 Structures	No limit
	Land Areas	No limit

NMRC Diving Buildings Radiological (D&D) Final Status Survey Report, June 2000

A NMRC survey unit was defined as a room or other discrete area. Every room and area in the NMRC Diving Buildings was examined for residual contamination. Large rooms or areas were sub-divided into smaller sections to facilitate surveying and monitoring. All known licensed-materials use and/or storage areas were classified as Class 3 survey units. Offices, hallways, lavatories, and other common non-licensematerial-use areas were treated as non-impacted areas. As a quality control (QC) measure, all survey units or areas received some level of survey effort.

I. QUALITY ASSURANCE AND QUALITY CONTROL

Quality Control (QC) and Quality Assurance (QA) measures were incorporated into the decommissioning efforts to ensure the security and integrity of the data and procedures. Maintenance and calibration of equipment were key considerations. In addition, daily and weekly quality control checks for all monitoring equipment and instruments were performed, analyzed and recorded. Periodic training was provided to maintain a competent work force. Redundant survey efforts, which included collecting and analyzing numerous additional swipe samples were planned, completed and documented.

Quality control measures and standard operating procedures were implemented before initiating decommissioning actions. Quality control procedures and data for ensuring the validity of survey results are included in this report in Appendix D. These procedures and plans included:

- (1) properly calibrated instrumentation,
- (2) necessary replicate, reference and blank measurements, and
- (3) comparison of field measurement results to laboratory sample analyses.

For comparison of survey data to DCGLs, the survey data obtained from swipe-analyses via laboratory measurements was reported in dpm. The swipe analyses results for H-3 and C-14 were compared directly to the MMRC Radiological action limits in Table 6.

Data verification ensured that operating procedures were implemented as prescribed. Data verification means that deficiencies or problems that occurred during implementation were documented and presented in this report. This also means that NMRC D&D efforts were assessed regularly and findings were documented. Corrective actions were reviewed for adequacy and appropriateness and documented in response to findings. Data verification activities included, but are not limited to, inspections, QC checks, surveillance, technical reviews, performance evaluations and audits.

To ensure that the conditions requiring corrective actions were identified and addressed promptly, data verification activities were initiated as part of sample collection and data preparation phases of the survey.

The performance of tasks by personnel was compared to standard procedures and was assessed using inspections, surveillance and audits. Selfassessments and independent assessments were planned, scheduled and performed as part of the survey. Selfassessment also means that personnel performing the work were required to document and report deficiencies or problems that they encountered to the D&D supervisor.

The performance of equipment such as radiation detectors or measurement was monitored using control charts. Control charts were be used to record the results of quantitative QC checks such as background and daily calibration or performance checks. Control charts were used to document instrument and measurement system performance on a daily or as-needed basis to identify conditions requiring real-time corrective actions. Periodic technical reviews may be requested as an independent-assessment measure.

Data validation activities ensured that the results of data collection activities supported the objectives of the survey. Data verification compared the collected data with the investigational levels (action limits) or DCGLs provided in Table 6.

J. FIELD MEASUREMENT METHODS AND INSTRUMENTATION

1. FIELD MEASUREMENTS

Surveys for decommissioning required the collection of two types of radiological data:

- (1) direct field measurements using portable instruments and
- (2) sample analyses using fixed laboratory equipment or systems.

For either type of measurement, the selection and proper use of appropriate instruments were the most critical factors in assuring that the survey accurately determined the radiological status of the site. Table 8 summarizes equipment and instrument sensitivities for laboratory and field measurements.

NMRC Diving Buildings Radiological (D&D) Final Status Survey Report, June 2000

Table 8. 7	Typical Measurer	ment Sensitivities for Laboratory Radi	ometric
I	Procedures (MARS	SSIM, Chapter 7)	
Sample	Radionuclides	Procedure	Approximate
Туре	or Radiation		Measurement
	Measured		Sensitivity
Smears	Gross beta	Gas-flow proportional counter;	10 dpm
(filter		5-min count	
paper)		End window GM with scaler; 5-min	80 dpm
		count (unshielded detector)	
	Low energy	Liquid scintillation spectrometer;	30 dpm
	beta	5-min count	
	(H-3, C-14)		
Soil	Co-60	Germanium detector (25% relative	0.04 - 0.1 Bq/g
Sediment		efficiency) with multi-channel	(1-3 pCi/g)
		analyzer; pulse height analyzer;	
		200-min count	
Water	Gross beta	Gas-flow proportional counter; 100-	0.04 Bg/L
		ml sample, 200-min count	(1 pCi/L)
	Co-60	Germanium detector (25% relative	0.4 Bq/L
		efficiency) with multi-channel	(10 pCi/L)
		analyzer; pulse height analyzer;	
		3.5L sample, 16-hour count	
	Н-3	Liquid scintillation spectrometer;	100 Bq/L
		5-min count	(300 pCi/L)

2. INSTRUMENTS, EQUIPMENT, MEASUREMENT AND SAMPLING TECHNIQUES

Instruments for alpha detection were not required. Instrumentation and equipment were selected such that radiation detection procedures (scanning, direct measurements, and sample analyses) could be completed for the radionuclides of interest. Specifications and calibrations of radioactive check sources and radiation-monitoring equipment and instruments are provided in Appendix D, sections 3 and 4, respectively.

For beta radiation detection, the following instruments and equipment were used:

- Samples were analyzed using standard procedures and calibrated liquid scintillation counters.
- (2) Scanning was performed using appropriate Geiger-Mueller detectors with pancake probes or digital-scaler, gas proportional counter with appropriate detectors.

For gamma radiation detection, scanning was performed using appropriate Geiger-Mueller detectors with appropriate probes.

Exposure rate measurements, when necessary, were obtained using appropriate instrumentation.

A combination of two approaches was used for scanning Class 3 areas. The two approaches included uniformly scanning for the radionuclides of interest and/or performing scans in areas with the greatest potential for residual contamination based on professional judgment and the objectives of the survey. The investigation level was set at an elevated activity reading that exceeded 2 to 3 times the background readings.

Appropriate calibration check sources were obtained to perform instrument background and check-source (quality assurance and quality control) checks. Instrument response (background and check source) was tested and recorded a minimum of once daily prior to beginning the day's measurements — to ensure continued acceptable operation. If an instrument's response did not satisfy the established acceptable range, the instrument was removed from use until the reason for the deviation could be determined and resolved and acceptable response again demonstrated. If repair and/or recalibration were necessary, acceptable response ranges were reestablished and documented.

K. SAMPLING AND PREPARATION FOR LABORATORY MEASUREMENTS

1. FIELD SAMPLE PREPARATION AND PRESERVATION

Proper sample preparation and preservation are essential parts of any radioactivity-sampling program. The sampling objectives were specified in the D&D Standard Operating Procedures (D&D SOP) manual and Appendix C. Precise records of sample collection and handling were performed to ensure that data obtained from different locations or timeframes were correctly compared.

According to MARSSIM, Chapter 7, field and preparation of building surfaces, including smear samples, is generally not required. Smear samples were collected from randomly selected locations and other areas, labeled, packaged, tracked, and delivered to the on-site counting laboratory for analyses.

2. OTHER MEASUREMENTS AND SAMPLING LOCATIONS

In addition to building surface areas, there were numerous other locations where measurements and/or sampling were performed. Examples include items of equipment and furnishings, workbenches, shelves, building fixtures, drains, ducts, and piping. Many of these items or locations had both internal and external surfaces, requiring evaluation. Each such location classified as impacted was scanned and individual measurements and/or sampling were performed. Non-impacted locations were also surveyed, but at lower frequencies, consistent with contamination potential and findings as the survey progressed.

IX. DECONTAMINATION PROCEDURES

Using the guidance, procedures, and checklists provided in the MARSSIM, characterization surveys, reclassification of survey areas, remedial actions, and remediation-action support surveys were performed and documented, when necessary. When contamination was identified in a survey unit, characterization surveys (via scanning or swipe-sample counting) were performed to determine the extent of the contamination. After a characterization survey was completed, remedial actions were completed. The remediation actions consisted of repeated washings or scrubbing or removal of the contaminated media. Waste generated from this process was treated as radioactive waste and disposed of accordingly. Remediation action surveys were completed following iterations of decontamination efforts to measure the effectiveness of the efforts. Final surveys of the contaminated areas were completed and documented.

X. FINAL STATUS SURVEY

A. OVERVIEW

The guidance provided in MARSSIM, Section 5.5, Final Status Surveys, was closely followed. The final status surveys were performed to demonstrate that residual radioactivity in each survey unit satisfied the predetermined release criteria for unrestricted future use. The survey provided data to demonstrate that all radiological conditions did not exceed the established DCGLs or NMRC action limits.

In demonstrating that the objective was met, the null hypothesis (H_o) tested was that residual contamination exceeded the release criterion; the alternate hypothesis (H_a) was that residual contamination met the release criterion. Two statistical tests could be used to evaluate data from final status surveys. For contaminants that are present in background, the Wilcoxon Rank Sum (WRS) test would be used. When contaminants are not present in background, the Sign test is used.

Since contaminants at the Diving Buildings were not present in background in significant amounts, the Sign test was used. Statistical parameters (α and β) were chosen such that there was less than a 5% probability that radioactivity would be reported present when it was really absent (Type I error) or reported absent when it was really present (Type II error).

B. SURVEY PREPARATIONS

Residual radioactivity limits for the radionuclides present at the site were selected. NMRC action limits or investigational levels were equivalent to a small fraction of D&D screening values and were set such that the NRC release criteria were satisfied. **Table 6** provides the NRC D&D screening values for NMRC isotopes of concern (ref 11.13), action plan values, and NMRC's residual radioactivity limits. Guidance in **Table 9** was used to interpret the survey results. NMRC Diving Buildings Radiological (D&D) Final Status Survey Report, June 2000

Table 9. Interpretation of Samp Reference Area Is Used	ole Measurements When No d (MARSSIM, Table 2.5)
SURVEY UNIT MEASUREMENT RESULTS	CONCLUSION
All concentrations are less	Survey unit meets release
than $DCGL_w$ (or NMRC action	criteria
limit)	
Average concentration is	Survey unit fails
greater than $ extsf{DCGL}_w$ (or $ extsf{NMRC}$	
action limit)	
Any concentration is greater	Conduct Sign test and elevated
than $DCGL_w$ and average	measurement comparison
concentration is less than $DCGL_w$	

Representative reference (background) areas were selected for the survey units. Reference areas were selected from non-impacted areas and were:

- (1) free of contamination from site operations,
- (2) exhibited similar physical, chemical, and biological characteristics as the survey unit, and
- (3) had similar construction, but had no history of radioactive operations.

If swipe samples were collected for impacted survey units on the first floor during a day's operations, background samples were collected on the first floor as well. Similarly, if swipe samples were collected in impacted areas on the second floor during a day's operations then background samples were collected from second floor locations.

Survey instruments were selected based upon the techniques to be employed (scanning, direct measurements, and sampling) and the radionuclides of concern. The selected instruments were capable of detecting the contamination at 10-50% of the DCGLs.

The survey unit was prepared before conducting the final survey by clearing and providing access to the areas to be surveyed, including the floor, lower walls, workbenches, shelves and fixtures. Reference coordinate systems using grid maps were established and followed.

C. Survey Design

Sample collection and analysis procedures were developed and implemented as D&D Standard Operating Procedures in Appendix C. All radiological workers received initial and periodic training.

The number of data points for statistical tests were determined using the assumption that the radionuclides were not present in background. The Sign test was used. The number of samples for Class 3 areas were determined using the following statistical values, Type I error where $\alpha = 0.025$, Type II error where $\beta =$ 0.025, and the relative shift, $\Delta/\sigma_s = 2$. The relative shift was determined by using data from a contaminated area and the formulae provided in MARSSIM. The data and the determination of the relative shift are provided in Table 10. The number of data points, N, was determined to be 21. Per MARSSIM, this N value added an extra 20% to account for unusable data or uncertainty in the value of N.

The number of random data samples collected for each Impacted Class 3 survey unit was 21. Additional samples/measurements were taken for quality control and to allow for possible loss of data and increased sampling with survey units. The additional data points were not used when conducting statistical tests of survey data. In conjunction with completing surveys for all impacted areas, reference background readings and samples were collected. Г

Table 10. Determination of using swipe data from a con MARSSIM, Section 5.5.2.3)	the Relative Shift, Δ/σ , taminated area (MARSSIM,
Let $\Delta = 2500$ dpm - 700 dpm	
Basis of Δ : one-half of the Guide 1.86 contamination li action limit of 700 dpm	conservative Regulatory mit of 5,000 dpm and NMRC
In this table, σ is calcula	ted using H-3 swipe data
from a contaminated area	
Sample Number	H-3, dpm
1	345
2	700
3	621
4	391
5	2884
6	160
7	205
8	203
9	153
10	2892
11	642
12	1009
13	74
14	43
15	104
16	67
17	1161
18	172
19	51
20	58
21	77
22	2795
23	110
24	183
Average	629
Standard deviation	913
Set $\Delta = 2500 - 700$ dpm	1800
Δ/σ , calculated	1.97 ~= 2

Table 11 provides recommended survey coverage for structures and land areas (MARSSIM, table 5.9).

Table 11. Reco	mmended Surv	vey Coverage For	r Structures	
and	Land Areas	(MARSSIM, Table	5.9)	
AREA	STRU	JCTURES	LANI	AREAS
CLASSIFICATION	Surface	Surface	Surface	Soil Samples
	Scans	Activity	Scans	
		Measurements		
Class 1	100%	Number of	100%	Number of
		data points		data points
		from		from
		statistical		statistical
		tests and		tests and
		additional		additional
		measurements		measurements
		may be		may be
		necessary for		necessary
		small areas		for small
		of elevated		areas of
		activity		elevated
Class 2	10 to 100%	Number of	10 to 100%	Number of
	(10 to 50%	data points	Systematic	data points
	for upper	from	and	from
	walls and	statistical	Judgmental	statistical
	ceilings)	tests		tests
	Systematic			
	and			1
	Judgmental			
Class 3	Judgmental	Number of	Judgmental	Number of
		data points		data points
		from		from
		statistical		statistical
		tests		tests

Sampling locations for Impacted Class 3 areas utilized a rectangular grid system that used a local landmark, such as doors in survey units. Grid systems were established for the impacted survey sites to:

- Facilitate systematic selection of measuring/sampling locations,
- Provide a mechanism for referencing a measurement/sample back to a specific location so that the same survey point can be relocated, and
- Provide a convenient means for determining average activity levels.

The grids consisted of a system of intersecting lines created by using colored paper or plastic adhesive tape and referenced to a fixed site location or benchmark. The grid lines were arranged in a perpendicular pattern, dividing the survey location into approximately one-meter squares or blocks. Grid patterns on horizontal surfaces were identified numerically on one axis and numerically on the other axis. Grids on vertical surfaces are shown as flat surfaces on the grid maps. Grid maps used during the survey efforts are provided with the data and summaries in Appendix XII. Gridding was limited to the floor and lower (up to 2 meters height) walls.

Because radioactive-material use during the last 10 years has involved the use of small quantities in liquid form in hoods or workbench areas, only the lower two meters of the walls in Impacted Class 3 areas will be surveyed. No airborne residual contamination is anticipated.

For the Class 3 survey units, the locations of the 21 random data points were determined using lists of random numbers. Additional measurements were selected to ensure good coverage of the survey unit and to survey selected areas within the survey unit.

Quality control procedures for ensuring the validity of survey results were included in D&D Standard Operating Procedures. These procedures included:

- (1) survey procedures and techniques,
- (2) sample collection and processing procedures,
- (3) data-analysis procedures,
- (4) equipment and instrument procedures, and
- (5) source-standards, background-standards and reference-background measurements.

D. CONDUCTING SURVEYS

The reference (background) area measurements and sampling were performed and documented.

Final status survey activities were conducted. The activities included:

- performing surface scans of Class 3 areas and a percentage of non-impacted areas,
- (2) conducting surface activity measurements and sampling at previously selected sampling locations, and
- (3) conducting additional direct measurements and sampling at locations based on professional judgment.

Necessary investigation activities, including remediation and re-survey were performed and documented. Reclassifications of survey units were not required.

Measurement and sample locations were documented on grid maps. The grids with sample locations and sample numbers are provided in Appendix E with survey data for each impacted survey unit.

Observations, abnormalities, and deviations from the standard operating procedures were documented.

E. DATA CONVERSION

Data conversion was not required for results obtained by evaluating swipe samples using the liquid scintillation counter. The results for each swipe sample were reported in dpm for both H-3 and C-14 for an area of approximately 100 cm². Data conversion was required for surface activity measurements that were obtained using a hand-held, gas proportional detector. Guidance provided in chapter 6 of MARSSIM was used to convert the meter readings to the conventional surface activity units of dpm per 100 cm². Table 12 provides the applicable MARSSIM equation and a sample calculation.

Table 12. Data Conversion for Direct Readings of Surface Activity Using a Gas Proportional Detector (MARSSIM, Section 6.6.1) MARSSIM formula 6-4: $\frac{dpm}{(100 \text{ cm}^2)} = \frac{(C_s/T_s) - (C_b/T_b)}{(E_i \text{ x} (A/100))}$ where, A = physical probe area in cm^2 , $E_i = instrument$ efficiency, sample and background count times, T_s and T_b in seconds, $C_s =$ sample counts during time T_s , and $C_b =$ background counts during time, Tb Note: Interpret negative values of the calculated dpm as equal to zero. Sample calculation: Let A = 126 cm²; $T_s = T_b = 1$ minute; $E_i = 15\%$; $C_{\rm s} = 361$ and $C_{\rm b} = 211$ ((361/1) - (211/1))dpm (0.15 * (126/100)) = 794 (100 cm^{-1})

F. EVALUATING AND DOCUMENTING SURVEY RESULTS

Samples were analyzed using the procedures detailed in the SOP and Appendix C. Swipe-sample results in dpm were compared directly to DCGL values and NMRC action limits. Individual measurements and sample concentrations were first compared to DCGL levels for evidence of small areas of elevated activity. No additional data comparisons were necessary. The initial swipe results or results following remediation efforts did not exceed NMRC action limits.

Verification of assumptions of statistical tests was not necessary. Per MARSSIM, Appendix B, "The results of the survey should be compared to derived concentration guideline levels (DCGLs) using an appropriate statistical test, such as Student's t test or Wilcoxon test. If all measurements are less than the DCGL_w, then statistics do not need to be addressed because the conclusions are obvious. If the mean of the measurements exceeds the DCGL_w, the survey unit obviously fails to demonstrate compliance and the statistics do not need to be addressed."

Survey results for each survey unit were compared with NMRC residual radioactivity investigational levels or action levels. These action levels represented small fractions of DCGLs (NRC D&D screening values). Elevated measurement comparisons were not performed and were not needed.

This final status survey report includes the standard operating procedures, quality assurance information and procedures, data, results, interpretations, deviations from standard procedures, and other applicable information.

XI. REFERENCES

- 11.1 NUREG-1575, MARSSIM, Multi-Agency Survey and Site Investigation Manual (December 1997).
- 11.2 NMRC Historical Site Assessment, Naval Medical Research Center (NMRC, formerly NMRI), June 1999.
- 11.3 ESA 1997, Environmental Site Assessment, Naval Medical Research Institute (NMRI), by the Environmental Company, Inc., May 1998.
- 11.4 10CFR20.1402, Title 10, Code of Federal Regulations, Part 20 - Standards for Protection Against Radiation, Subpart E.
- 11.5 NUREG-1500, Working Draft Regulatory Guide on Release Criteria for Decommissioning: NRC's Staff Draft for Comment, August 1994.
- 11.6 NUREG-1501, Background as a Residual Radioactivity Criterion for Decommissioning, Draft Report for Comment, August 1994.
- 11.7 NUREG-1505, A Non-parametric Statistical Methodology for Design and Analysis of Final Status Decommissioning Surveys.
- 11.8 NUREG-1507, Minimum Detectable Concentrations with Typical Radiation Survey Instruments for Various Contaminants and Field Conditions, December 1997.
- 11.9 NUREG-1549, Decision Methods for Dose Assessment to Comply With Radiological Criteria for License Termination (Draft, July 1998).
- 11.10 NUREG/CR-5512, Residual Radioactive Contamination from Decommissioning (Draft, April 1996).
- 11.11 NUREG/CR-5849, Manual for Conducting Radiological Surveys in Support of License Termination, Draft for Comment June 1992 (superseded by NUREG-1575, MARSSIM).
- 11.12 Draft Regulatory Guide DG-4006, Demonstrating Compliance with the Radiological Criteria for License Termination (supersedes NUREG 1500), August 1998.

- 11.13 Federal Register, Volume 63, Number 222, November 18, 1998, page 64134.
- 11.14 REG GUIDE 1.86, Nuclear Regulatory Commission Regulatory Guide 1.86 (1974).

XII. APPENDICES

- A. Conceptual Model and Site Diagram
- B. Documents, Figures and Photographs
- C. Standard Operating Procedures and Quality Assurance Program
- D. Survey Unit Release Reports (Narratives, Data, Floor Plans, and Grid Maps)
- E. Quality Assurance Information and Data

APPENDIX A. CONCEPTUAL MODEL AND FLOOR DIAGRAMS

Appendix A-1. CONCEPTUAL MODEL

NMRC Bethesda Di	ving Buildings (28 , 53, 59, 69, 79
Contaminants of con	icern:	
H-3, C-14, P-32, P-3	33, Ca-45, Ba-1	33, Co-57
BUTLDING SUPP	PACE CLASSI	FICATIONS
BUILDING SURE	FACE CLASSI Number	FICATIONS Total
BUILDING SURE	FACE CLASSI Number Of Survey	FICATIONS Total Area
BUILDING SURE	FACE CLASSI Number Of Survey Units	FICATIONS Total Area (SF)
BUILDING SURE	FACE CLASSI Number Of Survey <u>Units</u> 78	FICATIONS Total Area (SF) 22,136
BUILDING SURE Non-impacted Impacted Class 3	FACE CLASSI Number Of Survey <u>Units</u> 78 12	FICATIONS Total Area (SF) 22,136 7.673

Appendix A-2. FLOOR DIAGRAM, FIRST FLOOR, BUILDING 28



BUILDING 28, FIRST FLOOR

Appendix A-3. FLOOR DIAGRAM, BASEMENT FLOOR, BUILDING 28



Appendix A-4. FLOOR DIAGRAM, FIRST FLOOR, BUILDING 53



Appendix A-5. FLOOR DIAGRAM, BASEMENT FLOOR, BUILDING 53



Appendix A-6. FLOOR DIAGRAM, FIRST FLOOR, BUILDING 59



~

•

..

- .

Appendix A-7. FLOOR DIAGRAM, BASEMENT FLOOR, BUILDING 59



. . . .

.

Appendix A-8. FLOOR DIAGRAM, BUILDING 69


Appendix A-9. FLOOR DIAGRAM, FIRST FLOOR, BUILDING 79



HALINGTON HA

•

Appendix A-10. FLOOR DIAGRAM, BASEMENT FLOOR, BUILDING 79



.

. .

• ...

.*

APPENDIX B. PHOTOGRAPHS

Appendix B-1. PHOTOGRAPHS

NumberNumberPhoto #1NMRC Building 28 - Front View69-1Photo #2NMRC Building 53 - Rear View69-2Photo #3NMRC Building 53 - Front View69-3Photo #4MMRC Building 53 - Front View69-4Photo #5NMRC Building 59 - Side View69-6Photo #6NMRC Building 59 - Rear View69-7Photo #7NMRC Building 69 - Front View (in background)69-7Photo #8NMRC Building 69 - Side View69-10Photo #10NMRC Building 59 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 53 - Typical Floor Grids69-12Photo #13NMRC Building 53 - Typical Floor Grids69-14AVesselVessel69-14AVesselVessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14CVesselPhoto #14DNMRC Building 53 - Lab 117 Pressure (pre-decon = before commencing decontamination efforts)69-14DPhoto #16Bench Area in Building 53, Lab 117 (pre-decon)69-1569-16Photo #17Bench and Floor Area in Building 53, Lab 122 (pre-decon)69-1769-18Photo #18Bench and Floor Area in Building 53, Lab 122 (pre-decon)69-1969-19Photo #14Bench in Building 53, Lab 122 (post- decon)69-1969-19Photo #14Bench in Building 53, Lab 122 (post- decon)69-19	Photo	Photo Name	Page
Photo #1NMRC Building 28 - Front View69-1Photo #2NMRC Building 53 - Front View69-3Photo #3NMRC Building 53 - Front View69-3Photo #4NMRC Building 53 - Front View69-4Photo #5NMRC Building 53 - Rear View69-5Photo #6NMRC Building 59 - Side View69-6Photo #7NMRC Building 69 - Front View (in background)69-7Photo #8NMRC Building 69 - Side View69-9Photo #10NMRC Building 79 - Side View69-10Photo #11NMRC Building 79 - Side View69-12Photo #12NMRC Building 53 - Typical Floor Grids69-12Photo #13NMRC Building 53 - Typical Floor Grids69-14AVesselVessel69-14APhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14APhoto #14DNMRC Building 53 - Lab 117 Pressure (pre-decon = before commencing decontamination efforts)69-15Photo #15Floor Area in Building 53, Lab 122 (pre-decon)69-16Photo #16Bench and Floor Area in Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (pre-decon)69-18Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #10Bench in Building 53, Lab 122 (post- decon)69-19	Number		Number
Photo #2NMRC Building 28 - Rear View69-2Photo #3NMRC Building 53 - Front View69-3Photo #4NMRC Building 53 - Front View69-4Photo #5NMRC Building 53 - Rear View69-5Photo #6NMRC Building 59 - Side View69-6Photo #7NMRC Building 69 - Front View (in background)69-7Photo #9NMRC Building 69 - Side View69-9Photo #10NMRC Building 69 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #11NMRC Building 59 - Side View69-12Photo #11NMRC Building 53 - Typical Floor Grids69-13Photo #13NMRC Building 53 - Typical Floor Grids69-14AVesselVessel69-14AVesselPhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure (pre-decon = before commencing decontamination efforts)69-16Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #18Bench and Floor Area in Building 53, Lab 122 (pre-decon)69-18Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19	Photo #1	NMRC Building 28 - Front View	69-1
Photo #3NMRC Building 53 - Front View69-3Photo #4NMRC Building 53 - Front View69-4Photo #5NMRC Building 53 - Rear View69-5Photo #6NMRC Building 59 - Side View69-6Photo #7NMRC Building 59 - Rear View69-7Photo #7NMRC Building 69 - Front View (in background)69-7Photo #10NMRC Building 69 - Side View69-9Photo #10NMRC Building 79 - Side View69-10Photo #11NMRC Building 79 - Side View69-11Photo #12NMRC Building 53 - Typical Floor Grids69-12Photo #13NMRC Building 53 - Lab 117 Pressure69-14AVesselVessel69-14APhoto #14BNMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVessel(pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 12269-16(pre-decon)(pre-decon)69-1869-18Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #10Bench in Building 53, Lab 122 (post- decon)69-19	Photo #2	NMRC Building 28 - Rear View	69-2
Photo #4NMRC Building 53 - Front View69-4Photo #5NMRC Building 53 - Rear View69-5Photo #6NMRC Building 59 - Side View69-6Photo #7NMRC Building 59 - Rear View69-7Photo #8NMRC Building 69 - Front View (in background)69-8Photo #9NMRC Building 69 - Side View69-10Photo #10NMRC Building 59 - Side View69-10Photo #11NMRC Building 59 - Side View69-12Photo #12NMRC Building 79 - Side and Rear Views69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure Vessel69-14APhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Area in Building 53, Lab 122 (pre-decon)69-18Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-19	Photo #3	NMRC Building 53 - Front View	69-3
Photo #5NMRC Building 53 - Rear View69-5Photo #6NMRC Building 59 - Side View69-6Photo #7NMRC Building 69 - Rear View69-7Photo #8NMRC Building 69 - Front View (in background)69-8Photo #9NMRC Building 69 - Side View69-10Photo #10NMRC Building 59 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 79 - Side View69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure69-14AVesselVessel69-14EPhoto #14BNMRC Building 53 - Lab 117 Pressure69-14CVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVessel99Photo #14BBench Area in Building 53, Lab 12269-16(pre-decon)999Photo #16Bench and Floor Area in Building 53, Lab 12269-17(pre-decon)9969-18Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-20	Photo #4	NMRC Building 53 - Front View	69-4
Photo #6NMRC Building 59 - Side View69-6Photo #7NMRC Building 59 - Rear View69-7Photo #8NMRC Building 69 - Front View (in background)69-8Photo #9NMRC Building 69 - Side View69-9Photo #10NMRC Building 79 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 59 - Side View69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure69-14AVesselVessel69-14BPhoto #142NMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #140NMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #145Floor Area in Building 53, Lab 11769-15Photo #16Bench Area in Building 53, Lab 12269-16(pre-decon)(pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 12269-17Photo #18Bench and Floor Area in Building 53, Lab 12269-18Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19	Photo #5	NMRC Building 53 - Rear View	69-5
Photo #7NMRC Building 59 - Rear View69-7Photo #8NMRC Building 69 - Front View (in background)69-8Photo #9NMRC Building 69 - Side View69-9Photo #10NMRC Building 79 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 53 - Side View69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure69-14AVesselVessel69-14BPhoto #14BNMRC Building 53 - Lab 117 Pressure69-14CVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #14NMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #145Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-16Photo #17Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19	Photo #6	NMRC Building 59 - Side View	69-6
Photo #8NMRC Building 69 - Front View (in background)69-8Photo #9NMRC Building 69 - Side View69-9Photo #10NMRC Building 79 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 59 - Side and Rear Views69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14NMRC Building 53 - Lab 117 Pressure Vessel69-14APhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14BPhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 112 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-18Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19	Photo #7	NMRC Building 59 - Rear View	69-7
background)background)Photo #9MMRC Building 69 - Side View69-9Photo #10MMRC Building 79 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 79 - Side and Rear Views69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure Vessel69-14APhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure (pre-decon = before commencing decontamination efforts)69-16Photo #15Floor Area in Building 53, Lab 112 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-18Photo #18Bench and Floor Area in Building 53, Lab 122 (pre-decon)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19	Photo #8	NMRC Building 69 - Front View (in	69-8
Photo #9NMRC Building 69 - Side View69-9Photo #10NMRC Building 79 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 79 - Side and Rear Views69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure69-14AVesselVessel69-14BPhoto #14BNMRC Building 53 - Lab 117 Pressure69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure69-14CVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #14BShort Area in Building 53, Lab 11769-15(pre-decon = before commencing decontamination efforts)69-16Photo #17Bench and Floor Building 53, Lab 12269-16(pre-decon)(pre-decon)69-18Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #10Bench in Building 53, Lab 122 (post- decon)69-19		background)	
Photo #10NMRC Building 79 - Side View69-10Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 79 - Side and Rear Views69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure69-14AVesselVessel69-14BPhoto #14BNMRC Building 53 - Lab 117 Pressure69-14BVesselVessel69-14CPhoto #14CNMRC Building 53 - Lab 117 Pressure69-14CVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #15Floor Area in Building 53, Lab 11769-15Photo #16Bench Area in Building 53, Lab 12269-16(pre-decon)(pre-decon)69-17Photo #17Bench and Floor Building 53, Lab 12269-17Photo #18Bench and Floor Area in Building 53, Lab 12269-18Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #9	NMRC Building 69 - Side View	69-9
Photo #11NMRC Building 59 - Side View69-11Photo #12NMRC Building 79 - Side and Rear Views69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure69-14AVesselVessel69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure69-14BVesselVessel69-14CPhoto #14CNMRC Building 53 - Lab 117 Pressure69-14CVesselVessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure69-14DVesselVessel69-14DPhoto #15Floor Area in Building 53, Lab 11769-15(pre-decon = before commencing decontamination efforts)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-19	Photo #10	NMRC Building 79 - Side View	69-10
Photo #12NMRC Building 79 - Side and Rear Views69-12Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure Vessel69-14APhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16 (pre-decon)Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19	Photo #11	NMRC Building 59 - Side View	69-11
Photo #13NMRC Building 53 - Typical Floor Grids69-13Photo #14ANMRC Building 53 - Lab 117 Pressure Vessel69-14APhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-16Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #12	NMRC Building 79 - Side and Rear Views	69-12
Photo #14ANMRC Building 53 - Lab 117 Pressure Vessel69-14APhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14BFloor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #13	NMRC Building 53 - Typical Floor Grids	69-13
VesselPhoto #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #14A	NMRC Building 53 - Lab 117 Pressure	69-14A
Photo #14BNMRC Building 53 - Lab 117 Pressure Vessel69-14BPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16 (pre-decon)Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20		Vessel	
VesselPhoto #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16 (97-16)Photo #17Bench And Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #14B	NMRC Building 53 - Lab 117 Pressure	69-14B
Photo #14CNMRC Building 53 - Lab 117 Pressure Vessel69-14CPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16 (9-16 (9-16)Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20		Vessel	
VesselPhoto #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #14C	NMRC Building 53 - Lab 117 Pressure	69-14C
Photo #14DNMRC Building 53 - Lab 117 Pressure Vessel69-14DPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20		Vessel	
VesselPhoto #15Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)69-15Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #14D	NMRC Building 53 - Lab 117 Pressure	69-14D
Photo #15Floor Area in Building 53, Lab 11769-15(pre-decon = before commencing decontamination efforts)69-16Photo #16Bench Area in Building 53, Lab 12269-16(pre-decon)(pre-decon)69-17Photo #17Bench and Floor Building 53, Lab 12269-17(pre-decon)(pre-decon)69-18Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-19Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20		Vessel	
(pre-decon = before commencing decontamination efforts)Photo #16Bench Area in Building 53, Lab 122 (pre-decon)Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)Photo #19Bench in Building 53, Lab 122 (post- decon)Photo #20Bench in Building 53, Lab 122 (post- decon)	Photo #15	Floor Area in Building 53, Lab 11/	69-15
Photo #16Bench Area in Building 53, Lab 122 (pre-decon)69-16Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20		(pre-decon = before commencing	
Photo #16Bench Area in Building 53, Lab 12269-16(pre-decon)(pre-decon)69-17Photo #17Bench and Floor Building 53, Lab 12269-17(pre-decon)(pre-decon)69-18Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	D1 1 // 1 C	decontamination efforts)	60.16
Photo #17Bench and Floor Building 53, Lab 122 (pre-decon)69-17Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Photo #16	Bench Area in Bullding 53, Lab 122	69-16
Photo #17Bench and Floor Building 53, Lab 12269-17(pre-decon)(pre-decon)69-18Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	DL . L . (117)	(pre-decon)	CO 17
Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	PNOTO #1/	Bench and Floor Building 53, Lab 122	69-17
Photo #18Bench and Floor Area in Building 53, Lab 122 (post-decon = after surveys following decontamination efforts)69-18Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20	Dhata #10	(pre-decon)	60.10
Lab 122 (post-decon - after surveys following decontamination efforts)Photo #19Bench in Building 53, Lab 122 (post- decon)Photo #20Bench in Building 53, Lab 122 (post- decon)	Photo #18	Bench and Floor Area in Building 55,	69-10
Photo #19Bench in Building 53, Lab 122 (post- decon)69-19Photo #20Bench in Building 53, Lab 122 (post- decon)69-20		following decontamination offerts)	
Photo #20 Bench in Building 53, Lab 122 (post- decon) Photo #20 Bench in Building 53, Lab 122 (post- decon)	Photo #19	Bonch in Building 53 Lab 122 (nost-	69-19
Photo #20 Bench in Building 53, Lab 122 (post- 69-20 decon)		decon)	
decon)	Photo #20	Bench in Building 53 Lab 122 (nost-	69-20
		decon)	

Photo #1 - Building 28, Front View



Photo #2 - Building 28, Rear View



Photo #3 - Building 53, Front View



Photo #4 - Building 53, Front View



Photo #5 - Building 53, Rear and Side Views



<u>9.6.2006</u>

Photo #6 - Building 59, Side View (building on left)



Photo #7 - Building 59, Rear View (building in center)

Photo #8 - Building 69, Front View (in background)



Photo #9 - Building 69, Side View



Photo #10 - Building 79, Side View (in background)



Photo #11 - Building 59, Side View



Photo #12 - Building 79, Side and Rear Views



Photo #13 - Building 53, Typical Floor Grids



Photo #14A - Building 53, Lab 117 Pressure Vessel (yellow-blocked areas were contaminated with C-14)



Photo #14B - Building 53, Lab 117 Pressure Vessel



Photo #14C - Building 53, Lab 117 Pressure Vessel



Photo #14D - Building 53, Lab 117 Pressure Vessel



Photo #15 - Floor Area in Building 53, Lab 117 (pre-decon = before commencing decontamination efforts)



Photo #16 - Bench Area in Building 53, Lab 122 (pre-decon)



Photo #17 - Bench and Floor Building 53, Lab 122 (pre-decon)



i.

Photo #18 - Bench and Floor Area in Building 53, Lab 122
(post-decon = after surveys following decontamination
efforts)



Photo #19 - Bench in Building 53, Lab 122 (post-decon)



Photo #20 - Bench in Building 53, Lab 122 (post-decon)



APPENDIX C. STANDARD OPERATING PROCEDURES AND QUALITY ASSURANCE PROGRAM

Appendix C-1. Radiological D&D Action Plan Checklist

1. Select appropriate monitoring instruments and equipment and ensure satisfactory daily and weekly calibration checks for your selected instruments. Record serial numbers. Choose from:

Ludlum Model 3 with pancake probe (beta) Ludlum Model 3 with linch diameter NaI probe (gamma) Ludlum with 2-inch diameter NaI probe (gamma) Cart-mounted gas proportional counter (beta; not H-3) Hand-held gas proportional counter (beta; not H-3) Liquid scintillation counter (beta and gamma) Gamma Counter (gamma)

2. Select materials to take to the scene. In addition to items you deem appropriate, the following items may be needed:

- ______ tape measure or meter stick or yard stick
- water in a spray bottle
- _____ zip lock bags to secure paper swipes after sampling
- _____ pens, note pad, chalk
- _____ D&D forms
- _____ paper swipes
- masking tape and reinforced tape
- _____ scissors
- tool kit
- 3. Pre-Survey Activities
- _____ Review room layouts and floor diagrams for survey units to be monitored.
- _____ Verify possession of room diagrams and generic grid diagrams.
- _____ Select a list of random numbers for each survey unit that is classified as Impacted Class 3 (known RAM-use area).
- Use a grid map of the size closely fitting the size of the survey unit, mark the locations of the random measurements. Use the digits of each random number to
 - simulate a position on the grid map.

- As soon as samples have been analyzed, assemble a D&D information package for submission to the Project Manager/Technical Writer. Your D&D information package should include:
 - (a) completed action checklist,
 - (b) floor plans and grid maps with locations of sample annotated on the grid,
 - (c) a completed summary sheet for the completed survey units,
 - (d) list or order of swipe sample loading and LSC data printouts, and
 - (e) a summary of deviations from procedures, problems encountered, and remediation actions, and other information (if applicable).
- Pre-label swipes for field sample collection; use a labeling technique that ensures the ability to track all samples throughout the collection and processing periods. Place the swipes for a particular area or room in a sealable plastic bag or fold the swipes into a latex glove after the swipe samples have been taken.
- 4. Survey and Monitoring Activities

Prepare the survey unit (room, office, or other area)

- (a) remedy safety or other items of concern that may impact your work,
- (b) if necessary remove excess materials,
- (c) if necessary, use masking tape, chalk lines or other means to mark grid lines within the survey unit

- Perform background data collection for each instrument that will be used; choose a non-radiation area where construction materials and environment is similar to the selected survey units; additional information:
 - (a) For scanning instruments, record the meters' responses after a 1-minute stability period; ensure proper battery check.
 - (b) For gamma and liquid scintillation counter background data collection, collect 21 samples for each counter, building and/or floor if necessary.
 - (c) For scaler counts (gas proportional or NaI), record a select number of 1-minute background counts and desired-location counts
 - (d) Perform background measurements on the day of selected measurements.
 - (e) The same background data can be used for multiple survey units.

Perform scanning operations using appropriate instruments; hold the probe at a one to three inch distance from the floor or wall; scan at a rate of approximately one foot per second; annotate all "hot" spots on the survey grid map by recording the maximum observed reading. Record any readings other than background. Appendix C-2. CHECKLIST FOR IMPACTED CLASS 3 AREAS OR SURVEY UNITS (known radiation use or storage locations)

Team Leader, Chief of Field Operations

- Prepare grid floor maps and select random swipe locations (using sets of random numbers) and additional (control) swipe locations (as needed).
- 2. Grid the room (survey unit) and perform the swipes at the designated locations.
- 3. Oversee operations and monitor the day's progress.
- Deliver swipe samples and survey summary sheets to the Chief of Office and Laboratory Operations for processing.

TEAM MEMBERS

- 1. Prepare the room or area for surveying. Remove debris and excess materials. Practice safety first. Remedy unsafe conditions before beginning survey efforts.
- Label the background swipes -> B1 B21. Perform the reference background swipes for each floor, if necessary. A set of background swipes is required for each floor and each day that swipes for impacted areas are collected. Collect the reference background samples from non-impacted areas in proximity to survey units to be surveyed.
- 3. Label the standard control swipes -> C1 C5. Perform the standard controls: C1-floor inside room at door; C-2-floor outside room at door, C3-inside doorknob, C4outside doorknob, and C5-inside light switch at door. Use additional controls as needed to ensure good swiping coverage in the survey unit.
- Prepare a sketch of the room with door(s) and draw in lab benches or other built-ins.
- Number the benches and proceed to perform additional control swipes -> staring with C6, C7, etc. or the consecutive C- number.

- 6. Using a survey meter, beta-gamma scan all drawers and shelves.
- 7. Using a survey meter, beta-gamma scan 50-100% of floor and wall areas.
- Document the room number for the above actions; include date, serial numbers of meters, room number, and the total number of control (C1-C5, or other). Annotate quality control concerns or issues on survey summary sheets.
- 9. Verify the quantity and sequence of the swipes. Place the swipes in a latex glove and then insert into a labeled plastic bag when transporting samples. Deliver the swipes and other information to the Laboratory and Office Team leader for processing.

Appendix C-3. Checklist for Non-Impacted Areas or Survey Units (radiation-free labs, offices, lavatories, storage areas, closets, hallways, stairways, and other common areas and rooms)

Team Leader, Chief of Field Operations

- 1. Oversee operations and monitor the day's progress.
- Deliver swipe samples and survey summary sheets to the Chief of Office and Laboratory Operations for processing.

Team Members

- 1. Prepare the room or area for surveying. Remove debris and excess materials. Practice safety first.
- 2. If impacted areas will be surveyed, label background swipes -> B1 - B21. Perform the reference background swipes for each floor, if necessary. Collect the reference background samples from non-impacted areas in proximity to survey units to be surveyed.
- 3. Label the standard control swipes -> C1 C5. Perform the standard controls: C1-floor inside room at door; C-2-floor outside room at door, C3-inside door knob, C4outside doorknob, and C5-inside light switch at door. Use additional controls as needed to ensure good swiping coverage in the survey unit.
- 4. Using a survey meter, beta-gamma scan 10-50% of floor and wall areas.
- Document the room number for the above actions; include date, serial numbers of meters, room number, and the total number of control (C1-C5, or other). Annotate quality control concerns or issues on survey summary sheets.
- 6. Verify the quantity and sequence of the swipes. Place the swipes in a latex glove and then insert into a labeled plastic bag when transporting samples. Deliver the swipes and other information to the Office and Laboratory Team leader for processing.
Appendix C-4. Sample Collection Procedures

Team Leader, Chief of Office and Laboratory Operations

- 1. Perform daily quality control equipment and instrument checks before initiating the day's field and laboratory operations. Record appropriate data. Investigate and document observations that are outside the norm.
- Ensure that adequate supplies are on hand, including swipes, liquid scintillation counter (LSC) fluid, batteries, calibrated meters and instruments, and other materials.
- 3. Coordinate the training of personnel who will assist in sample-analysis operations
- 4. Coordinate the servicing of equipment and instruments.
- 5. Audit swipe-sample packages and survey-summary sheets received from the Chief of Field Operations. Resolve inconsistencies or discrepancies before analyzing the samples using the laboratory liquid scintillation counter (LSC) or gamma counter.
- Prepare the samples for evaluation by loading 6. individual swipes into 7-milli-liter glass vials and topping off the vials with biodegradable liquid scintillation fluid or appropriate vials for gamma counting without topping off with fluid. Label the caps of each vial by assigning sequential numbers. Load the vials into racks into the LSC or gamma counter. For LSC evaluations, to prevent elevated results due to chemi-luminescence, store the samples for two to three hours out of ambient light before counting the samples. Chemi-luminescence may be attributed to the swipe or its contents interacting with the LSC fluid. For LSC evaluations, select and initiate the dual-dpm (H-3 and C-14) protocol for evaluating the swipe samples. For gamma counting samples, follow procedures provided in the vendor's operating manual.
- 7. Monitor the progress of the LSC operations during the swipe evaluations.

- 8. Remove the computer printout of the swipe evaluations. For each survey unit or data set, make appropriate annotations on the computer printout.
- 9. Review the data and the survey summary sheets to verify that they match. Resolve discrepancies.
- 10. Report all questionable observations, discrepancies and deviations from procedures.
- 11. Submit the LSC printout, grid maps, and survey summaries to the Quality Control Officer for review.
- 12. Assist the Quality Control Officer as needed.

Appendix C-5. Sample Control. Sample control procedures were implemented to ensure the integrity of samples, sample identification, collection, chain of custody, and retention.

- 1. Samples collected were traceable from the time collected through the analysis phase, during the retention period, and until they were disposed of properly.
 - Sample Custody: All samples collected were maintained under secure conditions. There were fewer than five individuals in the chain-of-custody.
 - Transfer of Custody and Shipment: Samples collected were delivered directly to the laboratory operations office.
 - Sample Packaging: Samples were collected, labeled, and placed in bags at the point of generation by trained personnel. These bags were brought directly to the laboratory operations office. The bags were opened and inventoried by the chief of laboratory operations before evaluation.
- 2. Samples were retained in storage for as long as feasible to support data analysis. Samples, when analyzed, were retained in secure storage until after final report review and then disposed.

Appendix C-6. Data Analysis Procedures

- Survey information packets were delivered to the Quality Control Officer for review and incorporation into the final status report. The survey information packets consisted of computer printouts of liquid scintillation (LSC) counter or gamma counter results, grid maps (for impacted survey units), survey summary sheets, and other supportive information.
- 2. The number of samples to be collected for impacted class 3 areas were determined using guidance provided in MARSSIM and analysis results for a known contaminated area. The number of random, statistical data samples collected for each impacted area was 21. An exception to the 21 samples taken for an impacted area was when 100% of the grid locations were sampled. For each impacted area surveyed, 21 reference background samples were also collected. Additional swipes, called control swipes, were collected to ensure good coverage of sampling within a survey unit and to sample selected locations.
- 3. At least five control samples were collected for each impacted and non-impacted survey unit. These five control swipes were collected using a pre-selected set of locations within each survey unit.
- For each impacted survey unit, applicable survey data was 4. assembled and analyzed. This data consisted of reference background data, direct surface readings, scanning results, and LSC swipe evaluations. The direct surface readings required conversion from cpm values to readings in dpm per 100 cm^2 . After the data conversion of the direct readings, all survey data was then compared directly to the NMRC direct-reading action limit of 2,000 dpm. The NMRC action limit represented a value that was a small fraction of the NMRC screening values or DCGLs. Scanning results were called satisfactory when results did not exceed two to three times the background readings. Any individual reading that exceeded the NMRC action limit or the scanning limit was further investigated. For any area exceeding the limits, remediation efforts were completed until subsequent surveys yielded results that were below NMRC action limits. Additional scanning and swipes were completed as deemed necessary. All results have been included in this final report.

- 5. For non-impacted survey units (common areas, offices, storage rooms, halls, and stairs), control swipes for selected areas were collected and analyzed. In addition, 10-50% of the floor area was beta-gamma scanned.
- 6. All original computer printouts, grid maps, survey summary sheets and applicable documents have been maintained for inclusion in permanent, radiation-safety-program files.
- 7. Grid maps with applicable sample locations have been included in this final report.
- All data entries in the final report were verified (for accuracy and completeness) by at least two persons on the D&D team.

Appendix C-7. Equipment and Instrument Procedures

- All equipment and instruments designated for use in NMRC radiological D&D operations were properly calibrated and kept in a secure location when not in use. Operators were trained in the proper use of the equipment and instruments.
- 2. Daily operational and check-source tests were performed and documented by trained personnel.
- 3. Radiation equipment and instruments were used for direct surveys, scanning, and sample analyses.
 - a. FIXED CONTAMINATION MEASUREMENTS. Direct surveys or measurements for fixed contamination were performed using a portable gas proportional detector. Prior to its daily quality control tests, the detector was purged with gas per the manufacturer's operating procedures. Readings for one-minute samples were collected for desired locations. The primary isotopes of concern were Carbon-14 and Sulfur-35.
 - b. WALK-THROUGH SURVEYS. Scan surveys were performed using portable, Geiger-Mueller (GM) meters with either pancake probes or gamma scintillation probes. The beta-gamma scanning consisted of walk-through surveys carried out by holding the detector probe approximately one to two inches above the floor or area and walking slowly (about one foot per second) through the specific site-stopping when increases in count rates were observed. Areas where the count rates exceeded two to three times background were investigated by a more detailed survey in the suspect area. Scanning practices included using the audible response of the meter for detection and, after stabilization, using the meter response for measurement. The primary isotopes of concern included Carbon-14, Cobalt-57, Barium-133 and Cesium-137.
 - c. <u>REMOVABLE SURFACE CONTAMINATION SURVEYS</u> (Swipes/Smears and Swabs). Swipes (swipes, smears and swabs) were collected as deemed necessary to determine if radioactivity above the limit was removable (or loose). Swipes were performed by lightly misting the area with a water spray and wiping a 100 cm² surface (applying moderate pressure) with a dry filter paper (swipe/smear/swab). Swipes were evaluated using a liquid scintillation or gamma counter, depending upon the isotopes of concern for the survey unit. For

liquid-scintillation-counter evaluations, tritium (H-3) and Carbon-14 were the principal isotopes of concern.

4. Information about the D&D radiological monitoring equipment and instruments is provided in Appendix D. Documents summarizing instrument calibrations and specifications are provided in Appendix D. Appendix C-8. Source-Standards, Background-Standards, and Reference-Background Measurements

- All available source standards (Hydrogen-3, Carbon-14, and Iodine-129) and background standards were collected for use during the duration of the radiological decontamination and decommissioning operations.
- 2. NIST-traceable, source standards (Carbon-14 and Cobalt-60) were purchased and kept stored in a secure location. Only trained and authorized personnel handled the check sources and used the sources to verify proper instrument response to radioactivity.
- 3. Daily-check-source and operational calibration checks were performed. Document results, maintaining logs and control charts as necessary.
- 4. Inventories and leak tests of check sources were performed, at least semi-annually. The leak tests and documentation satisfied regulatory requirements.
- 5. Information about the various check sources used during the D&D effort is provided in Appendix D.

APPENDIX D. QUALITY ASSURANCE INFORMATION AND DATA

Appendix D-1. Training Information and Certificates

All personnel who participated in radiological decontamination and decommissioning (D&D) efforts received a level of training commensurate with the work assignments and level of participation. The incumbent Radiation Safety Officer, LCDR S. L. Gaiter, held the primary D&D positions of Decommissioning Officer and Quality Control Officer. The Chief of Office and Laboratory Operations was the incumbent Assistant Radiation Safety Officer, Mr. Saifeddin Ahmad. The Chief of Field Operations, Mr. Willie Deadwyler, had previously held numerous positions as a Navy radiation health technician. The personnel listed above received basic radiation safety training and instrument training in conjunction with their participation in NMRC D&D efforts. Mr. Glen Garth and Mr. Rodney McGhee received basic radiation safety training and on-the-job training. Mr. Garth and Mr. McGhee worked under the direct supervision of Mr. Ahmad or Mr. Deadwyler. Ms. Carolyn Hall served as the Project Coordinator and Editor. Her assignments were completed in an office environment or infrequent visits to the field. LCDR Gaiter or Mr. Ahmad performed all radiological materialshandling and decontamination work.

Document D-1 provides training and experience information for LCDR Gaiter. Documents D-2 through D-11 provide pertinent radiation safety training certificates for the participants in the NMRC radiological D&D efforts. Table D-1 lists the training documents included in this final report.

Table D-1	. TRAINING INFORMATION AND CERTIFICATES
Document	Training Certificate or Form
Number	
D-1-1	Training and Experience of Authorized Radioisotope User,
200	LCDR Gaiter
D-1-2	Radiation Safety Training, LCDR Gaiter
D-1-3	Ludlum Measurements Instruments, LCDR Gaiter
D-1-4	Radiation Safety Training, Mr. Ahmad
D-1-5	Ludlum Measurements Instruments, Mr. Ahmad
D-1-6	Radiation Safety Training, Mr. Deadwyler
D-1-7	Ludlum Measurements Instruments, Mr. Deadwyler
D-1-8	Radiation Safety Training, Ms. Hall
D-1-9	Ludlum Measurements Instruments, Ms. Hall
D-1-10	Radiation Safety Training, Mr. Garth
D-1-11	Radiation Safety Training, Mr. McGhee

	OCUMENT D-1-1. TRAINING AND EX OF AUTHORIZED RADIOISOTOPE (TYPE OR PRINT LEGIBLY)	(PERIENC USERS	E	
	A GENERAL INFORMATION	N		
AUTHORIZATION NUMBER:		DATE O	F APPLICATION: 31 Ma	rch 1999
NAME OF AUTHORIZED USER (LAST, FIRST, MI)		RANK/C	GS GRADE:	
Gaiter, Schleurious L.		T.CDR, M	ISC,USN	
ORGANIZATION / ORGANIZATION DIVISION:		BUILDIN	IG AND ROOM NUMBE	R: All NMRC Bethesda and
Naval Medical Research Center, Safety Program		Rockville	Buildings	
	B. CERTIFICATIONS N/A			
SPECIALTY BOARD	CATH	GORY	0	ATE CERTIFIED
American Academy of Health Physics	Part I of II (Associate Member st	atus)	7/	1995
	C. FORMAL EDUCATION			
TYPES OF LICENSES : None				
TERRITORY OR STATE IN WILCH LICENSED M	D, DDS. DVM, ETC.): N/A			
HIGHER EDUCATIONAL, INSTITUTIONS	MAJOR COURSE OF STUDY		DEGREE, DIPLO	MA OR CERIFICATE
ATTENDED	AND DATES ATTENDED		RECEIVED	AND DATE
Florida Community College (FCC), Jacksonville, FL	Pro-Engineering	Associate in	n Arts, 4/1978	
University of Florida (UF), Gainesville, FL	Chemical Engineering	B.S. Chem	ical longmeening, 12/1981	
Georgestuum University (GLD Washington, DC	Padiation Science	Manuer of	Science 7/1985	
	PECEIVED IN BASIC PADIOISOTOPE I		TECHNIOLES	
	NACIAVED IN DABIC NODOBOTORE I		TYPE AND LEN	GTU OF TRAINING
	LOCATION AND DATES OF TRAIN	NING .		
FIELD OF TRAINING	ANCLUDE COURSE TITLE JF KNO	WN)	LECTURE/	SUPERVISED
A	В		LABORATORY	LABORATORY
			COURSES	EXPERIENCE
1			(IIOURS)	(IJOURS)
			С	D
	Navy Radiation Health Officer Course, 1988	3	160 hours	
1. RADIATION PHYSICS AND	Radiologic Physics I, GU, 1991		30 hours	
INSTRUMENTATION	Health Physics, GU, 1992		30 hours	
	Radiation Detection, GU, 1993		20 hours lectures; 20 hou	rs ab experience
	Radiation Dosimetry, GU, 1993		20 hours lectures; 20 hou	rs lab experience
	Health Physicist/Instructor, Armed Forces	at hand a	work experience (1990-1	990)
	Radiobiology Research Institute (AFICIC), B	etnesda,	work ormanianas (1905)	200
	MD, Padiation Safaty Officer NMPC Batharda		work experience (1995-1	,
	Name Radiation Health Officer Course 1988	2	160 hours	-
2 RADIATION PROTECTION	Operational Realth Physics GU 1992	,	10 hours	
	Radiation Biology, GU, 1993		30 hours	1
	Health Physicist/Instructor, AFRRI		work experience (1990-1	9951
	Radiation Safety Officer, NMRC, Bethesda		work experience (1995-1	999)
	Navy Radiation Health Officer Course, 1988	}	160 hours	
3. MATTIEMATICS PERTAINING TO	Radiologic Physics I, GU, 1991		30 hours	
THE USE AND MEASUREMENT	Ilcalth Physics, GU, 1992		30 hours	
OF RADIOACTIVITY	Health Physics Internations 1 and 11		150 hours laboratory exp	erience
	Radiation Detection, GU, 1993		20 hours lectures; 20 hou	rs lab experience
	Radiation Dosimetry, GU, 1993		20 hours lectures; 20 hou	rs lab experience
	Ilealth Physicist/Instructor, AFRRI		work experience (1990-1)	995)
	Radiation Sofety Officer, NMRC, Bethesda		work experience (1995-1	999)
	Radiation Biology, GU, 1993	_	30 hours	
	AEDPL 1000-1005	s,	SOU TOULS	
	λι/λ		NIA	·
5 RADIOPHARMACEUTICAL	19/74		11/21	
CHEMISTRY				

PAGE 1

	Ē.	ACTUAL EXPE	RJENCE WITH RADIOISOTOPE	S (SEALE	D AND/OR UNSEALED S	OURCE)		
ISOTOPE	MAXIM	UM AMOUNT	WHERE EXPERIENCE GAINED	WAS	DURATION OF EXPERI	ENCE	TYPE OF USE		
Cobalt-60 (soaled) Cosium-137 (sealed)	150,000 (14,500 C	Ci i	AFRRI. Bethesda, MD NMRC, Bethesda, MD		1990-1995 1995-1999		Open-air irradiations Closed-chamber irradiations; provided training to users		
H-3 C-14 P-32 P-33 Cu-45 Cr-51 I-125	3 50 Ci 14 10 Ci 2 10 Ci 3 25 mCi 45 10 mCi 51 5 Ci 25 1 Ci		NMRC, Bethesda, MD		1995-1999	Receipt, handling, monitoring and disposal of unscaled materials for use by Authorized Users performing biomedical research efforts			
F. EXPERIENCE WITH RADIATION PRODUCING DEVICE					RAY MACHINES, IRRADIAI	ORS, ET	Υ.)		
DEVICES		WHERE EX	APERIENCE WAS GAINED	DURA	TION OF EXPERIENCE		TYPE OF USE		
Gamma Irradiators (Cesium-137) Shepherd Mark I (9000 Ci)		a, MD	1995-199	99	Train personnel and maintain units for use by staff scientist, technicians, and other researchers to irradiate specimens, non-human animals, cells, etc.				

G. CERTIFICATION

I CERTIFY THAT THE ABOVE INFORMATION PROVIDED HEREON IS TRUE AND COMPLETE TO THE BEST OF MY KNOWLEDGE

SIGNATURE OF APPLICANT

Best Industries Model (2500 Ci) Isomodix Gammator Unit (300

DATE SIGNED

· · -

PAGE

2

Ci)

TRAININING AND EXPERIENCE OF AUTHORIZED RADIOISOTOPE USERS 的 医白豆状的 网络

This Certifies That

Schleurious Gaiter

has successfully completed the 1-day course of instruction for

Radiation Safety Training

February 23, 2000

Presented By

CSI-Radiation Safety 481 North Frederick Avenue, Suite 302 Gaithersburg, Maryland 20877

For

Naval Medical Research Center

Bethesda, MD

Ray Johnson

Raymond Johnson, C.H.P., P.E. Training Director





This Certifies That

Saifeddin Ahmad

has successfully completed the 1-day course of instruction for

Radiation Safety Training

February 23, 2000

Presented By

CSI-Radiation Safety 481 North Frederick Avenue, Suite 302 Gaithersburg, Maryland 20877

For

Naval Medical Research Center

Bethesda, MD

Cart Johnson

Raymond Johnson, C.H.P., P.E. Training Director





This Certifies That

Willie Deadwyler

has successfully completed the 16-hour course of instruction for

Fundamentals of Radiation Safety

November 9-10, 1999

Presented By

CSI-Radiation Safety Training 481 North Frederick Avenue. Suite 302 Gaithersburg, Maryland 20877

in association with

Radiation Service Organization. Inc. Post Office Box 1526 Laurel, Maryland 20725

Raymood Jønson, C.H.P., P.E. Training Director





This Certifies That

Carolyn Hall

has successfully completed the 16-hour course of instruction for

Fundamentals of Radiation Safety

November 9-10, 1999

Presented By

CSI-Radiation Safety Training 481 North Frederick Avenue, Suite 302 Gaithersburg, Maryland 20877

in association with

Radiation Service Organization. Inc. Post Office Box 1526 Laurel. Maryland 20725

Raymond Jøhnson, C.H.P., P.E.

Training Director





This Certifies That

Glen Garth

has successfully completed the 1-day course of instruction for

Radiation Safety Training

February 23, 2000

Presented By

CSI-Radiation Safety 481 North Frederick Avenue, Suite 302 Gaithersburg, Maryland 20877

For

Naval Medical Research Center

Bethesda, MD

Ray Johnson

Raymond Johnson, C.H.P., P.E. Training Director



This Certifies That

Rodney McGhee

has successfully completed the 1-day course of instruction for

Radiation Safety Training

February 23, 2000

Presented By

CSI-Radiation Safety 481 North Frederick Avenue, Suite 302 Gaithersburg, Maryland 20877

For

Naval Medical Research Center

Bethesda, MD

Ray Johnson

Raymond Johnson, C.H.P., P.E. Training Director



Appendix D-2. Service and Calibration Certificates for Instruments, Equipment, and Check Sources

Table D-2 lists the service and calibration certificates for the instruments, equipment and check sources used in the D&D effort described in this final report.

Table D-2	. SERVICE AND CALIBRATION CERTIFICATES FOR INSTRUMENTS, EQUIPMENT AND CHECK SOURCES
Document Number	Service and/or Calibration Certificate
D-2-1	Packard Liquid Scintillation Counter, S/N 403848, 4/5/99
D-2-2	Packard Liquid Scintillation Counter, S/N 403848, 3/10/00
D-2-3	Packard Liquid Scintillation Counter, S/N 403848, 4/6/00
D-2-4	Statement of Data Analysis by Packard, 4/7/00
D-2-5	EG&G Wallac, Inc. Gamma Counter, S/N 82-324, 8/23/99
D-2-6	Bicron Meter, S/N B602G, 7/22/99
D-2-7	Ludlum Model 3, S/N 54537, 7/22/99
D-2-8	Ludlum Model 3, S/N 114886, 7/22/99
D-2-9	Ludlum Model 3, S/N 121461, 8/12/99
D-2-10	Ludlum Model 3, S/N 121816, 2/22/00
D-2-11	Ludlum Model 3, 5/N 132946, 2/22/00
D-2-12	Certificate of Participation, Reports of Traceability to
	NIST, Analytics, Inc., 1999
D-2-13	Certificate of Calibration, Analytics, Inc., Carbon-14
	check source, S/N 58806A-530
D-2-14	Certificate of Calibration, Analytics, Inc., Carbon-14
	check source, S/N 58807-530
D-2-15	Certificate of Calibration, Analytics, Inc., Cobalt-60
	check source, S/N 58808-530
D-2-16	Certificate of Calibration, Analytics, Inc., Cobalt-60
	check source, S/N 58809-530
D-2-17	Certificate of Calibration, Ludlum Measurements, Inc.,
	Model 2350-1, S/N 157674, Gas Proportional Unit with
	Detectors
D-2-18	Certificate of Calibration, Ludlum Measurements, Inc.,
	Model 2350-1, S/N 157669, Gas Proportional Unit with
	Detectors

Par 'ard CUSTOMER NAME		CHELL - HITEL	290-0		~)	7					
RUL TO:						ISTOKE					
BILL TO.	1/1/2/17		MALI	ر آگا		I			1151	12	7
DEPARTMENT	-1-1-2-		EM SERIAL NU			DATE					-1
Decentra Control Pertino Philos	R1215	TOPII	-	572	PH4	MOL	DAY YEA	ROUR	is	HOUR	15
INSTITUTION'S, NAME		RIC SYSTEM	SERIAL NUME	<u>Z I – –</u> BER		74	< 99		0	7	10
Maral Medical Research (enter		-	- - -	al		- I					
ADDRESS		CALI	NUMBER								
8901 Wisconsin Avenue	BRA	331		1							
GITY STATE COUNTRY ZIP		· cc	PY TO:							_	
Bethesdg, MD 20889-5607						TOTAL	S		0	Z	2
		דואט	TOTAL		S	SUMMAR	Y OF CHAR	GES TO CUS	TOMER		
# REPLACED ITEM		PAICE	PRICE		ITEM		HOURS	RATES	CH	ARGES	;
1 16550 75 Printer ribbon cartridge		27 00	<u>27 þ</u>	\hat{O}	ZONE	MIN.	10	350,00	<u>کک ا</u>	50	20
	·				LABOR		t_				
					SUB TOTAL				<u></u>	50	20
				_	MATÉRIAL					/ Ç	20
	<u> </u>	<u> </u>			SHIPPING CH	ARGES					
					VENDOR CHA	HGES			7	7 1	00
	<u> </u>					_					50
9					TAX						
10				- -	MISC.						
PARTS TRANSFER ITEM PAYEE #	TOTAL	MATERIAL	270	0	TOTAL				37	77	00
EXPLANATION: THETTIMENT CALIFICATION	DIA CIL	NR SA	MATAT	în	CUSTOMER'S	P.O. (IE A	AISSING EXP	LAIN		C	
WARITED OUD (P Alarmalize)	-lihm	Ter	Land	<u> </u>	CUSTOMED SI		TK	$-\frac{1}{3}$	\underline{c}	_/_	
Non Provice, Tronsinger	411019	Lo. C	1CATIC			Scrif	ddi.	in Ah	mod	Ś	
and inducates hunder, ighigh		Doon, (Jane	2	TSESIGNATUE	HEF/	$V_{\sim}V_{\sim}$				
tan tilter, checkee cooling	TONL	, Islen	1 disi		_/Cove	EASON FOR	A SERVICE			, , ,	
From Keyboard & detector P.	STOTIC	CONTR	ler_		PRIMARY		SECOND	DARY	ITPE OF	- SERV	ICE
Cleanez, ranne changer deck	C+ M	onitor	face.		DOPERATOR EAROR		INCORRECT OF	PERATION	11041		
Adjusted down Alate Tellion	PTOIT	or RINZ	Th		2 MECHANICALI ELECTROMECHANCI	AL	2 AOJUST 2 REPAIR	EM SPECS	SU WAARAANIY SI ENG CHANGE! 52 TRAVING/API	S PLICATION S	uppcai
ACTIVICE OF A DISIC TODOR D		Tin			J ANALOG DIGITAL		A REPLACE		T21222A 23 LAT 23	DEMO / SI	ŵ.
Improve low the changer open	9110	$\sum = E \sum \{F \in \mathcal{F} \}$	ELEC		6 UPDATE 7 NO PROBLEM FOUND	D	7 DIGITAL 8 SOFT & FIRMW	ARE	COMPACIEISI COMPACIEISI COMPACIEISI	WE AGREEM	ENT (2)
GN& JUbriCale	E CX	ernal.	<u>115 </u>	_			9 OTHER (PLEAS)	EEXPLAIN	PHA CHILY AGRE	GREEMENT	
HEFF CLOITOBKG. ICPM (hecked and	- rese	t envi	-109:						IS ALE	ANTENANCE	E
14C EFF. 9642 BKG23 CPM ADDOX 3,319	deva	stor cy	der								
1123 EFF BKG SINCE 27 AC	r 9£						7			6	8

Pac Ind CUSTOMER NAME:												
APacard DoScurice y D. /-IIIMAD		30 2	15-0002		<u> </u>	ECEI	<u>PT FC</u>	DF . '	RVI	CE		
BILL TO: RECEIPT LEXTRUL SECTION	DIST	TERA KOME	PAYEE NO). cus	STOMER I			SERVIC	ORDER	Ħ		
DEPARTMENT SAMELICE AVE	722		M SERIAL NUM			SERVICE				6		
BLOG 17, RM. 132 BETTESDA AD 20839-5600	BIZIS		1-149	3181418	MOLD		B HOU	AS	HOUR	S		
INSTITUTION'S NAME		RIC SYSTEM	SERIAL NUMBE	R	21	6 00	···	50	2	15		
NAVAL MEDICAL RESEARCH CTR.		- -	9									
ADDRESS BEOIL WISCOPSIN ANE		CALL	NUMBER									
KAPINTION SFTY. ULDG 11 KOUM 10	L			<u></u>	_ _					-		
BETHERAA MID 20884-51	37					<u> </u>						
							SES TO CU			i		
REPLACED ITEM PART DESCRIPTION	QUANTITY	PRICE	PRICE	ITEM		HOURS	BATES		HARGES			
1					AIN.	10	390		390	00		
2				LABOR		50	195		975	2		
3			· ·	SUB TOTAL				1.	365	00		
			<u> </u>	VENDOR CHAR	GES							
7				SUB TOTAL		{						
В				TRAVEL								
9								ļ				
				MISC.				<u> </u>	200	-		
EXPLANATION: The OF A COMPANY O			1	CUSTOMER'S P.	D. (IF MIS	SING EXP		L_/.	<u>3651</u>			
CUENTED COUNTER, CLEADE	EDUE	LIELEYC M	ISSEMBL	N30	2398	OO KO	<u> 726</u>	23		-		
(IFANEL) AND LUBILICATED SAMPLE (HANGE	K. C.ALI	3 RATED	CUSTOMER SIGI	NATURE	Saifed	Lin Ab	md	3/3/00			
HIGH VOLTACE (HACKED 317 = (-14	FFFIC	<u>LENCIE</u>	5	TSE SIGNATURE	180-	+				_		
NOTE:				A JOG J T/C	SON FOR SE	AVICE				CE		
ALTHOUGH THISECUNTER IS NOT YO	2K Co.	MPLIAN	T (CORREC		_	SECOND	ARY)					
PATE CANNOT BE ENTERED. J. THE CO	DATE CARNOF BE ENTERED. THE CORRECTED DAM VALUES ODERATOR EAROR ODERATION ON WARMANY											
FOR 311 AND CIL ARE CORRECT.	HE 3	HAND (.14	2 MECHANICALI ELECTROMECHANCIAL JANALOG	- 35	NDJUST REPAIR REPLACE		SI EHG CHAN S2 TRAINING / S5 INSTALLATIO	GES Application sua Ni	IROPP		
QUENCY CURVES INSTALLED AREL	207 DO	FRENDAN	NT GPel	S SOFTWARE & FIRMWARE	E 67			63 SALES ASSIS	SE / DEMO / SHOW	N 41 (1) 41 (2)		
A DATE BUTARE DEAENDENT UPO,	NTHE	F ACTUA	2 DPM'S	I NO PROBLEM FOUND	0 S 9 C	OFT & FIRMWA	RE EXPLANN	60 DEPOT REP 61 LIMIED USE	AGREEMENT			
THEFF 100,97 CBKG. 16 IN THE QUENIN	5785	AT THE	TIME					68 BILLADLE SE	RVICE			
INCEFF 9/0. 3 TUBKG 20 THE CURVES LU	ERE I	INSTAL	LIED	г		Т		IT SALE				
1123 EFF BKG									68	5		

~		Pa		ar	d		CUSTOMER NAME:			CU	JST P	TEL:	:				7	REC	CEI				uch
BILL	. TO;		-	_					DIST		TEN	#DNF	Г	PAYEE	NO		CUSTOM		BER		5		BA
								4	3	3	32	/	2	31316	4	2				6-	Ab	458	Î
DEP	ARTME	INT			_			1		PA	CKARD	SYSTE	ĒM	SERIAL N	NUME	BER	DATE	OF SE	RVICE	SER	/ICE	TRA	VEL
														1-141	0[3181418	мо	DAY	YEA	R HOU	JAS	нои	IRS
INST	ITUTIC	N'S	MAM	E							RIC SYS	TEM	SE	RIAL NUN	/8EF	a	04_	06	0	2	2Ω)	LX
													Ŀ		9		_		Ļ				_ <u> </u>
	RE55							0	ماس ا		-01	CALL I.I	. N 		ı.	1 1 1					_		
CITY			ST	ATE			COUNTRY ZIP		2 [2	12			P	/ TO:									
																	TOTA	LS	!				
TEM	PAR	IT NU	MBE	RO	F	Γ				1	UNIT		Τ	TOTAL			SUMMA	RY OF (CHARC	SES TO CL	STOME	R	
#	RE	PLAC	ED	тем				QUA			PRICE			PRICE		ITEN	1	HOL	JAS	RATES		CHARGE	S
1		_		_	_	$ \rightarrow$				-						ZONE	<u> MIN.</u>						
2	$\left\{ \right\}$			_	\downarrow	_				Ĺ		<u> </u>	L			LABOR							
3	┢		-		_					<u> </u>			L			SUB TOTAL							ļ
4.	┼-┼-	_ _	\vdash	_	_	_				╞		—				MATERIAL							
5	\vdash	+-		-+	+	-				┢		<u> </u>	┢			SHIPPING C	HARGES						
7	+- -	+	┟──┤	+	- -	+	·			⊢			┝			VENDOR CH	ARGES	+			\rightarrow		
8	\vdash	+		+	-	+		+-		┢╴		—	⊢			TDAVEL							
-					-+-	+				┢─			┢			TAY							
10					╋	1		1					-			MISC					-		1
PARTS	S TRAN	VSFE	1. A ITI	EM_				1	TOTAL	. м/	ATERIAL	<u> </u>				TOTAL							
EXPLA	NATIO	IN:	C	101	102	/	sample changer deck	Å	mar	· ra	Lang	Lc.	1	ched	/	CUSTOMER'	s p.o. (If	MISSIN	G EXP	LAIN)			
PL	174	*		Ces	20	4	1 PMT's in sack	ie fo	<u>; (</u>	4	ected			dasica	Д	CUSTOMER	SIGNATU	RE					```
m	21/24	irn ,	<u> </u>		£		churchar Charon co	$c \epsilon_4$	'a./_,	£	can (<u>?/a</u>	22	tac		TSE SIGNAT	JRE		1/1-				
60	dy		X	ei	Â		Contector ins	c <i>h</i> ru	1120	₁∠́	/	2 A	6	cuical		(k_	AEASON FO	DB SERVIC	fair	ala_		``	
Cii	stor	14	15		\mathcal{A}	\sim	tost & SAIC . /s	25:4	145		lacke	\checkmark	a	and.		PRIMARY	Ø_	S	ECEND	AAY	TYPE	OF SER	IVICE
												` ب 	/	-9		0 OPERATOR ERROF I POWER SUPPLY 7 MECHANICALI ELECTROMECHAN 3 ANALOG 4 DIGITAL	CIAL	0 INCOR 1 DUTSI 2 ADJUS 3 REPAIR 4 REPLA 5 MECHA	IRECT OPI DE SYSTE IT R ICE AMICAL	ERATION M SPECS	41 EDAN So WARRAN SI ENG CH S2 TRAININ SS INSTALL 63 SALES A	ITY KHIGES G / APPLICATION NION SSIST / OEMO / S	I SUPPORI ŠI KOW
						ภ										S SOFTWARE & FIRM & UPDAIE 7 NO PROBLEM FOU	ND	5 ANALO 7 DIGISAI 8 SOFT 8 9 OTHER	NG L & FIAAIWA I IPLEASE	ዋጀ E አድር ላየብ	6C COMPRE 61 COMPRE 60 DEPOT A 61 THATED 64 PM OFAY 68 BILLABLE	DENSIVE AGREE DENSIVE AGREE EPAIR VSE AGREELIEN AGREEMENT SERVICE	2MENT (1) 2MENT (2) 1T
IN EF	<u>r -6</u>	$\frac{\rho}{\frac{1}{2}}$	<u></u>)	uilli.	 								_		_						69 PREVENT	NE MAINTENAN	Cε
C El	FF. <u>74</u>	<u>a</u> (/		BKG.	<u> </u>	2	/								-		7					1	10
1125 E	F.F			BKG.	_	_						_						<u>Э _</u>				0	0



,

.

10

۰.

. • ^

DATE:04/07/2000

FAX# 301-295-5651

TO: Dr. Jaifeddin Ahmad

FROM: TOM GERBER

RE: Packard B2500 LSA

: . .

. · ·

1

I have analyzed your data and have found the data to be in control. The procedure that I used was to take 2 square roots of the mean and look for random events on both sides of the mean. The data should stay within 677 counts 95 % of the time. It passes this test quite well. An assumption that I made is that the CPM value was taken at 1 minute.

.

. .

. •

Thank you, Tom Gerber

A FRAR MALLAC INC

1286 Beller Floi The aburg, Maylend 20877 (800) 888-770 1286 Beller Floi Texamon and the method of the	KEBEB• WALLAC, INC.	SERVICE REPORT
In this his is Single Singl	3238 Galiher Rol althersburg, Maryland 20877 (800) 638-671 USER	
1000000000000000000000000000000000000	Rudiulium Sulety. (301) 2950002	MO. DAY VR. 1156159663 39915
BID 17 Rest 546 BID 17 Res 546 CUN TH	M NAVAL Madrice Research	MODEL AS AN ANTAL AS A COMPLETE A PROBLEM SREPAIR
$ \begin{array}{c c c c c c c c c c c c c c c c c c c $	BID 17 Ron 5-16	1282 820329 800 900 1104 21 21
$\begin{array}{ c c c c c c c c c c c c c c c c c c c$	C STREET 896 Wisconsin Ave	NUMBER NUMBER DI NO ACTION DI ACTION DI NO A
$\begin{array}{ c c c c c c c c c c c c c c c c c c c$	N BETROSDA IM) 20889	CONTRACT RENTAL CONTRACT CONTR
	<u>8</u> Pm	DEPART ARRIVE COMPLETE 20 ELECTRICAL ADJUSTMENT FARURE 30 REBOOT INSTRUMENT 21 ELECTRICAL COMPONENT FAILURE 31 REBOOT COMPUTERS 23 REPLOT COMPUTERS 24 REPLOT COMPUTERS 25 REPLOT SOFTWARE
$\begin{array}{c c c c c c c c c c c c c c c c c c c $	WILL NOT INITIALIZE!	31 SOFTWARE FAILURE INTERNAL 40 REPLACE ASSAY 40 READENT FAILURE 41 REPLACE READENT
$\frac{\left \begin{array}{c} \left(\begin{array}{c} \left(\right(\left(\begin{array}{c} \left(\begin{array}{c} \left(\begin{array}{c} \left(\begin{array}{c} \left(\right(\left(\begin{array}{c} \left(\left(\begin{array}{c} \left(\right(\left(\begin{array}{c} \left(\left(\begin{array}{c} \left(\left(\begin{array}{c} \left(\right(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left(\left$	E (RAD CPM Bd-ECC)	PARTS USED
$\frac{1}{2} = \frac{1}{2} = \frac{1}$	M	QUAN!
$\begin{array}{c c c c c c c c c c c c c c c c c c c $		1 ECC Bd CPU 10847527 605 00 605 00
$\begin{array}{c c c c c c c c c c c c c c c c c c c $	E Replaced ECCRd BAID Power	1 Power Switch 1/84/5021 12:00 12:00
$\begin{array}{c c c c c c c c c c c c c c c c c c c $	Switch, charked Aliquments,	
$\begin{array}{c c c c c c c c c c c c c c c c c c c $	CONVOYUR, TESTA (LeAnd	
$\begin{array}{c c c c c c c c c c c c c c c c c c c $	i countine chanles cherlest	
$\frac{B \operatorname{Art} g \operatorname{\mu ond} = 92}{\operatorname{SOFTWARE VERSION: } \operatorname{\mu (TMAT P TE V) S}}$ $\frac{\operatorname{SOFTWARE VERSION: } \operatorname{\mu (TMAT P TE V) S}}{\operatorname{DVM}}$ $\frac{\operatorname{I/O} \operatorname{DeVICES AND BETA PLATE ACCESSORIES}}{\operatorname{I/O} \operatorname{DeVICES AND BETA PLATE ACCESSORIES}}$ $\frac{\operatorname{I/O} \operatorname{DeVICES AND BETA PLATE ACCESSORIES}}{\operatorname{I/O} \operatorname{DAL LABOR}}$ $\frac{\operatorname{S}}{\operatorname{I/O} \operatorname{LABOR}}$ $\frac{\operatorname{S}}{\operatorname{I/O} \operatorname{LABOR}}$ $\frac{\operatorname{S}}{\operatorname{I/O} \operatorname{LABOR}}$ $\frac{\operatorname{S}}{\operatorname{I/O} \operatorname{LABOR}}$ $\frac{\operatorname{S}}{\operatorname{I/O} \operatorname{I/O} I/O$	N EFFICIENCY F129: 64.57-	
SOFTWARE VERSION: $UITATAT TE VIS$ DVM	R BArighond = 42	
SOFTWARE VERSION: (LILTATIM TEULS) DVM	E	
SOFTWARE VERSION: (1/1/1/17/17/17/17/17/17/17/17/17/17/17/1		
I/O DEVICES AND BETA PLATE ACCESSORIES I/O DEVICES ACCESSORIES I/O DEVICE	SOFTWARE VERSION: <u>LICTATOR</u> THE U.I.S	
PART NUMBER S/N LABOR HOURS Z @ \$ TOTAL LABOR \$ ('OIM PHTM II SEX OTHER CHARGES (Description) \$ \$ \$ \$ \$ \$		TOTAL PARTS \$: 6/7.00
('OINPUTM USER OTHER CHARGES (Description) TRAVEL ZONE CHARGE \$ 111 000 / TG1 USER USER Image: Charge in the second sec	PART NUMBER S/N	LABOR HOURS 2 @\$ 175 TOTAL LABOR \$ 3500
$\frac{111001721}{11000}$ $\frac{111000}{11000}$ $\frac{111000}{11000}$ $\frac{111000}{11000}$ $\frac{111000}{11000}$ $\frac{111000}{11000}$ $\frac{111000}{11000}$ $\frac{111000}{11000}$	(OMPHTM USER	OTHER CHARGES (Description)
TOTAL CHARGE \$ 11420	111 avitan USER.	TRAVEL ZONE / CHARGE \$ 175,00
ALL WORK HAS BEEN DONE TO MY		TOTAL CHARGE ► \$ 114 200
SATISFACTION AND THE EQUIPMENT		ALL WORK HAS BEEN DONE TO MY SATISFACTION AND THE EQUIPMENT
IS IN ACCEPTABLE WORKING IS IN ACCEPTABL		IS IN ACCEPTABLE WORKING CONDITION UNLESS CALL IS MARKED INCOMPLETE

.

.

Gamma Preventive Maintenance

¢,

يري				1		- //	1	
Wallac Inc.	Date \$.23-99	Engine	er	HAR	RI (n Ll.	NS	
9238 Gaither Road	dob # //	5615		Contract	#	Non	2	
Gaithersburg, MD 20877	Model 12	87		Serial #	8	$\sum_{i=1}^{n}$	324	
(800) 638-6710	Instrument Softwa		00 (1)	TLATIN	- n	FT_	VerIS	
	inistrament optiva		011 <u>72</u>		. ::			
		уешот сору с	of this PM Jorn to	и гресским ри				
BEFORE PM		AFTER	RPM	0/ 1100	SP	ECTRU	MPLOI	
I-129 Efficiency (27) Backareuad		iency _	- <u>640</u>		37			0
								9
MECHANICAL TEST	TS TS		ELEC	TRICAL TH	ESTS			
Counterweights		(Pa-	Sensors		-	-		A
Tube Shields		A	Display/Keyb	oard				Ø
Elevator Tips/Forks		DEL	ID Decoder					X
Conveyor Belts	· ·		Power Line				1 W vac	হ
Chains	Teeth	F	Low Volt DC		÷/- 5	+/- 12		<u>م</u>
Clutch	es	¢ ا	Battery	-				য
Motors	Brakes	لعر العر	Master Clear				15,75 vdc	A
Lubrication Elevate	or	<u>A</u>	Power Up				9.57 vdc	a.
vsors Levers			High Voltage					
Positio	n	B	1/033	2	3	4	5	
Test Meter	/		6	7	8	9	10	
Model KLuike	25		11	12	13	. 14	15	
Serial # 5560	60519		16	17	18	. 19	20	
	× - /	I						!
1260/1261			1	470				
Sample Switch		D	Fan /	/				
Detector LED Grounds	-	, D	Video					
1282		a	Power Switc	n /			\sim	
Slide Arm			Pobotics 77			/.		ā
			NODUICO (/		~			
VO DEVICES AND	SOFTWARE			ESTS				
Printer			Printout Clar	rity				
Models	Serial #		Ribbon ·	,				
			Mechanical	(gears, plate	en, moto	ors, print	head)	0
			Cables					
Model		0	Fan					
Monitor		/	Disk Drives					<u>-</u> _
Model	Serial # USSR		Video					D
			Cables		_			
Computer Software //								
	LIRA IDRM	14	Version		1. < -	<u> </u>		
Customer's	10	Λ		-	/		n 17 11	••
Signature Sourpri	di Alsur	J	Signature	2	fans	1. 1		
	and i pippid	<u> </u>	Gigitature	/ ,				
DOC 4.09.035 Rev. 2						-		
White Copy - Сизгото	2	Y еШот Сору -	With Instrument			Pi	іль Сору - Ілгенты	ni File

Certificate of Calibration

ISSUED TO: Naval Medical Research Bethesda, MD 20889-5607 CONTACT: HMB AHMAD PO NO: PHONE: (301) 295-0001

RSO, Inc.

Laurel, MD 20725 (301) 953-2482

P.O. Box 1526

INSTRUMENT: BICRON MODEL: SURVEYOR 50 TYPE: RATEMETER SN: B602G

°.,.,

RSO, Inc. certifies that on 07/22/1999 the above described instrument was calibrated using a radioactive source to determine the efficiency for a specific radionuclide(s) and using electronically generated pulse for the linearity.

The results are tabulated below. Calibration is traceable to NIST.

	Calibrat	tion Data			
RANGE	EXPECTED	OBSE	RVED	C.E.	
1.00	100.00	100.000	cpm	1.000	
	500.00	500.000	cpm	1.000	
10.00	1000.00	1000.000	cpm	1.000	
	5000.00	5000.000	срп	1.000	
100.00	10000.00	10000.000	срп	1.000	
	50000.00	50000.000	cpm	1.000	
			C.F. AVERAGE	1.000	
	RANGE 1.00 10.00 100.00	Calibra RANGE EXPECTED 1.00 100.00 500.00 500.00 10.00 1000.00 5000.00 50000.00 100.00 50000.00	Calibration Data RANGE EXPECTED OBSE 1.00 100.000 100.000 500.00 500.000 500.000 10.00 1000.000 1000.000 10.00 1000.000 1000.000 100.00 10000.000 50000.000 100.00 10000.000 50000.000 100.00 50000.000 50000.000	Calibration Data RANGE EXPECTED OBSERVED 1.00 100.00 500.00 cpm 10.00 1000.00 1000.000 cpm 10.00 1000.00 5000.000 cpm 10.00 10000.000 cpm cpm 100.00 10000.000 cpm cpm 50000.00 50000.000 cpm cpm 100.00 100000.000 cpm c.F. AVERAGE	Calibration Data RANGE EXPECTED OBSERVED C.F. 1.00 100.00 100.000 cpm 1.000 500.00 500.000 cpm 1.000 10.00 1000.000 cpm 1.000 10.00 1000.000 cpm 1.000 100.00 1000.000 cpm 1.000 100.00 10000.000 cpm 1.000 100.00 50000.000 cpm 1.000 50000.00 50000.000 cpm 1.000 C.F. AVERAGE 1.000 C.F. AVERAGE 1.000

	Probe D	/p c (s) Pro	bel: PANCA	KE 1	Probe2:			Probe3:	
MODEL	L SER	WINDOW	GEOMETRY	VOLT	ISOTOPE 1 EFF.(%)	ISOTOPE 2 E	eff.(%)	ISOTOPE 3 EFF.(%)	ISOTOPE 4 EFF.(%)
РСМ	B716G	OPEN	CONTACT	900	32P- 31	355	7	33P 14	4503 ** 7

· Simulated using 90Sr with source at constant with probe,

== Simulated using 14C with source at contact with probe. . . *** Simulated using 99Te with source at contact with probe.

** Simulated using 14C with source at contact with probe.

-	
INSTRUMENT	CHECKS

BATTERY CHECK: NORMAL

CHECK SOURCE 1: NA

CHECK SOURCE 2:

ENVIRONMENTAL

TEMP: 25 C PRESS: 755 mmHg HUMID: 65 %

Cal Date: 07/22/1999

THE SUGGESTED RECALIBRATION DATE FOR THIS INSTRUMENT IS 🚽	07/22/2000
---	------------

READING:

READING:

Calibrated By:

Reviewed By:

Maryland License MD-33-021-01

RSO, Inc. P.O. Box 1526 Certificate of Calibration Laurel, MD 20725 (301) 953-2482

ISSUED TO: Navai Medical Research Bethesda, MD 20889-5607 CONTACT: HM3 AHMAD PHONE: (301) 295-0001

PO NO:

INSTRUMENT: LUDLUM MODEL: 3 TYPE: RATEMETER SN: 54537

332--- 14

45Ca**

7

· • •

RSO, Inc. certifies that on 07/22/1999 the above described instrument was calibrated using a radioactive source to determine the efficiency for a specific radionuclide(s) and using electronically generated pulse for the linearity.

The results are tabulated below. Calibration is traceable to NIST.

		Calib	ration Data			
H	RANGE	EXPECTED	OBS	ERVED	C.E.	
x	0.10	100.00	100.000	cpm	1.000	
		300.00	300.000	cpm	1.000	
x	1.00	1000.00	1000.000	cpm	1.000	
		3000.00	3000.000	cpm	1.000	
х	10.00	10000.00	10000.000	срл	1.000	
		30000.00	30000.000	срт	1.000	
х	100.00	100000.00	100000.000	срт	1.000	
		300000.00	300000.000	cpm	1.000	
		~		C.F. AVERAGE	1.000	
Probe type(s)	Probe1:	PANCAKE	Probe2:		Probe3:	

MODEL SERJ WINDOW GEOMETRY VOLT ISOTOPE 1 EFF. (%) ISOTOPE 2 EFF. (%) ISOTOPE 3 EFF. (%) ISOTOPE 4 EFF. (%)

355**

2

29

32P-

44-9 39815 OPEN

CONTACT 900 - Simulated using 90Sr with source at contact with probe."

** Simulated using 14C with source at conduct with probe.

*** Simulated using 99Te with source at contact with probe.

** Simulated using 14C with source at contact with probe.

INSTRUMENT CHECKS

BATTERY CHECK: NORMAL CHECK SOURCE 1: NA READING: CHECK SOURCE 2: READING:

Alexin Austin

ENVIRONMENTAL

TEMP: 25'C PRESS: 755 mmHg HUMID: 65 %

THE SUGGESTED RECALIBRATION DATE FOR THIS INSTRUMENT IS 07/22/2000

. '

Calibrated By:

Reviewed By:

Cal Date: 07/22/1999

Maryland License MD-33-021-01



ISSUED TO:	Naval Medical Research		INSTRUMENT:	LUDLUM
	Bethesda, MD 20889-5607		MODEL:	3
CONTACT:	HM3 AHMAD		TYPE:	RATEMETER
PHONE:	(301) 295-0001	PO NO:	SN:	114886

RSO, Inc. certifies that on 07/22/1999 the above described instrument was calibrated using a radioactive source to determine the efficiency for a specific radionuclide(s) and using electronically generated pulse for the linearity.

The results are tabulated below. Calibration is traceable to NIST.

		Calib	ration Data			`.
	RANGE	EXPECTED	OBSI	ERVED	C.E.	
х	0.10	100.00	100.000	cpm	1.000	
		400.00	400.000	cpm	1.000	
Х	1.00	1000.00	1000.000	cpm	1.000	
		4000.00	4000.000	cpm	1.000	
х	10.00	10000.00	10000.000	срт	1.000	
		40000.00	40000.000	срт	1.000	
х	100.00	100000.00	100000.000	cpm	1.000	
		400000.00	400000.000	cpm	1.000	
		-		C.F. AVERAGE	1.000	
Probe type(s) Probel:	SCINTILATOR	Probe2:		Probe3:	

MODEL SER WINDOW GEOMETRY VOLT ISOTOPE 1 EFF. (%) ISOTOPE 2 EFF. (%) ISOTOPE 3 EFF. (%) ISOTOPE 4 EFF. (%)

750 44-21 132842 OPEN CONTACT 32P-34 1251--18 51Cr*** 1

. Simulated using 90Sr with source at contact with probe.

... Simulated using 1291 with source at contact with probe.

*** Simulated using 133Ba with source at contact with probe.

INSTRUMENT CHECKS

BATTERY CHECK: NORMAL CHECK SOURCE 1: NA READING: CHECK SOURCE 2: READING:

ENVIRONMENTAL

TEMP: 25°C PRESS: 755 mmHg HUMID: 65 %

THE SUGGESTED RECALIBRATION DATE FOR THIS INSTRUMENT IS 07/22/2000

Calibrated By:

Reviewed By:

6-D)

Cal Date: 07/22/1999

Maryland License MD-33-021-01

·



Certificate of Calibration

INSTRUMENT: LUDLUM ISSUED TO: Naval Medical Research Bethesda, MD 20889-5607 MODEL: 3 CONTACT: HM3 Ahmad TYPE: RATEMETER PHONE: (301) 295-0001 PO NO: CC SN: 121461

RSO, Inc. certifies that on 08/12/1999 the above described instrument was calibrated using a radioactive source to determine the efficiency for a specific radionuclide(s) and using electronically generated pulse for the linearity.

The results are tabulated below. Calibration is traceable to NIST.

		Calibr	ration Data			
	RANGE	EXPECTED	OBSE	RVED	<u>C.F.</u>	
х	0.10	100.00	100.000	cpm	1.000	
		400.00	400.000	cpm	1.000	
х	1.00	1000.00	1000.000	cpm	1.000	
		4000.00	4000.000	cpm	1.000	
х	10.00	10000.00	10000.000	cpm	1.000	
		40000.00	40000.000	срт	1.000	
х	100.00	100000.00	100000.000	cpm	1.000	
		400000.00	400000.000	cpm	(.000	
				C.F. AVERAGE	1.000	
		~				
Probe type(s) Probel:	SCINTILATOR	Probe2:		Probe3:	

32P--

۰.

29

MODEL SERJ WINDOW GEOMETRY VOLT ISOTOPE I EFF. (%) ISOTOPE 2 EFF. (%) ISOTOPE 3 EFF. (%) ISOTOPE 4 EFF. (%)

17

1251-

44-21 121786 OPEN CONTACT 1050

· Simulated using 1291 with source at contact with probe...

** Simulated using 90Sr with source at contact with probe.

*** Simulated using 133B4 with source at contact with probe.

INSTRUMENT CHECKS

.

ENVIRONMENTAL

51Cr****

BATTERY CHECK: NORMAL CHECK SOURCE 1: NA READING: CHECK SOURCE 2: READING:

TEMP: 25°C PRESS: 750 mmHg HUMID: 65 %

THE SUGGESTED	RECALIPRATION	DATE FOR THIS	INSTRUMENT IS	08/12/2000
Calibrated By:	Dorfy Auctus	Reviewed By:	Thinkin	Cal Date: 08/12/1999

Maryland License MD-33-021-01

1361

· : •

.

RSO, Inc. P.O. Box 1526 Laurel, MD 20725 (301) 953-2482

.....

Certificate of Calibration

ISSUED TO:	Naval Medical Research
	8901 Wisconsin Ave.
	Bethesda, MD 20889-5607
CONTACT:	S. Ahmad
PHONE:	(301) 295-0002

INSTRUMENT: LUDLUM MODEL: 3 TYPE: RATEMETER SN: 121816

X X x x

1

RSO, Inc. certifies that on 02/22/2000 the above described instrument was calibrated using a radioactive source to determine the efficiency for a specific radionuclide(s) and using electronically generated pulse for the linearity. The results are tabulated below. Calibration is traceable to NIST.

PO NO:

Calibration Data

RANGE	EXPECTED	OBSERVED	C.F.
.1 I	100 500 1000	100 срт 500 срт 1000 срт	1.00 1.00 1.00
10	5000 10000 5000	5000 cpm 10000 cpm 5000 cpm	1.00 1.00 1.00
100	100000 500000	100000 cpm 500000 cpm C.F. AVERAGE	1.00 1.00 1.00

Probe type(s) Probe1: PANCAKE

Probe2:

Probe3:

33P--- 14

. .

MODEL SERJ WINDOW GEOMETRY VOLT ISOTOPE | EFF. (%) ISOTOPE 2 EFF. (%) ISOTOPE 3 EFF. (%) ISOTOPE 4 EFF. (%)

32P**

۰.

35

7

44-9 119401 FIXED CONTACT 906 14C-

= JSS=7%, 45Ca=7%; Simulated using 14C with source at contact with probe.

. • • .

-- Simulated using 90Sr with source at contact with probe.

--- Simulated using 99Tc with source at contact with probe.

INSTRUMENT CHECKS		ENVIRONMENTAL
BATTERY CHECK: NORMAL CHECK SOURCE 1: CHECK SOURCE 2:	READING: READING:	TEMP: 26°C PRESS: 768 mmHg HUMID: 55 %
THE SUGGESTED RECAL Calibrated By: Autor	TERATION DATE FOR THIS INSTRUMENT IS Reviewed By: 23	02/22/2001 Cal Date: 02/22/2000

Maryland License MD-33-021-01

1440 . • •

RSO, Inc. P.O. Box 1526 Laurel, MD 20725 (301) 953-2482

Certificate of Calibration

ISSUED TO:	Naval Medical Research		INSTRUMENT:	LUDLUM
	8901 Wisconsin Ave.		MODEL:	3
	Bethesda, MD 20889-5607		TYPE:	RATEMETER
CONTACT:	S. Ahmad		SN:	132946
PHONE:	(301) 295-0002	PO NO:		

RSO, Inc. certifies that on 02/22/2000 the above described instrument was calibrated using a radioactive source to determine the efficiency for a specific radionuclide(s) and using electronically generated pulse for the linearity.

The results are tabulated below. Calibration is traceable to NIST.

		Calibrati	on Data		
	RANGE	EXPECTED	OBSER	VED	C.F.
x	.1	100 400	100 400	cpm cpm	1.00
x	1	1000 4000	1000 4000	cpm cpm	1.00 1.00
x	10	10000 40000	10000 40000	cpm cpm	00.1 1.00
x	100	100000 400000	100000 400000 C.F	cpm cpm AVERAGE	1.00 1.00 1.00

Probe type(s) Probel: SCINTILATOR

Probe3:

.

MODEL SERJ WINDOW GEOMETRY VOLT ISOTOPE 1 EFF. (%) ISOTOPE 2 EFF. (%) ISOTOPE 3 EFF. (%) ISOTOPE 4 EFF. (%)

Probe2:

44-21 135670 FIXED CONTACT 756 1251- 16 51Cr-+ 2.6

· Simulated using 1291 with sourse at contact with probe.

: `

== Simulated using 133Ba with source at contact with probe, using correction factor of X0.1

INSTRUMENT CHECKS	· ·	ENVIRONMENTAL
BATTERY CHECK: NORM	AL	TEMP: 27°C
CHECK SOURCE 1:	READING:	PRESS: 768 mmHg
CHECK SOURCE 2:	READING:	HUMID: 55 \$
THE SUGGESTED REC	ALIBRATION DATE FOR THIS INSTRUMENT IS	02/22/2001
Calibrated By: Michael	Reviewed By: <u>28</u>	Cal Date: 02/22/2000

Maryland License MD-33-021-01

...


U.S. DEPARTMENT OF COMMERCE

National Institute of Standards & Technology Gaithersburg, MD 20899

Certificate of Participation

Analytics, Incorporated Atlanta, Georgia

is a participant for the period January 1, 1999, through December 31, 1999, in a radioactivity measurements assurance program conducted by the National Institute of Standards and Technology, in cooperation with the Nuclear Energy Institute. Continued participation is evidenced by dated Reports of Traceability issued for particular radionuclides, which indicate the deviation of the participant's reported value from that measured by the National Institute of Standards and Technology. The significance of these Reports is addressed below."

For the Director,

Ancerano

Lisa R. Karam, Group Leader Radioactivity Group Physics Laboratory

As guidance for the proper use of Reports of Traceability, it should be emphasized that the National Institute of Standards and Technology is concerned only with fostering good measurements capability and consistency with the national measurements system. The assurance of the proper application of that capability to the ultimate consumer products is the responsibility of each manufacturer of these products and of the Federal regulatory agencies.

A continuing traceability program in radioactivity demonstrates, to the degree established by the periodic assays of calibrated radioactivity samples, a continuing competence to maintain the instrument systems and standards necessary for accurate measurement. Such a program cannot, however, endorse each and every measurement nor the final product, any more than a spot check can vouch for every unchecked item. Care should be taken, therefore, not to imply such endorsement. The proper use of this Report is governed by section 200.114 of Title 15 of the Code of Federal Regulations. These regulations may be met if Reports are quoted only in their entirety. Excerpts out of context may be misleading.



Phone (404) 352-8677 Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

58806A-530

C-14 32 mm Diameter x 3 mm Thick Plastic Disk

This standard radionuclide source was prepared using an aliquot measured gravimetrically from a master radionuclide solution The master radionuclide solution standard was standard. calibrated by the Department Des Applications Et De La Metrologie Des Rayonnements Ionisants (DAMRI), Paris, France, as number ANALYTICS maintains traceability to the National 22453. Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1. After source preparation the calibration was checked by beta counting using a gas flow proportional counter.

ISOTOPE:	C-14
ACTIVITY (dps):	5049
HALF-LIFE:	5730 years
CALIBRATION DATE:	November 18, 1999 12:00 EST
TOTAL UNCERTAINTY*:	5.0%

*99% Confidence Level

Diameter of active area: 16 mm. Source covering 0.5 mg/cm² mylar.

No expiration date has been given for this source due to the fragile nature of the mylar covering. This source should be carefully tested for leakage at least every six months. If leakage is detected this source should be disposed of by approved radioactive waste disposal procedures.

P O NUMBER N32398-99-M-0228, Item 1

SOURCE PREPARED BY: Magraere

M. Taskaeva, Radiochemist

Q A APPROVED:

· ·

2 11-19-59

1380 Seaboard Industrial Blvd. Atlanta, Georgia 30318 - U.S.A.



Phone (404) 352-8677 Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

58807-530

C-14 32 mm Diameter x 3 mm Thick Plastic Disk

This standard radionuclide source was prepared using an aliquot measured gravimetrically from a master radionuclide solution standard. The master radionuclide solution standard was calibrated by the Department Des Applications Et De La Metrologie Des Rayonnements Ionisants (DAMRI), Paris, France, as number 22453. ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1. After source preparation the calibration was checked by beta counting using a gas flow proportional counter.

TOTAL UNCERTAINTY*:	5.0%
CALIBRATION DATE:	November 18, 1999 12:00 EST
HALF-LIFE:	5730 years
ACTIVITY (dps):	5347
ISOTOPE:	C-14

*99% Confidence Level

Diameter of active area: 16 mm. Source covering 0.5 mg/cm^2 mylar.

No expiration date has been given for this source due to the fragile nature of the mylar covering. This source should be carefully tested for leakage at least every six months. If leakage is detected this source should be disposed of by approved radioactive waste disposal procedures.

P O NUMBER N32398-99-M-0228, Item 1

SOURCE PREPARED BY: Maericera

M. Taskaeva, Radiochemist

Q A APPROVED:

. . .

NM Mot 3-11-19-95

1380 Seaboard Industrial Blvd. Atlanta, Georgia 30318 · U.S.A.

R

Phone (404) 352-8677 Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

58808-530

Co-60 32 mm Diameter x 3 mm Thick Plastic Disk

This standard radionuclide source was prepared using aliquots measured gravimetrically from calibrated master liquid radionuclide solutions. The master sources were calibrated with an ion chamber that was calibrated by the National Physical Laboratory, Teddington, U.K., and is directly traceable to national standards.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

ISOTOPE:	Co-60
ACTIVITY (dps):	3.599 E+04
HALF-LIFE:	5.2714 years
CALIBRATION DATE:	November 18, 1999 12:00 EST
TOTAL UNCERTAINTY *:	4.8%
SYSTEMATIC:	4.5%
RANDOM :	0.3%

*99% confidence level.

Diameter of active area: 25 mm. Source covering 0.85 mg/cm² mylar.

No expiration date has been given for this source due to the fragile nature of the mylar covering. This source should be carefully tested for leakage at least every six months. If leakage is detected this source should be disposed of by approved radioactive waste disposal procedures.

P O NUMBER N32398-99-M-0228, Item 2

SOURCE PREPAR	ED BY:	libaen	care	
	ľ	4. Taskaeva,	Radiochem	ist
Q A APPROVED:	-	M Mity		-19-91

1380 Seaboard Industrial Blvd. Atlanta, Georgia 30318 - U.S.A.



Phone (404) 352-8677 Fax (404) 352-2837

CERTIFICATE OF CALIBRATION Standard Radionuclide Source

58809-530

Co-60 32 mm Diameter x 3 mm Thick Plastic Disk

This standard radionuclide source was prepared using aliquots measured gravimetrically from calibrated master liquid radionuclide solutions. The master sources were calibrated with an ion chamber that was calibrated by the National Physical Laboratory, Teddington, U.K., and is directly traceable to national standards.

Radionuclide purity and calibration were checked using a germanium gamma spectrometer system. The nuclear decay rate and assay date for this source are given below.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

ISOTOPE :	Co-60
ACTIVITY (dps):	3.730 E+04
HALF-LIFE:	5.2714 years
CALIBRATION DATE:	November 18, 1999 12:00 EST
TOTAL UNCERTAINTY*:	4.8%
SYSTEMATIC:	4.5%
RANDOM :	0.3%

*99% confidence level.

Diameter of active area: 25 mm. Source covering 0.85 mg/cm² mylar.

No expiration date has been given for this source due to the fragile nature of the mylar covering. This source should be carefully tested for leakage at least every six months. If leakage is detected this source should be disposed of by approved radioactive waste disposal procedures.

P O NUMBER N32398-99-M-0228, Item 2

SOURCE PREPARED	BY:	Ichaercaera
		M. Taskaeva, Radiochemist
Q A APPROVED:		NM MA 11-19-94

IM	Designer and N at	lanulacturer I				POST OFFICE BOX 8	UKEMENIS, IN 10 PH. 915-235-5	C. 494
	Scientific and Instrum	d Industrial ents	CERTIFICATE	OF CALIBRA	IION	501 OAK STREET SWEETWATER, TEXAS	FAX NO. 915 79556, U.S.A.	-235-4672
OMER	NAVAL MED	ICAL RESEARCH C	TR			ORDER NO.	237400	<u> </u>
	Ludium Measu	rements, Inc.	_ Model	2350	-1	Seriol No	157674	
Cal. Date	27-\$0	<u>ep-99</u> Co	I Due Date	<u>27-Sep-0</u>	0 Cal. Int	ervol <u>1Year</u>	Meterlace	<u>N/A</u>
Check mark	$\underline{\checkmark}$ applies to ap	plicable instr. and	d/or detector IAW	míg. spec.	ſ76_ °F	RH56_9	6 All 700.8	_ mm Hg
V New In:	strument Instru	ument Received	🔄 Within Toler, +	-10% 🔄 10-20%	🕻 📋 Out of Tol.	Requiring Repair	Other-See co	omments
🖌 Mecho	inical check	,		,		🗹 Ing	out Sens. Linearity	
√ F/S Res	p. check	Reset c	heck	Vindo	ow Operation			
✓ Audio d	check eter Linearity che		ted Dose check		y check (Min, V de Mode check	011) <u>4.4</u> VDC		
	og check		ad check		Readoul check	Dial R	nold Iatio <u>, 100 =</u>	_4 mV
Calibrat	ted in accordance	ce with LMI SOP 1	4.8 rev 12/05/89.		ated in accordance	e with LMI SOP 14.9	rev 12/19/89.	_
∵ н∨ і	Readout (2 poin	is) Ref./Inst	498	/500	V Ref./inst	2002	/2000	V
COMMEN	TS: Firmwo	re: 37122N27						
I/O firmu	ware : 371231	NO5 Resolutio	n for 44-10 ≈	9.36 % for	Cs137. Model	2350-1 and all	l probes	
Calibrate Gamma Calibratio	ed using 6 for on GM detectors position	t. cable only Netperpendicular to source	. See ATTACHM except for M 44-9 in which	IENT 1 for e the front of probe faces s	fficiencies. curce		1	
	Frobe		Higu		Units/	Dead Lime	Canoration	Linearity
	Model	Serial #	Voltage	Threshold	Time Base	Correction Factor	Constant	=10%*
	LMI43-68	PR149314	1750	100		2.342273E-03	1.000000E+00	
Detector # 2	LMI43-68	PR149314	1/50	100	<u> </u>	2.342273E-05	1.000000E-00	
Detector # 3	LMI43-58	PR145314	1/50	100	/ / 1	2.342273E-05	1.0000002~00	
Detector # 4	LMI43-37	PR148947	1800	100	7 / 1	2.361271E-05	1.000000E-00	
Detertor # 5	LMI43-37	PR148947	1800	100	7 / 1	2.361271E-05	1.00000000000	
∩≏ #6	LMI43-37	PR148947	1800	100	7 / 1	2.361271E-05	1.000000E-00	
.or # 7	LMI44-10	PR164841	1200	250	4 / 2	1.577578E-05	5.406332E+10	
Detector # 8	LMI44-10	PR164841	1200	250	4 / 2	1.5775785-05	5.406332E+10	
Detector # 9	LMI44-10	PR164841	1200	250	4 / 2	1.577578E-05	5.405332E+10	
Detector # 10	PEAK	CO-60	660	560	7 / 1	0.000000E+00	1.000000E+00	
Detector # 11	PEAK	06-00	660	560	7 / 1	0.000000E+00	1.0000002+00	
Detector # 12	PEAK	CO-60	660	550	7 / 1	0.000000E+00	1.000000E+00	
Detector # 13	PEAK	CS-137	660	321	7 / 1	0.000000E+00	1.000000E+00	
Detector # 14	LMI44-10	PR154841	1200	250	7 / 1	1.577578E-05	1.000000E+00	
Detector #								
Detector #								
Units: 0 - a	rad, 1 - Gray, 2 - rem, 3	3 - Sv. 4 - R. S - C/Kg. 6	- Disintegrations, 7 - Cou	ints, 8 - Ci/cm sq., 9	- Bq/cm sq.			
Time Base; 0 – 3	Seconds, 1 - Minutes.	2 - Hours				• See a	Nached delector documentat	ion, if applicable.
	REFERENCE		NT INSTR				NT INSTRU METER	READING
Digilal Readout	400kcpm		399	969	400cs			400
	40kcpm		- 40	014	4001	<u>2m</u>		40
	4kcpm	<u> </u>	<u> </u>	009				
Officer Internation	ements, Inc. centiles ti anal Standards Organ: system conforms to In	hai the coove instrumer zation members, or havi the requirements of ANSI/	nt has been calibrated to been derived from ac NCSL 2540-1-1994 and A	cepted values of not NNSEN323-1978,	ie to the National Institut ural physical constants d	e of Standards and Techno x have been derived by th State of Texas	plogy, at to the calibration e ratio type of calibration Calibration License N	n lectiniques. 0. LO-1963
Reference	Instruments ar	nd/or Sources: o	S-137 Gamma S/N					
_ 1162	G112 M565	5 🛄 5105 🛄 T100	8 🟹 1879 🗹 E552	E551 Neu	fron Am-241 Be S/N I	-304		
Alp	ha S/N		🗹 Beta S/N	⊂ 9 Ý <u>635/8</u> 3 ,	1-659, 422-44	_ 📝 Other <u>Am</u> 2	241 - 1.73µCi/1 129	0.076 uCi
- 7 m :	500 S/N	141233	a (\frown	Г./ M	ullimeter S/N	59120574	
Colibrator		Oin Ch	to alla	Ý.		JJ-CE	12.99	
Calibidied	юу	erry Co	YVYY Y	- <i>H</i>	Date	$-\alpha c \rightarrow c$	/ <u>^) \</u>	

Reviewed	Ву:
FORM C14A	07/30/99

Ins certificate shall not be reproduced except in full, without the written approval of Ludium Measurements, Inc.

Z

LUDLUM MEASUREMENTS, INC. 501 OAK ST./PO BOX 810 SWEETWATER, TEXAS 79558 Phone: 915/235-5494 800/622-0828(USA) F.ax: 915/235-4572

M

DESIGNER AND MANUFACTURER OF Soluntific and Industrial Instruments

ATTACHMENT 1

 4π efficiencies for 43-68:

Efficiencies for 43-68 taken at 1750v.

Factory background ≈ 261cpm

Ca45 and P33 were substituted with Tc99 Tc99 is \approx 38.55% net Tc99 reading \approx 8568 net

S35 was substituted with C14 C14 is \approx 10.82% net C14 reading \approx 33740 net

I 125 was substituted with I 129 I 129 is $\approx 0.25\%$ net I 129 reading is ≈ 428 net

P 32 was substituted with Si 32 Si 32 is \approx 35% net Si 32 reading is \approx 34460 net

 4π efficiency for 44-10:

Efficiencies for 44-10 taken at 1200v. using Det.14 and 1 minute counts.

Factory background ≈ 7508 cpm

Co 57 is ≈ 28.70 % net Co 57 reading is ≈ 25801 cpm net

Co 60 is ≈ 22.17 % net Co 60 reading is ≈ 9045 cpm net

Cs137 is \approx 12.13 % net Cs137 reading is \approx 1854 cpm net

4 π efficiencies for 43-37 :

Efficiencies for 43-37 taken at 1800v.

Factory background ≈ 908cpm

Ca45 and P33 were substituted with Tc99 Tc99 is \approx 35.9% net Tc99 reading \approx 8223 net

S35 was substituted with C14 C14 is \approx 10.42% net C14 reading is \approx 32507 net

I 125 was substituted with I 129 I 129 is $\approx 0.44\%$ net I 129 reading is ≈ 738 net

P 32 was substituted with Si 32 Si 32 is \approx 35.16% net Si 32 reading is \approx 34521 net

27-SEP-P9

Serving The Austean Industry Stree 1962 .

DETECTOR SETUP CHECK LIST REPORT The following list is stored as detector setup D14 in the Model 2350. Today's date is 09/27/1999. The current time of day is: 09:36:33.

-

.

I ...ve verified the list below has NO discrepancies with the DETECTOR SETTINGS TABLE: ______

Comments:

`

_ ·

.

·

Model 2350 Serial # =	157674.
High Voltage = Threshold =	1200 volts. 250.
Window =	100,Off.
Overload Current =	40.0 micro amperes.
Scaler Count Time =	60 seconds.
Readout Units =	counts.
Readout Time Base =	min.
Readout Range Multiplier =	auto.
Defector Dead Time =	1.577578E-05.
<pre>:ctor Calibration Constant =</pre>	1.000000E+00.
L_lector Model =	LMI44-10.
Detector Serial # =	PR164841.
Ratemeter Alarm Setting =	1.000000E+09.
Scaler Alarm Setting =	1000000.
Integrated Dose Alarm Setting =	1.000000E+09.
Low Count Alarm Setting =	х.
Operating Battery Voltage =	6.1 volts.

. . .



.

Designer and Manufacturer of Scientific and Industrial Instruments POST OFFICE BOX 810 PH, 915-235-5494 501 OAK STREET FAX NO. 915-235-4672 SWEETWATER, TEXAS 79556, U.S.A.

Model 2350	Bench Test Data	
------------	-----------------	--

Customer <u>NAVAL N</u>	EDICAL RESEARCH CTR	Date	25-Sep-99	Order #, '	237400
Model <u>2350-1</u>	Serial No. <u>157674</u>	Detector _	44-10	Serial No	PR164841
Source <u>CS137</u>	21mCi 2.4	12mCi			
High Voltage	1200 V As Found	V, Input	10.00 mV	As Found	mV.
Cal. Constant	5,406332E+10	as found			
Dead Time	1.577578E-05	as found			
Alarm Setting: Rat	iemeter <u>2(100000000.00</u>	0000 as found			
Sco	ler 100000.00	10000 as found			
Inte	egrated dose <u>100000000</u>	<u>.0000</u> cs found			
Overload 20n 3	ZOff as found ⊇On ⊇O	ff Window	_100 (eff) as fo	bund	
Detector Re	eceived: 🗌 Within Toler. +-10	9% — 10-20% 🗌 Out o	f Tol. 🔄 Requiri	ng Repair 🗌 Ott	ner-See comments
	Reference Point	"As Found" Readings: Meter Reading	After Adjustm Meter	nent Readings: Reading	
	2.0 mR/hr 1.5 /		<u> </u>	<u>3 mR/h</u> r	
	1.0		1.03	3)	
	500 HR/hr		508	<u>uR/hr</u>	
	200		204		
	150		149		
	100		102	,)	
	/				
)	
			,	\mathbf{i}	
Other	Calibrated	Using 6	H. za	ple.	
Signature	Chix Cha	vert-	U	Date 7	5-SER-91
·		<i>X</i>			•.
FORM C6-1 09/08/97	• Serv	ing The Nuclear Industry	Since 1962 •		

•••



Bench Test Data For Detector

Detector	44-10	_ Serial No.	PR164841	_	
Customer NA		RESEARCH CTR		Order #.	237400
Counter	2350-1	Senal No.	157674	_ Counter Input Sensitivity	10.00 mV
Count Time _		SEL.		_ Distance Source to Detector _	Sempre
Other Cal C	Constant = 1.00	0000 <u>E+00</u> Deac	<u>1.5775788</u>	-05	/
High		Isotope An	1241 isotope_	lsotope	lsotope
Voltage	Background	Size PI.	<u>الخير</u> ر Size		
1000	573	11013			
1050	609	1 12508	5		
1100	636	12519	1		1
1150	567	12753			* •
- 1200	516	1 2879	1		
1250	582	12803			
1300	618	12834	i		
1350	5: 608	12954			i
1411): 622	12883			
	:				
		:			1 1
		1			
	1				
	i	1		1	
	<u> </u>				
				i	
	:				
	i		1		
	1	;		ł	
			11	`	
Signature	El	ior C	have	L Date	15-5EP.89
	· ·	- ,	\bigcirc		
DRM C4A 02/23/94		● Se	rving The Nuclea	r Industry Since 1962 •	

Designer and Manufacturer
oí
Scientific and Industrial

Instruments

Reviewed By:

FORM C14A 07/30/99

U

m.

lon

CERTIFICATE OF CALIBRATION

LUDLUM MEASUREMENTS, INC.

 POST OFFICE BOX 810
 PH. 915-235-5494

 SÔI OAK STREET
 FAX NO. 915-235-4672

 SWEETWATER, TEXAS 79556, U.S.A.

C OMER		DICAL RESEARCH CTR				ORDER NO	23740	00
ł	Ludlum Meas	urements, Inc.	Model	2350	-1	Serial No	157669	
Cal. Date	_ 27-		ue Date	<u>27-Sep-00</u>	D Cal. Int	erval <u>Year</u>	Meterface	N/A
Theck mark	v applies to a	pplicable instr. and/c	or detector IAV	V míg. spec. 1	. <u>76</u> °F	RH56_	% All 700	.8 mm Hg
V: New In	strument Inst	trument Received	Within Toler.	+-10%10-20%	Out of Tol.		ir 🗍 Other-See d	
	anical check	_	-	_	_	√ In	put Sens, Linearity	
F/S Res	sp. check	🏹 Reset che	ck	🟹 Windo	w Operation		, ,	
Audio	check	Alarm Set	ling check	Botter	ycheck (Min.V	olt] 4.4 VDC		
	eter Linearity Cr	neck <u>v</u> integrated	obeck		Readout check	Three	shold Ratio 100 =	1 ~~~
	ted in accorda	nce with LMI SOP 14.8	rev 12/05/89.	Calibra	ited in accordance	e with LMI SOP 14.5	9 rev 12/19/89.	<u> </u>
. <u>√</u> H∨	Readout (2 poi	nts) Ref./Inst	499	/500	V Ref./Inst	1998		0V
COMMEN	ITS: Firmw	are: 37122N27						
I/O firm	ware : 3712	3N05 Resolution	for 44-10	≈ 9.36 % for	Cs137. Model	2350-1 and al	l probes	
calibrat Gamma Calibrat	ed using 6 :	ft. cable only.	See ATTACH	MENT 1 for e	fficiencies.			
00	Prope		nign		บถแร/	Uead I Ime	Calibration	Lineanty
	Model	Serial #	Voltage	Threshold	Time Base	Correction Factor	Constant	=10%*
Detector # 1	LMI43-68	PR149314	1/50			2.3422/3E-05	1.000000E+00	
	LMI43-68		1750	100	7 / 1	2.342273E-05	1.000000E-00	·
Detector # 3	LM143-68	PR 149314	- 1750			2.3422/32-05	1.0000005-00	
Detector # 4	LMI43-37	PR148947	1800			2.361271E-03	1.000000E+00	·
Detector # 5	LMI43-37	PR148947	1800	100		2.3612/1E-05	1.000000E-00	
De	LMI43-37	PR148947	1800	100	7/1	2.361271E-05	1.000000E+00	
Den7	LMI44-10	PR164841	1200	250	4 / 2	1.577578E-05	5.406332E+10	
Delector # 8	LM144-10 .	PR164841	1200	250	4 / 2 ·	1.577578E-05	5.406332E+10	
Detector # 9	LMI44-10	PR164841	1200	250	4 / 2	1.577578E-05	5.406332E+10	<u> </u>
Detector # 10	PEAK	CO-50	6ā0	560	7 / 1	0.0000008+00	1.000000E+00	
Delector # 11	PEAK	CO-60	660	560	7 / 1	0.000000E+00	1.000000E-00	
Detector # 12	PEAK	CO-60	660	560	7 / 1	0.000000E-00	1.000000E-00	
Delector # 13	PEAK	CS-137	660	321	7 / 1	0.0000000E+00	1.000C00E-00	
Detector # 14	LMI44-10	PR164841	1200	250	7 / 1	1.577578E-05	1.000000E+00	
Detector #							······	
Delector #								
Units: 0 - Time Base: 0 -	rad, 1 – Gray, 2 – rem, Seconds, 1 – Minutes,	, 3 - Sv. 4 - R, 5 - C/Kg, 6 - C 2 - Hours	Disintegrations, 7 - Co	ounts. 8 - Cilomisq., 9 -	~ Ва/ст sq.	* See	attached detector document	tauon, if applicable
	REFERENCE	INSTRUMENT	INST	RUMENT	REFERENCE	INSTRUME	INSTR	UMENT
Digilol	CAL. POINT	RECEIVED	MET	ER READING"	CAL. POINT	RECEIVED		R READING" リロロ
REGOOUI		<u>m</u>	<u> ז</u>	9912	<u>40000</u>	<u></u>		40
	4kcpr	m		4000				
Ludium Measu other Internation	rements, Inc. certifies onal Standards Organ	that the above instrument his nization members, or have be	as been colibrated sen derived from 0 \$1,2540-1-1994 and	by standards traceco accepted volves of national	e to the National Institut wat physical constants a	e of Standards and Tech r have been derived by t	nology, or to the colibration license	tion facilities of ion techniques.
Reference	e Instruments c		37 Gamma C/M	دي / 77 ° تي پر لي ۳۰ ∎ي - د .				
- 2	□ G112 □ м58	45	1879 X E552	E551 Neu	Iron Am-241 Be S/N I	-304		
	oha S/N		V Belo S/N	V 635/83	C14 5132 1-659, 422-44	V Olher An	n241 ~1.73µCi/I 12	<u>9 0.076 µÇi</u>
V.m	500 S/N	141233			Г м	ullimeter S/N	59120574	
Calibrates	50	in the	\sim	.		77-5	CP- 89	

2

This certificate shall not be reproduced except in full, without the written approval of Ludium Measurements, Inc.

Dole 27-Sent-99

LUDLUM MEASUREMENTS, INC. 501 OAK ST./PO BOX 810 SWEETWATER, TEXAS 79558 Phone: 915/235-5494 800/622-0828(USA) c: 915/235-4672



DESIGNER AND MANUFACTURER OF Solentifis and Industrial Instruments

ATTACHMENT 1

 4π efficiencies for 43-68 :

Efficiencies for 43-68 taken at 1750v.

Factory background ≈ 261cpm

Ca45 and P33 were substituted with Tc99 Tc99 is \approx 38.55% net Tc99 reading \approx 8568 net

S35 was substituted with C14 C14 is $\approx 10.82\%$ net C14 reading ≈ 33740 net

I 125 was substituted with I 129 I 129 is $\approx 0.25\%$ net I 129 reading is ≈ 428 net

P 32 was substituted with Si 32 Si 32 is \approx 35% net Si 32 reading is \approx 34460 net

 4π efficiency for 44-10:

Efficiencies for 44-10 taken at 1200v. using Det.14 and 1 minute counts.

Factory background ≈ 7508 cpm

Co 57 is ≈ 28.70 % net Co 57 reading is ≈ 25801 cpm net

Co 60 is \approx 22.17 % net Co 60 reading is \approx 9045 cpm net

Cs137 is \approx 12.13 % net Cs137 reading is \approx 1854 cpm net

Elie 1

 4π efficiencies for 43-37 :

Efficiencies for 43-37 taken at 1800v.

Factory background ≈ 908cpm

Ca45 and P33 were substituted with Tc99 Tc99 is \approx 35.9% net Tc99 reading \approx 8223 net

S35 was substituted with C14 C14 is \approx 10.42% net C14 reading is \approx 32507 net

I 125 was substituted with 1 129 I 129 is ≈ 0.44% net I 129 reading is ≈ 738 net

P 32 was substituted with Si 32 Si 32 is = 35.16% net Si 32 reading is \approx 34521 net

27-SEP-PP

Scroing The Avalence Industry Stree 1962

DETECTOR SETUP CHECK LIST REPORT The following list is stored as detector setup D14 in the Model 2350. Today's date is 09/27/1999. The current time of day is: 09:33:46.

I have verified the list below has NO discrepancies with the DETECTOR SETTINGS TABLE: ______

Comments:

157669. Model 2350 Serial # = User I.D. = . 1200 volts. High Voltage = Threshold = 250. Window = 100,Off. Overload Current = 40.0 micro amperes. 60 seconds. Scaler Count Time = Readout Units = counts. Readout Time Base = mín. Readout Range Multiplier = _____ auto. D' actor Dead Time = 1.577578E-05. \therefore calibration Constant \Rightarrow 1.000000E+00. Lutector Model = LMI44-10. Detector Serial # = PR164841. Scaler Alarm Setting = 1.000000E+09. Integrated Doce 2010 Integrated Dose Alarm Setting = 1.000000E+09. Low Count Alarm Setting = X. Operating Battery Voltage = 6.2 volts. Χ.

. • .



Designer and Manulacturer of Scientific and Industrial Instruments

.

LUDLUM MEASUREMENTS, INC. POST OFFICE BOX 810 PH. 915-235-5494 .-501 OAK STREET FAX NO. 915-235-4672 SWEETWATER, TEXAS 79556, U.S.A.

Customer <u>NAVAL MEDICAL RESEARCH CTR</u>	Date	25-Sep-99	Order #	237400
Model <u>2350-1</u> Serial No. <u>157669</u>	Detector _	44-10	Serial No	PR164841
source <u>CS137</u> 21m(i, 2	47 mC1'			
High Voltage 1200 V As Found	V. Input	<u>10.00</u> mV	As Found	mV.
Cal. Constant 5.406332E+10	as found			
Dead Time 1.577578E-05	as found			
Alarm Setting: Ratemeter 0.00 Scaler 1000000.00 Integrated dose 1000000000	2000 cs found 2000 as found 2000 as found			
OverloadOnOff as foundOnOn	ff Window	_100L0F7.)as fi	ound	
Reference Point 2.0 mR/Ar 1.5 1.0 500 mR/Ar 200 150 150 100 Other Cultured ignature Line Ch	"As Found" Readings: Meter Reading	After Adjustin Meter 1.93 1.50 1.02 508 204 149 102 102	Date _2	<u>5-5EP-9</u> }

-	1
ï	
1.5	 Ш.
a second	5

Designer and Manufacturer of Scientific and Industrial Instruments LUDLUM MEASUREMENTS, INC. POST OFFICE BOX 810 PH. 915-235-5494 501 OAK STREET FAX NO. 915-235-4672 SWEETWATER, TEXAS 79556, U.S.A.

Bench	Test	Data
-------	------	------

tector	43-68	Serial No	PR149314	Ord	der #,	237400
istomer NAV	AL MEDICAL RES					
ounter	2350-1 Ser	rial No	157669	Counter Input Ser	nsitivity	4.00 mV
ount Time	1m	in		Distance Source to Det	ector S	nface
her <u>Cal Ca</u>	onstant = 1.00000	00E+00 Dead	Time = 2.342273E-05	Plateaud	w/6/	A. cable.
High Voltage	Background	Isotope 729 Size 22.9	9 Isorope C Kapm Size 0,1	14 Isotope Size 98	32 Isoto	ppe I_{129} Size 0.076 μ G
1550	1146 :	7106		12,794		310
1600	192	8138	32,18	35 22419		495
1550	235	8420	- 33,8'	18 32,00		658
1700	23/	8349	<u> </u>	33.724		662
- 1750	261	8829	34,50	4534721		689
1800	265		2-8-37,8	97 38,06		924
1850	2198	3127	8019	125 66973	6	1073
	·				!	
	!					
	•					
				:		
	•		;		1	
				·!		
				ı 		
				· · · · · · · · · · · · · · · · · · ·		
			1			
				·		
	!				· · · · · · · · · · · · · · · · · · ·	
J Gas prov	Dortional derec	-	decreased ans	ter 15 hour crosse	20*	Cable
gnature _	Elio	_ Cha	very list		Date	5-5EP-9)

Sonving The Must

·



Designer and Monufacturer oľ

Scientific and Industrial Instruments

LUDLUM MEASUREMENTS, INC. POST OFFICE BOX 810 PH. 915-235-5494 501 OAK STREET FAX NO. 915-235-467 SWEETWATER, TEXAS 79556, U.S.A. .

Bench Test Data

Detector <u>43-37</u>	Serial No	PR148947	Order #.	237400
Customer NAVAL MEE	DICAL RESEARCH CTR			
Counter2350-1	Serial No.	157669	Counter Input Sensitivity	4.00 mV
Count Time	minute	D	istance Source to Detector	Sunpa
Other <u>Cal Constant</u>	= 1.000000E+00 Dead	Time = 2.361271E-05	Plateour "	/6ft. cable.
High Voltoge Backs	Isotope <u>TC9</u>	3isotope_ <u></u> WdpySize_ <u>0.14</u>	14_ Isotope <u>5/32</u> 10. C. Size <u>98,168</u>	Isotope 129
1550: 11	8 2787	/353	7 3985	294
1200 : 24	2 4330	1954	8: 7559	511
1650.38	8 6228	2746	8 13230	694
1700 63	5 7837	30858	21851	1027
1750 8:	14 8760	3232	3 32023	1505
1800 90	×8 9131	3341	5 35429	1646
1850 95	54 9120	3360	<u>5 . 35693</u>	1805
1900 9	9/ 9214	3431	<u>8 35922</u>	774
				·
<u> </u>				
<u> </u>				
• i		:		
			: 	1
	l	1]
\overline{V} Gas proportion ignature	al detector count rate	decreased \$ 103 att	er 15 hour static test us. Dote	25-56P-99
dium Measurements, Inc. certifies I	hai the above instrument has been a	alibroled by slandords tracecos	le to the National Institute of Standards on	a technology, or to the collocation loc

Ludium Measurements. Inc. certifies that the above instrument has been calibrated by standards tra-other international Standards Organization members, or have been derived from accepted values of the celibration system conforms to the redurements of ANSI/NCSL 2540-1-1994 and ANSI N323-1997. State of Texas Calibration License No. L .

Appendix D-3. Instrument and Equipment Quality Control Data

Table D-3-1 provides the daily quality-control data obtained by evaluating tritium (H-3), Carbon-14 (C-14) and background standards on the in-house Packard Liquid Scintillation Counter that was used to evaluate all swipes for H-3 and C-14 contamination.

Other pertinent equipment information is as follows:

Packard Liquid Scintillation Counter Model: 2500TR Serial Number: 403848 Calibration date: 10 March 2000 Efficiencies: H-3 = 60.9%; C-14 = 96.3% Packard Check (standard) sources: H-3 (activity < 0.2µCi = 267600 dpm, date 17 May 93) C-14 (activity < 0.1µCi = 136200 dpm, date 17 May 93) Background (May 1993)

Table D-3-1 provides quality-control data for liquidscintillation-counter evaluations in the form of dpm values which were obtained by evaluating H-3 and C-14 radioactive (check source) standards and a background standard. QC data was collected, on a minimum, each day that samples were evaluated and/or equipment tests performed. The upper warning limit (UWL) and lower warning limits (LWL) for H-3 and C-14 were obtained by calculating the mean and adding or subtracting 5% of the mean values for H-13 and C-14. The first ten sets of data (1/19/00-2/4/00) were used to calculate the UWL and LWL for H-3 and C-14. When dpm readings were obtained that exceeded the UWL or LWL, investigations were conducted. If inconsistent operation was evident, the instrument would have been taken out of service and replaced with another instrument or the vendor called for assistance.

Note: Instrument calibration performed by vendor on 3/10/00; the vendor on 4/6/00 performed maintenance service; QC checks were performed by the vendor and it was determined that the instrument was working properly.

Table D-3-1. Quality Control Data - Packard Liquid Scintillation Counter Daily Evaluations of H-3, C-14 and Background Standards

Instrument Use: Beta-Gamma Counting for loose surface contamination

DPM Values: Mean value for H-3 = 170631; Mean value for C-14 = 135664

QC DATA SUMMARY: Daily QC readings did not exceed Upper or Lower Warning Limits (UWL, LWL)

QC		H-3			C-14		Background
Date	DPM	ITWT.	T.WT.	DPM	TIWT.	T.WT.	(C-14
	DII		2002	DIII	UNL	DAD	channel)
				Í			
							DPM
1/19/00	170573	179162	142447	135519	142447	128881	9
1/20/00	170156	179162	142447	135595	142447	128881	.5
1/21/00	169786	179162	142447	135742	142447	128881	11
1/24/00	169724	179162	142447	135937	142447	128881	3
1/27/00	169876	179162	142447	134805	142447	128881	3
1/28/00	169708	179162	142447	135914	142447	128881	7
2/1/00	170667	179162	142447	136135	142447	128881	6
2/2/00	172562	179162	142447	135692	142447	128881	8
2/3/00	171476	179162	142447	135380	142447	128881	б
2/4/00	171783	179162	142447	135919	142447	128881	5
2/7/00	171460	179162	142447	135801	142447	128881	2
2/8/00	171475	179162	142447	135478	142447	128881	7
2/9/00	170774	179162	142447	135808	142447	128881	9
2/10/00	171733	179162	142447	135762	142447	128881	11
2/11/00	171433	179162	142447	135625	142447	128881	4
2/14/00	171341	179162	142447	135616	142447	128881	5
2/15/00	170767	179162	142447	136250	142447	128881	1
2/16/00	171125	179162	142447	136030	142447	128881	6
2/17/00	170122	179162	142447	135130	142447	128881	7
2/18/00	171538	179162	142447	136421	142447	128881	2
2/22/00	171187	179162	142447	1:35679	142447	128881	8
2/23/00	170449	179162	142447	135378	142447	128881	7
2/24/00	170161	179162	142447	135978	142447	128881	10
2/25/00	170708	179162	142447	136330	142447	128881	5
2/28/00	170363	179162	142447	136677	142447	128881	1
2/29/00	169704	179162	142447	136105	142447	128881	7
3/1/00	170894	179162	142447	136667	142447	128881	3
3/2/00	170568	179162	142447	135452	142447	128881	6
3/3/00	169353	179162	142447	135742	142447	128881	11
3/6/00	169402	179162	142447	135795	142447	128881	7
3/7/00	170117	179162	142447	135779	142447	128881	8
3/8/00	170261	179162	142447	136068	142447	128881	1.0
3/9/00	169343	179162	142447	136548	142447	128881	5
3/10/00	170124	179162	142447	136041	142447	128881	6
3/13/00	170482	179162	142447	135716	142447	128881	7

3/14/00	168454	179162	142447	135604	142447	128881	4
3/15/00	169482	179162	142447	136194	142447	128881	4
3/16/00	169274	179162	142447	135433	142447	128881	2
3/17/00	169113	179162	142447	136035	142447	128881	6
3/20/00	168952	179162	142447	135484	142447	128881	11
3/21/00	170135	179162	142447	136005	142447	128881	6
3/22/00	169263	179162	142447	136441	142447	128881	3
3/23/00	169197	1791.62	142447	135920	142447	128881	1
3/24/00	169311	179162	142447	136396	142447	128881	8
3/27/00	169187	179162	142447	134581	142447	128881	5
3/28/00	170543	179162	142447	135647	142447	128881	9
3/29/00	169356	179162	142447	135799	142447	128881	6
3/30/00	168609	179162	142447	136585	142447	128881	6
3/31/00	168114	179162	142447	135115	142447	128881	11
4/3/00	168368	179162	142447	135449	142447	128881	9
4/4/00	169815	179162	142447	136440	142447	128881	1
4/5/00	170876	179162	142447	135859	142447	128881	4
4/6/00	169462	179162	142447	136061	142447	128881	13
4/7/00	169412	179162	142447	135993	142447	128881	1
4/11/00	167894	179162	142447	135638	142447	128881	4
4/12/00	168861	179162	142447	135409	142447	128881	11
4/13/00	167436	179162	142447	135514	142447	128881	6
4/14/00	167649	179162	142447	135896	142447	128881	
4/17/00	168492	179162	142447	135267	142447	128881	2
4/18/00	167391	179162	142447	135261	142447	128881	3
4/19/00	168016	179162	142447	135729	142447	128881	8
4/20/00	168549	179162	142447	134965	142447	128881	8
4/21/00	168412	179162	142447	135453	142447	128881	7
4/24/00	167776	179162	142447	135662	142447	128881	11
4/25/00	167959	179162	142447	136325	142447	128881	12
4/26/00	167591	179162	142447	135617	142447	128881	4
4/27/00	165803	179162	142447	135901	142447	128881	4
4/28/00	167727	179162	142447	135095	142447	128881	6
5/1/00	166078	179162	142447	135681	142447	128881	4
5/2/00	167768	179162	142447	135673	142447	128881	11
5/3/00	1668.59	179162	142447	136478	142447	128881	1
5/4/00	165253	179162	142447	135087	142447	128881	12
5/5/00	166271	179162	142447	134972	142447	128881	4
5/8/00	166602	179162	142447	134720	142447	128881	2
6/9/00	164570	179162	142447	135904	142447	128881	5

Table D-3-2 provides the daily quality-control data obtained by evaluating an Iodine-129 (sealed-source) standard using a Wallac Gamma Counter. Table D-3-2 provides daily quality-control data for the gamma counter. The data was obtained by evaluating a I-129 (sealed-source) standard. The upper warning limit (UWL) and lower warning limits (LWL) for I-129 cpm values were obtained by calculating the mean and adding or subtracting 5% of the mean for I-129 cpm values. The first 30 sets of data (3/23/00) were used to calculate the UWL and LWL for I-129. If dpm readings were obtained that exceeded the UWL or LWL, investigations would have been conducted. If inconsistent operation was evident, the instrument would have been taken out of service and replaced with another instrument or the vendor called for assistance.

Other pertinent equipment information is as follows:

Instrument: Wallac LKB Model 1282 Gamma Counter Serial Number: 324 Calibration Date: 23 Aug 99 Efficiencies: Cr-51 = 3.7 % Check Source: I-129 (rod-shaped; batch number: 8203, catalog number: 1270-102; activity: 0.0221 µCi; calibration date: July 1982; half life: 15.7 million years, 40 keV gamma at 7.5%)

Table D-3-2. Quality Control Data - Daily Evaluations of I-129 Standard Using the Wallac LKB Model 1282 Gamma Counter Instrument Use: Gamma Counting for loose surface contamination CPM Values: Mean = 30261, UWL = 31774, LWL = QC DATA SUMMARY: Daily QC readings did not exceed Upper or Lower Warning Limits (UWL, LWL) Iodine-129 QC Date CPM UWL LWL 3/23/00 3/24/00 3/27/00 3/28/00 3/29/00 3/30/00 3/31/00 4/3/00 4/4/00 4/5/00 4/6/00 4/7/00 4/10/00

NMRC	Di	ving	Bu	ildings	Radiolo	gical	(D&D)
Fin	al	Stat	us	Survey	Report,	June	2000

4/11/00	30431	31774	28748
4/12/00	30305	31774	28748
4/13/00	30385	31774	28748
4/14/00	30127	31774	28748
4/18/00	30366	31774	28748
4/19/00	30307	31774	28748
4/20/00	30572	31774	28748
4/21/00	30099	31774	28748
4/24/00	30320	31774	28748
4/25/00	30233	31774	28748
4/26/00	30649	31774	28748
4/27/00	30340	31774	28748
4/28/00	30307	31774	28748
5/1/00	30366	31774	28748
5/2/00	30053	31774	28748
5/3/00	30202	31774	28748
5/4/00	30235	31774	28748
5/5/00	30612	31774	28748

Tables D-3-3 through D-3-12 provide the daily quality control data obtained by evaluating Carbon-14 or Cobalt-60 check sources using portable beta and beta-gamma detection meters.

Table D-3-3. DAILY QUALITY CONTROL PORTABLE METER CHECK								
SHEET AN	SHEET AND DATA							
Instrument: Ludlum Serial Number: Calibration Date:								
Model 3		114886		7/22/99				
Meter Us	e: Beta-G	amma Scannin	ng for fixed	d or loose	surface			
contamin	ation							
Check So	urce: Cob	alt-60; 1.00)8 μCi; cali	lbration da	te: 18 Nov			
99; Seri	al number	: 58807-530		1400				
Meter Ca	libration	Isotope/Eff	iciencies:	P-32, I-12	5, Cr-51;			
348, 188	, 18							
UWL = Up	per Warnı	ng Limit = m	1ean + 5% of	t mean (mea	n was			
calculat	ed using	the first IC	cpm readin	lgs = 15200	0)			
$\Box W \Box = \Box O$	wer warni	ng Limit = m	iean - 5% Of	mean	17			
QC DATA	SUMMARY:	Daily QC re	adings did	not exceed	upper or			
האתב	DATTERV	PACKEPOUND		COUDCE DE	ADING			
DALE	CHECK	(CPM)		(CPM)	ADING			
	CHECK		CPM	TIMI	T.WT.			
1/19/00	OK	< 300	150000	159600	144400			
1/20/00	OK	< 350	150000	159600	144400			
1/21/00	OK	< 350	150000	159600	144400			
1/24/00	OK	< 300	150000	159600	144400			
1/27/00	OK	< 350	150000	159600	144400			
1/28/00	OK	< 350	150000	159600	144400			
2/1/00	OK	< 400	160000	159600	144400			
2/2/00	OK	< 350	150000	159600	144400			
2/3/00	OK	< 350	150000	159600	144400			
2/4/00	OK	< 300	160000	159600	144400			
2/7/00	OK	< 300	150000	159600	144400			
2/8/00	OK	< 350	150000	159600	144400			
2/9/00	OK	< 350	150000	159600	144400			
2/10/00	OK	< 350	150000	159600	144400			
2/11/00	OK	< 350	150000	159600	144400			
2/14/00	OK	< 350	150000	159600	144400			
2/15/00	OK	< 350	150000	159600	144400			
2/16/00	OK	< 350	150000	159600	144400			
2/17/00	OK	< 300	150000	159600	144400			
2/18/00	OK	< 300	150000	159600	144400			
2/22/00	OK	< 300	150000	159600	144400			
2/23/00	OK	< 350	150000	159600	144400			

2/24/00	OK	< 350	150000	159600	144400
2/25/00	OK	< 350	150000	159600	144400
2/28/00	OK	< 300	150000	159600	144400
2/29/00	OK	< 300	150000	159600	144400
3/1/00	OK	< 350	150000	159600	144400
3/2/00	OK	< 350	150000	159600	144400
3/3/00	OK	< 300	150000	159600	144400
3/6/00	OK	< 350	150000	159600	144400
3/7/00	OK	< 350	150000	159600	144400
3/8/00	OK	< 350	150000	159600	144400
3/9/00	OK	< 350	150000	159600	144400
3/10/00	OK	< 350	150000	159600	144400
3/13/00	OK	< 350	150000	159600	144400
3/14/00	OK	< 350	150000	159600	144400
3/15/00	OK	< 350	150000	159600	144400
3/16/00	OK	< 350	150000	159600	144400
3/17/00	OK	< 350	150000	159600	144400
3/20/00	OK	< 350	150000	159600	144400
3/21/00	OK	< 350	150000	159600	144400
3/22/00	OK	< 300	150000	159600	144400
3/23/00	OK	< 350	150000	159600	144400
3/24/00	OK	< 350	150000	159600	144400
3/27/00	OK	< 350	150000	159600	144400
3/28/00	OK	< 350	150000	159600	144400
3/29/00	OK	< 350	150000	159600	144400
3/30/00	OK	< 350	150000	159600	144400
3/31/00	OK	< 350	150000	159600	144400
4/3/00	OK	< 350	150000	159600	144400
4/4/00	OK	< 400	150000	159600	144400
4/5/00	OK	< 350	150000	159600	144400
4/6/00	OK	< 350	150000	159600	144400
4/7/00	OK	< 350	150000	159600	144400
4/10/00	OK	< 350	150000	159600	144400
4/11/00	OK	< 350	150000	159600	144400
4/12/00	OK	< 350	150000	159600	144400
4/13/00	OK	< 350	150000	159600	144400
4/14/00	OK	< 350	150000	159600	144400
4/17/00	OK	< 350	150000	159600	144400
4/18/00	OK	< 350	150000	159600	144400
4/19/00	OK	< 350	150000	159600	144400
4/20/00	OK	< 350	150000	159600	144400
4/21/00	OK	< 350	150000	159600	144400
4/24/00	OK	< 350	150000	159600	144400
4/25/00	OK	< 350	150000	159600	144400
4/26/00	OK	< 350	150000	159600	144400
					C. Statement of the local division of the lo

4/27/00	OK	< 350	150000	159600	144400
4/28/00	OK	< 350	150000	159600	144400
5/1/00	OK	< 350	150000	159600	144400
5/2/00	OK	< 350	150000	159600	144400
5/3/00	OK	< 350	150000	159600	144400
5/4/00	OK	< 350	150000	159600	144400
5/5/00	OK	< 350	150000	159600	144400

Table D-3-4. DAILY QUALITY CONTROL PORTABLE METER CHECK						
SHEET AN	D DATA		1			
Instrume Model 3	Model 3 132946 2/22/00					
Instrume	nt Use: E	Seta-Gamma Sc	canning for	fixed or 1	oose	
surface	contamina	ition				
Check So	urce: Cob	alt-60; 1.00)8 µCi; cal.	ibration da	te: 18 Nov	
99; Seri	al number	: 58807-530				
Meter Ca	libration	Isotope/Eff	iciencies:	I-125, Cr-	51: 16%,	
2.68						
UWL = Up	per Warni	ng Limit = m	tean $+ 58$	(mean was c	alculated	
using th	e first 1	0 cpm readin	lgs = 16000	0)		
LWL = LO	wer Warni	ng Limit = n	<u>nean - 5%</u>		27	
QC DATA	SUMMARY :	Daily QC re	adings did	not exceed	Upper or	
	LOI	wer warning	Limits (UWL	·		
DATE	BATTERY	BACKGROUND	CHECK	SOURCE RE	ADING	
	CHECK	(CPM)	CEN	(CPM)	7.527	
2/15/00	01/	1 200	CPM 1 COOCIO			
3/15/00	OK	< 300	160000	168000	152000	
3/17/00	OK	< 300	160000	160000	152000	
3/20/00	OK	< 300	160000	168000	152000	
3/21/00	OK	< 350	160000	168000	152000	
3/22/00	OK	< 300	160000	168000	152000	
3/23/00	OK	< 350	160000	168000	152000	
3/24/00	OK	< 350	160000	168000	152000	
3/27/00	OK	< 350	160000	168000	152000	
3/28/00	OK	< 350	160000	168000	152000	
3/29/00	OK	< 350	160000	168000	152000	
3/30/00	OK	< 350	160000	168000	152000	
3/31/00	OK	< 350	160000	168000	152000	
4/3/00	OK	< 350	160000	168000	152000	
4/4/00	OK	< 400	160000	168000	152000	
4/5/00	OK	< 350	160000	168000	152000	
4/6/00	OK	< 350	160000	168000	152000	
4/7/00	OK	< 350	160000	168000	1.52000	
4/10/00	OK	< 300	160000	168000	152000	
4/11/00	OK	< 350	160000	168000	152000	
4/12/00	ÛК	< 350	160000	168000	152000	
4/13/00	OK	< 350	160000	168000	152000	
4/14/00	OK	< 350	160000	168000	152000	
4/17/00	OK	< 300	160000	168000	152000	
4/18/00	OK.	< 350	160000	168000	152000	
4/19/00	OK	< 350	160000	168000	152000	

4/20/00	OK	< 300	160000	168000	152000
4/21/00	OK	< 300	160000	168000	152000
4/24/00	OK	< 350	160000	168000	152000
4/25/00	OK	< 300	160000	168000	152000
4/26/00	OK	< 300	160000	168000	152000
4/27/00	OK	< 300	160000	168000	152000
4/28/00	OK	< 300	160000	168000	152000
5/1/00	OK	< 350	160000	168000	152000
5/2/00	OK	< 350	160000	168000	152000
5/3/00	OK	< 350	160000	168000	152000
5/4/00	OK	< 350	160000	168000	152000
5/5/00	OK	< 300	160000	168000	152000

Table D-	3-5. DAI	LY QUALITY C	CONTROL POR	TABLE METER	CHECK
SHEET AN	D DATA				
Instrume	nt: Ludlu	m Serial N	umber:	Calibrati	on Date:
Model 3		121461		8/12/99	
Instrume	nt Use: E	Seta-Gamma Sc	anning for	fixed or l	oose
surface	contamina	tion			
Check So	urce: Cob	alt-60; 1.00)8 µCi; cal	ibration da	te: 18 Nov
99; Seri	al number	: 58807-530			
Meter Ca	libration	Isotope/Eff	iciencies;	P-32, I-12	5, Cr-51:
178, 298	, 18				
UWL = Up	per Warni	.ng Limit = n	nean + 5%	(mean was c	alculated
using th	e first 1	0 cpm readin	igs = 150, 50	00)	
TMT = TO	wer Warni	ng Limit = m	1ean - 5%		
QC DATA	SUMMARY:	Daily QC re	adings did	not exceed	Upper or
DAME	LOL	wer warning	Limits (UWL	, LWL)	5 5 T) 1 4
DATE	BATTERY	BACKGROUND	CHECK	SOURCE RE	ADING
	CHECK	(CPM)	CDM		T 5.71
1/10/00	01	< 200			142075
1/19/00	OK	< 300	150000	158025	142975
1/20/00	OK	< 350	145000	158025	142975
1/21/00	OK	< 300	145000	158025	142975
1/24/00	OK	< 400	150000	158025	142975
1/29/00	OK	< 350	150000	158025	142975
2/1/00	OK	< 350	160000	158025	142975
2/1/00	OK	< 300	160000	158025	142975
2/2/00	OK	< 350	150000	159025	142975
2/4/00	OK	< 350	150000	158025	142975
2/7/00	OK	< 350	160000	158025	142975
2/8/00	OK	< 350	150000	158025	142975
2/9/00	OK OK	< 300	150000	158025	142975
2/10/00	OK	< 300	150000	158025	142975
2/11/00	OK	< 350	150000	158025	142975
2/14/00	OK	< 350	150000	158025	142975
2/15/00	OK	< 350	150000	158025	142975
2/16/00	OK	< 350	150000	158025	142975
2/17/00	OK	< 300	150000	158025	142975
2/18/00	OK	< 300	150000	158025	142975
2/22/00	OK	< 300	150000	158025	142975
2/23/00	OK	< 350	150000	158025	142975
2/24/00	OK	< 350	1.50000	158025	142975
2/25/00	OK	< 350	150000	158025	142975
2/28/00	OK	< 300	150000	158025	142975
2/29/00	OK	< 350	150000	158025	142975

3/1/00	OK	< 350	150000	158025	142975
3/2/00	OK	< 350	150000	158025	142975
3/3/00	OK	< 300	150000	158025	142975
3/6/00	OK	< 350	150000	158025	142975
3/7/00	OK	< 350	150000	158025	142975
3/8/00	OK	< 350	150000	158025	142975
3/9/00	OK	< 350	150000	158025	142975
3/10/00	OK	< 300	150000	158025	142975
3/13/00	OK	< 350	150000	158025	142975
3/14/00	OK	< 350	150000	158025	142975
3/15/00	OK	< 350	150000	158025	142975
3/16/00	OK	< 350	150000	158025	142975
3/17/00	OK	< 350	150000	158025	142975
3/20/00	OK	< 350	150000	158025	142975
3/21/00	OK	< 300	150000	158025	142975
3/22/00	OK	< 300	150000	158025	142975
3/23/00	OK	< 350	150000	158025	142975
3/24/00	OK	< 350	150000	158025	142975
3/27/00	OK	< 350	150000	158025	142975
3/28/00	OK	< 350	150000	158025	142975
3/29/00	OK	< 350	150000	158025	142975
3/30/00	OK	< 300	150000	158025	142975
3/31/00	OK	< 300	150000	158025	142975
4/3/00	OK	< 300	150000	158025	142975
4/4/00	OK	< 350	150000	158025	142975
4/5/00	OK	< 350	150000	158025	142975
4/6/00	OK	< 350	150000	158025	142975
4/7/00	OK	< 350	150000	158025	142975
4/10/00	OK	< 350	150000	158025	142975
4/11/00	OK	< 350	150000	158025	142975
4/12/00	OK	< 350	150000	158025	142975
4/13/00	OK	< 300	150000	158025	142975
4/14/00	ОК	< 300	150000	158025	142975
4/17/00	OK	< 350	150000	158025	142975
4/18/00	OK	< 350	150000	158025	142975
4/19/00	OK	< 350	150000	158025	142975
4/20/00	OK	< 350	150000	158025	142975
4/21/00	OK	< 350	150000	158025	142975
4/24/00	OK	< 350	150000	158025	142975
4/25/00	OK	< 350	150000	158025	142975
4/26/00	OK	< 350	150000	158025	142975
4/27/00	OK	< 350	150000	158025	142975
4/28/00	OK	< 300	150000	158025	142975
5/1/00	OK	< 300	150000	158025	142975
5/2/00	OK	< 350	150000	158025	142975

5/3/00	OK	< 300	150000	158025	142975
5/4/00	OK	< 350	150000	158025	142975
5/5/00	OK	< 350	150000	158025	142975

.

Table D-3-6. DAILY QUALITY CONTROL PORTABLE METER CHECK							
SHEET AND DATA							
Instrume	Instrument: Ludlum Serial Number: Calibration Date:						
Model 3		121730		7/22/99			
Instrume	nt Use: B	eta-Gamma Sc	anning for	fixed or l	oose		
surface	contamina	tion					
Check So	urce: Cob	alt-60; 1.00)8 μCi; cal:	ibration da	te: 18 Nov		
99; Seri	al number	: 58807-530		1111			
Meter Ca	libration	Isotope/Eff	liciencies:	P-32, I-12	5, Cr-51:		
328, 168	, 18						
UWL = Up	per Warni	ng Limit = m	nean + 5%	(mean was c	alculated		
using th	e first 1	0 cpm readin	igs = 155,00	00)			
LWL = LO	wer Warni	ng Limit = m	1ean - 58				
QC DATA	SUMMARY:	Daily QC re	adings did	not exceed	Upper or		
	Lov	wer Warning :	Límits (UWL	, LWL)			
DATE	BATTERY	BACKGROUND	CHECK	SOURCE RE	ADING		
	СНЕСК	(CPM)	The second s	(CPM)			
			CPM	UWL	LWL		
1/19/00	OK	< 300	150000	162750	147250		
1/20/00	OK	< 350	150000	162750	147250		
1/21/00	OK	< 300	150000	162750	147250		
1/24/00	OK	< 350	150000	162750	147250		
1/27/00	OK	< 350	1.50000	162750	147250		
1/28/00	OK	< 350	160000	162750	147250		
2/1/00	OK	< 300	160000	162750	147250		
2/2/00	OK	< 300	160000	162750	147250		
2/3/00	OK	< 350	160000	162750	147250		
2/4/00	OK	< 300	160000	162750	147250		
2/7/00	OK	< 350	160000	162750	147250		
2/8/00	OK	< 350	160000	162750	147250		
2/9/00	OK	< 350	160000	162750	147250		
2/10/00	OK	< 350	160000	162750	147250		
2/11/00	OK	< 350	150000	162750	147250		
2/14/00	OK	< 350	160000	162750	147250		
2/15/00	OK	< 350	160000	162750	147250		
2/16/00	OK	< 350	160000	162750	147250		
2/17/00	OK	< 350	150000	162750	147250		
2/18/00	OK	< 300	160000	162750	147250		
2/22/00	OK	< 350	160000	162750	147250		
2/23/00	OK	< 300	160000	162750	147250		
2/24/00	OK	< 300	160000	162750	147250		
2/25/00	OK	< 300	160000	162750	147250		
2/28/00	OK	< 300	160000	162750	147250		
2/29/00	OK	< 350	160000	162750	147250		

3/1/00	OK	< 350	160000	162750	147250
3/2/00	OK	< 350	160000	162750	147250
3/3/00	OK	< 350	160000	162750	147250
3/6/00	OK	< 350	160000	162750	147250
3/7/00	OK	< 350	160000	162750	147250
3/8/00	OK	< 350	160000	162750	147250
3/9/00	OK	< 350	160000	162750	147250
3/10/00	OK	< 300	160000	162750	147250
3/13/00	OK	< 350	160000	162750	147250
3/14/00	OK	< 350	160000	162750	147250
3/15/00	PROBE	BROKEN	ON	THIS	DATE

Table D-3-7. DAILY QUALITY CONTROL PORTABLE METER CHECK SHEET AND DATA							
Instrument: Ludlum Serial Number: Calibration Date:							
Model 3	Model 3 121816 2/22/00						
Instrume	nt Use: B	eta Scanning	for fixed	or loose s	urface		
contamin	ation						
Check So 99; Seri	urce: CAR al number	BON-14; 0.14 : 58809-530	l5 μCi; cal:	ibration da	te: 18 Nov		
Meter Ca	libration	Isotope/Eff	iciencies:	C-14, P-32	, P-33,		
S-35, CA	-45: 78,	35%, 14%, 7%	5, 78				
UWL = Up	per Warni	ng Limit = m	nean + 5%	(mean was c	alculated		
using th	e first 1	0 cpm readin	ngs = 26,000))			
IWL = LO	wer Warni	ng Limit = n	nean - 5%				
QC DATA	SUMMARY ;	Daily QC re	adings did	not exceed	Upper or		
	Lou	wer Warning	Limits (UWL	, LWL)			
DATE	BATTERY	BACKGROUND	CHECK	SOURCE REA	ADING		
	CHECK	(CPM)		(CPM)			
2/27/00	077	1.00	СРМ		LWL		
3/1//00	OK	< 100	26000	27300	24700		
3/20/00	OK	< 100	26000	27300	24700		
3/21/00	OK	< 100	26000	27300	24700		
3/22/00	OK	< 100	26000	27300	24700		
3/23/00	OK	< 100	26000	27300	24700		
3/24/00	OK	< 100	26000	27300	24700		
3/29/00	OK	< 100	26000	27300	24700		
3/29/00	OK	< 100	26000	27300	24700		
3/30/00	OK	< 100	26000	27300	24700		
3/31/00	OK	< 100	26000	27300	24700		
4/3/00	OK	< 100	26000	27300	24700		
4/4/00	OK OK	< 100	26000	27300	24700		
4/5/00	ОК	< 100	26000	27300	24700		
4/6/00	OK	< 100	26000	27300	24700		
4/7/00	OK	< 100	26000	27300	24700		
4/10/00	OK	< 100	26000	27300	24700		
4/11/00	OK	< 100	26000	27300	24700		
4/12/00	OK	< 100	26000	27300	24700		
4/13/00	OK	< 100	26000	27300	24700		
4/14/00	OK	< 1.00	26000	27300	24700		
4/17/00	OK	< 100	26000	27300	24700		
4/18/00	OK	< 100	26000	27300	24700		
4/19/00	OK	< 100	26000	27300	24700		
4/20/00	OK	< 100	26000	27300	24700		
4/21/00	OK	< 100	26000	27300	24700		

4/24/00	OK	< 100	26000	27300	24700
4/25/00	OK	< 100	26000	27300	24700
4/26/00	OK	< 100	26000	27300	24700
4/27/00	OK	< 100	26000	27300	24700
4/28/00	OK	< 100	26000	27300	24700
5/1/00	OK	< 100	26000	27300	24700
5/2/00	OK	< 100	26000	27300	24700
5/3/00	OK	< 100	26000	27300	24700
5/4/00	OK	< 100	26000	27300	24700
5/5/00	OK	< 100	26000	27300	24700

•

Table D-3-8. DAILY QUALITY CONTROL PORTABLE METER CHECK									
SHEET AND DATA									
Instrument: Ludlum Model 3		m Serial Nu 54537	Serial Number: 54537		Calibration Date: 7/22/99				
Instrument Use: Beta Scanning for fixed or loose surface									
contamination									
Check Source: CARBON-14; 0.145 µCi; calibration date: 18 Nov									
99; Serial number: 58809-530									
Meter Calibration Isotope/Efficiencies: C-14, P-32, P-33,									
S-35, CA-45: 78, 358, 148, 78, 78									
UWL = Upper Warning Limit = mean + 7% (mean was calculated									
using the first 10 cpm readings = 20600)									
LWL = Lower Warning Limit = mean - 78									
QC DATA SUMMARY: Daily QC readings did not exceed Upper or									
Lower Warning Limits (UWL, LWL)									
DATE	BATTERY	BACKGROUND	CHECK SOURCE READING						
	CHECK	(CPM)		(CPM)					
			CPM	UWL	LŴL				
1/19/00	OK	< 50	19000	22042	19570				
1/20/00	OK	< 100	18000	22042	19570				
1/21/00	OK	< 100	21000	22042	19570				
1/24/00	OK	< 100	21000	22042	19570				
1/27/00	OK	< 50	21000	22042	19570				
1/28/00	OK	< 50	21000	22042	19570				
2/1/00	OK	< 50	22000	22042	19570				
2/2/00	OK	< 50	21000	22042	19570				
2/3/00	OK	< 50	21000	22042	19570				
2/4/00	OK	< 50	21000	22042	19570				
2/7/00	OK	< 100	20000	22042	19570				
2/8/00	OK	< 100	21000	22042	19570				
2/9/00	OK	< 100	21000	22042	19570				
2/10/00	OK	< 100	21000	22042	19570				
2/11/00	OK	< 100	21000	22042	19570				
2/14/00	OK	< 100	22000	22042	19570				
2/15/00	OK	< 150	21000	22042	19570				
2/16/00	OK	< 100	22000	22042	19570				
2/17/00	OK	< 100	21000	22042	19570				
2/18/00	OK	< 150	21000	22042	19570				
2/22/00	OK	< 100	21000	22042	19570				
2/23/00	OK	< 100	22000	22042	19570				
2/24/00	OK	< 100	22000	22042	19570				
2/25/00	OK	< 100	21000	22042	19570				
2/28/00	OK	< 100	21000	22042	19570				
2/29/00	OK	< 100	21000	22042	19570				

3/1/00	OK	< 100	20000	22042	19570
3/2/00	OK	< 100	21000	22042	19570
3/3/00	OK	< 100	20000	22042	19570
3/6/00	OK	< 100	20000	22042	19570
3/7/00	OK	< 100	22000	22042	19570
3/8/00	OK	< 100	21000	22042	19570
3/9/00	OK	< 100	21000	22042	19570
3/10/00	OK	< 100	21000	22042	19570
3/13/00	OK	< 100	21000	22042	19570
3/14/00	OK	< 100	22000	22042	19570
3/15/00	OK	< 100	21000	22042	19570
3/16/00	OK	< 100	22000	22042	19570
3/17/00	OK	< 100	22000	22042	19570
3/20/00	OK	< 100	22000	22042	19570
3/21/00	OK	< 100	22000	22042	19570
3/22/00	OK	< 100	22000	22042	19570
3/23/00	OK	< 100	22000	22042	19570
3/24/00	OK	< 100	22000	22042	19570
3/27/00	OK	< 100	22000	22042	19570
3/28/00	OK	< 100	21000	22042	19570
3/29/00	OK	< 100	22000	22042	19570
3/30/00	OK	< 100	22000	22042	19570
3/31/00	OK	< 100	22000	22042	19570
4/3/00	OK	< 100	22000	22042	19570
4/4/00	OK	< 100	22000	22042	19570
4/5/00	OK	< 100	22000	22042	19570
4/6/00	OK	< 100	21000	22042	19570
4/7/00	OK	< 100	21000	22042	19570
4/10/00	OK	< 100	22000	22042	19570
4/11/00	OK	< 100	21000	22042	19570
4/12/00	OK	< 100	21000	22042	19570
4/13/00	OK	< 100	22000	22042	19570
4/14/00	OK	< 100	22000	22042	19570
4/17/00	OK	< 100	22000	22042	19570
4/18/00	l ok	< 100	22000	22042	19570
4/19/00	OK	< 100	22000	22042	19570
4/20/00	OK	< 100	22000	220.42	19570
4/21/00	OK	< 100	22000	22042	19570
4/24/00	OK	< 100	22000	22042	19570
4/25/00	OK	< 100	21000	22042	19570
4/26/00	OK	< 100	22000	22042	19570
4/27/00	OK	< 100	22000	22042	19570
4/28/00	OK	< 100	21000	22042	19570
5/1/00	OK	< 100	22000	22042	19570
5/2/00	OK	< 100	21000	22042	19570
5/3/00	OK	< 100	22000	22042	19570
--------	----	-------	-------	-------	-------
5/4/00	OK	< 100	22000	22042	19570
5/5/00	OK	< 100	22000	22042	19570

Table D-	3-9. DAI	LY QUALITY C	CONTROL POR	TABLE METER	CHECK	
SHEET AN	D DATA					
Instrume	nt: Bicro	n Serial N	umber:	Calibrati	on Date:	
Survey 5	0	B602G		7/22/99		
Instrume	nt Use: B	eta Scanning	for fixed	or loose s	urface	
contamin	ation					
Check So	Check Source: CARBON-14; 0.145 µCi; calibration date: 18 Nov					
99; Seri	al number	: 58809-530				
Meter Ca	libration	Isotope/Eff	iciencies:	P-32, P-33	, s-35,	
CA-45: 3	18, /8, 1	48, 78				
dn = 1mn	per Warnı	ng Limit =	mean + 5%	(mean was	<u></u>	
	ed using	the first 10	cpm readin	ngs = 2150	0)	
LWL = LO	wer warni	ng Limit = n	iean - 5%	and an and	Trooper and	
QC DATA	SUMMARY:	Daily Qu re	adings did	NOL exceed	upper or	
	VOL Vaammka	PACKCROUND	LIMILS (UWL	(CONDOE DE:	ADING	
DATE	CHECK	BACKGROUND	CHECr	(CDM)	ADING	
	CHECK	(CPM)	CDM		TGT	
1/10/00	OK	< 50	20000	22575		
1/20/00	OK	< 50	20000	22.575	20425	
1/21/00	OK	< 80	21000	22575	20425	
1/21/00	OK	< 50	22000	22575	20425	
1/27/00	OK	< 50	22000	22575	20425	
1/29/00	OK	< 50	22000	22575	20425	
2/1/00	OK	< 60	21000	22575	20425	
2/2/00	OK	< 50	21000	22575	20425	
2/3/00	OK	< 60	21000	22575	20425	
2/4/00	OK	< 60	21000	22575	20425	
2/7/00	OK	< 80	21000	22575	20425	
2/8/00	OK	< 100	20000	22575	20425	
2/9/00	OK	< 100	21000	22575	20425	
2/10/00	OK	< 100	21000	22575	20425	
2/11/00	OK	< 60	21000	22575	20425	
2/14/00	OK	< 100	20000	22575	20425	
2/15/00	OK	< 100	20000	22575	20425	
2/16/00	OK	< 100	21000	22575	20425	
2/17/00	OK.	< 100	21000	2257.5	20425	
2/18/00	OK	< 100	21000	22575	20425	
2/22/00	OK	< 160	21000	22575	20425	
2/23/00	OK	< 100	21000	22575	20425	
2/24/00	OK	< 80	21000	22575	20425	
2/25/00	OK	< 80	21000	2257.5	20425	
2/28/00	OK	< 100	21.000	22575	20425	
2/29/00	OK	< 80	21000	22575	20425	

3/1/00	OK	< 80	20000	22575	20425
3/2/00	OK	< 100	20000	22575	20425
3/3/00	OK	< 100	20000	22575	20425
3/6/00	OK	< 100	21000	22575	20425
3/7/00	OK	< 100	22000	22575	20425
3/8/00	OK	< 80	22000	22575	20425
3/9/00	OK	< 80	21000	22575	20425
3/10/00	OK	< 80	22000	22575	20425
3/13/00	OK	< 50	21000	22575	20425
3/14/00	OK	< 50	21000	22575	20425
3/15/00	OK	< 80	22000	22575	20425
3/16/00	OK	< 50	21000	22575	20425
3/17/00	OK	< 50	21000	22575	20425
3/20/00	OK	< 80	20000	22575	20425
3/21/00	OK	< 80	22000	22575	20425
3/22/00	OK	< 80	22000	22575	20425
3/23/00	OK	< 100	22000	22575	20425
3/24/00	OK	< 80	21000	22575	20425
3/27/00	OK	< 80	22000	22575	20425
3/28/00	OK	< 80	22000	22575	20425
3/29/00	OK	< 100	21000	22575	20425
3/30/00	OK	< 80	21000	22575	20425
3/31/00	OK	< 80	22000	22575	20425
4/3/00	OK	< 80	21000	22575	20425
4/4/00	OK	< 80	22000	22575	20425
4/5/00	OK	< 80	22000	22575	20425
4/6/00	OK	< 80	22000	22575	20425
4/7/00	OK	< 80	22000	22575	20425
4/10/00	OK	< 60	21000	22575	20425
4/11/00	OK	< 80	21000	2257:5	20425
4/12/00	OK	< 80	21000	22575	20425
4/13/00	OK	< 80	21000	22575	20425
4/14/00	OK	< 80	22000	2257.5	20425
4/17/00	OK	< 60	22000	22575	20425
4/18/00	OK	< :30	22000	22575	20425
4/19/00	OK	< 80	21000	22575	20425
4/20/00	OK	08 >	21000	22575	20425
4/21/00	OK	< 80	21000	22575	20425
4/24/00	OK	< 80	22000	22575	20425
4/25/00	OK	< 80	22000	22575	20425
4/26/00	OK.	< 180	21000	22 57.5	20425
4/27/00	OK	< 80	22000	22575	20425
4/28/00	OK	< 80	22000	22575	20425
5/1/00	OK	< 80	21000	22575	20425
5/2/00	OK	< 80	22000	22575	20425

5/3/00	OK	< 80	22000	22575	20425
5/4/00	OK	< 80	22000	22575	20425
5/5/00	OK	< 80	22000	22575	20425

Table D- SHEET AN	3-10. DA D DATA	ILY QUALITY	CONTROL POP	RTABLE METE	R CHECK
Instrume Model 3	nt: Ludlu	m Serial N 121816	umber:	Calibrati 2/22/00	on Date:
Instrume	nt Use: B	eta Scanning	for fixed	or loose s	urface
contamination					
Check So	urce: CAR	BON-14; 0.14	5 µCi; cali	ibration da	te: 18 Nov
99; Seri	al number	: 58809-530	-		
Meter Ca	libration	Isotope/Eff	iciencies:	C-14, P-32	, P-33,
S-35, CA	-45: 78,	35%, 14%, 7%	5, 78		
UWL = Up	per Warni	ng Limit =	mean + 3 s:	igma = 261	39
(mean wa	s calcula	ted using al	l cpm read:	ings = 213	44; sigma
= standa	rd deviat	ion = 1600)			5.4.0
LWL = LO	wer Warni	ng Limit = m	iean - 3 *	sigma = 16	548
QC DATA	SUMMARY:	Jaily QC re	adings did Limits (UWL	not exceed LWL)	upper or
DATE	BATTERY	BACKGROUND	CHECK	SOURCE REA	ADING
21112	CHECK	(CPM)		(CPM)	
		1-11-2	CPM	UWL	LWL
1/19/00	OK	< 100	20000	26139	16548
1/20/00	OK	< 100	15000	26139	16548
1/21/00	OK	< 100	21000	26139	16548
1/24/00	OK	< 100	24000	26139	16548
1/27/00	OK	< 50	25000	26139	16548
1/28/00	OK	< 50	21000	26139	16548
2/1/00	OK	< 50	21000	26139	16548
2/2/00	OK	< 50	21000	26139	16548
2/3/00	OK	< 50	21000	26139	16548
2/4/00	OK	< 50	22000	26139	16548
2/7/00	OK	< 1 <u>00</u>	21000	26139	16548
2/8/00	OK	< 100	22000	26139	16548
2/9/00	OK	< 100	21000	26139	16548
2/10/00	OK	< 100	21000	26139	16548
2/11/00	OK	< 100	20000	26139	16548
2/14/00	OK	< 100	21000	26139	16548
2/15/00	OK	< 100	21000	26139	16548
2/16/00	OK	< 100	22000	26139	16548
2/17/00	OK	< 80	21000	26139	16548
2/18/00	OK	< 100	21000	26139	16548
2/22/00	OK	< 100	22000	26139	16548
2/23/00	OK	< 100	22000	26139	16548
2/24/00	OK	< 100	22000	26139	16548
2/25/00	OK	< 100	22000	26139	16548
2/28/00	OK	< 100	20000	26139	16548

2/29/00	OK	< 100	22000	26139	16548
3/1/00	OK	< 100	21000	26139	16548
3/2/00	OK	< 100	22000	26139	16548
3/3/00	OK	< 100	21000	26139	16548
3/6/00	OK	< 100	22000	26139	16548
3/7/00	OK	< 100	21000	26139	16548
3/8/00	OK	< 100	24000	26139	16548
3/9/00	METER	TAKEN	OUT	OF	SERVICE

Table D-	3-11. DA	ILY QUALITY	CONTROL POP	RTABLE METE	R CHECK
SHEET AN	D DATA			1	
Model 2350-1 157674 11/18/99					on Date:
Probe Mo	del/Seria	1 Number: 43	- Probe	$e: 126 \text{cm}^2$ g	as
37/14931	4		dete	ctor	
Instrume	nt Use: B	eta Scanning	for fixed	surface	
contamin	ation	157			
Check So 99; Seri	urce: CAR al number	BON-14; 0.14 : 58809-530	15 μCi; cali	bration da	te: 18 Nov
Notes:	Mid-day a	nd End-of-da	y readings	were recor	ded on
days tha	t the ins	trument was	used for ra	adiological	
monitori	ng of sur	vey units.	Daily QC ch	necks consi	sted of
comparin	g same-da	y CPM values	5.		
DATE	BATTERY	CPM, POST	CPM, POST	CPM, MID	CPM, END
	VOLTAGE	PURGING	PURGING	DAY	OF DAY
		WITH P-10	WITH P-10	READING	READING
		GAS	GAS		
		(T-O)	(T-10)		
3/28/00	5.9	44922	44740		
3/29/00	5.5	43781	44538		
3/30/00	5.7	44433	44391		
3/31/00	5.6	44214	44052	44992	43262
4/3/00	5.5	45342	45686		
4/4/00	5.6	44675	44888		
4/5/00	5.5	44515	44781	44529	
4/6/00	5.5	45079	45010	44935	45651
4/7/00	5.4	44616	44779	46105	
4/10/00	5.4	44584	44126		
4/11/00	5.4	45057	45266	45183	
4/12/00	5.3	44310	45043	44624	
4/13/00	5.3	45633	45571		
4/14/00	5.3	44848	44448		42976
4/17/00	5.2	44850	44678		
4/18/00	5.2	45564	45241	45614	44376
4/19/00	5.2	45752	45743		45756
4/20/00	5.1	44782	44600	45013	
4/21/00	5.1	44832	45178		
4/24/00	6.3	45560	45554	44437	
4/25/00	6.1	45446	45212		
4/26/00	6.0	44910	45422	44943	44122
4/27/00	6.0	45002	45138	44.387	
4/28/00	6.0	44118	44554		43771
5/1/00	5.9	46016	46555	46466	45494

5/2/00	5.8	44891	44695	44040	43221
5/3/00	5.7	46076	45761	43370	
5/4/00	5.6	44819	44616		
5/5/00	5.6	42944	NOT	USED	

Table D-	3-12. DA	ILY QUALITY	CONTROL POI	RTABLE METE	R CHECK
SHEET AN	D DATA				
Instrume	nt: Ludlu	m Serial N	umber:	Calibrati	on Date:
Model 23	50-1	157674		11/18/99	
Probe Mo	del/Seria	1 Number: 43	3- Prob	e: 582cm ² g	as
37/14894	7		dete	ctor	
Instrume	nt Use: B	eta Scanning	g for fixed	surface	
contamin	ation				
Check So	urce: CAR	BON-14; 0.14	15 µCi; cal:	ibration da	te: 18 Nov
99; Seri	al number	: 58809-530			
Notes:	Mid-day a	nd end-of-da	y readings	were recor	ded on
days tha	t the ins	trument was	used for ra	adiological	
monitori	ng of sur	vey units.	Daily QC ch	necks consi	sted of
comparin	g same-da	y_CPM values	5.		,
DATE	BATTERY	CPM, POST	CPM, POST	CPM, MID	CPM, END
	VOLTAGE	PURGING	PURGING	DAY	OF DAY
		WITH P-10	WITH P-10	READING	READING
		GAS	GAS		
		<u>(T-</u> 0)	(T-10)		l
3/28/00	5.9	44837	44672		
3/29/00	5.4	43313	43322	_	
3/30/00	5.7	45122	43346		
3/31/00	5.6	43939	43725	43839	42103
4/3/00	5.6	43532	43488		
4/4/00	5.6	42930	42763		39020
4/5/00	5.5	44658	45066	44073	
4/6/00	5.5	43345	43194	44745	43264
4/7/00	5.4	48062	47715	44503	
4/10/00	5.4	43402	43626		
4/11/00	5.4	44316	43874	42530	
4/12/00	5.4	44847	43458	41096	40547
4/13/00	5.3	43555	43591		
4/14/00_	5.3	44004	43881		43996
4/17/00	5.2	44876	44291		
4/18/00	5.2	45312	45565	42341	44979
4/19/00	5.2	47748	48067		43140
4/20/00	5.1	47982	48215	4.5984	
4/21/00	5.1	47204	46576		
4/24/00	6.3	45551	45282		
4/25/00	6.1	45396	46749		
4/26/00	6.0	44868	44387		
4/27/00	6.0	45295	45009	42548	
4/28/00	6.0	45525	45424		43220
5/1/00	5.9	44254	43733	43161	432.50

5/2/00	5.8	47994	47003	43598	
5/3/00	5.7	45166	45155	39031	
5/4/00	5.6	45259	45494		
5/5/00	5.7	42944	41404		
5/8/00	5.7	45554	44264		
5/9/00	5.7	44210	44036		

Appendix D-4. Quality Assurance Checklist (included in sections of the final status survey report)
Building number or location:
Monitoring Date(s): Appropriate and satisfactory daily and weekly calibration checks were completed for the selected survey monitoring equipment and instruments. Selected meters with current calibrations for the isotopes of concern. Survey maps or floor plans (for all survey units) and grid maps (for known RAM-use areas) were available. Background checks were performed when impacted areas were monitored. For known RAM-use survey units, additional checks were performed (workbenches, shelves, sinks, basins, traps, floor at sinks and entrances, wall at hoods and sinks, and/or other areas as deemed appropriate). Mean values, single sample values, and background values were below the action limits, indicating no contamination was found. Representative floor plans, sketches, and photographs were available. Deviation from survey plans and objectives were documented.
All selected survey units passed D&D tests; units are recommended for unrestricted release.
Additional Checks/Data passed D&D tests; unit is recommended for unrestricted release.

Notes: Tests include complying with applicable elements of the MARSSIM Final Status Survey checklist, mean-value and singlevalue comparisons of survey-unit data to action limits (DCGLs), mean-value comparison to background, completion of required monitoring, and satisfactory scan monitoring results.

Quality Assurance/Control Leader:

APPENDIX E. SURVEY UNIT RELEASE REPORTS (NARRATIVES, DATA, FLOOR PLANS, AND GRID MAPS)

XII. APPENDIX E

Appendix E-1. List of Survey Units, Data Tables, and Figures

A NMRC survey unit was defined as a room or other discrete area. Every room in every NMRC building was first examined for consideration as a survey unit. After review of current and historical survey monitoring results and licensed-material use, NMRC Diving-Building survey units were classified as either Non-impacted or Impacted Class 3. No NMRC Diving-Building survey unit was classified as Class 1 or Class 2.

Per MARSSIM, survey units should be limited in size based on classification, exposure pathway modeling assumptions, and site-specific conditions. In some instances, large rooms were subdivided into sections; each section was then treated as a unique survey unit.

Table 12-1-1 provides a list of tables associated with the completion of the final status surveys for the Non-impacted areas and Impacted Class 3 areas. These tables provide the actual data, reported in dpm per 100 cm^2 , obtained by evaluating swipe samples using a liquid scintillation counter.

Table 12-1-2 provides a list of figures associated with the survey-unit data tables and the completion of the final status surveys for the Impacted-Class 3 areas. These figures depict the grid map and survey sampling locations for the impacted areas.

Table 12-1-3 provides a summary of the NMRC Diving-Building survey units. Each survey unit was classified as either Non-impacted or Impacted Class 3. Table 12-1-3 lists discrete rooms or areas. In some instances, a room may have been sub-divided to accomplish a workable survey unit. These sub-divisions are indicated in column three of Table 12-1-3 as parenthesized comments or annotated using the room number followed by an alphabetic character.

Table 12-	1-1. List of Tables	
NUMBER	DESCRIPTION	PAGE
		NUMBER
12-1-3	NMRC Diving Building Survey Units	127
12-2-1	RADIOLOGICAL FINAL STATUS SURVEY SUMMARY -	131
	NMRC Diving Building, Non-impacted Areas	
12-2-2	Control Swipe Results, Building 28, Several	132
	Areas	
12-2-3	Control Swipe Results, Building 28, Several	133
	Areas	
12-2-4	Control Swipe Results, Building 53, Several	134
	Areas	
12-2-5	Control Swipe Results, Building 53, Several	135
	Areas	
12-2-6	Control Swipe Results, Building 53, Several	136
	Areas	
12-2-7	Control Swipe Results, Building 53, Several	138
	Areas	
12-2-8	Control Swipe Results, Building 53, Several	139
	Areas	100
12-2-9	Control Swine Results Building 53, Several	140
12 2 2	Areas	110
12-2-10	Control Swine Posulte Building 59 Several	141
12 2 10	Aroas	141
12-2-11	Control Swine Results Building 69 Several	142
	Areas	1.2
12-2-12	Control Swine Results Building 69 Several	143
16 6 16	Arpas	110
12-2-13	Control Swine Results Building 79 Several	144
	Arpag	<i></i>
12-2-14	Control Swine Results Building 79 Several	145
	Areas	110
	Aleas	
12-3-1	PADIOLOGICAL FINAL STATUS SURVEY SUMMARY -	148
12. 0 1	NMBC Diving Building Impacted Areas	140
12-3-2	Final Status Survey Summary For Direct Readings	150
12-3-2	of Impacted Survey Units Using a Cas	130
	Droportional Counter	
10_2_2	Final Status Survey Survey For Direct Boadings	161
12-3-3	of Impacted Survey Units Using a Cas	
	Droportional Counter	
10 3 4	Figel Chatus Cumunu Cummenu Fer Direct Desdinger	150
12-3-4	Final Status Survey Summary For Direct Keadings	132
	OL Impacted Survey Units Using a Gas	
10 0 5	Proportional Counter	
12-3-5a	Room 101-1, Building 53, Random Statistical	154
	Swipe Results	

NMRC Diving Buildings Radiological (D&D) Final Status Survey Report, June 2000

12-3-5b	Room 101-1, Building 53, Control Swipe Results	155
12-3-6a	Room 101-2, Building 53, Random Statistical	156
	Swipe Results	
12-3-6b	Room 101-2, Building 53, Control Swipe Results	157
12-3-7	Room 102, Building 53, Random Statistical Swipe	160
	Results	
12-3-8	Room 102, Building 53, Control Swipe Results	161
12-3-9	Room 103, Building 53, Random Statistical Swipe	163
	Results	
12-3-10	Room 103, Building 53, Control Swipe Results	164
12-3-11	Room 111, Building 53, Random Statistical Swipe	166
	Results	
12-3-12	Room 111, Building 53, Control Swipe Results	167
12-3-13	Room 112, Building 53, Random Statistical Swipe	169
	Results	
12-3-14	Room 112, Building 53, Control Swipe Results	170
12-3-15	Room 112A, Building 53, Random Statistical Swipe	172
	Results	
12-3-16	Room 112A, Building 53, Control Swipe Results	173
12-3-17	Room 113, Building 53, Random Statistical Swipe	175
	Results	
12-3-18	Room 113, Building 53, Control Swipe Results	176
12-3-19	Room 117, Building 53, Random Statistical Swipe	178
	Results	
12-3-20	Room 117, Building 53, Control Swipe Results	179
12-3-21	Room 118, Building 53, Random Statistical Swipe	181
	Results	
12-3-22	Room 118, Building 53, Control Swipe Results	182
12-3-23	Room 119, Building 53, Random Statistical Swipe	184
	Results	
12-3-24	Room 119, Building 53, Control Swipe Results	185
12-3-25	Room 120, Building 53, Random Statistical Swipe	187
	Results	
12-3-26a	Room 120, Building 53, Control Swipe Results	188
12-3-26b	Room 120D, Building 53, Control Swipe Results	189
12-3-27	Room 121, Building 53, Random Statistical Swipe	193
	Results	
12-3-28	Room 121, Building 53, Control Swipe Results	194
12-3-29	Room 122, Building 53, Random Statistical	196
	Samples	
12-3-30	Room 122, Building 53, Control Swipe Results	197

Table 12	-1-2. List of Figures (Grid Maps annotated with s	survey-
	data locations)	
NUMBER	DESCRIPTION	PAGE
		NUMBER
12-3-1a	Grid Map for Building 53, Room 101-1	158
12-3-1b	Grid Map for Building 53, Room 101-2	159
12-3-2	Grid Map for Building 53, Room 102	162
12-3-3	Grid Map for Building 53, Room 103	165
12-3-4	Grid Map for Building 53, Room 111	168
12-3-5	Grid Map for Building 53, Room 112	171
12-3-6	Grid Map for Building 53, Room 112A	174
12-3-7	Grid Map for Building 53, Room 113	177
12-3-8	Grid Map for Building 53, Room 117	180
12-3-9	Grid Map for Building 53, Room 118	183
12-3-10	Grid Map for Building 53, Room 119	186
12-3-11	Grid Map for Building 53, Room 120	191
12-3-12	Grid Map for Building 53, Room 120D	192
12-3-13	Grid Map for Building 53, Room 121	195
12-3-14	Grid Map for Building 53, Room 122	199

Table 12	-1-3. NM	RC Divin	g-Buildings	Survey Uni	ts
DATE	BUILDING	ROOM OR	DESCRIPTION	ISOTOPES OF	D&D
FINAL		AREA		CONCERN	CLASSIFICATION
STATUS					IMP = Impacted,
SURVEY					Class 3
COMPLETED					NON = NOn-
			-		impacted
	28	100	Supply		Non-impacted
	28	101	Mechanical		Non-impacted
	28	102	Storage		Non-impacted
	28	103	Storage		Non-impacted
	28	106	Office		Non-impacted
	28	106 A	Laboratory		Non-impacted
	28	107	Office		Non-impacted
	28	108	Office		Non-impacted
	28	200	Office		Non-impacted
5/2/00	28	201	Office		Non-impacted
	28	201 A	Office		Non-impacted
5/2/00	28	202 A	Office		Non-impacted
5/2/00	28	202 B	Storage		Non-impacted
	28	203	Supply		Non-impacted
5/2/00	28	204	Shop		Non-impacted
	28	205	Entryway		Non-impacted
	53	S-1	Entryway		Non-impacted
	53	10	Shop		Non-impacted
	53	11	Shop		Non-impacted
	53	13	Office		Non-impacted
	53	14	Storage		Non-impacted
	53	15	Storage		Non-impacted
	53	22	Galley		Non-impacted
	53	23	Compressors		Non-impacted
	53	24	NRCC		Non-impacted
5/2/00	53	25	Office		Non-impacted
5/2/00	53	25 A	Office		Non-impacted
5/2/00	53	25 B	Office		Non-impacted
	53	26	Office		Non-impacted
	53	26 A	Sickbay		Non-impacted
	53	26 B	Office		Non-impacted
	53	28	Sickbay		Non-impacted
	53	29	Office		Non-impacted
	53	30	Office		Non-impacted
	53	31	Instrument		Non-impacted
			room		
	53	32	Mechanical		Non-impacted
	53	33	Mechanical		Non-impacted
	53	100	Fover		Non-impacted
	53	100 B	Office		Non-impacted
5/9/00	53	101	Laboratory	H3. C14	Impacted Clase?
575700	55	101	Laboracory	CoS7	Impacted, Classs
3/23/00	53	102	Laboratory	H3, C14, Co57	Impacted,Class3
3/23/00	53	103	Laboratory		Non-impacted
3/23/00	53	104	Office		Non-impacted

3/23/00	53	105	Office		Non-impacted
3/15/00	53	106	Office		Non-impacted
3/15/00	53	107	Office		Non-impacted
	53	108	Conference		Non-impacted
3/15/00	53	109	Office		Non-impacted
	53	110	Office		Non-impacted
4/26/00	53	111	Laboratory	H3. C14.	Impacted. Class3
			Daboratory	Co57	
3/23/00	53	112	Office/Lab	H3, C14, Co57, Ba133, Cs137	Impacted,Class3
3/23/00	53	112A	Laboratory	H3, C14, Co57	Impacted,Class3
3/23/00	53	113	Laboratory	H3, C14, Co57	Impacted, Class3
3/15/0	53	114	Office		Non-impacted
3/15/00	53	115	Gear Lab		Non-impacted
3/15/00	53	116	Office		Non-impacted
3/23/00	53	117	Laboratory	H3, C14, Co57	Impacted, Class3
3/23/00	53	118	Laboratory	H3, C14, Co57	Impacted, Class3
5/9/00	53	119	Laboratory	H3, C14, Co57	Impacted,Class3
5/9/00	53	120 120D	Office	H3, C14, Co57	Impacted, Class3
3/23/00	53	120 A	Office		Non-impacted
3/23/00	53	120 B	Office		Non-impacted
3/23/00	53	120 C	Office		Non-impacted
4/19/00	53	121	Laboratory	H3, C14, Co57	Impacted, Class3
4/19/00	53	122	Laboratory	H3, C14, Co57, Ba133, Cs137	Impacted, Class3
	53	123	Office		Non-impacted
	53	124	Office		Non-impacted
	53	124 A	Office		Non-impacted
	53	125	Office		Non-impacted
	53	126	Office		Non-impacted
	53	126 A	Office		Non-impacted
	53	126 B	Office		Non-impacted
	53	127	Office		Non-impacted
	53	128	Telephone		Non-impacted
	59	11	Scuba Locker		Non-impacted
5/2/00	59	100	Research tank		Non-impacted
5/2/00	59	101	Laboratory		Non-impacted
5/2/00	59	102	Laboratory		Non-impacted
	59	103	Electrical room		Non-impacted

	59	104	Closet	Non-impacted
5/2/00	59	106	Laboratory	Non-impacted
	69	100	Shop/mech.	Non-impacted
			room	
5/11/00	69	101	Laboratory	Non-impacted
5/11/00	69	102	Laboratory	Non-impacted
5/11/00	69	103	Laboratory	Non-impacted
5/11/00	69	104	Control	Non-impacted
			room	
	79	S-1	Entryway	Non-impacted
	79	100	Storage	Non-impacted
5/11/00	79	101	Mechanical	Non-impacted
			shop	
5/11/00	79	102	Office	Non-impacted
5/11/00	79	103	Office	Non-impacted
5/11/00	79	104	Laboratory	Non-impacted
5/11/00	79	107	Laboratory	Non-impacted
	79	120A	Office	Non-impacted

Appendix E-2. Diving Buildings, Non-impacted Areas

- Location: NMRC Diving Buildings 28, 53, 59, 69, and 79 at NNMC Bethesda, Maryland
- Monitoring Date(s): 3/15/00, 3/23/00, 5/2/00, 5/11/00
- > Radioactive Contaminants of Concern: None
- Minimum Control Beta Swipe Locations (performed for every survey unit): (1) floor at door in room; (2) floor at door outside room; (3) doorknob in room; (4) doorknob outside room; and (5) light switch inside room. Additional control samples were taken at the discretion of the D&D team.

> Survey Instrument Information:

- Beta meters -- serial numbers (background CPM): 11030 (<100CPM), 54537 (<100 CPM)
- Gamma meters -- serial numbers (background CPM): 114886 (<300 CPM), 121730 (<200 CPM), 121886 (<300 CPM)
- Liquid Scintillation Counter: Packard 2500TR, S/N 403848
- ➢ Radiation-detection Action Limits:
- 2 to 3 times background for beta and gamma scans
- 500 700 dpm or higher for either H-3 or C-14 beta swipes counted using a liquid scintillation counter
- > Summary of Radiological Scans and Swipe Results:
- Background level for all beta and gamma scans
- All H-3 and C-14 beta swipe results below Action Limits
- > Contamination and Remedial Actions: None
- Conclusion: The radiological monitoring results for all survey units addressed in this section meet the criteria for release for unrestricted use.

.

TABLE 12-2-1. RADIOLOGICAL FINAL STATUS SURVEY										
SUMMARY - Non-impacted Areas of NMRC										
Diving Buildings										
Date	Building	Room	Beta	Maximum	Recommend					
Final	Number	number	and/or	Random or	Survey					
Status		or area	Gamma Scan	Control	Unit for					
Survey		(Survey	Monitoring	beta	radio-					
Completed		Unit)		(H-3 or	logical					
			(check if	C-14)	clearance					
		* = at	yes)	swipe	(Y/N)					
		another		result						
		door		(DPM)	Action					
					Limit =					
F / P / P P		0.01			700 dpm					
5/2/00	28	201	Y	38	Y					
5/2/00	28	202 *	Y	29	Y					
5/2/00	28	202	<u>Y</u>	37	Y					
5/2/00	28	204	Y	24	Y					
5/2/00	53	25 *	Y	24	Y					
5/2/00	53	25 *	Y	25	Y					
5/2/00	53	25	Y	28	Y					
3/23/00	53	104	Y	34	Y					
3/23/00	53	105	Y	46	Y					
3/15/00	53	100	I V	24	Y					
3/15/00	53	107	Y Y	27	Y					
3/15/00	53	109	I V	21	I V					
3/15/00	53	115	I V	17	I V					
3/15/00	53	116	I V	32	I V					
3/23/00	53	120 7	I V	21	I V					
3/23/00	53	120 A	v	31	v					
3/23/00	53	120 B	v	17	v					
3/23/00	53	120 C		23	v v					
3/23/00	53	125	v v	25	Y Y					
3/23/00	53	127	v v	24	Y Y					
5/2/00	59	100	Y Y	22	v v					
5/2/00	59	101	Y	44	Y					
5/2/00	59	102	Y Y	32	Y					
5/2/00	59	106	Y	62	Y					
5/11/00	69	101	Ŷ	41	Y					
5/11/00	69	102	Ŷ	79	Ŷ					
5/11/00	69	103	Y	24	Y					
5/11/00	69	104	Y	17	Y					
5/11/00	79	101	Y	30	Y					
5/11/00	79	102	Y	24	Y					
5/11/00	79	103	Y	30	Y					
5/11/00	79	104	Y	22	Y					
5/11/00	79	107	Y	24	Y					

	Table 12	-2-2.	Final	Status	Survey	Summary	For	Non-Impacted	Survey	Units
--	----------	-------	-------	--------	--------	---------	-----	--------------	--------	-------

Building Number: 28

Monitoring date(s): 5/2/00

Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for Residual Contamination (Count Times: 1 minute)

Data is reported in DPM for a survey unit (building number/room number)

Ducu 10 10pc	eported in bin ior a survey unit (buriding number/toom number)								
SAMPLE	ROOM N	UMBER:	ROOM NUMBER: 202		ROOM NUMBER:		ROOM NUMBER: 204		
NUMBER	20	01				202*		1	
	H-3	C-14	H-3	C-14	H-3	C-14	Н-3	C-14	
1	38	10	18	11	26	13	15	6	
2	18	11	16	4	29	6	8	7	
3	21	11	12	4	28	10	24	8	
4	24	11	27	13	0	12	16	4	
5	23	8	37	2	7	7	11	11	
6	20	9	35	0			10	7	
7	15	4	2	12			19	6	
8	3	7	1	17			9	7	
9	10	12	17	11			9	9	
10	33	10	2	9			0	2	
AVERAGE	21	9	17	8	18	10	12	7	
STANDARD		_							
DEVIATION	10	2	13	5	14	3	7	2	
MAXIMUM									
VALUE	38	12	37	17	29	13	24	11	
ACTION									
LIMIT	700	700	700	700	700	700	700	700	

Table 12-2-3. Final Status Survey Summary For Non-Impacted Survey Units								
Building Number: 28								
Monitoring date(s): 5/2/00								
Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for								
Residual Contamination (Count Times: 1 minute)								
Data is reported in DPM for a survey unit (building number/room number)								
SAMPLE	ROOM N	UMBER:	ROOM N	UMBER:				
NUMBER	Basement	Hallway	Basemer	nt Door				
	H-3	C-14	H-3	C-14				
1	12	11	12	7				
2	3	7	34	10				
3	8	4	6	12				
4	28	8	15	9				
5	20	8	10	11				
6	0	13						
7	17	6						
8	22	8						
9	52	4						
10	13	6						
AVERAGE	18	8	15	10				
STANDARD								
DEVIATION	15	3	11	2				
MAXIMUM								
VALUE	52	13	34	12				
ACTION								
LIMIT	700	700	700	700				

Table 12-2-4. Final Status Survey Summary For Non-Impacted Survey Units								
Building Number: 53								
Monitoring date(s): 5/2/00								
Liquid Scin	tillation	Swipe Res	sults for	Backgrou	nd, H-3 ai	nd C-14 A	nalyses f	or
Residual Contamination (Count Times: 1 minute)								
Data is repo	orted in D	PM for a	survey un	nit (build	ding numbe	er/room n	umber)	
SAMPLE	ROOM N	UMBER:	ROOM N	UMBER:	ROOM N	UMBER:	ROOM NUN	4BER: 25
NUMBER	Base	ment	Basemer	nt Door	Basemer	nt Door		
	H-3	C - 14	C-14	H-3	H-3	C-14	H-3	C-14
1	14	2	29	14	2	29	6	2
2	9	4	12	9	4	12	27	1
3	32	0	5	32	0	5	18	13
4	24	8	7	24	8	7	26	0
5	28	5	57	28	5	57	23	8
6							17	6
7							15	9
8							0	12
9							28	3
10							2	12
AVERAGE	21	4	22	21	4	22	16	7
STANDARD								
DEVIATION	10	3	22	10	3	22	10	5
MAXIMUM								
VALUE	32	8	57	32	8	57	28	13
ACTION								
LIMIT	700	700	700	700	700	700	700	700

Table 12-2-5. Final Status Survey Summary For Non-Impacted Survey Units								
Building Number: 53								
Monitoring date(s): 5/2/00								
Liquid Scint	illation	Swipe Res	sults for	Backgrou	nd, H-3 an	nd C-14 A	nalyses f	or
Residual Con	ntaminatic	on (Count	Times: 1	l minute)				
Data is repo	orted in D	OPM for a	survey u	nit (build	ding numbe	er/room n	umber)	
SAMPLE	ROOM NUN	MBER:25*	ROOM NUI	MBER:25*	ROOM NUN	4BER:115	ROOM NUI	ABER:116
NUMBER	H-3	C-14	H-3	C-14	H-3	C-14	H-3	C-14
1	2	10	19	13	8	4	7	9
2	24	8	8	9	17	6	8	4
3	0	22	23	11	12	4	13	6
4	6	12	12	4	8	4	2	9
5	12	4	25	1	8	2	32	12
6					17	9	3	5
7					13	4	0	24
8					13	4	4	5
9					7	7	22	6
10					12	4	1	12
AVERAGE	9	11	17	8	12	5	9	9
STANDARD								
DEVIATION	10	7	7	5	4	2	10	6
MAXIMUM								
VALUE	24	22	25	13	17	9	32	24
ACTION								
LIMIT	700	700	700	700	700	700	700	700

Table 12-2-6. Fir	hal Status	5 5	Survey Summ	ary for CON	TROL
SAMPLES FOR NON-	IMPACTED S	ហ	RVEY UNITS		
Building: 053		Su	rvey Unit or Roc	om Number: 104	and 105
Survey Date: 3/23	3/00 0	Со	unt Times:	1 minute	
Liquid Scintillation Swipe Re	sults for Backgr	ou	nd, H-3 and C-14 A	nalyses for Residua	1 Contamination
Sample Number	DPM,	Ro	00m 104	DPM, Ro	oom 105
	Н-З		C-14	Н-3	C-14
1	10		7	6	14
2	3		7	16	9
3	11		7	16	6
4	0		7	11	11
5	9		7	8	4
6	23		6	8	2
7	5		2	14	0
8	16		14	10	14
9	30		10	7	7
10	11		4	7	9
11	8		4	7	9
12	13		6	4	2
13	5		17	0	19
14	3		7	0	5
15	11		9	8	7
16	0		12	46	0
17	4		7	6	2
18	34		3	0	5
19	8		4	3	10
20	17		4	14	9
21	6		9	12	9

22	11	7	12	9
23	3	10	0	12
24	26	3	3	7
25	4	19	22	18
26	28	13	15	9
27	4	22	14	11
28	20	11	18	13
29	3	7	12	6
30	21	4	11	9
31	0	19	6	12
32	26	10	8	4
33	8	19	22	0
34	2	9	5	16
35	24	11	6	9
36	2	9	12	4
37	7	9	16	4
38	2	9	16	4
Average DPM	11	9	11	8
Standard				
Deviation, DPM	10	5	8	5
Maximum DPM	34	22	46	19
Action Limit,				
DPM	700	700	700	700

Table 12-2-	7. Final S	Status Sur	evey Summa	ary For No	on-Impacte	ed Survey	Units			
Building Nur	mber: 53	·								
Monitoring of	Monitoring date(s):03/15/00									
Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for										
Residual Contamination (Count Times: 1 minute)										
Data is repo	orted in D	PM for a	survey ur	<u>it (builc</u>	ling numbe	er/room nu	umber)	_		
SAMPLE	ROOM NUN	MBER:106	ROOM NUM	IBER: 107	ROOM NUM	BER: 109	ROOM NUM	IBER: 114		
NUMBER	H-3	C-14	H-3	C-14	H-3	C-14	H-3	C-14		
1	16	9	4	2	13	4	31	5		
2	13	0	0	12	6	12	12	4		
3	8	7	17	6	0	17	29	5		
4	11	9	2	12	1	12	12	6		
5	8	4	8	7	8	4	3	7		
6	16	6	2	10	17	18	14	6		
7	8	4	20	6	4	5	22	6		
8	3	7	27	8	18	4	0	12		
9	24	11	16	9	0	10	7	9		
10	24	11	4	5	13	4	0	7		
AVERAGE	13	7	10	8	8	9	13	7		
STANDARD	7	3	9	3	7	б	11	2		
DEVIATION			 							
MAXIMUM	24	11	27	12	18	18	31	12		
VALUE										
ACTION	700	700	700	700	700	700	700	700		
LIMIT										

Table 12-2-8. Final Status Survey Summary For Non-Impacted Survey Units

Building Num	nber: 53										
Monitoring of	date(s):03	3/15/00									
Liquid Scint	illation	Swipe Res	sults for	Backgrour	nd, H-3 ar	nd C-14 Ar	alyses fo)r			
Residual Contamination (Count Times: 1 minute)											
Data is repo	orted in D	OPM for a	<u>survey</u> un	it (build	ling numbe	er/room nu	imber)				
SAMPLE	ROOM NUM	IBER:120A	ROOM NUM	IBER:120B	ROOM NUM	BER:120C	ROOM NUM	BER: 123			
NUMBER	H-3	C-14	H-3	C-14	H-3	C-14	H-3	C-14			
1	4	2	24	8	12	4	23	1			
2	5	17	17	4	12	11	9	2			
3	17	4	6	17	7	14	15	4			
4	5	2	11	9	6	9	12	2			
5	0	10	12	6	11	6	15	9			
6	7	7	14	2	0	10	16	6			
7	9	7	5	19	17	11	15	14			
8	8	9	1	14	12	6	21	4			
9	1	12	31	3	8	4	8	7			
10	21	4	31	5	9	4	12	9			
AVERAGE	8	7	15	9	9	8	15	6			
STANDARD	7	5	11	6	5	4	5	4			
DEVIATION											
MAXIMUM	21	17	31	19	17	14	23	14			
VALUE											
ACTION	700	700	700	700	700	700	700	700			
LIMIT											

Table	12 - 2 - 9.	Final	Status	Survey	Summary	For	Non-Tmr	pacted	Survey	Units
			~~~~~			~ ~ ~		Jaccoa	Curvey	

Building Number: 53

Monitoring date(s):03/15/00

Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for Residual Contamination (Count Times: 1 minute)

Data is reported in DPM for a survey unit (building number/room number)

SAMPLE	ROOM NUN	1BER: 125	ROOM NUN	ROOM NUMBER: 127		UMBER:	ROOM N	UMBER:
NUMBER	H-3	C-14	H-3	C-14	H-3	C-14	Н-З	C-14
1	13	9	18	13				
2	39	2	3	7				
3	4	5	8	7				
4	6	12	12	4				
5	18	13	24	8				
6	11	9	3	5				
7	8	7	7	12				
8	13	4	1	12				
9	13	9	12	14				
10	1	14	16	11				
AVERAGE	13	8	10	9				
STANDARD	11	4	7	4				
DEVIATION								
MAX IMUM	39	14	24	14				
VALUE								
ACTION	700	700	700	700				
LIMIT								

Table 12-2-1	Table 12-2-10. Final Status Survey Summary For Non-Impacted Survey Units										
Building Num	ber: 59										
Monitoring d	late(s): 5	5/2/00									
Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for											
Residual Contamination (Count Times: 1 minute)											
Data is repo	rted in D	PM for a	survey ur	nit (build	ding numb	er/room n	umber)				
SAMPLE	ROOM N	UMBER:	ROOM NUM	BER: 101	ROOM NUM	BER: 102	ROOM NUM	BER: 106			
NUMBER	Pool										
	H-3	C-14	H-3	C-14	H-3	C-14	H-3	C-14			
1	21 9 0 10 10 9 18 14										
2	0 10 10 9 24 1 11 7										
3	21         6         31         0         23         18         10         16										
4	22 1 21 6 32 17 3 7										
5	12	4	18	6	7	24	62	1			
6	4	10	10	15	6	5	14	7			
7	0	10	5	7	4	10	0	26			
8	13	4	22	4	12	9	11	9			
9	14	4	44	5	13	22	4	10			
10	10	15	28	3	21	6	15	7			
AVERAGE	12	7	19	7	15	12	15	10			
STANDARD											
DEVIATION	8	4	13	4	9	8	18	7			
MAXIMUM											
VALUE	22	15	44	15	32	24	62	26			
ACTION											
LIMIT	700	700	700	700	700	700	700	700			

Table 12-2-11. Final Status Survey Summary For Non-Impacted Survey Units											
Building Num	ber: 69				<u> </u>		<u>.</u>				
Monitoring d	Monitoring date(s): 5/11/00										
Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for											
Residual Contamination (Count Times: 1 minute)											
Data is repo	Data is reported in DPM for a survey unit (building number/room number)										
SAMPLE ROOM NUMBER: 101 ROOM NUMBER: 102 ROOM NUMBER: 103 ROOM NUMBER:											
NUMBER							10	3*			
	H-3	C-14	H-3	C-14	H-3	C-14	H-3	C-14			
]	41	35	0	8	7	19	2	10			
2	12	16	0	13	0	10	0	17			
3	9	29	4	2	0	7	0	13			
4	7	12	45	7	13	6	3	7			
5	19	33	6	12	8	24	2	12			
6	4	30	0	10	0	3					
7	5	7	4	10	13	9					
8	20	12	11	4	3	12					
9	0	18	254	15	11	4					
10	0	22	0		3	12					
AVERAGE	12	21	32	9	6	11	1	12			
STANDARD											
DEVIATION	12	10	79	4	5	7	1	4			
MAXIMUM											
VALUE	41	35	254	15	13	24	3	17			
ACTION											
LIMIT	700	700	700	700	700	700	700	700			

Table 12-2-1	l2. Final	Status Si	irvey Sum	nary For No	on-Impac	ted Survey Un	nits	
Building Num	nber: 69							
Monitoring o	date(s): 5	5/11/00						
Liquid Scint	illation	Swipe Res	sults for	Background	1, H-3 a	nd C-14 Analy	yses for	
Residual Cor	ntaminatic	on (Count	Times: 3	l minute)				
Data is repo	prted in D	PM for a	survey u	nit (buildi	ing numb	er/room numbe	er)	
SAMPLE	ROOM NUM	BER: 104	ROOM NUM	IBER: 104*	ROOM N	UMBER: 104*		
NUMBER	H-3	C-14	H-3	C-14	H-3	C-14		
1	10	4	14	7	6	2		
2	1	17	10	9	0	15		
3	29	6	2	15	22	16		
4	8	9	8	7	8	7		
5	4	5	8	4	4	5		
6	23	0						
7	14	4						
8	14	14						
9	4	10		1				
10	14	12						
AVERAGE	12	8	8	8	8	9		
STANDARD								
DEVIATION	9	5	4	4	8	6		
MAXIMUM								
VALUE	29	17	14	15	22	16		
ACTION								
LIMIT	700	700	700	700	700	700		

Table 12-2-1	Table 12-2-13. Final Status Survey Summary For Non-Impacted Survey Units										
Building Nur	nber: 79	_									
Monitoring o	Monitoring date(s): 5/11/00										
Liquid Scintíllation Swipe Results for Background, H-3 and C-14 Analyses for											
Residual Contamination (Count Times: 1 minute)											
Data is repo	orted in I	OPM for a	survey un	nit (build	ding numb	er/room n	umber)				
SAMPLE	ROOM NUM	BER: 101	ROOM NUM	IBER: 102	ROOM NUM	BER: 103	ROOM NUM	BER: 104			
NUMBER	н-з	C-14	<u>H-</u> 3	C-14	H-3	C-14	H-3	C-14			
1	3	10	8	17	0	10	7	0			
2	4	10	2	12	12	7	5	5			
3	0	2	22	6	22	3	20	8			
4	30 8 16 11 30 13 22 3										
5	1	14	17	4	21	6	11	11			
6	15	7	3	7	4	7	0	8			
7	7	2	24	4	19	1	0	20			
8	25	4	10	9	0	7	5	5			
9	14	7	3	12	0	5	4	10			
10	0	20	9	12	11	7	10	7			
AVERAGE	10	8	11	9	12	7	8	8			
STANDARD											
DEVIATION	11	6	8	4	11	3	8	5			
MAXIMUM											
VALUE	30	20	24	17	30	13	22	20			
ACTION											
LIMIT	700	700	700	700	700	700	700	700			

Table 12-2-14. Final Status Survey S	Summary For	Non-Impacted	Survey	Units
--------------------------------------	-------------	--------------	--------	-------

Building Number: 79

Monitoring date(s): 5/11/00

Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for Residual Contamination (Count Times: 1 minute)

Data is reported in DPM for a survey unit (building number/room number)

		•					,	
SAMPLE	ROOM NUME	BER: 104*	ROOM NUM	BER: 107	ROOM NUM	BER: 107*		
NUMBER	Н-3	C-14	H-3	C-14	H-3	C-14		
1	14	9	7	14	5	7		
2	5	22	9	9	13	9		
3	17	4	11	9	10	4		
4	8	4	20	11	9	4		
5	1	12	12	6	24	11		
6			16	11				
7	}		0	13	-			
8			3	12				
9			0	12	_			
10			9	17				
AVERAGE	9	10	9	11	12	7		
STANDARD								
DEVIATION	7	7	7	3	7	3		
MAXIMUM								
VALUE	17	22	20	17	24	11		
ACTION								
LIMIT	700	700	700	700	700	700		

Appendix E-3. Diving Buildings, Impacted Class 3 Areas

- Location: NMRC Diving Buildings, Impacted Class 3 Areas
- Monitoring Date(s): 3/23/00, 4/19/00, 4/26/00
- Radioactive Contaminants of Concern: Beta- and gamma-radiation emitters, Tritium (H-3), Carbon-14 (C-14), Cobalt-57 (Co-57), Barium-133 (Ba-133), and Cesium-137 (Cs-137)
- Minimum Control Beta Swipe Locations (performed for every survey unit): (1) floor at door in room; (2) floor at door outside room; (3) doorknob in room; (4) doorknob outside room; and (5) light switch inside room.
- Survey Instrument Information:
- Beta meters -- serial numbers (background CPM): 11030 (<100CPM), 54537 (<100 CPM)
- Gamma meters -- serial numbers (background CPM): 114886 (<300 CPM), 121730 (<200 CPM), 121886 (<300 CPM)
- Liquid Scintillation Counter: Packard 2500TR, S/N 403848
- Radiation-detection Action Limits:
- 2 to 3 times background for beta and gamma scans
- 500 700 dpm or higher for either H-3 or C-14 beta swipes counted using a liquid scintillation counter
- > Summary of Radiological Scans and Swipe Results:
- Background level for all beta and gamma scans
- All H-3 and C-14 beta swipe results below Action Limits
- Contamination and Remedial Actions: Residual carbon-14 contamination was detected in rooms 117 and 122 of building 53. Photographs are included in Appendix B that show the affected areas before and after decontamination efforts were completed. Swipe samples collected at these sites and evaluated using a liquid scintillation counter revealed that the contamination was fixed. The contamination was located on floor tile and upon a high-pressure vessel in room 117. In room 122, the contamination
was limited to four areas on the laminated surface of workbenches. The contamination was detected during the performance of direct readings as part of NMRC's final status surveys. The direct readings at the contaminated areas exceeded two to three times the background readings. Elevated readings ranged from several hundred to several thousand disintegrations per second. Decontamination efforts included spraying the areas with radiological cleaners and washing down the area before remonitoring. After repeated washings of the areas with no decrease in radioactivity, other measures were taken to decontaminate the affected areas. These measures included removing vinyl floor tiles, removing sections of laminate covering from workbenches, removing paint from metallic surfaces using a liquid paint remover, and sawing off sections of contaminated wood from the stand supporting the high-pressure vessel. All wastes generated from these decontamination efforts were disposed of as radioactive wastes. The affected areas were resurveyed following completion of decontamination efforts.

Conclusion: All contaminated areas were successfully decontaminated. Remediation surveys following the completion of decontamination efforts revealed that background-radiation levels had been achieved. The radiological monitoring results for all survey units addressed in this section met the criteria for release for unrestricted use. Þ

TABLE 12-3-1. RADIOLOGICAL FINAL STATUS SURVEY SUMMARIES - Diving Building 53, Impacted Class 3 Areas								
Date	Room number or area (Survey Unit)	Beta and/or Gamma Scan Monitoring (check if yes)	Maximum Background beta (H-3 or C-14) swipe result (DPM)	Maximum Random or Control beta (H-3 or C-14) swipe result (DPM)	Recommend Survey Unit for radiological clearance (Y/N)			
5/9/00	53	101 (101-1, 101-2)	28	424	Y			
3/23/00	53	102	31	30	Y			
3/23/00	53	103	31	31	Y			
4/26/00	53	111	36	35	Y			
3/23/00	53	112	31	35	Y			
3/23/00	53	112A	31	35	Y			
3/23/00	53	113	31	24	Ŷ			
3/23/00	53	117	31	30	Y			
3/23/00	53	118	31	63	Y			
6/9/00	53	119	28	125	Y			
6/9/00	53	120, 120D	28	33	Y			
4/19/00	53	121	23	76	Y			
4/19/00	53	122	23	21	Y			

#### Quality Assurance Checklist

Building number or location: NMRC Diving Buildings Impacted Areas Monitoring Date(s): 3/23/00, 4/19/00, 4/26/00, 6/9/00 Y - Appropriate and satisfactory daily and weekly calibration checks were completed for the selected survey monitoring equipment and instruments. Y - Selected meters with current calibrations for the isotopes of concern Y - Survey maps or floor plans (for all survey units) and grid maps (for known RAM-use areas) were available Y - Background checks were performed when impacted areas were monitored. Y - For known RAM-use survey units, additional checks were performed (workbenches, shelves, sinks, basins, traps, floor at sinks and entrances, wall at hoods and sinks, and/or other areas as deemed appropriate) Y - Mean values, single sample values, and background values were below the action limits, indicating no contamination was found. Y - Representative floor plans, sketches, and photographs were available. Y - Deviation from survey plans and objectives were documented. Y - Extent of remediation actions documented. Y - All selected survey units passed D&D tests; units are recommended for unrestricted release. Y - Additional Checks/Data passed D&D tests; unit is recommended for unrestricted release Notes: Tests included complying with applicable elements of the

Notes: Tests included complying with applicable elements of the MARSSIM Final Status Survey checklist, mean-value and singlevalue comparisons of survey-unit data to action limits (DCGLs), mean-value comparison to background, completion of required monitoring, and satisfactory scan monitoring results.

Quality Assurance Leader: Schleurious L. Gaiter, LCDR, MSC, USN

Table 12-3-2. Final Status Survey Summary For Direct Readings of									
Impacted Survey Units Using a Gas Proportional Counter									
DATA CONVERSION WAS PERFORMED AS DESCRIBED BELOW									
Building Number: 53									
Monitoring date(s): 5/2/00									
Data is reported	Data is reported in DPM for direct readings for Carbon-14								
MARSSIM formula 6	-4:								
	dpm		$(C_s/)$	$T_s$ ) - ( $C_b/T_b$	)				
$(100 \text{ cm}^2) = (E_1 \times (A/100))$									
where, A=126 cm ² , $E_i$ = instrument efficiency = 0.15; sample and									
background count	times = 3	$T_s = T_b = 1$	min, $C_s$	= sample	counts di	uring			
time $T_s$ , and $C_b$ =	backgrou	nd counts	during t	ime T _b					
					_				
Note: Negative va	lues of t	the calcul	Lated dpm	were set	equal to	o zero.			
All areas with ca	lculated	apm value	exceed	ing the N	MRC ACTIO	on Limit			
were remediated (	aecontam.	inated or	materiai	s removed	) and re-	- 7 Nation			
Limit of 2.000 dr	mai suive m per 100	) cm ² for	direct r	eadings		ACCION			
SAMPLE NUMBER	ROOM NU	MBER: 101	ROOM NU	MBER:102	ROOM NU	MBER:103			
	CPM	DPM/	CPM	DPM /	CPM	DPM/			
	CIH	$100 \text{ CM}^2$	CTH	$100 \text{ CM}^2$	CT H	$100 \text{ CM}^2$			
1	294	138	272	0	280	116			
2	348	423	306	164	286	148			
3	283	79	301	138	318	317			
4	273	26	281	32	307	259			
5	313	238	307	169	313	291			
6	316	254	274	0	288	159			
7	296	148	268	0	285	143			
8	270	11	289	74	289	164			
9	281	69	252	0	284	138			
10	327	312	272	0	266	42			
11					373	608			
12					314	296			
AVERAGE	1	70	5	8	2	23			
STANDARD									
DEVIATION	1	34	7	3	1	48			
MAXIMUM VALUE	4:	23	1	69	6	08			
BACKGROUND, CPM	2	68	2	75	2	58			
ACTION LIMIT, DPM/100 CM ²	2,	000	2,	000	2,	000			

Table 12-3-3. Final Status Survey Summary For Direct Readings of Impacted Survey Units Using a Gas Proportional Counter Building Number: 53

Monitoring date(s): 5/2/00

Data is reported in DPM for direct readings for Carbon-14

See Table 12-3-2 for the application of MARSSIM formula 6-4 in converting the gas proportional counter data to dpm per 100 cm2.

The final surveys revealed readings below the Action Limit.

SAMPLE NUMBER	ROOM NUN	4BER:111	ROOM N	UMBERS:	ROOM NUI	MBER:117
			112,	112A		
	CPM	DPM/	CPM	DPM/	CPM	DPM/
		100 CM ²		100 CM ²		100 CM ²
1	284	26	265	0	249	0
2	287	42	255	0	276	63
3	286	37	249	0	252	0
4	285	32	240	0	247	0
5	255	0	2:83	90	280	85
6	271	0	245	0	262	0
7	228	0	254	0	258	0
8	260	0	262	0	292	148
9	268	0	.267	5	287	122
10	263	0	241	0	452	995
11	279	0	2:90	127	289	132
12			269	16	310	243
13			237	0	362	519
14			269	16	389	661
15			254	0	321	302
16		0			351	460
17					310	243
18					467	1074
19	,	ĺ		l	370	561
20				_	401	725
AVERAGE	1	3	1	7	3	17
STANDARD	1	8	3	38	3.	38
DEVIATION					ļ	
MAXIMUM VALUE	4	2	1.	27	10	74
BACKGROUND, CPM	21	79	2	66	2:	58
ACTION LIMIT,					1	
$DPM/100 CM^2$	2,0	000	2,	000	2,000	

Table 12-3-4. Final Status Survey Summary For Direct Readings of								
Impacted Survey Units Using a Gas Proportional Counter								
Building Number: 53								
Monitoring date(s): 5/2/00								
Data is reported in DPM for direct readings for Carbon-14								
See Table 12-3-2 for the application of MARSSIM formula 6-4 in								
converting the gas proportional counter data to dpm per 100 cm ² .								
After data conv	ersion of	the surfa	ace activ	ity counts	s, all are	eas with		
calculated dpm values exceeding the NMRC Action Limit were								
remediated (decontaminated or materials removed) and re-surveyed.								
The final surve	ys reveal	ed reading	JS DELOW	the Action	DOOM NUM	DED: 100		
SAMPLE NUMBER	KOOM NUM	BEK: 110	ROOM NU	MBERS:	ROOM NUM	BER: 122		
	CPM		CPM		CPM	DPM/		
	CFM	$100 \text{ CM}^2$	CFM	$100 \text{ CM}^2$	CFM	$100 \text{ CM}^2$		
1	254	0	239	0	247	0		
2	273	48	312	254	305	217		
3	299	185	282	95	284	106		
4	312	254	272	42	291	143		
5	261	0	281	90	295	164		
6	234	0	282	95	277	69		
7	280	85	281	90	284	106		
8	268	21	272	42	305	217		
9	257	0	296	169	251	0		
10	248	0	251	0	298	180		
11	232	0	268	21.	327	333		
12	286	0	255	0	329	344		
13	300	190	287	122	483	1159		
14			257	0	335	376		
15			240	0	309	238		
16			284	106	299	185		
17			248	0	279	79		
18					291	143		
19					294	159		
20	· · · · · · · · · · · · · · · · · · ·				298	180		
21					320	296		
22					281	90		
23					252	0		
24					289	132		
25					291	143		
AVERAGE	3	0	5	9	.21	02		

STANDARD	54	68	224
DEVIATION			
MAXIMUM VALUE	164	243	1159
BACKGROUND,	281	266	264
CPM			
ACTION LIMIT,			
DPM/100 $CM^2$	2,000	2,000	2,000

Table 12-3-5a. Final Status Survey Summary for RANDOM								
STATISTICAL SAME	LES FOR .	IMPACTED	SURVEY UNIT					
Bullaing: 055								
A. Survey	' Unit or	Room Num	ber: 101-1					
Survey Date: 06/	09/00		Count Times:	1 minute				
Liquid Scintillation Swipe F	Results for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination				
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, C-14				
	H-3	<u>C-14</u>		1.0				
	4	5	2	10				
2	20	1	0	5				
3	2	10	15	14				
4	1	17	25	4				
5	0	3	0	22				
6	2	12	1	14				
7	4	7	13	4				
8	28 1 24 13							
9	14 2 4 7							
10	7	9	20	11				
11	2	12	14	7				
12	6	14	0	7				
13	8	7	7	9				
14	19	4	12	11				
15	19	6	0	5				
16	7	12	25	4				
17	6	14	19	16				
18	6	14	17	4				
19	0	10	1	17				
20	0	7	0	15				
21	24	8	0	30				
Average DPM	9	8	9	11				
Standard								
Deviation, DPM	9	5	9	7				
Maximum DPM	28	17	25	30				
Action Limit, DPM	700	700	700	700				

Table 12-3-5b. Final Status Survey Summary for CONTROL								
STATISTICAL SAMP	LES FOR	IMPACTED	SURVEY UNIT					
Building: 053								
B. Survey Unit or Room Number: 101-1								
Survey Date: 06/	09/00		Count Times:	1 minute				
Liquid Scintillation Swipe F	Results for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination				
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	4	5	7	14				
2	20	1	17	6				
3	2	10	22	0				
4	6	17	2	10				
5	1	17	2	10				
6	0	3	13	11				
7	2	12	0	13				
8	4 5 18 6							
9	28 1 1 15							
10	14 2 3 10							
11	7	9	11	7				
12	2	12	4	7				
13	6	14	11	2				
14	8	7	39	5				
15	19	4	11	4				
16	19	6	1	17				
17	7	12	36	5				
18	6	14	17	19				
19	6	14		1.				
20	0	10						
21	0	7						
Average DPM	8	9	12	9				
Standard								
Deviation, DPM	8	5	12	5				
Maximum DPM	28	17	39	19				
Action Limit,	700	700	700	700				
	,00	, , , , , , , , , , , , , , , , , , , ,	, , , , , , , , , , , , , , , , , , , ,	,				

Table 12-3-6a. F STATISTICAL SAME	LES FOR	tus Surve IMPACTED	y Summary for 1 SURVEY UNIT	RANDOM			
Building: 053							
C. Survey	Unit or	Room Num	ber: 101-2				
Survey Date: 06/	09/00	_	Count Times:	1 minute			
Liquid Scintillation Swipe R	Results for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination			
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14			
	H-3	C-14					
1	4	5	2	10			
2	20	1	5	5			
3	2	10	0	13			
4	1	17	21	9			
5	0	3	15	14			
6	2	12	6	12			
7	4	5	5	7			
8	28	1	0	25			
9	14 2 0 8						
10	7	9	6	12			
11	2	12	14	6			
12	6	14	7	0			
13	8	7	5	5			
14	19	4	91	5			
15	19	6	6	14			
16	7	12	12	9			
17	6	14	8	9			
18	6	14	1	18			
19	0	10	11	9			
20	0	7	25	15			
21	24	8	11	7			
Average DPM	9	8	12	10			
Standard							
Deviation, DPM	9	5	19	5			
Maximum DPM	28	17	91	25			
Action Limit, DPM	700	700	700	700			

Table 12-3-6b. F	inal Stat	tus Surve	y Summary for (	CONTROL
Building: 053	LEG FOR .		SORVEI ONII	
D Survey	Unit or	Room Num	ber: 101-2	
Survey Date: 06/	09/00	Room Ruin	Count Times:	1 minute
Liquid Scintillation Swipe R	esults for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14
	H-3	C-14		
1	4	5	19	0
2	20	1	7	9
3	2	10	2	10
4	1	17	9	2
5	0	5	14	6
6	2	12	0	28
7	4	5	8	12
8	28	1	424	0
9	14	2	14	19
10	7	9	2	15
11	2	12	3	10
12	6	14	18	6
13	8	7		
14	19	4		
15	19	6	5	
16	7	12		
17	6	14		
18	6	14		
19	0	10		
20	0	7		
21	24	8		
Average DPM	9	8	43	10
Standard				
Deviation, DPM	9	5	120	8
Maximum DPM	28	17	424	28
Action Limit, DPM	700	700	700	700

		TO THI 101	E OTHI	ER HAI	JF OF	ROOM			
			C6		C7				R1
R2	R3	R4		C8	С9		R5		R6
	R7	R8	R9		R10			R11	R12
R13	R14		C10		R15	R16	C11		
R17		R18			R19				
		R20		R21		C12		C13	
		C14			C15		C16		
				1					

Figure 12-3-1a. Grid Map for Building 53, Room 101-1

Legend:

→ = Main Entrance C17 = Left Sink C18 = Right Sink

	1	TO O	THER	HALF	OF ROO	OM 101		1	1
1			R2	83		anna an th	RA		

R9

R16

↑

С6

R13

R21

R14

C8

C10

R17

R10

R15

R8

R20

С9

C11

Figure 12-3-1b. Grid Map for Building 53, Room 101-2 Legend:

$$\rightarrow$$

**R**5

R11

R7

C7

R18

R12

R19

R6

= Main Entrance C12 = Sink

Table 12-3-7. Final Status Survey Summary for RANDOM									
STATISTICAL SAME	LES FOR	IMPACTED	SURVEY UNIT						
Building: 053		Survey U	nit or Room Nu	mber: 102					
Survey Date: 03/	23/00	Count Ti	mes: 1 minute						
Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for Residual Contamir									
Sample Number	DPM, Bad	Background DPM, H-3 DPM, C							
	H-3	H-3 C-14							
1	3	7	9	4					
2	14	9	0	8					
3	11	14	7	7					
4	8	2	17	б					
5	12	4	8	4					
6	3	5	0	8					
7	0	17	5	17					
8	8	7 6 5							
9	3	3 10 6 7							
10	13	7	30	14					
11	16	14	9	20					
12	9	4	0	6					
13	7	9	26	12					
14	6	12	0	2					
15	29	5	13	7					
16	0	19	4	19					
17	3	10	17	15					
18	15	4	1	18					
19	10	2	23	12					
20	2	10	2	18					
21	31	8	9	11					
Average DPM	10	9	9	10					
Standard									
Deviation, DPM	8	5	9	6					
Maximum DPM	31	19	30	20					
Action Limit,									
DPM	700	700	700	700					

Table 12-3-8. Final Status Survey Summary for CONTROL								
SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053		Survey Unit or Koom Number: 102						
Survey Date: 03/2	23/00	Count Times: 1 minute						
Liquid Scintillation Swipe Re	esults for Backg	round, H-3 and C-14 Analyses for Residual Contamination						
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, C-14				
	2 2 1	7		E				
	3	/						
2	14	9	0					
<u>_</u>	11	14	<u> </u>	14				
4	8	2	<u> </u>	10				
5	12	4	4	/				
0	3	5	0	22				
	0		0	18				
8	8		0	5				
9	3	10	15	16				
10	13	7	7	9				
11	16	14	7	7				
12	9	4	13	11				
13	7	9	10	14				
14	6	12	9	12				
15	29	5	24	11				
16	0	19	6	17				
17	3	10	3	7				
18	15	4	1	15				
19	10	2	25	6				
20	2	10	13	4				
21	31	8	2	12				
22			11	4				
23			3	10				
24	_		7	7				
25			11	9				
26			0	2				
27			0	5				
28			2	12				
Average DPM	10	9	7	10				
Standard								
Deviation, DPM	8	5	7	5				
Maximum DPM	31	19	25	22				
Action Limit,								
DPM	700	700	700	700				

	Rl	R2				$\square$	
					R3		
R4				R5	R6		
	R7		C6		R8	R9	R10
R11		с7			R12		
R13		R14	н				
		C8	R15	С9	R16		R17
		R18	C10				
C11			R19			R20	
		C12			R21		
					C13		
	C14			C15	G		
	1		C16		C17		

Figure 12-3-2. Grid Map for Building 53, Room 102 Legend: —> Main Entrance; C18 = Sink #1 and Counter; C19 = Sink #2 and Counter; G Gamma Counter; C20 = Sink #3 and Counter; C21- C25 = Secondary Entrance; I = Incubator; C26 = Floor Centrifuge; C27 = Gamma Counter; C28 = Desk Centrifuge

Table 12-3-9. Final Status Survey Summary for RANDOM								
STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053 Survey Unit or Room Number: 103								
Survey Date: 03/23/00 Count Times: 1 minute								
Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for Residual Contaminati								
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	3	7	15	6				
2	14	9	22	13				
3	11	14	12	11				
4	8	2	3	5				
5	12	4	4	5				
6	3	5	6	14				
7	0	17	30	13				
8	8	7	11	2				
9	3	10	9	2				
10	13	7	14	14				
11	16	16 14 13		9				
12	9	4	7	9				
13	7	9	21	13				
14	6	12	1	17				
15	29	5	12	11				
16	0	19	7	7				
17	3	10	13	2				
18	15	4	31	3				
19	10	2	31	1				
20	2	10	31	6				
21	31	8	20	6				
Average DPM	10	9	15	8				
Standard								
Deviation, DPM	8	5	10	5				
Maximum DPM	31	19	31	17				
Action Limit,								
DPM	700	700	700	700				

Table 12-3-10. Final Status Survey Summary for CONTROL								
SAMPLES FOR IMPACTED SU	JRVEY UNIT							
Building: 053	Survey Unit or Room Number: 103							
Survey Date: 03/23/00	Count Time	s: 1 mir	nute					
Liquid Scintillation Swipe Results for B	ackground, H-3 an	d C-14 Analyse	s for Residual C	Contamination				
Sample Number	DPM, Bacl	kground	DPM,	DPM,				
	H-3	C-14	H-3	C-14				
1	3	7	22	8				
2	14	9	0	20				
3	11	14	17	9				
4	8	2	13	4				
5	12	4	4	7				
6	3	5	3	7				
7	0	17	4	7				
8	8	7	4	22				
9	3	10	14	6				
10	13	7	3	10				
11	16	14	3	10				
12	9	4	11	14				
13	7	9	10	12				
14	б	12	10	14				
15	29	5	17	16				
16	0	19	2	10				
17	3	10	0	13				
18	15	4	0	13				
19	10	2	35	0				
20	2	10	0	5				
21	31	8	15	4				
22			14	14				
23			4	5				
24			7	7				
25			3	24				
26			6	12				
27			12	9				
28			17	11				
29			13	9				
Average DPM	10	9	9	10				
Stand. Deviation, DPM	8	5	8	5				
Maximum DPM	31	19	35	24				
Action Limit, DPM	700	700	700	700				

$\overline{\}$								$\square$
				R1	R2	R3		
			R4					
	C6				R5			
R6			R7				R8	
R9				R10				R11
R12		C7		R13			C8	
R14					R15			
R16			R17			C9 G	R18	
		С10 Н		R19		C11		
C12		R20	R.21			C13		
		C14			C15			
				C16			$\sum$	
	×							

Figure 12-3-3. Grid Map for Building 53, Room 103

Legend: — Main Entrance; C18 = Left Rear Counter; C19 = Left Front Counter; G = Liquid Scintillation Counter; C20 = Right Front Counter; C21 = Right Rear Counter; H = Hyperbaric Chamber; C22 = Left Sink; C23 = Right Front Sink; C24 = Right Rear Sink

Table 12-3-11. Final Status Survey Summary for RANDOM								
STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053 Survey Unit or Room Number: 111								
Survey Date: 4/2	6/00	Count Ti	mes: 1 minute					
Liquid Scintillation Swipe R	esults for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination				
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	22	13	7	24				
2	26	1	11	9				
3	7	12	17	4				
4	16	16	14	14				
5	15	0	17	9				
6	14	6	34	5				
7	8	7	16	6				
8	11	0	15	14				
9	7	12	9	2				
10	23	3	35	3				
11	14	4	12	6				
12	13	б	11	4				
13	13	6	17	4				
14	36	5	13	9				
15	13	9	10	4				
16	4	5	4	7				
17	9	4	3	7				
18	21	4	3	10				
19	23	11	0	10				
20	24	1	11	9				
21	7	14	24	3				
Average DPM	16	7	13	8				
Standard								
Deviation, DPM	8	5	9	5				
Maximum DPM	36	16	35	24				
Action Limit, DPM	700	700	700	700				

Table 12-3-12. Final Status Survey Summary for CONTROL								
SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053 Survey Unit or Room Number: 111								
Survey Date: 4/26/00 Count Times: 1 minute								
Liquid Scintillation Swipe R	esults for Back	ground, H-3 an	d C-14 Analyses for Resid	lual Contamination				
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	22	13	0	12				
2	26	1	14	18				
3	7	12	15	9				
4	16	16	0	0				
5	15	0	15	9				
6	14	6	0	7				
7	8	7	2	12				
8	11	0	9	12				
9	7	12	33	12				
10	23	3	11	11				
11	14	4	3	12				
12	13	6	21	6				
13	13	6	3	7				
14	36	5	9	0				
15	13	9	12	6				
16	4	5	8	4				
17	9	4	2	12				
18	21	4	0	10				
19	23	11	21	6				
20	24	1	11	9				
21	7	14	19	13				
22			8	4				
23			7	7				
Average DPM	16	7	10	9				
Standard								
Deviation, DPM	8	5	8	4				
Maximum DPM	36	16	33	18				
Action Limit, DPM	700	700	700	700				

		R1	R2		R3		
		R4	R5	R6	R7		
R8	R9		C6		R10		
R11			R12		R13		
			R14	R15	R16		
R17		с7			R18		
			C8				
С9					C11		C10
		R19		R20		C12	
		R21	1				
			C13				

Figure 12-3-4. Grid Map for Building 53, Room 111

Legend:

→ = Main Entrance

Table 12-3-13. Final Status Survey Summary for RANDOM								
STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053 Survey Unit or Room Number: 112								
Survey Date: 03/23/00 Count Times: 1 minute								
Liquid Scintillation Swipe I	Results for Back	ground, H-3 an	d C-14 Analyses for Resid	lual Contamination				
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	3	7	21	8				
2	14	9	18	4				
3	11	14	17	11				
4	8	2	3	5				
5	12	4	14	14				
6	3	5	3	7				
7	0	17	10	0				
8	8	7	10	2				
9	3	10 3 10						
10	13	7 2		12				
11	16	14 8		7				
12	9	4	20	4				
13	7	9	3	10				
14	6	12	10	11				
15	29	5	34	0				
16	0	19	4	2				
17	3	10	8	9				
18	15	4	25	8				
19	10	2	6	14				
20	2	10	13	9				
21	31	8	2	10				
Average DPM	10	9	11	7				
Standard								
Deviation, DPM	8	5	9	4				
Maximum DPM	31	19	.34	14				
Action Limit,	200	200	700	7.0.0				

Table 12-3-14. Final Status Survey Summary for CONTROL								
Building: 053 Survey Unit or Room Number: 112								
Survey Date: 03/	23/00	Count Ti	mes: 1 minute					
Liquid Sciptillation Swine Results for Background H-3 and C-14 Analyses for Residual Contamination								
Sample Number	DPM Background DPM H-3 DPM							
	H-3	C-14						
1	3	7	35	5				
2	14	9	1	15				
3	11	14	30	15				
4	8	2	7	9				
5	12	4	6	7				
6	3	5	13	4				
7	0	17	19	1				
8	8	7	19	13				
9	3	3 10 17 9						
10	13	7	17	9				
11	16	14	27	6				
12	9	4	0	5				
13	7	9	0	5				
14	6	12	0	8				
15	29	5						
16	0	19						
17	3	10						
18	15	4						
19	10	2						
20	2	10						
21	31	8						
Average DPM	10	9	14	8				
Standard								
Deviation, DPM	8	5	12	4				
Maximum DPM	31	19	35	15				
Action Limit, DPM	700	700	700	700				

				C6				
		R1	R2		R3	R4		
			R5		C7			
R6		C8		С9			R7	
	R8		C10			R9		
R10	R11	R12	R13			R14		
	R15			C11	C12			
		R16	R17	R18	R19			
				R20	R21			

Figure 12-3-5. Grid Map for Building 53, Room 112 Legend:

→ = Main Entrance; C13 = Sink; C14 = Counter

Table 12-3-15. Final Status Survey Summary for RANDOM								
STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053		Survey U	nit or Room Nu	mber: 112A				
Survey Date: 03/23/00 Count Times: 1 minute								
Liquid Scintillation Swipe R	csults for Back	ground, H-3 an	d C-14 Analyses for Resid	fual Contamination				
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	3	7	3	7				
2	14	9	22	18				
3	11	14	35	5				
4	8	2	10	11				
5	12	4	13	4				
б	З	5	7	9				
7	0	17	9	9				
8	8	7	10	9				
9	3	10	14	9				
10	13	7	23	13				
11	16	14	12	7				
12	9	4	2	10				
13	7	9	6	14				
14	6	12	5	7				
15	29	5	2	12				
16	0	19	7	17				
17	3	10	14	4				
18	15	4	10	9				
19	10	2	34	6				
20	2	10	18	1				
21	31	8	10	7				
Average DPM	10	9	13	9				
Standard								
Deviation, DPM	8	5	9	4				
Maximum DPM	31	19	35	18				
Action Limit, DPM	700	700	700	700				

Table 12-3-16. Final Status Survey Summary for CONTROL							
SAMPLES FOR IMPA	CTED SURV	VEY UNIT					
Building: 053		Survey Unit or Room Number: 112A					
Survey Date: 03/	23/00	Count Ti	mes: 1 minute				
Liquid Scintillation Swipe R	esults for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination			
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14			
	H-3	C-14					
1	3	7	6	12			
2	14	9	18	0			
3	11	14	4	19			
4	8	2	3	7			
5	12	4	18	9			
6	3	5	7	7			
7	0	17	27	1			
8	8	7	2	10			
9	3	10	20	8			
10	13	7	12	7			
11	16	14	9	12			
12	9	4	5	2			
13	7	9	14	4			
14	6	12	24	1			
15	29	5	21	6			
16	0	19					
17	3	10					
18	15	4					
19	10	2					
20	2	10					
21	31	8					
Average DPM	10	9	13	7			
Standard							
Deviation, DPM	8	5	8	5			
Maximum DPM	31	19	27	19			
Action Limit, DPM	700	700	700	700			

		R1	R2	R3			
		Г Т	, P 1	C13	R5	+	R6
R7	R8	a n k					R9
R10	R11	R12	C11	C12			
					R13	R14	
	C10	С9	C8			R15	R16
		<u>с</u> 7	C6	R17	R18		
			R19	R20	R21		

Figure 12-3-6. Grid Map for Building 53, Room 112A Legend: = Main Entrance; C14 = Sink; C15 = Counter

Table 12-3-17. Final Status Survey Summary for RANDOM								
Building: 053	LES FOR .	IMPACTED	SURVEY UNIT					
A Survey	Init or	Doom Num	hor: 112					
A. Survey	A. Survey Unit or Room Number: 113							
Liquid Sciptillation Swipe Results for Background H-3 and C-14 Analyses for Residual Contamination								
Sample Number	ample Number DPM, Background DPM, H-3 DPM, C-14							
	H-3	C-14						
1	3	7	7	7				
2	14	9	16	4				
3	11	14	12	7				
4	8	2	12	7				
5	12	4	4	2				
6	3	5	22	1				
7	0	17	0	5				
8	8	7	23	9				
9	3	10	14	14				
10	13	7	11	9				
11	16	14	13	18				
12	9	4	17	4				
13	7	9	11	11				
14	6	12	4	7				
15	29	5	5	0				
16	0	19	5	14				
17	3	10	9	12				
18	15	4	13	0				
19	10	2	9	9				
20	2	10	7	9				
21	31	8	11	2				
Average DPM	10	9	11	7				
Standard								
Deviation, DPM	8	5	6	5				
Maximum DPM	31	19	23	18				
Action Limit, DPM	700	700	700	700				

.

Table 12-3-18. Final Status Survey Summary for CONTROL								
SAMPLES FOR IMPA	CTED SURV	VEY UNIT						
Building. 000								
B. Survey	Unit or	Room Num	ber: 113					
Survey Date: 03/	Survey Date: 03/23/00 Count Times: 1 minute							
Liquid Scintillation Swipe R	Results for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Containination				
Sample Number	DPM, Bac	ckgrouna	DPM, H-3	DPM, C-14				
1	н-3 З	$\frac{-14}{7}$	13	1				
2	1.4	, 	5	17				
2	11	1.4						
		14	7	9				
4	8	2	8	2				
5	12	4	12	9				
6	3	5	3	7				
7	0	17	8	4				
8	8	7	16	9				
9	3	10	23	4				
10	13	7	4	5				
11	16	14	24	5				
12	9	4	18	9				
13	7	9	11	14				
14	6	12	20	11				
15	29	5	8	9				
16	0	19	6	14				
17	3	10	5	2				
18	15	4	8	7				
19	10	2	0	12				
20	2	10	2	10				
21	31	8	8	2				
22			0	10				
Average DPM	10	9	10	8				
Standard								
Deviation, DPM	8 5 7 4							
Maximum DPM	31	19	24	17				
Action Limit, DPM	700	700	700	700				

		C6					
			R1	R2			
R3	R4			R5	R6		
R7			R8		С9	R9	
	С7	Е		R10			
	R11			C8		R12	
C10		<u> </u>	R13			R14	
	R15	l	R16	R17			
R18			C11		R19	R20	
R21				-			
		C13					
			1	C12			
					,		

Figure 12-3-7. Grid Map for Building 53, Room 113

Legend:

Table 12-3-19. Final Status Survey Summary for RANDOM STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT						
Building: 053		Survey U	nit or Room Nu	mber: 117		
Survey Date: 03/	23/00	Count Ti	mes: 1 minute			
Liquid Scintillation Swipe F	esults for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination		
Sample Number	DPM, Bac	ckground DPM, H-3		DPM, C-14		
	H-3	C-14				
1	3	7	12	9		
2	14	9	10	2		
3	11	14	14	6		
4	8	2	2	5		
5	12	4	22	1		
6	3	5	17	4		
7	0	17	26	8		
8	8	7	4	7		
9	3	10	13	б		
10	13	7	9	7		
11	16	14	10	4		
12	9	4	26	11		
13	7	9	12	11		
14	6	12	2	10		
15	29	5	15	11		
16	0	19	16	4		
17	3	10	21	2		
18	15	4	0	7		
19	10	2	8	7		
20	2	10	13	7		
21	31	8	12	14		
Average DPM	10	9	13	7		
Standard						
Deviation, DPM	8	5	7	3		
Maximum DPM	31	19	26	14		
Action Limit,						
DPM	700	700	700	700		

Table 12-3-20. Final Status Survey Summary for CONTROL SAMPLES FOR IMPACTED SURVEY UNIT							
Building: 053		Survey U	nit or Room Nu	mber: 117			
Survey Date: 03/	23/00	Count Ti	mes: 1 minute				
Liquid Scintillation Swipe F	Results for Back	ground, H-3 an	d C-14 Analyses for Resid	dual Contamination			
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14			
	H-3	C-14					
1	3	7	21	131			
2	14	9	5	17			
3	11	14	17	16			
4	8	2	21	8			
5	12	4	26	18			
6	3	5	6	12			
7	0	17	7	7			
8	8	7	17	4			
9	3	10	2	12			
10	13	7	0	13			
11	16	14	8	20			
12	9	4	8	7			
13	7	9	12	9			
14	6	12	7	7			
15	29	5	30	6			
16	0	19	0	3			
17	3	10	5	2			
18	15	4	23	15			
19	10	2	4	7			
20	2	10	13	14			
21	31	8	0	7			
22			4	5			
23			9	2			
24			8	4			
Average DPM	10	9	11	14			
Standard							
Deviation, DPM	8	5	9	25			
Maximum DPM	31	19		131			
Action Limit, DPM	700	700	700	700			

			R1	R2				
R3		C6			C4			
	R5	С7			C8			
R8				C9			R7	R8
	R9		R10	:	R11		R12	R13
	R14	R15		C10		C11		R16
			C12	817	н		R18	
C13			C14	1	R19			
		C15				R20		R21
				C16				
				1		C17		
			C18					

Figure 12-3-8. Grid Map for Building 53, Room 117

Legend:

→ = Main entrance; H = Hyperbaric Chamber; Extra Controls = Sink, Counters and Benches

Table 12-3-21. Final Status Survey Summary for RANDOM						
Building: 053		Survey U	nit or Room Nur	mber: 118		
Survey Date: 03/	23/00	Count Times: 1 minute				
Liquid Scintillation Swipe R	esults for Back	eround H-3 an	d C-14 Analyses for Resid	lual Contamination		
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, $C-14$		
_	H-3	C-14				
1	3	7	62	0		
2	14	9	63	6		
3	11	14	5	17		
4	8	2	18	11		
5	12	4	32	6		
6	3	5	0	3		
7	0	17	21	14		
8	8	.7	2	12		
9	3	10	3	10		
10	13	7	11	9		
11	16	14	9	7		
12	9	4	28	18		
13	7	9	20	4		
14	6	12	8	2		
15	29	5	11	9		
16	0	19	15	4		
17	3	10	13	7		
18	15	14	20	13		
19	10	2	15	16		
20	2	10	20	б		
21	31	8	4	5		
Average DPM	10	9	18	9		
Standard						
Deviation, DPM	8	5	17	5		
Maximum DPM	31	19	63	18		
Action Limit, DPM	700	700	700	700		

Table 12-3-22. Final Status Survey Summary for CONTROL								
STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053		Survey Unit or Room Number: 118						
Survey Date: 03/	23/00	Count Ti	mes: 1 minute					
Liquid Scintillation Swipe R	lesults for Back	ground, H-3 an	d C-14 Analyses for Resid	tual Contamination				
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	3	7	7	9				
2	14	9	6	14				
3	11	14	8	4				
4	8	2	11	2				
5	12	4	5	17				
6	3	5	11	2				
7	0	17	6	17				
8	8	7	21	14				
9	3	10	8	9				
10	13	7	8	14				
11	16	14	0	15				
12	9	4	7	31				
13	7	9	7	4				
14	б	12	3	5				
15	29	5	6	17				
16	0	19						
17	3	10						
18	15	14						
19	10	2						
20	2	10						
21	31	8						
Average DPM	10	9	8	12				
Standard								
Deviation, DPM	8	5	5	8				
Maximum DPM	31	19	21	31				
Action Limit, DPM	700	700	700	700				
						R1		
-----	-----	-----	-----	-----	----	-----	-----	--
			C6					
R2	R3			R4		R5		
		C7			R6		R7	
				R8	R9	R10	R11	
R12	R13	R14		R15				
	R16	R17		R18			C8	
С9				C10				
		R19	R20					
				1		R21		
				C11				

Figure 12-3-9. Grid Map for Building 53, Room 118

Legend: —> = Main entrance C12 = Left Counter C13 = Right Counter C14 = Left Sink C15 = Right Sink

Table 12-3-23. Final Status Survey Summary for RANDOM								
STATISTICAL SAMP	LES FOR	IMPACTED	SURVEY UNIT					
Building: 053								
C. Survey	Unit or	Room Num	ber: 119					
Survey Date: 06/	09/00		Count Times:	1 minute				
Liquid Scintillation Swipe R	cesults for Back	ground, H-3 an	d C-14 Analyses for Resi	dual Contamination				
Sample Number	DPM, Bad	DPM, Background DPM, H-3 DPM,						
	H-3	C-14						
1	4	5	16	6				
2	20	1	26	3				
3	2	10	8	7				
4	1	17	0	10				
5	0	3	10	14				
6	2	12	7	9				
7	4	5	9	4				
8	28	28 1 14 21						
9	14	14 2 9 19						
10	7	9	11	125				
11	2	12	4	5				
12	6	14	15	2				
13	8	7	8	7				
14	19	4	0	0				
15	19	6	8	2				
16	7	12	13	7				
17	6	14	11	11				
18	6	14	1	12				
19	0	10	0	20				
20	0	7	11	21				
21	24	8	8	7				
Average DPM	9	8	9	15				
Standard								
Deviation, DPM	9	5	6	26				
Maximum DPM	28	17	26	125				
Action Limit, DPM	700	700	700	700				

Table 12-3-24. Final Status Survey Summary for CONTROL STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT									
Building: 053	Building: 053								
D. Survey Unit or Room Number: 119									
Survey Date: 06/09/00 Count Times: 1 minute									
Liquid Scintillation Swipe Results for Background, H-3 and C-14 Analyses for Residual Contamination									
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, C-14					
	H-3	C-14							
1	4	5	29	8					
2	20	1	25	11					
3	2	10	2	10					
4	1	17	11	11					
5	0	3	5	17					
6	2	12	24	18					
7	4	5	10	4					
8	28	1	29	6					
9	14	2	14	6					
10	7	9	31	1					
11	2	12	31	10					
12	6	14	19	1					
13	8	7	29	13					
14	19	4	32	8					
15	19	6	18	6					
16	7	12	14	0					
17	6	14	8	4					
18	6	14	52	11					
19	0	10	21	13					
20	0	7							
21	24	8							
Average DPM	9	8	21	8					
Standard									
Deviation, DPM	9	5	12	5					
Maximum DPM	28	17	52	18					
Action Limit, DPM	700	700	700	700					

$\backslash$			R1	R2	R3	R4			
					R5				
R6		C6	R7		R8	C7	R9		
R10		C8		R11	C9		R12		
R13	R14	C10		R15			R16	R17	R18
					R19	C11		R20	
			C12				R21		
		C13		C14		C15			

Figure 12-3-10. Grid Map for Building 53, Room 119

Legend:

 $\rightarrow$ 

= Main Entrance ; C16 = Sink; C17 = Hood; C18 = Center Counter; C19 Middle Counter

Table 12-3-25. Final Status Survey Summary for RANDOM								
STATISTICAL SAMP	STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT							
Building: 053								
E. Survey Unit or Room Number: 120								
Survey Date: 06/	09/00		Count Times:	1 minute				
Liquid Scintillation Swipe Results for Background, H-3 and								
C-14 Analyses fo	or Residual Contamination							
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14				
	H-3_	C-14						
1	4	5	3	7				
2	20	1	8	7				
3	2	10	33	5				
4	1	17	1	15				
5	0	3	4	5				
6	2	12	2	10				
7	4	5	0	7				
8	28	1	13	9				
9	14	2	0	10				
10	7	9	9	4				
11	2	12	0	7				
12	б	14	7	12				
13	8	7	0	17				
14	19	4	14	2				
15	19	6	12	7				
16	7	12	5	19				
17	6	14	24	8				
18	6	14	4	2				
19	0	10	3	7				
20	0	7	9	7				
21	24	8	21	9				
Average DPM	9	8	8	8				
Standard								
Deviation, DPM	9	5	9	4				
Maximum DPM	28	17	33	19				
Action Limit, DPM	700	700	700	700				

Table 12-3-26a. Final Status Survey Summary for CONTROL										
STATISTICAL SAME	PLES FOR :	IMPACTED	SURVEY UNIT							
Building: 053										
Survey Unit or Room Number: 120										
Survey Date: 06/	Survey Date: 06/09/00 Count Times: 1 minute									
Liquid Scintillation Swipe Results for Background, H-3 and										
C-14 Analyses for Residual Contamination										
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, C-14						
	H-3	C-14								
1	44	5	13	7						
2	20	1	0	18						
3	2	10	15	4						
4	1	17	7	9						
5	0	3	0	5						
6	2	12	16	11						
7	4	5	6	17						
8	28	1	0	7						
9	14	2	3	10						
10	7	9								
11	2	12								
12	б	14								
13	8	7								
14	19	4								
15	19	6								
16	7	12								
17	6	14								
18	6	14								
19	0	10								
20	0	7								
21	24	8								
Average DPM	9	8	7	10						
Standard		-								
Deviation, DPM	9	5	7	5						
Maximum DPM	28	17	16	18						
Action Limit, DFM	700	700	700	700						

Table 12-3-26b. Final Status Survey Summary for CONTROL STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053	Building: 053							
Survey Unit or R	oom Numbe	er: 120D						
Survey Date: 06/	09/00		Count Times:	1 minute				
Liquid Scintilla	tion Swip	pe Result	s for Backgrou	nd, H-3 and				
C-14 Analyses fo	r Residua	al Contam	ination					
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14				
	H-3	C-14						
1	4	5	3	7				
2	20	1	3	7				
3	2	10	2	12				
4	1	17	3	7				
5	0	3	0	10				
6	2	12	12	9				
7	4	5	6	14				
8	28	11	1	14				
9	14	2	10	19				
10	7	9	17	9				
11	2	12	22	6				
12	6	14	13	4				
13	8	7	13	9				
14	19	4	13	4				
15	19	6	0	10				
16	7	12	21	4				
17	6	14	25	6				
18	6	14	14	16				
19	0	10	11	11				
20	0	7	4	5				
21	24	8	0	15				
22			18	4				
23			9	2				
24			18	4				
25			2	12				
26			0	12				
27			9	4				
28			18	6				
29			17	14				
30			13	7				
31			0	15				
Average DPM	9	8	10	9				
Standard								
Deviation, DPM	9	5	8	4				

## NMRC Diving Buildings Radiological (D&D) Final Status Survey Report, June 2000

Maximum DPM	28	17	25	19
Action Limit,				
DPM	700	700	700	700

$\square$		R1				
R2	R3		R4	R5		
			C6		R6	R7
	R8	R9	R10	R11		
R12			С7			
		C8		R13	R14	R15
	R16		С9	R17	R18	
		R19		T _{R20}		
			R21			

Figure 12-3-11. Grid Map for Building 53, Room 120 Legend:

 $\rightarrow$ 

= Main Entrance

		6	8		
		7	9		$\sum$
24	25	26	27	11	10
22	23	28	29	13	12
20	21	30	31	15	14
		19	17		
		18	16		

Figure 12-3-12. Grid Map for Building 53, Room 120D

Legend:

-> = Main Entrance

Table 12-3-27. Final Status Survey Summary for RANDOM							
STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT							
Building: 053 Survey Unit or Room Number: 121							
Survey Date: 04/	Survey Date: 04/21/00 Count Times: 1 minute						
Liquid Scintilla	tion Swi	pe Result	s for Backgroun	nd, H-3 and			
C-14 Analyses fo	r Residu	al Contam	ination				
Sample Number	DPM, Ba	ckground	DPM, H-3	DPM, C-14			
	H-3	C-14					
1	19	6	76	0			
2	12	7	16	9			
3	16	9	3	7			
4	13	7	8	4			
5	17	9	12	11			
б	2	10	37	0			
7	23	6	3	10			
8	7	7 9 26 11					
9	12	12 11 7 14					
10	13	13 9 3		10			
11	3	7	3	10			
12	23	1	2	10			
13	12	7	22	4			
14	18	9	21	8			
15	2	10	0	7			
16	0	7	15	14			
17	12	11	16	9			
18	7	12	0	10			
19	3	7	8	4			
20	22	6	5	5			
21	12	16	20	4			
Average DPM	12	8	14	8			
Standard							
Deviation, DPM	7	3	17	4			
Maximum DPM	23	16	76	14			
Action Limit, DPM	700	700	700	700			

Table 12-3-28. Final Status Survey Summary for CONTROL								
SAMPLES FOR IMPA	SAMPLES FOR IMPACTED SURVEY UNIT							
Building: 053 Survey Unit or Room Number: 121								
Survey Date: 04/	21/00		Count Times:	1 minute				
Liquid Scintilla	tion Swip	pe Result	s for Backgrou	nd, H-3 and				
C-14 Analyses fo	C-14 Analyses for Residual Contamination							
Sample Number	DPM, Bad	ckground	DPM, H-3	DPM, C-14				
1	H-3	C-14						
1	19	0	9	/				
2	12	/	11	14				
3	16	9	31	3				
4	13	/	5	0				
5	17	9	25	8				
6	2	10	7	20				
7	23	6	17	1				
8	7	9	3	7				
9	12	11	4	5				
10	13	9	14	4				
11	ო	7	25	1				
12	23	1	7	12				
13	12	7	11	11				
14	18	9	7	14				
15	2	10	0	7				
16	0	7	25	13				
17	12	11	17	4				
18	7	12	0	10				
19	3	7	19	4				
20	22	6	3	7				
21	12	16	12	9				
22			8	4				
23			0	12				
24			3	7				
25			17	6				
26			6	17				
27			2	10				
28			26	11				
29			16	6				
Average DPM	12	8	11	8				
Standard		~						
Deviation, DPM	7	3	9	5				
Maximum DPM	23	16	31	20				
Action Limit,				-				
DPM	700	700	700	700				

		R1	R2		R3			
		R4						
		R5	R6		R7			
				C6				
	R8		R9			C19		С7
R10				R11		C18		C8
	R12	R13				С9		
R14	R15			R16		C20	C10	
R17	R18			C11		C12	C13	
R19		R20	R21			C14		
				-	C15			
				<b>↑</b> ^{C17}				
					C16			

Figure 12-3-13. Grid Map for Building 53, Room 121

## Legend:

→ = Main entrance; C18= LEFT (L) COUNTER; C19= L COUNTER, FRONT; C20=FRONT COUNTER; C21=RIGHT (R) COUNTER; C22=L SMALL SINK; C23=L LARGE SINK; C24=L SMALL SINK; C25=L FRONT SINK; C26=R SINK; C27=HOOD; C28= HOOD; C29= BIOHOOD

## NMRC Diving Buildings Radiological (D&D) Final Status Survey Report, June 2000

Table 12-3-29. Final Status Survey Summary for RANDOM STATISTICAL SAMPLES FOR IMPACTED SURVEY UNIT							
Building: 053		Survey Unit or Room Number: 122					
Survey Date: 04/	21/00	Count Times: 1 minute					
Liquid Scintilla	tion Swip	e Results for Background, H-3 and					
C-14 Analyses for Residual Contamination							
Sample Number	DPM, Bac	ckground	DPM, H-3	DPM, C-14			
	H-3	C-14					
1	19	б	18	6			
2	12	7	7	12			
3	16	9	9	4			
4	13	7	0	3			
5	17	9	0	5			
6	2	10	0	15			
7	23	6	13	4			
8	7	9	10	12			
9	12	11	8	4			
10	13	9	8	9			
11	3	7	0	13			
12	23	1	12	2			
13	12	7	21	6			
14	18	9	11	9			
15	2	10	16	6			
16	0	7	14	2			
17	12	11	4	5			
18	7	12	6	12			
19	3	7	11	9			
20 22		6	21	6			
21 12		16	1	17			
Average DPM 12		8	9	8			
Standard							
Deviation, DPM 7		3	7	4			
Maximum DPM 23		16	21	17			
Action Limit,							
DPM	700	700	700	700			

Table 12-3-30. Final Status Survey Summary for CONTROL								
SAMPLES FOR IMPACTED SURVEY UNIT								
Building: 053	Survey Unit or Room Number: 122							
Survey Date: 04/21/00 Count Times: 1 minute								
Liquid Scintillation Swipe Results for Background, H-3 and								
C-14 Analyses for Residual Contamination								
Sample Number	DPM, Bac	ckground	DPM,	DPM, C-				
	H-3	C-14	H-3	14				
1	19	6	15	9				
2	12	7	10	2				
3	16	9	21	6				
4	13	7	17	4				
5	17	9	10	11				
6	2	10	18	1				
7	23	6	9	14				
8	7	9	17	6				
9	12	11	13	2				
10	13	9	0	10				
11	3	7	11	4				
12	23	1	12	7				
13	1.2	7	13	11				
14	18	9	10	9				
15	2	10	16	11				
16	0	7	4	10				
17	12	11	14	2				
18	7	12	14	0				
19	3	7	9	7				
20	22	6	19	6				
21	12	16	10	2				
22			14	9				
23			9	ŗ.				
24			15	16				
25			17	6				
26			2	10				
27			1	12				
28			20	8				
29			12	9				
30			11	9				
31			0	2				
32			2	10				
33			1	15				
34	_		13	4				

35			0	17
Average DPM	12	8	11	8
Standard Deviation,				
DPM	7	3	6	4
Maximum DPM	23	16	21	17
Action Limit, DPM	700	700	700	700

## NMRC Diving Buildings Radiological (D&D) Final Status Survey Report, June 2000

		R1		C6				
		R2						
		R3			R4			
		R5	R6			С7		С8
С9				R7				
	Armony		R8		C10			
	R9	R10	R11	R12				
R13	R14	R15	R16			C11		
R17	R18				C12		C14	
R19	R20		R21			C13		
				C15	C16			
			C18	1				
					C17			



Legend:

,