

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Hillsdale Community Health Center 168 S. Howell Street Hillsdale, MI 49242 REPORT NUMBER(S)	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351
--	---

3. DOCKET NUMBER(S) 030-29588	4. LICENSEE NUMBER(S) 21-24813-01	5. DATE(S) OF INSPECTION July 29, 2008
---	---	--

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		7/29/08

**SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION**

1. LICENSEE Hillsdale Community Health Center		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
REPORT NUMBER(S) 2008-001			
3. DOCKET NUMBER(S) 03029588	4. LICENSE NUMBER(S) 21-24813-01	5. DATE(S) OF INSPECTION June 9, 2008	
6. INSPECTION PROCEDURES USED 87131 (10/24/02)	7. INSPECTION FOCUS AREAS 03.01-03.07		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02120	2. PRIORITY 3	3. LICENSEE CONTACT Cynthia H. Barker, RSO	4. TELEPHONE NUMBER 517-437-5153
------------------------------------	-------------------------	--	--

Main Office Inspection Next Inspection Date: **June 2011**

Field Office **216 Olds Street, Jonesville, MI 49250**

Temporary Job Site Inspection

PROGRAM SCOPE

The licensee was a medical clinic with locations as specified on the license in Hillsdale and Jonesville, Michigan. At Jonesville, MI, no licensed activities have been conducted since 2001 with close-out surveys to release the facility for unrestricted use. The RSO will submit an amendment request to remove this location from the license. At Hillsdale, MI, the nuclear medicine department was staffed with two nuclear medicine technologists (NMT) who routinely conducted an average of 250 procedures each month with the majority of the studies being cardiac. The licensee is not authorized to use I-131 that requires a written directive. The licensee receives a generator and other licensed material is received from a nearby Grand Rapids, MI pharmacy. No change in RSO or NMT staff since the previous inspection. The licensee is also authorized for permanent brachytherapy implants, but none have been performed since the previous inspection.

Performance Observations

During the inspection, the licensee's RSO/NMT demonstrated/discussed: (1) unit dose prep, labeling, and safe use; (2) package check-in and return procedures; (3) wipe test counting; (4) dose calibrator tests; (5) security of license materials; (6) radiation safety program reviews; (7) surveys; (8) sealed source inventory; (9) generator elutions; and (10) minor contamination events.