

RJ TOKARZ MEDICAL IMAGING RADIATION SAFETY CORP  
44 STELTON ROAD, SUITE 115  
PISCATAWAY, NJ 08854  
732-424-0909  
732-424-3715

Q-9

MS-16

August 11, 2008

Stephen Hammann  
Health Physicist  
Commercial and R&D Branch  
United States  
Nuclear Regulatory Commission  
475 Allendale Road  
King of Prussia, PA 19406-1415

29-30809-01

03036233

RECEIVED  
REGION 1  
2008 AUG 13 AM 11:08


RE: Docket 142513

Dear Mr. Hammann:

Enclosed is the report of the closeout survey for ambient exposure rate and surface contamination. The facility performed cardiac imaging with technetium-99m.

Please direct all questions to me at 732-424-0909.

Sincerely,



Robert J. Tokarz  
Medical Physicist  
*Radiation Safety Officer*

Cc:  
Dr. Buck

142513  
NRC/RGNI MATER. ALS-002

## Close-out Survey for Menlo Park Medical at 111 James Street

Menlo Park Medical Group  
111 James Street  
Edison, NJ 08820

Date: 3-Jan-2008

Area 1 Lead L  
Area 2 Dose Calibrator  
Area 3 Hot Lab Floor  
Area 4 Gamma Camera  
Area 5 Gamma Camera Floor  
Area 6 Treadmill Tread  
Area 7 Treadmill Rails  
Area 8 Treadmill Floor  
Area 9 Injection Chair

1.06E-07 Ci assay  
11/11/03 assay  
1/3/08 now  
1514 days  
4.14521958 years  
0.13771494 half-lives  
9.635E-08 Ci now  
213895.926 dpm

Efficiency 0.13939022  
CF 7.17410452

Survey Meter: Ludlum 14C SN 203199 (calibrated 6/14/07)  
Well-Counter: Diagnostic Services; Model 500

Survey Results	
Area	Exposure (mR/h)
HOT LAB BKG	0.05
1	0.05
2	0.05
3	0.05
SCAN ROOM BKG	0.02
4	0.02
5	0.02
6	0.02
7	0.02
8	0.02
9	0.02

Wipe Results			
Area	CPM	NET CPM	DPM
BKG	277		
1	246	-31	-222.4
2	255	-22	-157.8
3	285	8	57.4
4	318	41	294.1
5	299	22	157.8
6	279	2	14.3
7	281	4	28.7
8	324	47	337.2
9	279	2	14.3

**NOTE:** All areas >> 100cm<sup>2</sup>

**NOTE:** <sup>57</sup>Co SOURCES TAKEN BY DIAGNOSTIC SERVICES. HOT LAB SECURED UPON COMPLETION OF EQUIPMENT REMOVAL BY DIAGNOSTIC SERVICES.

Prepared By:

  
Robert J. Tokarz, Medical Physicist

44 Stelton Rd., Suite 115  
Piscataway, NJ 08854  
732-424-0909