GEOVision
geophysical services

SEISMOGRAPH CALIBRATION DATA SHEET REV 4/6/06

INSTRUMENT DATA

SYSTEM MFR:	OYO	MODEL NO.:	3331A
SERIAL NO.:	19029	CALIBRATION DATE:	4/21/2006
BY:	ROBERT STELLER	DUE DATE:	4/21/2007
COUNTER MFR:	HEWLETT PACKARD	MODEL NO.:	5335A
SERIAL NO.:	2626A09881	CALIBRATION DATE:	12/9/2005
BY:	SCE #S1-01252	DUE DATE:	6/9/2006
FCTN GEN MFR:	HEWLETT PACKARD	MODEL NO.:	3325B
SERIAL NO.:	2847A14447	CALIBRATION DATE:	11/3/2005
BY:	SCE #S1-03355	DUE DATE:	11/3/2006

SYSTEM SETTINGS:

GAIN:	10
FILTER:	20 KHZ
RANGE:	100 MILLISEC
DELAY:	0
STACK: 1 (STD)	1
PULSE:	1.6
DISPLAY:	NA
SYSTEM: DATE = CORRECT DATE & TIME	4/21/2006, 10:30AM

PROCEDURE:

SET FREQUENCY TO 100.0HZ SQUAREWAVE WITH AMPLITUDE APPROXIMATELY 0.25 VOLT PEAK. RECORD BOTH ON DISK AND PAPER TAPE, IF AVAILABLE. ANALYZE AND PRINT WAVEFORMS FROM ANALYSIS UTILITY. ATTACH PAPER COPIES OF PRINTOUT AND PAPER TAPES, IF AVAILABLE, TO THIS FORM. AVERAGE FREQUENCY MUST BE BETWEEN 99.0 AND 101.0 HZ.

AS FOUND 100.0 AS LEFT 100.0

WAVEFORM	FILE NO	FREQUENCY	TIME FOR 9 CYCLES Hn	TIME FOR 9 CYCLES Hr	TIME FOR 9 CYCLES V	AVERAGE FREQ.
SQUARE	201	100.0	90.0	90.0	90.0	100.0
SQUARE	202	100.0	90.0	90.0	90.0	100.0
SINE	203	100.0	89.9	89.9	89.9	100.1
SINE	204	100.0	90.0	90.0	89.9	100.0

CALIBRATED BY: ROBERT STELLER 4/21/2006 Robert Steller
NAME DATE SIGNATURE

Seismic recorder/Logger Calibration Data Sheet Rev 1.30 4-6-06

OYO S/N 1902a

Suspension 170 4.25

ID_NO. : 201
HOLE NO. : 0
DEPTH : 0.0 [m]
DATE : 21/04/06 10:41:08 AM
H-SAMPLE RATE: 100 [μSEC]
V-SAMPLE RATE: 100 [μSEC]
PULSE WIDTH : 1.6 [mSEC]
DELAY TIME : 3 [mSEC]

	H1	/H1	V1	H2	/H2	V2
GAIN	X 10	X 10	X 10	X 10	X 10	X 10
LCF[Hz]	5	5	5	5	5	5
HCF[Hz]	20K	20K	20K	20K	20K	20K
STACK	1	1	1	1	1	1

TRACE SIZE : 1
H-TIME SCALE: 1.00 [mSEC/LINE]
V-TIME SCALE: 1.00 [mSEC/LINE]



OYO S/N 1902a

Suspension 170 4.25

ID_NO. : 202
HOLE NO. : 0
DEPTH : 0.0 [m]
DATE : 21/04/06 10:45:05 AM
H-SAMPLE RATE: 100 [μSEC]
V-SAMPLE RATE: 100 [μSEC]
PULSE WIDTH : 1.6 [mSEC]
DELAY TIME : 3 [mSEC]

	H1	/H1	V1	H2	/H2	V2
GAIN	X 10	X 10	X 10	X 10	X 10	X 10
LCF[Hz]	5	5	5	5	5	5
HCF[Hz]	20K	20K	20K	20K	20K	20K
STACK	1	1	1	1	1	1

TRACE SIZE : 1
H-TIME SCALE: 1.00 [mSEC/LINE]
V-TIME SCALE: 1.00 [mSEC/LINE]



OYO S/N 19029

Suspension 170 4.25

ID_NO. : 203
HOLE NO. : 0
DEPTH : 0.0 [m]
DATE : 21/04/06 10:46:15 AM
H-SAMPLE RATE: 100 [μSEC]
V-SAMPLE RATE: 100 [μSEC]
PULSE WIDTH : 1.6 [mSEC]
DELAY TIME : 3 [mSEC]

	H1	/H1	V1	H2	/H2	V2
GAIN	X 10	X 10	X 10	X 10	X 10	X 10
LCF [Hz]	5	5	5	5	5	5
HCF [Hz]	20K	20K	20K	20K	20K	20K
STACK	1	1	1	1	1	1

TRACE SIZE : 1
H-TIME SCALE: 1.00 [mSEC/LINE]
V-TIME SCALE: 1.00 [mSEC/LINE]



OYO S/N 19024

Suspension 170 4.25

ID_NO. : 204
HOLE NO. : 0
DEPTH : 0.0 [m]
DATE : 21/04/06 10:47:47 AM
H-SAMPLE RATE: 100 [μSEC]
V-SAMPLE RATE: 100 [μSEC]
PULSE WIDTH : 1.6 [mSEC]
DELAY TIME : 3 [mSEC]

	H1	/H1	V1	H2	/H2	V2
GAIN	X 10	X 10	X 10	X 10	X 10	X 10
LCF [Hz]	5	5	5	5	5	5
HCF [Hz]	20K	20K	20K	20K	20K	20K
STACK	1	1	1	1	1	1

TRACE SIZE : 1
H-TIME SCALE: 1.00 [mSEC/LINE]
V-TIME SCALE: 1.00 [mSEC/LINE]



TEST NUMBER
501203

EDISON ESI®
A SOUTHERN CALIFORNIA EDISON® Company

Calibration Report

METROLOGY
7300 Fenwick Lane
Westminster, CA 92683
866-723-2257
edisonmetrology.com

GEOVision Geophysical Services
1151 Pomona Road, Unit P
Corona, CA 92882
P.O. No.: 6162-060414-01

Manufacturer: Oyo
Model Number: 3403
Description: Unit, Suspension Telemetry,
Asset Number: 160023
Serial Number: 160023

Calibration Date: 04/21/2006
Calibration Due Date: 04/21/2007
Calibration Interval: 12 Months
Condition As Found: In Tolerance
Condition As Left: In Tolerance

Remarks:
The UUT (unit under test) was calibrated using the customer's procedure. The UUT was operated by the customer's personnel and data collection was observed by SCE personnel. The UUT was found to be in tolerance to customer supplied specifications. The reference standards used are in compliance with ISO/IEC 17025:1999 and laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0. Frequency is accredited.
Please see attached data.

Standards Utilized

I.D. No.	Mfg.	Model No.	Description	Cal. Date	Due Date
S1-01252	Hewlett Packard	5335A OPT 010,203040	Counter, Universal	12/09/2005	06/09/2006
S1-03355	Hewlett Packard	3325B OPT 001, 002	Generator, Function, Synthesizer	11/03/2005	11/03/2006
S1-03686	Fluke	910	Standard, Frequency, Controlled, Gps	01/16/2006	01/16/2007

Procedure: Customer
Temperature: 23° C
Humidity: 40% RH
Test No.: 501203

Calibration Performed By:			Quality Reviewer:	
Branson, Craig A CAB	Metrologist	714-895-0714		04-21-06
Name	Title	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of this laboratory. This report may not be used to claim product endorsement by NVLAP or any agency of the US Government. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, and ISO 9002-94.

Page 2 of 2
MMD
4-21-06

GEOVision
geophysical services

SEISMOGRAPH CALIBRATION DATA SHEET REV 4/6/06

INSTRUMENT DATA

SYSTEM MFR:	OYO	MODEL NO.:	3403
SERIAL NO.:	160023	CALIBRATION DATE:	4/21/2006
BY:	ROBERT STELLER	DUE DATE:	4/21/2007
COUNTER MFR:	HEWLETT PACKARD	MODEL NO.:	5335A
SERIAL NO.:	2626A09881	CALIBRATION DATE:	12/9/2005
BY:	SCE #S1-01252	DUE DATE:	6/9/2006
FCTN GEN MFR:	HEWLETT PACKARD	MODEL NO.:	3325B
SERIAL NO.:	2847A14447	CALIBRATION DATE:	11/3/2005
BY:	SCE #S1-03355	DUE DATE:	11/3/2006

SYSTEM SETTINGS:

GAIN:	2
FILTER:	10 KHZ
RANGE:	100 MILLISEC, 100 MICROSECOND SAMPLE RATE
DELAY:	0
STACK: 1 (STD)	1
PULSE:	1.6
DISPLAY:	NA
SYSTEM: DATE = CORRECT DATE & TIME	4/21/2006, 11:07AM

PROCEDURE:

SET FREQUENCY TO 100.0HZ SQUAREWAVE WITH AMPLITUDE APPROXIMATELY 0.25 VOLT PEAK. RECORD BOTH ON DISK AND PAPER TAPE, IF AVAILABLE. ANALYZE AND PRINT WAVEFORMS FROM ANALYSIS UTILITY. ATTACH PAPER COPIES OF PRINTOUT AND PAPER TAPES, IF AVAILABLE, TO THIS FORM. AVERAGE FREQUENCY MUST BE BETWEEN 99.0 AND 101.0 HZ.

AS FOUND 100.0 AS LEFT 100.0

WAVEFORM	FILE NO	FREQUENCY	TIME FOR 9 CYCLES Hn	TIME FOR 9 CYCLES Hr	TIME FOR 9 CYCLES V	AVERAGE FREQ.
SQUARE	305	100.0	90.0	90.0	90.0	100.0
SQUARE	306	100.0	90.0	90.0	90.0	100.0
SINE	307	100.0	90.0	90.0	90.0	100.0
SINE	308	100.0	90.1	90.0	90.0	100.0

CALIBRATED BY: ROBERT STELLER 4/21/2006 Robert Steller
NAME DATE SIGNATURE

Seismic recorder/Logger Calibration Data Sheet Rev 1.30 4-6-06

TEST NUMBER
501204

EDISON ESI®
A SOUTHERN CALIFORNIA EDISON® Company

Calibration Report

METROLOGY
7300 Fenwick Lane
Westminster, CA 92683
866-723-2257
edisonmetrology.com

GEOVision Geophysical Services
1151 Pomona Road, Unit P
Corona, CA 92882
P.O. No.: 6162-060414-01

Manufacturer: Oyo
Model Number: 3403
Description: Unit, Suspension Telemetry,
Asset Number: 160024
Serial Number: 160024

Calibration Date: 04/21/2006
Calibration Due Date: 04/21/2007
Calibration Interval: 12 Months
Condition As Found: In Tolerance
Condition As Left: In Tolerance

Remarks:
The UUT (unit under test) was calibrated using the customer's procedure. The UUT was operated by the customer's personnel and data collection was observed by SCE personnel. The UUT was found to be in tolerance to customer supplied specifications. The reference standards used are in compliance with ISO/IEC 17025:1999 and laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0. Frequency is accredited.
Please see attached data.

Standards Utilized

I.D. No.	Mfg.	Model No.	Description	Cal. Date	Due Date
S1-01252	Hewlett Packard	5335A OPT 010,203040	Counter, Universal	12/09/2005	06/09/2006
S1-03355	Hewlett Packard	3325B OPT 001, 002	Generator, Function, Synthesizer	11/03/2005	11/03/2006
S1-03686	Fluke	910	Standard, Frequency, Controlled, Gps	01/16/2006	01/16/2007

Procedure: Customer
Temperature: 23° C
Humidity: 40% RH
Test No.: 501204

Calibration Performed By:			Quality Reviewer:	
Branson, Craig A CAB	Metrologist	714-895-0714		04-21-06
Name	Title	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of this laboratory. This report may not be used to claim product endorsement by NVLAP or any agency of the US Government. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, and ISO 9002-94.

GEOVision
geophysical services

SEISMOGRAPH CALIBRATION DATA SHEET REV 4/6/06

INSTRUMENT DATA

SYSTEM MFR:	OYO	MODEL NO.:	3403
SERIAL NO.:	160024	CALIBRATION DATE:	4/21/2006
BY:	ROBERT STELLER	DUE DATE:	4/21/2007
COUNTER MFR:	HEWLETT PACKARD	MODEL NO.:	5335A
SERIAL NO.:	2626A09881	CALIBRATION DATE:	12/9/2005
BY:	SCE #S1-01252	DUE DATE:	6/9/2006
FCTN GEN MFR:	HEWLETT PACKARD	MODEL NO.:	3325B
SERIAL NO.:	2847A14447	CALIBRATION DATE:	11/3/2005
BY:	SCE #S1-03355	DUE DATE:	11/3/2006

SYSTEM SETTINGS:

GAIN:	2
FILTER:	10 KHZ
RANGE:	100 MILLISEC, 100 MICROSECOND SAMPLE RATE
DELAY:	0
STACK: 1 (STD)	1
PULSE:	1.6
DISPLAY:	NA
SYSTEM: DATE = CORRECT DATE & TIME	4/21/2006, 11:30AM

PROCEDURE:

SET FREQUENCY TO 100.0HZ SQUAREWAVE WITH AMPLITUDE APPROXIMATELY 0.25 VOLT PEAK. RECORD BOTH ON DISK AND PAPER TAPE, IF AVAILABLE. ANALYZE AND PRINT WAVEFORMS FROM ANALYSIS UTILITY. ATTACH PAPER COPIES OF PRINTOUT AND PAPER TAPES, IF AVAILABLE, TO THIS FORM. AVERAGE FREQUENCY MUST BE BETWEEN 99.0 AND 101.0 HZ.

AS FOUND 100.0 AS LEFT 100.0

WAVEFORM	FILE NO	FREQUENCY	TIME FOR 9 CYCLES Hn	TIME FOR 9 CYCLES Hr	TIME FOR 9 CYCLES V	AVERAGE FREQ.
SQUARE	401	100.0	90.0	90.0	90.0	100.0
SQUARE	402	100.0	90.0	90.0	90.0	100.0
SINE	403	100.0	89.9	90.0	90.1	100.0
SINE	404	100.0	90.0	90.1	90.1	99.9

CALIBRATED BY: ROBERT STELLER 4/21/2006 Robert Steller
NAME DATE SIGNATURE

Seismic recorder/Logger Calibration Data Sheet Rev 1.30 4-6-06

TEST NUMBER
501205

EDISON ESI®
A SOUTHERN CALIFORNIA EDISON® Company

Calibration Report

METROLOGY
7300 Fenwick Lane
Westminster, CA 92683
866-723-2257
edisonmetrology.com

GEOVision Geophysical Services
1151 Pomona Road, Unit P
Corona, CA 92882
P.O. No.: 6162-060414-01

Manufacturer:	Geometrics	**Calibration Date:**	04/21/2006
Model Number:	STRATAVIEW	**Calibration Due Date:**	04/21/2007
Description:	Siesmograph,	**Calibration Interval:**	12 Months
Asset Number:	75299	**Condition As Found:**	In Tolerance
Serial Number:	75299	**Condition As Left:**	In Tolerance

Remarks:
The UUT (unit under test) was calibrated using the customer's procedure. The UUT was operated by the customer's personnel and data collection was observed by SCE personnel. The UUT was found to be in tolerance to customer supplied specifications. The reference standards used are in compliance with ISO/IEC 17025:1999 and laboratory accreditation criteria established by NIST/NVLAP under the specific scope of accreditation for lab code 105014-0. Frequency is accredited.
Please see attached data.

Standards Utilized

I.D. No.	Mfg.	Model No.	Description	Cal. Date	Due Date
S1-01252	Hewlett Packard	5335A OPT 010,203040	Counter, Universal	12/09/2005	06/09/2006
S1-03355	Hewlett Packard	3325B OPT 001, 002	Generator, Function, Synthesizer	11/03/2005	11/03/2006
S1-03686	Fluke	910	Standard, Frequency, Controlled, Gps	01/16/2006	01/16/2007

Procedure: Customer
Temperature: 23° C
Humidity: 40% RH
Test No.: 501205

Calibration Performed By:			Quality Reviewer:	
Branson, Craig A CAB	Metrologist	714-895-0714	[signature]	04-21-06
Name	Title	Phone	Name	Date

This report may not be reproduced, except in full, without written permission of this laboratory. This report may not be used to claim product endorsement by NVLAP or any agency of the US Government. The results stated in this report relate only to the items tested or calibrated. Measurements reported herein are traceable to SI units via national standards maintained by NIST and were performed in compliance with MIL-STD-45662A, ANSI/NCSL Z540-1-1994, 10CFR50, Appendix B, and ISO 9002-94.

GEOVision
geophysical services

SEISMOGRAPH CALIBRATION DATA SHEET REV 4/6/06

INSTRUMENT DATA

SYSTEM MFR:	GEOMETRICS	MODEL NO.:	STRATAVIEW
SERIAL NO.:	75299	CALIBRATION DATE:	4/21/2006
BY:	ROBERT STELLER	DUE DATE:	4/21/2007
COUNTER MFR:	HEWLETT PACKARD	MODEL NO.:	5335A
SERIAL NO.:	2626A09881	CALIBRATION DATE:	12/9/2005
BY:	SCE #S1-01252	DUE DATE:	6/9/2006
FCTN GEN MFR:	HEWLETT PACKARD	MODEL NO.:	3325B
SERIAL NO.:	2847A14447	CALIBRATION DATE:	11/3/2005
BY:	SCE #S1-03355	DUE DATE:	11/3/2006

SYSTEM SETTINGS:

GAIN:	15 DB
FILTER:	NONE
RANGE:	256 MILLISEC, 31 MICROSECOND SAMPLE RATE
DELAY:	0
STACK: 1 (STD)	1
PULSE:	NA
DISPLAY:	NA
SYSTEM: DATE = CORRECT DATE & TIME	4/21/2006, 12:09PM

PROCEDURE:

SET FREQUENCY TO 100.0HZ SQUAREWAVE WITH AMPLITUDE APPROXIMATELY 0.25 VOLT PEAK. RECORD BOTH ON DISK AND PAPER TAPE, IF AVAILABLE. ANALYZE AND PRINT WAVEFORMS FROM ANALYSIS UTILITY. ATTACH PAPER COPIES OF PRINTOUT AND PAPER TAPES, IF AVAILABLE, TO THIS FORM. AVERAGE FREQUENCY MUST BE BETWEEN 99.0 AND 101.0 HZ.

AS FOUND 100.0 AS LEFT 100.0

WAVEFORM	FILE NO	FREQUENCY	TIME FOR 9 CYCLES Hn	TIME FOR 9 CYCLES Hr	TIME FOR 9 CYCLES V	AVERAGE FREQ.
SQUARE	501	100.0	90.0	90.0	90.0	100.0
SQUARE	502	100.0	90.0	90.0	90.0	100.0
SINE	503	100.0	90.0	90.0	90.0	100.0
SINE	504	100.0	90.0	90.0	90.0	100.0

CALIBRATED BY: ROBERT STELLER 4/21/2006 Robt Steller
NAME DATE SIGNATURE

Seismic recorder/Logger Calibration Data Sheet Rev 1.30 4-8-06

EOMETRICS S/N 75299 StrataView

READ FROM 501.DAT 15:14:03 21/APR/2006
LINE NUMBER 00-00 GROUP INTERVAL 3.28
SHOT LOC 1876.72 PHONE 1 LOC 1800.00 PHONE 6 LOC 1816.40
SAMPLE INTERVAL 031 uS RECORD LEN 256 MS DELAY 0 MS
ACQ FILT LO CUT 0HZ NOTCH 0HZ STACKS 1
DISP FILT HI CUT 250HZ OUT FIXED GAIN



S/N 75299

SAMPLING 0.031 ms LENGTH 256 ms DELAY 0 ms READ FROM 501.DAT
LO CUT 0HZ NOTCH 0HZ DISP FILT PREVIEW STACK 1



Shot point is 1876.72

Channel	1	Phone location 1800.00	Arrival time	9.8 msec
Channel	2	Phone location 1803.28	Arrival time	9.8 msec
Channel	3	Phone location 1806.56	Arrival time	8.9 msec
Channel	4	Phone location 1809.84	Arrival time	189.0 msec
Channel	5	Phone location 1813.12	Arrival time	188.9 msec
Channel	6	Phone location 1816.40	Arrival time	188.5 msec

GEOMETRICS StrataView

READ FROM 502.DAT 15:15:49 21/APR/2006
LINE NUMBER 00-00 GROUP INTERVAL 3.28
SHOT LOC 1876.72 PHONE 1 LOC 1800.00 PHONE 6 LOC 1816.40
SAMPLE INTERVAL 031 uS RECORD LEN 256 MS DELAY 0 MS
ACQ FILT LO CUT 0HZ NOTCH 0HZ STACKS 1
DISP FILT HI CUT 250HZ OUT FIXED GAIN



S/N 75249

SAMPLING 0.031 ms LENGTH 256 ms DELAY 0 ms READ FROM 502.DAT
LO CUT 0HZ NOTCH 0HZ DISP FILT PREVIEW STACK 1



Shot point is 1876.72

Channel	1	Phone location 1800.00	Arrival time	6.4	msec
Channel	2	Phone location 1803.28	Arrival time	6.4	msec
Channel	3	Phone location 1806.56	Arrival time	6.4	msec
Channel	4	Phone location 1809.84	Arrival time	96.4	msec
Channel	5	Phone location 1813.12	Arrival time	96.4	msec
Channel	6	Phone location 1816.40	Arrival time	96.4	msec

GEOMETRICS S/N 75299 StrataView

page 5 of 6

READ FROM 503.DAT 15:17:14 21/APR/2006
LINE NUMBER 00-00 GROUP INTERVAL 3.28
SHOT LOC 1876.72 PHONE 1 LOC 1800.00 PHONE 6 LOC 1816.40
SAMPLE INTERVAL 031 uS RECORD LEN 256 MS DELAY 0 MS
ACQ FILT LO CUT 0HZ NOTCH 0HZ STACKS 1
DISP FILT HI CUT 250HZ OUT FIXED GAIN



S/N 75299

SAMPLING 0.031 ms LENGTH 256 ms DELAY 0 ms READ FROM 503.DAT
LO CUT 0HZ NOTCH 0HZ DISP FILT PREVIEW STACK 1



Shot point is 1876.72

Channel	1	Phone location	1800.00	Arrival time	13.8 msec
Channel	2	Phone location	1803.28	Arrival time	13.8 msec
Channel	3	Phone location	1806.56	Arrival time	13.8 msec
Channel	4	Phone location	1809.84	Arrival time	183.8 msec
Channel	5	Phone location	1813.12	Arrival time	183.8 msec
Channel	6	Phone location	1816.40	Arrival time	183.8 msec

GEOMETRICS S/N 75299 StrataView

READ FROM 504.DAT 15:19:34 21/APR/2006
LINE NUMBER 00-00 GROUP INTERVAL 3.28
SHOT LOC 1876.72 PHONE 1 LOC 1880.00 PHONE 6 LOC 1916.40
SAMPLE INTERVAL 031 uS RECORD LEN 256 MS DELAY 0 MS
ACQ FILT LO CUT 0HZ NOTCH 0HZ STACKS 1
DISP FILT HI CUT 250HZ OUT FIXED GAIN



S/N 75299

SAMPLING 0.031 ms LENGTH 256 ms DELAY 0 ms
OUT OUT DISP FILT PREVIEW STACK 0



Shot point is 1876.72

Channel	1	Phone location 1880.00	Arrival time	6.0 msec
Channel	2	Phone location 1883.28	Arrival time	6.0 msec
Channel	3	Phone location 1886.56	Arrival time	6.0 msec
Channel	4	Phone location 1889.84	Arrival time	96.0 msec
Channel	5	Phone location 1913.12	Arrival time	96.0 msec
Channel	6	Phone location 1916.40	Arrival time	96.0 msec

United States Department of Commerce
National Institute of Standards and Technology

NVLAP®

Certificate of Accreditation to ISO/IEC 17025:1999

NVLAP LAB CODE: 105014-0

Southern California Edison Company
Westminster, CA

is recognized by the National Voluntary Laboratory Accreditation Program for conformance with criteria set forth in NIST Handbook 150:2001 and all requirements of ISO/IEC 17025:1999.
Accreditation is granted for specific services, listed on the Scope of Accreditation, for:

CALIBRATION LABORATORIES

2006-04-01 through 2007-03-31
Effective dates

For the National Institute of Standards and Technology

National Voluntary Laboratory Accreditation Program

SCOPE OF ACCREDITATION TO ISO/IEC 17025:1999

Southern California Edison Company
7300 Fenwick Lane
Westminster, CA 92683
Ms. Jennifer E. Smith
Phone: 714-895-0133 Fax: 714-895-0781
E-mail: Jennifer.Smith@sce.com
URL: http://www.edisonmetrology.com

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

NVLAP Code: 20/A01 ANSI/NCSL Z540-1-1994; Part 1 Compliant

DIMENSIONAL

NVLAP Code: 20/D03
Gage Blocks

Nominal Length in in	*Best Uncertainty (±) in µin* [note 1]
0.01 to < 0.05	1.9
0.05 to < 0.1	1.7
0.1 to < 1.0	1.2
1.0	1.4
2.0	1.8
3.0	2.2
4.0	2.9
5.0	5.4
6.0	5.6
7.0	5.8
8.0	6.0
10.0	6.8
12.0	7.2
16.0	8.1
20.0	9.4

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

Nominal Length in mm	*Best Uncertainty (±) in nm* note 1
0.5 to < 1.0	52
1.0 to < 2.5	44
2.5 to < 25.0	39
25.0	44
50.0	47
75.0	60
100.0	80

NVLAP Code: 20/D11
Spherical Diameter; Ring Gages

Range in inches	*Best Uncertainty (±) in µin* note 1	*Remarks*
0.040 to 0.825	6	Comparison to gage blocks
> 0.825 to 1.510	7	Comparison to gage blocks
> 1.510 to 2.510	8	Comparison to gage blocks
> 2.510 to 4.510	12	Comparison to gage blocks
> 4.510 to 6.510	14	Comparison to gage blocks
> 6.510 to 9.010	16	Comparison to gage blocks
> 9.010 to 12.010	19	Comparison to gage blocks
> 12.010 to 13.25	31	Comparison to gage blocks

ELECTROMAGNETICS - DC/LOW FREQUENCY

NVLAP Code: 20/E02
AC Current

Range	*Best Uncertainty (±) in ppm* note 1 *Frequency in Hz*			
	10	*20*	*40*	*400 to 10 k*
10 mA	270	199	127	116
20 mA	270	199	127	116
30 mA	270	199	127	116
50 mA	286	208	141	130
100 mA	270	199	127	116

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

200 mA		270		199	127		116
300 mA		270		199	127		116
500 mA		270		208	141		130
	10	*20*	*40*	*400 to 5 k*	*10 k*		
1A	270	199	127	116	130		
	10	*20*	*40*	*400 to 10 k*			
2A	271	200	129	118			
3A	271	200	129	118			
	10	*20*	*40*	*400 to 5 k*	*10 k*		
5A	286	209	142	132	148		
	10	*20*	*40*	*400*	*1 k*	*5 k*	*10 k*
10A	273	233	132	121	121	143	143
							400 to 10 k
20A							144

NVLAP Code: 20/E05
DC Current

Range	*Best Uncertainty (±) in ppm* [note 1]
10 nA	2.9
100 nA	2.3
1 µA	2.0
10 µA	2.0
100 µA	2.0
1 mA	1.9
10 mA	1.9
100 mA	1.9
1 A	10.4
10 A	10.4
30 A	20.6

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

DC Resistance

Nominal Value in Ω	*Best Uncertainty (±) in ppm* [note 1]	*Remarks*
100 μ	8.20	Automated DC Resistance Calibration System
1 m	5.50	Automated DC Resistance Calibration System
10 m	3.70	Automated DC Resistance Calibration System
100 m	2.10	Automated DC Resistance Calibration System
1	0.40	Automated DC Resistance Calibration System
10	0.40	Automated DC Resistance Calibration System
25	0.50	Automated DC Resistance Calibration System
100	0.50	Automated DC Resistance Calibration System
1 k	0.50	Automated DC Resistance Calibration System
10 k	0.50	Automated DC Resistance Calibration System
100 k	1.50	Automated DC Resistance Calibration System
1 M	2.30	Automated DC Resistance Calibration System
10 M	3.30	Automated DC Resistance Calibration System
100 M	4.00	Automated DC Resistance Calibration System

NVLAP Code: 20/E06

DC Voltage

Range	*Best Uncertainty (±) in ppm* [notes 1,2]	*Remarks*
1.018 V	0.80	Automated DC Calibration System
10.00 V	0.20	Automated DC Calibration System
1.000 V	0.80	Automated DC Calibration System
1 mV to 100 mV	1.3 [note 6]	Ratiometric Measurement Techniques performed by voltage transfer utilizing a high precision voltage
100 mV	0.7	Ratiometric Measurement Techniques performed by voltage transfer utilizing a high precision voltage
1.0 V	0.3	Ratiometric Measurement Techniques performed by voltage transfer utilizing a high precision voltage

2006-04-01 through 2007-03-31

Effective dates *For the National Institute of Standards and Technology*

NVLAP® National Voluntary Laboratory Accreditation Program

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

10.0 V	0.3	Ratiometric Measurement Techniques performed by voltage transfer utilizing a high precision voltage
20.0 V	0.5	Ratiometric Measurement Techniques performed by voltage transfer utilizing a high precision voltage
100.0 V	0.3	Ratiometric Measurement Techniques performed by voltage transfer utilizing a high precision voltage
1000.0 V	0.7	Ratiometric Measurement Techniques performed by voltage transfer utilizing a high precision voltage

NVLAP Code: 20/E09
LF AC Voltage

Best Uncertainty (±) in ppm *note 1,3,4*
Frequency in Hz

Range	*10*	*20*	*40*	*100*	*1k*	*10k*	*20k*	*50k*	*100k*	*300k*	*500k*	*800k*	*1M*
2 mV	448	912	889	969	379	865	1073	405	1131	1265	2116	2595	2938
10 mV	119	230	102	177	245	169	180	220	343	243	676	425	488
20 mV	83	70	67	67	66	76	76	165	261	361	521	372	442
30 mV	134	111	80	78	62	63	71	133	219	345	535	688	791
100 mV	36	72	23	42	34	35	34	43	77	169	220	287	225
190 mV	36	31	22	20	21	26	21	42	80	136	124	264	215
300 mV	46	61	30	32	34	19	28	36	59	116	143	189	205
1 V	120	36	18	10	13	12	11	25	14	87	102	104	98
1.9 V	36	22	22	9	9	9	8	18	11	94	101	85	89
3 V	26	34	23	17	14	14	13	27	14	100	108	95	97
10 V	20	42	19	10	10	9	10	11	16	80	100	111	100
19 V	26	23	20	11	9	9	10	11	16	98	109	82	82
30 V	30	37	26	19	15	16	19	37	44	118			
100 V	140	46	20	16	15	19	11	40	22				
190 V	47	27	20	20	13	13	13	41	26				
300 V			37	29	18	27	22	29	46				
500 V			33	25	17	20	19	38	32				

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

700 V	29	23	18	17	19	44	54
1000 V	22	23	21	19	22		

TIME AND FREQUENCY

NVLAP Code: **20/F01**
Frequency Dissemination

Range	*Best Uncertainty (±)* [note 1]	*Remarks*
10 MHz	1.2×10^{-12}	GPS Receiver

MECHANICAL

NVLAP Code: **20/M05**
Flow Rate

Nominal Flow Rate	*Best Uncertainty (±) in percent* [notes 1, 2]
(0.8 to 30) L/s	0.3
(0.1 to 800) mL/s	0.4
(0.006 to 0.1) mL/s	0.7

NVLAP Code: **20/M06**
Force

Nominal Force in lb	*Best Uncertainty (±)* [note 1]	*Remarks*
2 to 200	0.025 %	Dead Weight
> 200 to 300	0.086 lb	Proving Ring
> 300 to 500	0.14 lb	Proving Ring
> 500 to 1000	0.28 lb	Proving Ring
> 1000 to 2000	0.55 lb	Proving Ring
> 2000 to 5000	0.84 lb	Proving Ring
> 5000 to 10 000	1.7 lb	Proving Ring
> 10 000 to 20 000	5.5 lb	Proving Ring
> 20 000 to 35 000	5.8 lb	Proving Ring
> 35 000 to 50 000	13 lb	Proving Ring

2006-04-01 through 2007-03-31

Effective dates

For the National Institute of Standards and Technology

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

> 50 000 to 60 000	16 lb	Proving Ring
> 60 000 to 100 000	26 lb	Proving Ring
> 100 000 to 300 000	113 lb	Proving Ring

NVLAP Code: 20/M08
Mass

Range	*Best Uncertainty (±) in mg* [notes 1,2]	*Remarks*
10 kg	2.3	Echelon I
5 kg	0.93	Echelon I
3 kg	0.65	Echelon I
2 kg	0.43	Echelon I
1 kg	0.052	Echelon I
500 g	0.043	Echelon I
300 g	0.041	Echelon I
200 g	0.034	Echelon I
100 g	0.020	Echelon I
50 g	0.013	Echelon I
30 g	0.013	Echelon I
20 g	0.0095	Echelon I
10 g	0.0073	Echelon I
5 g	0.0048	Echelon I
3 g	0.0038	Echelon I
2 g	0.0029	Echelon I
1 g	0.0030	Echelon I
500 mg	0.0017	Echelon I
300 mg	0.0013	Echelon I
200 mg	0.0010	Echelon I
100 mg	0.0009	Echelon I
50 mg	0.0007	Echelon I
30 mg	0.0007	Echelon I
20 mg	0.0005	Echelon I
10 mg	0.0005	Echelon I
5 mg	0.0006	Echelon I
3 mg	0.0006	Echelon I

2006-04-01 through 2007-03-31

Effective dates *For the National Institute of Standards and Technology*

CALIBRATION LABORATORIES **NVLAP LAB CODE 105014-0**

2 mg	0.0005	Echelon I
1 mg	0.0005	Echelon I
30 kg	56	Echelon II
20 kg	22	Echelon II

THERMODYNAMIC

NVLAP Code: 20/T05
Pressure

Range	*Best Uncertainty (±) in ppm* [note 1]	*Remarks*
> 1.5 to 50	20	Gas
> 50 to 1450	45	Gas
> 1450 to 16 000	90	Gas
> 1000 to 10 000	60	Oil
> 10 000 to 30 000	110	Oil
> 30 000 to 50 000	210	Oil

1. Represents an expanded uncertainty using a coverage factor, k = 2, at an approximate level of confidence of 95 %.
2. Approximate value. Actual value determined by the test statistics.
3. All ACV measurements performed via AC/DC transfer system.
4. Uncertainties listed are representative of the laboratory's accredited capabilities within the stated ranges. Accreditation is not limited to only those fixed values shown.
5. Dependent upon principle of operation of device being calibrated and its performance relative to standards at the time of the test.
6. The equation: uncert. = $(A + B/mVDC^2)^{0.5}$ (where A = 0.16 and B = 0.013333) is provided in order for potential customers to calculate approximate uncertainties for values down to 1 mV. Example: uncertainty at 1 mVDC would calculate to approximately ±115.47 ppm.
7. The laboratory maintains Echelon II capability for ranges 20 kg to 1 mg and separate Echelon III for all ranges.
8. Avoirdupois mass calibration services are available by comparison to equivalent metric standards. Uncertainties may be appropriately larger.

2006-04-01 through 2007-03-31

Effective dates *For the National Institute of Standards and Technology*

Calibration Report

MICRO PRECISION CALIBRATION INC.
12686 Hoover Street, Garden Grove, CA 92841
Ph. (714) 901-5659 Fax (714) 901-5649

Customer: **GEOVISION**
Corona CA 92882
Account#: **15214**
Cust.PO#:
Page 1 of 2

MPC Ctrl#: AM6766
Report#: **199974**
Print Date: **041006**
MPC Job#: **L25384**

Instrument: **Caliper Calibration Plate**

Mfg: **Robertson Geo Logging**
Model: **N/A**
Size:
Res.:

Serial#: **201**
Cust Ctrl#:
Location:
Department:

Work Performed: **Inspected, cleaned, and calibrated.**
Parts Replaced: **None**

Calibration Condition as Received: **In tolerance**
Calibration Condition as Returned: **In tolerance**

Functions/Parameters Tested

Actual Values (inch)	As Measured
1.969	1.965
3.937	3.939
8.000	7.995
12.00	11.9965

Unless noted otherwise, Pass/Fail criteria is based on published manufacturer specifications and, unless noted otherwise, this instrument meets these specifications.
Services provided comply with ISO 17025:1999, ISO 9001:2000, MPC QM rev.3, MPC CSD rev.2 and customer purchase order requirements as required.

Calibration standards used for performance testing:

MPC#	Instrument	Due Date	Traceability
K3263	Pratt & Whitney C Super Micrometer	060706	192068
I2111	Mitutoyo 516-126 Gage Block Set	082406	397060

Environmental: 69 Deg / 40% Rh
Accuracy Ratio: 4:1
Cal Procedure: 33K6-4-552-1
Technician: CHRIS SPANGLER

Cal Int.: **12**
Cal Date: **040606**
Due Date: **040607**
Quality Approval: _____

QC 32

Form Cert 04-25-05

All standards used are either traceable to the National Institute of Standards and Technology or have intrinsic accuracy. All services performed have u[illegible] and are warranted for no less than (30) days. This report may not be reproduced in part without written permission of Micro Precision's Quality Assurance Manager.