



SCHNECK MEDICAL CENTER

Date: 8-27-08

Please Deliver To: NRC

Fax Number: 630-515-1078

Attention: Bill Reichhold

From:	Diagnostic Imaging Services	Phone: 812-522-0144
	Schneck Medical Center	Fax: 812-522-0466
	411 West Tipton	
	Seymour, IN 47274	

Sender: Lisa Cosby, RSO

Number of Pages Including this one: 3

Memo: Control # 317380

Please call if there is any problem with this transmission.

Confidentiality Notice

The documents accompanying this telecopy transmission contain confidential information. The information is intended only for the use of the individual(s) or entity named above. If you are not that intended recipient, you are notified that any disclosure copying, distribution or the taking of any action in reliance on the contents of this telecopy information is not permissible. If you have received this telecopy in error, please immediately notify us by telephone to arrange for return of the original documents. Thank you.

STAN A. HUBER CONSULTANTS, INC.

200 NORTH CEDAR ROAD -- NEW LENOX IL 60451-1751 -- PHONE 815-485-6161

SEALED GAMMA/BETA SOURCES - LEAK TEST CERTIFICATE

REPORT DATE: 04/01/2008

FACILITY: JACKSON COUNTY MEMORIAL HOSPITAL
SEYMOUR, IN

NRC LICENSE NUMBER: 13-05605-01

STATE LICENSE NUMBER:

SOURCE IDENTIFICATION:

RADIONUCLIDE: Co-57
 ACTIVITY: 10 mCi
 CALIBR. DATE: 06/14/2006
 MANUFACTURER: IPL
 MODEL NUMBER: BM01-10
 SERIAL NUMBER: BM01102295
 OTHER DESCRIPTION: Flood Source

ASSAY RESULTS:

COUNTING EFFICIENCY: 0.842

BACKGROUND AT TIME OF ASSAY: 54

	MAX CPM	NET CPM	DPM	ACTIVITY uCi
WET WIPE	54	0	0	< 0.000001
DRY WIPE	64	10	12	0.000005

ANALYSIS PERFORMED USING:
 Packard Auto-Gamma Scintillation Spectrometer Model #D5003 Serial #406282 AND/OR
 Packard 1900CA Tri-Carb Liquid Scintillation Analyzer Model #A1900 Serial #101464

ANALYSIS OF RESULTS:
 Sources are not leaking at this time
 Removable contamination is less than 5E-3 uCi. (.005)

NEXT LEAK TEST DUE BY: September 22, 2008

LEAK TEST PERFORMED BY: Jason A. Behling DATE: 03/26/2008

ANALYSIS PERFORMED BY: S. A. Huber Consultants, Inc. DATE: 04/01/2008

RADIATION SAFETY OFFICER SIGNATURE:

Lisa Cosby, RSO

STAN A. HUBER CONSULTANTS, INC.

200 NORTH CEDAR ROAD -- NEW LENOX IL 60451-1751 -- PHONE 815-485-6161

SEALED GAMMA/BETA SOURCES - LEAK TEST CERTIFICATE

REPORT DATE: 04/01/2008
 FACILITY: JACKSON COUNTY MEMORIAL HOSPITAL
 SEYMOUR, IN

NRC LICENSE NUMBER: 13-05605-01
 STATE LICENSE NUMBER:

SOURCE IDENTIFICATION:

RADIONUCLIDE: Cs-137
 ACTIVITY: 0.201 mCi
 CALIBR. DATE: 6-4-76
 MANUFACTURER: NEN
 MODEL NUMBER: NES-356
 SERIAL NUMBER: 2292925
 OTHER DESCRIPTION:

ASSAY RESULTS:
 COUNTING EFFICIENCY: 0.213
 BACKGROUND AT TIME OF ASSAY: 117

	MAX CPM	NET CPM	DPM	ACTIVITY uCi
WET WIPE	106	0	0	< 0.000001
DRY WIPE	116	0	0	< 0.000001

ANALYSIS PERFORMED USING:
 Packard Auto-Gamma Scintillation Spectrometer Model #D5003 Serial #406282 AND/OR
 Packard 1900CA Tri-Carb Liquid Scintillation Analyzer Model #A1900 Serial #101464

ANALYSIS OF RESULTS:
 Sources are not leaking at this time
 Removable contamination is less than 5E-3 uCi. (.005)

NEXT LEAK TEST DUE BY: September 22, 2008

LEAK TEST PERFORMED BY: Jason A. Behling DATE: 03/26/2008

ANALYSIS PERFORMED BY: S. A. Huber Consultants, Inc. DATE: 04/01/2008

RADIATION SAFETY OFFICER SIGNATURE:

Lisa Cosby, RSO