

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Hospital Dr. Cayetano Coll y Toste Nuclear Medicine Division P.O. Box 659 Arecibo, Puerto Rico 00613		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 475 Allendale Road King of Prussia, Pennsylvania 19406-1415	
REPORT Nos 2008-001			
3. DOCKET NUMBER(S) 030-35990	4. LICENSE NUMBER(S) 52-25589-01.	5. DATE(S) OF INSPECTION May 29, 2008	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.
- Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
35.13 states that a licensee shall apply for and must receive a license amendment before it changes RSO.

Contrary to the above, the licensee did not apply for and receive an amendment before changing the Radiation Safety Officer (RSO). Specifically, on September 6, 2006, Dr. Vaquer was terminated as RSO and Dr. Diaz was appointed RSO effective January 2, 2007. The NRC received an amendment request to change RSO on January 16, 2007. However, Dr. Diaz was listed as a RSO on another NRC license.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Tomar Perez		5/29/08
NRC INSPECTOR	Michelle Simmons		5/29/08