

CORRECTED COPY

CONVERSATION RECORD

TIME | DATE

8/21/08

VISIT

CONFERENCE

TELEPHONE

INCOMING

OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT

Jim Schuh, RSO

ORGANIZATION (OFFICE, DEPT. ETC.)

Mallinckrodt, Inc.

TELEPHONE NO.

314-654-7981

SUBJECT

Follow-up to letter from Mallinckrodt to NRC dated 8/18/08

SUMMARY

The inspector informed the RSO that, contrary to Item 13 of the Attachment to the letter dated 8/18/08 from the licensee to the NRC, the inspector did not "approve" the proposed label for the top of technetium-99m generators reminding users to perform a molybdenum-99 breakthrough test on each elution. Rather, the inspector informed the licensee that he had no questions or concerns about the label. In addition, the inspector informed the licensee that the NRC plans to issue a letter to the licensee acknowledging its receipt of the letter. The letter will include verbiage to clarify that the NRC did not approve the label.

The RSO apologized to the inspector for misinterpreting the inspector's statements as "approval."


ACTION REQUIRED

Place in ADAMS.

NAME OF PERSON DOCUMENTING CONVERSATION

Bob Gattone

SIGNATURE



DATE

9/4/08

ACTION TAKEN

SIGNATURE

TITLE

DATE