

Change of Heart Cardiology, LLC
2130 Route 35, Building C, Suite 321
Sea Grit, New Jersey 08570

September 3, 2008

Dennis Lawyer, License Reviewer
License Assistance Section
Nuclear Medicine Safety Branch
Division of Radiation Safety & Safeguards
U.S. Nuclear Regulatory Commission, Region 1
475 Allendale Road
King of Prussia, PA 19406-1415

J-6
29-31323-01
03037804

2008 SEP 15 PM 2:11

RECEIVED
REGION 1

RE: Supplemental Information - New License Application

Dear Mr. Lawyer:


Pursuant to your conversation with Michael Lairmore on September 3, 2008, the following information is provided in support of our pending new license application.

We confirm, that Change of Heart will take ownership and possession of all sealed sources and radioactive waste transferred from MIS. A listing of the transferred sources are provided on the attached page. Please refer to this document for details.

If you have additional questions, please contact Michael W. Lairmore or myself. Mr. Lairmore may be reached at (201) 693-2277.

We thank you in advance for your assistance with this pending licensing action.

Sincerely,



Administrative Representative

REC'D IN LAT 9/12/08

142638
NRCSS/RGNI MATER. ALS-002

**MEDICAL PHYSICS SURVEY
LEAK TEST ANALYSIS REPORT**

FACILITY: Change of Heart Cardiology DEPARTMENT: Nuclear Cardiology
LOCATION: Sea Girt, New Jersey ATTENTION: Frank DiGregorio, RSO

Analysis of the WIPE used to conduct a leak test on the sealed sources identified below was done by gas or scintillation detection and reveals removable contamination was less than 0.001 uCi, unless otherwise noted, when compared against NIST traceable standards.

Source: Cs-137 Vial; 225 uCi on 01/18/2008; RadQual; S/N BM06E37-021-31; Model #: BM06

Date:	04/06/08	06/26/08
Wipe No.:	COH1	COH4
MDA (uCi):	5.6E-5	5.6E-5
Net CPM:	26	8
Net uCi:	< MDA	< MDA
CPM for 0.001 uCi:	236	235
Analyst:	M. Lairmore	M. Lairmore

Source: Co-57 Flood Disk Source; 10.0 mCi on 04/18/2008; RadQual; S/N: BM05100861; BM05-10

Date:	04/06/2008	06/26/08
Wipe No.:	COH2	COH5
MDA (uCi):	9.90E-5	9.90E-5
Net CPM:	13	9
Net uCi:	< MDA	< MDA
CPM for 0.001 uCi:	1538	1526
Analyst:	M. Lairmore	M. Lairmore

RSO Signature

