

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: September 25, 2008

**TELEPHONE CONVERSATION RECORD**

Time: 12:45

Mail Control 2008-001 License No(s). 29-30697-01 Docket No(s). 03035878  
or Report No(s).

Name of Licensee: Mulkay Cardiology Consultants, P.C.

Name of Participant(s): Myra Quinones, Office Manager  
Steven R. Courtemanche, Health Physicist NRC RITelephone No. 201-996-9244  
610-337-5075

Subject: Exit Meeting

(NOTE: This will be used as the  
Documents Title in ADAMS)

I was informed by Ms. Quinones that Angel Mulkay, M.D. was unavailable. He was at a conference and would not be back until the beginning of next week. I stated that I could conduct an exit interview with her provided that she passed the information on to Dr. Mulkay. She stated that she would pass on the information. I informed her that, after management review, it was determined that a clear inspection would be issued. The licensee had requested, in 2001, that the range be 20 percent of the prescribed dose but guidance at that time suggested only a 10 percent range. Changes in Part 35 allowed for a 20 percent range subsequent to the issuance of the license and seem to have been adopted by a ministerial change, as evidenced by the technologist's belief that a 20 percent range was appropriate. I requested that the licensee make note of this conversation, place it on file, and to make appropriate changes on the form posted in the hot laboratory which states, in part, that the licensee shall not administer doses which deviate from the prescribed dose by more than ten percent.

Action Required: Place in ADAMS, Issue a Clear 591 from NRC office.

Document Availability:  Publicly Available  Non-Publicly Available  
 Non-Sensitive  Non-Sensitive Copyright  Sensitive  Sensitive Copyright  
 Immediate Release  Normal Release  Delay Release Date

Prepared &amp; SUNSI Review Completed By: Steven R. Courtemanche / RA / Date: 9/25/2008