

Materials Licensing Branch
United States Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, IL 60532-4352

September 22, 2008

Re: USNRC Materials License No. 13-32555-01

Dear Sir or Madam:

I am writing to request an amendment to our USNRC Materials License No. 13-32555-01.

American Health Network of Indiana, LLC, is planning to open an additional imaging site in order to perform diagnostic imaging which is located at 3601 North Morrison Road Muncie, IN 47304. Enclosed with this correspondence is a scaled floor plan which demonstrates the new areas of use (e.g., imaging, hot lab).

Also enclosed with this correspondence is an equipment list which our facility intends to obtain and place in use. The radioactive materials sources to be used at this site include:

- Approx. 200 uCi Cs-137 as a dose calibrator reference vial source for daily dose calibrator constancy tests and yearly accuracy tests;
- Approx. 0.5 uCi Cs-137 rod source for daily QC on well counter;
- Approx. 5.0 mCi Co-57 reference vial source for dose calibrator testing.
- ^{99m}Tc

EQUIPMENT LIST:

GE Millennium MG

Dose Calibrator: Capintec CRC-15R

Survey Meter: Ludlum Model 14C End Window with a range of 0-2000 mR/hr
(w/ check source)

Other: Capintec Caprac Well Counter or Ludlum 2200 Well counter

Syringe Shield(s)

Lead Glass Face Shield

2 Lead Lined RAM Waste

Lead Syringe Holders

Containers

Double Stack Lead Lined Cab.

Decontamination Kit

Sharps Containers

Lead Bricks

Long Handled Tongs

If you have any questions, please feel free to contact me at 317/755-9688.

Sincerely,



Edward E. Wroblewski, MA

Diplomate, ABSNM

Radiation Safety Officer

Enclosure

cc: USNRCfile

RECEIVED SEP 29 2008

Edward E. Wroblewski, MA
PHOTON MEASUREMENTS PLUS
1427 West 86th Street, Suite 282
Indianapolis, IN 46260

INDIANAPOLIS IN 462

24 SEP 2008 PM 3 T

"LET US DARE TO
THINK, SPEAK AND
John Adams, 1765
poweroftheletter.com



USA FIRST CLASS FOREVER

USNRC
Materials Licensing Section
Region III
2443 WARRENVILLE ROAD
Suite 210
Lisle, IL
60532-4352

60532+4352

