

MidMichigan
Medical Center
Midland



RADIATION ONCOLOGY
4005 Orchard Drive
Midland, MI 48670

Phone: (989) 839-3450
Fax: (989) 839-1347

FAX TRANSMITTAL

Date: 9-29-08
To: COLLEEN C. CASEY NRC REGION III
Fax No.: 630-515-1078
From: LARRY LANGRILL
No. of Pages: 2

****CONFIDENTIAL****

If this document is not received in its entirety, please call (989)839-3450.

The documents accompanying this transmission contain confidential information that is legally privileged. This information is intended only for the use of the individual or entity named above. The authorized recipient of this information is prohibited from disclosing this information to any other party unless required to do so by law or regulation and is required to destroy the information after its stated need has been fulfilled.

If you are not the intended recipient, you are hereby notified that any disclosure, copying, distribution or action taken in reliance on the contents of these documents, is strictly prohibited. If you have received this information in error, please notify the sender immediately and arrange for the return or destruction of these documents.

9-29-08 page 1 of 1

MidMichigan
Medical Center
Midland

September 29, 2008

Colleen C. Casey, Health Physicist
Materials Licensing
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road, Suite 210
Lisle, Illinois 60532-4352

Re: Control No. 317361
Additional Information


Dear Ms. Casey,

This letter supplies additional information and request to the pending amendment number 64, for NRC license number 21-01549-02, MidMichigan Medical Center, Midland, Michigan.

1. Please discontinue authorization for use of part 35.300 (as specified under license condition 10.E.) at our 703 n. McEwan Street, Clare, Michigan address. Please, however, Retain use for parts 35.100 and 35.200 at this location. The last patient done with 35.300 materials was in the second quarter of year 2002. Paitents were administered dose only via dry capsule so that no waste was generated which required storage, decay or disposal. The closeout survey of the old hot lab, submitted with the letter of May 8, 2008, showed no remaining radioactivity in the area of use of the 35.300 material.
2. Also, under this same condition 10.E., please note and correct the typographical error in the phrase "Subitems 6.A., 6.8.,". The "6.8." should be "6.B."

For further information or questions, please contact me directly at 517-839-3450 or at fax number 989-839-1347.

Sincerely,



Larry Langrill, RSO
Perfexion AMP
MidMichigan Medical Center
4005 Orchard Dr.
Midland, MI 48670