

U.S. NUCLEAR REGULATORY COMMISSION

10/08/08

Date:

TELEPHONE CONVERSATION RECORD

Time:

10:15 AM

Mail Control 142790
or Report No(s).

License No(s). 29-00018-02

Docket No(s). 03005210

Name of Licensee: Hoffmann-La Roche, Inc.

Name of Participant(s): Michael Drzyzga, Manager/RSO
Dennis Lawyer, R1, DNMS

Telephone No. 973-235-3418

Subject: RAI, License Quantities

(NOTE: This will be used as the
Documents Title in ADAMS)

Need to establish the quantities of material for the sealed sources on your license. During the discussion, they believe they will remove the Ni-63 as they don't use a gas chromatograph any longer. Need a Certification of Financial Assurance to complete the changes in the material amount.

Action Required: The RSO will send in: 1) the amount of material you would like to be licensed for both Cs-137 line items in the maximum per source and maximum on the license for each type source.

2) possibly requesting the removal of the Nickel 63 material. If so, include a statement that they have all been removed and transferred to XXX, include leak tests and state for previous devices removed that they were no leaking sources. (If not the item #1 applies)

3) a new Certification of Financial Assurance for the changes in the license as described in #1 and #2 above.

Document Availability: Publicly Available Non-Publicly Available

Non-Sensitive Non-Sensitive Copyright Sensitive Sensitive Copyright

Immediate Release Normal Release Delay Release Date

Prepared & SUNSI Review Completed By: Dennis Lawyer / RA / Date: 10/08/2008