

U.S. NUCLEAR REGULATORY COMMISSION

Date: 10/22/08

TELEPHONE CONVERSATION RECORD

Time: 2:35 pm

Mail Control 142784 License No(s) 01-31331-01 Docket No(s) 030-37827
or Report No(s).

Name of Licensee: Center for Domestic Preparedness
(Department of Homeland Security)

Name of Participant(s): Betsy Ullrich, NRC
Michael Vice, CDP

Telephone No. 256-847-2266

Subject: RAI dated 10/1/08

(NOTE: This will be used as the
Documents Title in ADAMS)

Summary: Mr. Vice had questions regarding some of the issues raised in the RAI. We discussed locations of use and temporary job sites; training for persons using material at temporary job sites; training for the RSO and Aus; instrumentation; and the CDP/ODP/FEMA/DHS organization.

Mr. Vice said that only the Ni63 detectors would be used for training at temporary job sites. His TJS procedures and training will take that into consideration. He will ask for flexibility in use of both the Ni63 and the Cs sources for training at their campus (the former Fort McClellan) so that specific buildings do not need to be named.

We also discussed the request to have repairs done by the manufacturer "under their license, at their site". Because of some of the other hazardous materials used during training, the Ni63 instruments become damaged but cannot be removed from the site due to contaminants; an trained repair technician, authorized by the manufacturer, performs repairs at the CPD facility. Most of the repairs (to the best of his knowledge, probably all) involve the electronics and not the licensed material. He will explain this in the response.

I confirmed with him that they do want the license issued to the CDP as the responsible licensee; although they are currently under DHS, this has not always been so. In case this may change, they prefer to have the license issued to CDP so that they may continue activities even if they are re-assigned to another department.

Michael Vice asked if a request for an extension was required to be in writing – I told him no.

Action Required: Telecom to LAT for ADAMS package

Document Availability: Publicly Available Non-Publicly Available

Non-Sensitive Non-Sensitive Copyright Sensitive Sensitive Copyright

Immediate Release Normal Release Delay Release Date

Prepared & SUNSI Review Completed By:

Betsy Ullrich

/ RA /

Date: 10/22/08