

Rulemaking Comments

From: Shanahan, Thomas [Shanahan.Thomas@mhsil.com]
Sent: Monday, October 27, 2008 11:55 AM
To: Rulemaking Comments
Subject: Comments on Proposed Rule for Medical Use of Byproduct Material-Amendments/Medical Event Definitions (RIN 3150-AI26, NRC-2008-0071) [See 73 FR 45635 (August 6, 2008)]
Attachments: NRC letter.doc

DOCKETED
USNRC

October 29, 2008 (12:45pm)

Annette L. Vietti-Cook
Secretary of the Commission
U.S. Nuclear Regulatory Commission
Washington, DC 20555-0001

OFFICE OF SECRETARY
RULEMAKINGS AND
ADJUDICATIONS STAFF

ATTN: Rulemakings and Adjudications Staff

Re: Comments on Proposed Rule for Medical Use of Byproduct Material—Amendments/Medical Event Definitions (RIN 3150-AI26, NRC-2008-0071) [See 73 FR 45635 (August 6, 2008)]

Dear Ms. Vietti-Cook:

My name is Thomas Shanahan MD. I have specializing in radiation therapy since 1990. I am currently practicing in Springfield, Illinois specializing in prostate brachytherapy. I have performed over 1500 prostate seed implants and currently perform over 130 implants per year. I use the Hybrid Interactive Mick applicator Technique and have published several articles on this technique. It combines intraoperative seed placement with accurate preimplant ordering of seeds.

The technique of prostate brachytherapy that I use may occasionally allows a single seed to “migrate” to the bladder or through the blood vessels along side of the prostate. I am concerned that the U.S. Nuclear Regulatory Commission’s (NRC’s) proposed modifications to 10 CFR 35.40 and 35.3045 to establish separate medical event criteria and written directive requirements for permanent implant brachytherapy would result in inappropriately categorizing some medically acceptable implants as “medical events” (ME’s).

1. TIMING OF WRITTEN DIRECTIVE AND MEDICAL EVENTS

The proposed rule language for § 35.40(b)(6) and § 35.3045(a)(2) does not take into account clinical practice realities. Many authorized users (AUs) perform real-time, adaptive, interactive planning, whereby the written directive and the source strength to be implanted are based on the actual volume dynamically determined during the procedure rather than based on the pre-implant volume.

Real-time planning is a more accurate method of implantation. It allows the physician to take into account any alterations in the organ volume and shape that occur between the time of the pre-plan and the implant procedure and therefore represents the actual organ volume and implant situation. For those performing real-time adaptive planning implantation, the total source strength to be implanted is determined intraoperatively during the implantation procedure and not pre-implant. Further, even those performing permanent brachytherapy using preplanned techniques will often modify their plan if intraoperatively they find major discrepancies in the gland or organ volume from the volumes determined during the preplan.

Template = SECY-067

I support ASTRO's suggested revisions to the proposed regulations. I believe this modification will clarify that the source strength implanted as stated in the WD refers to the source strength implanted after administration but before the patient leaves the post-treatment recovery area.

2. DEFINITION OF TREATMENT SITE

The definition of "treatment site" described in § 35.2 as "the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive" leads to some ambiguity regarding the exact volume that "treatment site" refers to in § 35.3045(a)(2)(ii). There are various standard volumes already defined in radiation oncology, including the gross tumor volume, which is the volume that contains tumor. Two other margins are added to the gross tumor volume during the brachytherapy planning process. One margin is added to account for the subclinical spread of tumor, which is termed the "clinical target volume," and a second margin is added to account for uncertainties in source positioning, tumor boundaries, isodose constrictions, etc., which is termed the "planning target volume."

These expansion margins are not constant but change for different clinical situations. Radiation oncologists use a larger margin if there is high degree of uncertainty and/or if there are no adjacent critical structures. Conversely, the margins are smaller if the boundary is distinct and/or if there are adjacent critical structures.

I believe that the proposed regulations cross into clinical decision-making by specifying margin parameters and the source strength to be placed in the margin. The NRC will be interfering into medical judgment if it dictates the amount of source strength the authorized user can place in the margins. Using the definition found at § 35.2 of "treatment site" as "the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive" raises ambiguities in terms of the proposed medical event reports and notifications as it is unclear whether the "treatment site" refers to the gross tumor volume or includes the margins in the clinical target volume or those in the planning target volume.

I support ASTRO's recommended changes to the definition of "treatment site" at § 35.2 be revised to reflect the distinct clinical areas - gross tumor, the clinical target volume, plus a variable planning target volume. Further, by following ASTRO's suggested alternative language, section § 35.3045 (a)(2)(iii) of the proposed rule would become superfluous and therefore could be eliminated.

I believe that these suggested modifications to the proposed rule language are necessary because in the normal course of some medically acceptable brachytherapy implant procedures, a few seeds may come to rest beyond 3 cm (1.2 in) from the outside boundary of the treatment site. This represents less than 1% of the radioactivity of the entire procedure.

Please feel free to contact me directly at:

Dr Thomas G Shanahan
Dept of Radiation Oncology
701 N First Street
Springfield, Illinois 62781-0001
217-788-3260
Shanahan.thomas@mhsil.com

Thank you for giving me this opportunity to provide comments on the NRC's proposed rule changes to 10 CFR 35.40 and 35.3045 related to medical events in permanent implant brachytherapy.

Sincerely,

Thomas G Shanahan

This message including any attachments contains confidential information intended for a specific individual and is prepared with the expectation that the intended recipient will not disclose its contents. Any disclosure, copying, or distribution of this message in the absence of the intended recipient is strictly prohibited.

VIA E-Mail to: rulemaking.comments@nrc.gov

Annette L. Vietti-Cook
Secretary of the Commission
U.S. Nuclear Regulatory Commission
Washington, DC 20555-0001

ATTN: Rulemakings and Adjudications Staff

Re: Comments on Proposed Rule for Medical Use of Byproduct Material—
Amendments/Medical Event Definitions (RIN 3150-A126, NRC-2008-0071) [*See* 73 FR
45635 (August 6, 2008)]

Dear Ms. Vietti-Cook:

My name is Thomas Shanahan MD. I have specializing in radiation therapy since 1990. I am currently practicing in Springfield, Illinois specializing in prostate brachytherapy. I have performed over 1500 prostate seed implants and currently perform over 130 implants per year. I use the Hybrid Interactive Mick applicator Technique and have published several articles on this technique. It combines intraoperative seed placement with accurate preimplant ordering of seeds.

The technique of prostate brachytherapy that I use may occasionally allows a single seed to “migrate” to the bladder or through the blood vessels along side of the prostate. I am concerned that the U.S. Nuclear Regulatory Commission’s (NRC’s) proposed modifications to 10 CFR 35.40 and 35.3045 to establish separate medical event criteria and written directive requirements for permanent implant brachytherapy would result in inappropriately categorizing some medically acceptable implants as “medical events” (ME’s).

1. TIMING OF WRITTEN DIRECTIVE AND MEDICAL EVENTS

The proposed rule language for § 35.40(b)(6) and § 35.3045(a)(2) does not take into account clinical practice realities. Many authorized users (AUs) perform real-time, adaptive, interactive planning, whereby the written directive and the source strength to be implanted are based on the actual volume dynamically determined during the procedure rather than based on the pre-implant volume.

Real-time planning is a more accurate method of implantation. It allows the physician to take into account any alterations in the organ volume and shape that occur between the time of the pre-plan and the implant procedure and therefore represents the actual organ volume and implant situation. For those performing real-time adaptive planning implantation, the total source strength to be implanted is determined intraoperatively during the implantation procedure and not pre-implant. Further, even those performing permanent brachytherapy using preplanned techniques will often modify their plan if

intraoperatively they find major discrepancies in the gland or organ volume from the volumes determined during the preplan.

I support ASTRO's suggested revisions to the proposed regulations. I believe this modification will clarify that the source strength implanted as stated in the WD refers to the source strength implanted after administration but before the patient leaves the post-treatment recovery area.

2. DEFINITION OF TREATMENT SITE

The definition of "treatment site" described in § 35.2 as "the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive" leads to some ambiguity regarding the exact volume that "treatment site" refers to in § 35.3045(a)(2)(ii). There are various standard volumes already defined in radiation oncology, including the gross tumor volume, which is the volume that contains tumor. Two other margins are added to the gross tumor volume during the brachytherapy planning process. One margin is added to account for the subclinical spread of tumor, which is termed the "clinical target volume," and a second margin is added to account for uncertainties in source positioning, tumor boundaries, isodose constrictions, etc., which is termed the "planning target volume."

These expansion margins are not constant but change for different clinical situations. Radiation oncologists use a larger margin if there is high degree of uncertainty and/or if there are no adjacent critical structures. Conversely, the margins are smaller if the boundary is distinct and/or if there are adjacent critical structures.

I believe that the proposed regulations cross into clinical decision-making by specifying margin parameters and the source strength to be placed in the margin. The NRC will be interfering into medical judgment if it dictates the amount of source strength the authorized user can place in the margins. Using the definition found at § 35.2 of "treatment site" as "the anatomical description of the tissue intended to receive a radiation dose, as described in a written directive" raises ambiguities in terms of the proposed medical event reports and notifications as it is unclear whether the "treatment site" refers to the gross tumor volume or includes the margins in the clinical target volume or those in the planning target volume.

I support ASTRO's recommended changes to the definition of "treatment site" at § 35.2 be revised to reflect the distinct clinical areas - gross tumor, the clinical target volume, plus a variable planning target volume. Further, by following ASTRO's suggested alternative language, section § 35.3045 (a)(2)(iii) of the proposed rule would become superfluous and therefore could be eliminated.

I believe that these suggested modifications to the proposed rule language are necessary because in the normal course of some medically acceptable brachytherapy implant procedures, a few seeds may come to rest beyond 3 cm (1.2 in) from the outside

boundary of the treatment site. This represents less than 1% of the radioactivity of the entire procedure.

Please feel free to contact me directly at:

Dr Thomas G Shanahan
Dept of Radiation Oncology
701 N First Street
Springfield, Illinois 62781-0001
217-788-3260
Shanahan.thomas@mhsil.com

Thank you for giving me this opportunity to provide comments on the NRC's proposed rule changes to 10 CFR 35.40 and 35.3045 related to medical events in permanent implant brachytherapy.

Sincerely,

Thomas G Shanahan

Received: from mail1.nrc.gov (148.184.176.41) by OWMS01.nrc.gov
(148.184.100.43) with Microsoft SMTP Server id 8.1.291.1; Mon, 27 Oct 2008
16:02:28 -0400

X-Ironport-ID: mail1

X-SBRS: 2.9

X-MID: 32051398

X-IronPort-Anti-Spam-Filtered: true

X-IronPort-Anti-Spam-Result:

ApkDAAW6BUmqG8sQgWdsb2JhbACCTC2RCgEBFiKpAAcBCIkugk0HAXo

X-IronPort-AV: E=Sophos;i="4.33,493,1220241600";

d="doc'32?scan'32,208,217,32";a="32051398"

Received: from e500-01.mhsil.com ([170.27.203.16]) by mail1.nrc.gov with
SMTP; 27 Oct 2008 16:02:22 -0400

Received: from (unknown [10.8.32.42]) by e500-01.mhsil.com with smtp id
27de_3508_934efe82_a437_11dd_a16c_001422223b7a; Mon, 27 Oct 2008 09:57:29
-0500

Received: from mmc-mail.ad.mhsil.com ([10.8.32.41]) by mmc-mail2.ad.mhsil.com
([10.8.32.42]) with mapi; Mon, 27 Oct 2008 10:55:06 -0500

From: "Shanahan, Thomas" <Shanahan.Thomas@mhsil.com>

To: "'rulemaking.comments@nrc.gov'" <rulemaking.comments@nrc.gov>

Date: Mon, 27 Oct 2008 10:55:05 -0500

Subject: Comments on Proposed Rule for Medical Use of Byproduct

Material-Amendments/Medical Event Definitions (RIN 3150-AI26, NRC-2008-0071)

[See 73 FR 45635 (August 6, 2008)]

Thread-Topic: Comments on Proposed Rule for Medical Use of Byproduct

Material-Amendments/Medical Event Definitions (RIN 3150-AI26, NRC-2008-0071)

[See 73 FR 45635 (August 6, 2008)]

Thread-Index: Ack4TGA6copR4QAuSYWUWSktdF5u8Q==

Message-ID: <24A4826E8EF0964D86BC5317306F58A52B9C0A0D1E@mmc-
mail.ad.mhsil.com>

Accept-Language: en-US

Content-Language: en-US

X-MS-Has-Attach: yes

X-MS-TNEF-Correlator:

acceptlanguage: en-US

Content-Type: multipart/mixed;

boundary="_004_24A4826E8EF0964D86BC5317306F58A52B9C0A0D1Emmcmailadm

hsi_ "

MIME-Version: 1.0

Return-Path: Shanahan.Thomas@mhsil.com