October 23, 2008

Licensing Assistant Section

Re: Terminate license and close out reports of Nuclear Medicine room

To Whom It May Concern:

CORRECT This letter is to inform the N.R.C. that on November 12, 2008 the Nuclear Medicine Department will close. We request that you terminate N.R.C license number 29-30895-01. A disposition of our CS 137 vial for dose calibrator and forms for close out surveys of Nuclear Medicine room.

Bra

Respectfully yours,

Dr. Ralph Dauito R.S.O

Evelyn O'Donnell RTR

2008 NOV 13 PM 2: 32

F. 01.S-CC2 NPSS/RGNI

October 23, 2008

÷

William Csaszar, Supervisor Radioactive Materials Section

Re: Terminate license and close out reports of nuclear room

Dear Mr. Csaszar,

This letter is to inform the state of N.J. that on November 12, 2008, the Nuclear Medicine Department will close. We request that you terminate N.J. State license NJSL 20804/01/000. A disposition of the Co57 sheet source and form for the close out surveys of Nuclear Medicine room.

Respectfully yours,

Dr. Dauit

Evelyn O'Donnell RTNR

1

### CERTIFICATE OF DISPOSITION OF MATERIALS

### PLEASE READ THESE INSTRUCTIONS BEFORE COMPLETING NRC FORM 314.

Subpart E of 10 CFR Part 20 establishes the radiological criteria for license terminations/decommissioning of facilities licensed under 10 CFR Parts 30, 40, 50, 60, 61, 70, and 72, as well as other facilities subject to the Commission's jurisdiction under the Atomic Energy Act of 1954, as amended, and the Energy Reorganization Act of 1974, as amended.

### INSTRUCTIONS

### Section B. Item 2.

Licensees should describe the specific radioactive material transfer actions. If radioactive wastes were generated in terminating this license, the licensee should describe the disposal actions taken, including the disposition of low-level radioactive waste, mixed waste, greater-than-Class-C waste, and sealed sources.

### Section B, Item 2.a.

The information provided concerning the transfer of radioactive material to another licensee should specify the date of the transfer, the name of the licensee recipient, an individual contact name and telephone number for the licensee recipient, and the recipient's NRC or Agreement State license number.

### Section B, Item 2.b.

For disposal of radioactive materials, licensees should describe the specific disposal method or procedure (e.g., decay-in-storage). For those cases when radioactive materials are disposed of by a licensed disposal site or by a waste contractor, the licensee should specify the name, address, and telephone number of the licensed disposal site operator or waste contractor.

<u>Section B, Item 2.c.</u> "Residual radioactivity," as defined in 10 CFR 20.1003, means radioactivity in 'areas' (structures, materials, soils, etc.) remaining as a result of activities (licensed and unlicensed) under the licensee's control from sources used by the licensee, excluding background radiation. ALARA is defined in 10 CFR 20,1003.

### FILE CERTIFICATES AS FOLLOWS:

#### IF YOU ARE LOCATED IN:

IF YOU ARE LOCATED IN:

CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, PENNSYLVANIA, RHODE ISLAND, OR VERMONT, SEND CERTIFICATES TO:

> LICENSING ASSISTANT SECTION NUCLEAR MATERIALS SAFETY BRANCH U.S. NUCLEAR REGULATORY COMMISSION, REGION I 475 ALLENDALE ROAD KING OF PRUSSIA, PA 19406-1415

ALABAMA, FLORIDA, GEORGIA, KENTUCKY, MISSISSIPPI, NORTH CAROLINA, PUERTO RICO, SOUTH CAROLINA, TENNESSEE, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND CERTIFICATES TO:

NUCLEAR MATERIALS SAFETY SECTION U.S. NUCLEAR REGULATORY COMMISSION, REGION II SAM NUNN ATLANTA FEDERAL CENTER 61 FORSYTH STREET, S.W., SUITE 23785 ATLANTA, GEORGIA 30303-8931

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND CERTIFICATES TO:

MATERIALS LICENSING SECTION U.S. NUCLEAR REGULATORY COMMISSION, REGION III 801 WARRENVILLE RD. LISLE, IL 60532-4351

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND CERTIFICATES TO:

MATERIAL RADIATION PROTECTION SECTION U. S. NUCLEAR REGULATORY COMMISSION, REGION IV 611 RYAN PLAZA DRIVE, SUITE 400 ARLINGTON, TX 76011-8064

RADIOACTIVE SEA	LED SOUR	RCE LEAK TEST REPORT
INSTITUTION: EASY /	antie	DiAgnostic Frist.
NRC LICENSE #:29-308	<u>95-0</u> / sta	TE LICENSE #: MJ52- 20804
sources tested: <u>Co</u>	57_	Bm01103681 Sheet Source
C/.	37 (2	8792) VIAL
DATE(S) OF TEST: // - /.	2-08	-
TYPE OF TEST: ALCOHOL S	WAB V	
DETECTION EQUIPMENT:	CAPR	AC CAPINTEC
APPROXIMATE EFFICIENCY	: <u>Cs 13</u>	7-2.87 × 105
RESULTS:	CO 57	7-1.22×100
Source	<u>Net CPM</u>	Removable Activity
Co 57	2	<u> </u>
<u>(s137</u>	0	20,001
Manada and a state of the state		

The above test satisfies the leak test requirements as specified under the conditions of the NRC and State Licenses.

.

Radiation Safety Officer

:



RadQual, LLC 114 Barrington Town Square #124 Aurora, Ohio 44202 (603) 513-1221• (603) 415-0160 Fax www.radqual.com

# Technical Data



### Test and Measurement Report

Model #:	BM01- 10	Serial #:	BM01103681	Reference Da	te: August 10, 2008
Radioactive	Isotope Information	at Reference Da	ate		
Primary N	uclide:	Co-57	10 mCi	370 MBq	
Principle (	Contaminants:	Co-56 Co-58	0.002 mCi 0.001 mCi	0.0555 MBq 0.037 MBq	0.02% 0.01%

### Leak Test Results

By:

The subject source was tested for radioactivity leakage and surface contamination as described by ANSI Standard N43.6-1997. Any leakage and/or contamination detected did not exceed  $5 \times 10^{-3}$  microcuries.,

Date: 7.14.08

### **Physical Properties** (for reference only)

Total Dimensions:	17.9" (45.5 cm) x 25.4" (64.5 cm)
Active Dimensions:	16.4" (41.7 cm) x 23.9" (60.7 cm)

### Physical Properties as Tested for Uniformity (for reference only)

Total Field of View:	15.2!! (39.6 arm) = 22.0!! (60.7 arm)
Useful Field of View	15.2 $(38.0 \text{ cm}) \times 23.9 (00.7 \text{ cm})$
esciul field of view.	13.7" (34.8 cm) x 21.5" (54.6 cm)

### Source Decay Characteristics<sup>1</sup>

Cobalt-57 half-life is 271.7 days

Photon Energies (keV)		Abundance (%)
$\gamma_1$	14.4	9.5
$\gamma_2$	122	85.6
<b>y</b> 3	136	10.6
y9	692	0.16

(1) "A Handbook of Radioactivity Measurement Procedures", NRCP Report No. 58, Second Edition, (February 1985).

Most gamma cameras tend to resolve the  $\gamma_2$  and  $\gamma_3$  as a single non-Gaussian peak for window alignment during quality control procedures. The medium and high-energy gamma emissions from the trace Co-56 and Co-58 (0.5 MeV — 3.0 MeV) impurities decrease relative to the Co-57 with an effective half-life of 105 days.

## RadQual

### **Decay Reference Chart for Cobalt-57 Flood Source**

### Source Serial No. BM01103681

Source Reference Date

10-Aug-08

Labeled Activity

mCi MBq

10

370

Date	mCi*	MBq*	
10-Aug-08	10.00	370.0	1
17-Aug-08	9.82	363.5	2
24-Aug-08	9.65	357.0	2
31-Aug-08	9.48	350.7	
7-Sep-08	9.31	344.5	1
14-Sep-08	9.15	338.4	1
21-Sep-08	8.98	332.4	2
28-Sep-08	8.83	326.5	3
5-Oct-08	8.67	320.8	1(
12-Oct-08	8.52	315.1	1
19-Oct-08	8.36	309.5	24
26-Oct-08	8.22	304.0	3
2-Nov-08	8.07	298.6	7
9-Nov-08	7.93	293.4	1.
16-Nov-08	7.79	288.2	2
23-Nov-08	7.65	283.1	2
30-Nov-08	7.52	278.1	
7-Dec-08	7.38	273.1	1
14-Dec-08	7.25	268.3	1
21-Dec-08	7.12	263.6	2
28-Dec-08	7.00	258.9	2
4-Jan-09	6.87	254.3	9
11-Jan-09	6.75	249.8	16
18-Jan-09	6.63	245.4	23
25-Jan-09	6.51	241.0	30
1-Feb-09	6.40	236.8	6
8-Feb-09	6.29	232.6	1:
15-Feb-09	6.18	228.5	20
22-Feb-09	6.07	224.4	2
1-Mar-09	5.96	220.5	4
8-Mar-09	5.85	216.6	1

Date	mCi*	MBq*	
15-Mar-09	5.75	212.7	
22-Mar-09	5.65	209.0	
29-Mar-09	5.55	205.3	1847 2014
5-Apr-09	5.45	201.6	
12-Apr-09	5.35	198.1	
19-Apr-09	5.26	194.6	
26-Apr-09	5.17	191.1	
3-May-09	5.07	187.7	
10-May-09	4.98	184.4	
17-May-09	4.90	181.2	
24-May-09	4.81	177.9	
31-May-09	4.72	174.8	
7-Jun-09	4.64	171.7	
14-Jun-09	4.56	168.7	
21-Jun-09	4.48	165.7	
28-Jun-09	4.40	162.7	
5-Jul-09	4.32	159.9	
12-Jul-09	4.24	157.0	
19-Jul-09	4.17	154.3	
26-Jul-09	4.10	151.5	
2-Aug-09	4.02	148.8	
9-Aug-09	3.95	146.2	
16-Aug-09	3.88	143.6	
23-Aug-09	3.81	141.1	
30-Aug-09	3.75	138.6	
6-Sep-09	3.68	136.1	
13-Sep-09	3.61	133.7	
20-Sep-09	3.55	131.4	
27-Sep-09	3.49	129.0	
4-Oct-09	3.43	126.8	
11-Oct-09	3.37	124.5	

Date	mCi*	MBq*
18-Oct-09	3.31	122.3
25-Oct-09	3.25	120.1
22-Nov-09	3.02	111.9
29-Nov-09	2.97	109.9
6-Dec-09	2.92	107.9
13-Dec-09	2.87	106.0
20-Dec-09	2.81	104.2
27-Dec-09	2.77	102.3
3-Jan-10	2.72	100.5
10-Jan-10	2.67	98.7
17-Jan-10	2.62	97.0
24-Jan-10	2.57	95.3
31-Jan-10	2.53	93.6
7-Feb-10	2.48	91.9
14-Feb-10	2.44	90.3
21-Feb-10	2.40	88.7
28-Feb-10	2.35	87.1
7-Mar-10	2.31	85.6
14-Mar-10	2.27	84.1
21-Mar-10	2.23	82.6
28-Mar-10	2.19	81.1
4-Apr-10	2.15	79.7
11-Apr-10	2.12	78.3
18-Apr-10	2.08	76.9
25-Apr-10	2.04	75.5
2-May-10	2.01	74.2
9-May-10	1.97	72.9
16-May-10	1.93	71.6

\* mCi and MBq values are reference values only, this is not a calibrated source

## RadQual

RadQual, LLC 114 Barrington Town Square #124 Aurora, Ohio 44202 (603) 513-1221• (603) 415-0160 Fax www.radqual.com

# Technical Data

### Radiation Safety Instructions and Recommendations For Use and Storage of Cobalt 57 (Co-57) Flood Sources

The Co-57 flood source has been manufactured to provide maximum safety and longevity of service. The flood source has been designed to satisfy the safety performance requirements of ANSI Standard N43.6-1997 for classification C22312, as recommended for calibration sources.

### **Recommended Use and Service**

The Co-57 flood source provides a uniform radiation field for performing quality control testing of a gamma camera flood field. A flood field study provides a record of camera field uniformity performance and subsequent need for correction of a less than optimal performing camera prior to use in a diagnostic procedure.

When performing a study using a Co-57 flood source, you must be sure that the energy window for the instrument includes the 122 keV gamma ray (100 keV - 150 keV is recommended). The use of a narrow Technetium 99 (Tc-99) window (140 keV) with Co-57 may result in your flood field image appearing faulty. Refer to your instrument manual to ascertain the correct number of counts to acquire for your flood study with the source directly on the collimator.

The Co-57 flood source should be used and stored at 10-40° C, ambient air pressure, and 10% to 80% relative humidity. Care should be taken to avoid contact with organic solvents, hot surfaces, and excessive mechanical stress.

### **Radiation Safety Recommendation**

Generally, radiation protection procedures used for the measurement and handling of radio-pharmaceutical products is applicable to the handling of a Co-57 flood source. The following chart is provided to aid in effective radiation exposure control, actual readings may vary depending upon measurement instrumentation and technique. Consult with your Radiation Protection Office for any additional handling requirements.

Model Number	Surface Dose Rate Unshielded	(1) Surface Dose Rate Shielded
BM01-10, BM01-15, BM01-20, BM01-99	Approx. 2.9 mR/hr/mCi	Approx. 0.10 mR/hr/mCi
BM02-05, BM02-10, BM02-99	Approx. 3.8 mR/hr/mCi	Approx. 0.15mR/hr/mCi
BM04-10, BM04-15, BM04-99	Approx 3.6 mR/hr/mCi	Approx. 0.15 mR/hr/mCi
BM05-10, BM05-99	Approx. 9.0 mR/hr/mCi	Approx. 0.10 mR/hr/mCi
BM07-10, BM07-75, BM07-99	Approx. 7.0 mR/hr/mCi	Approx. 0.10 mR/hr/mCi

(1) Shielded with Rad Shield



RadQual, LLC 114 Barrington Town Square #124 Aurora, Ohio 44202 (603) 513-1221• (603) 415-0160 Fax www.radqual.com

# Technical Data

Bv:

### **Test and Measurement Report**

Model #:	BM01- 10	Serial #:	BM01103681	Reference Da	te: August 10, 2008
Radioactive	Isotope Informati	on at Reference Da	ate		
Primary N	uclide:	Co-57	10 mCi	370 MBq	
Principle (	Contaminants:	Co-56 Co-58	0.002 mCi 0.001 mCi	0.0555 MBq 0.037 MBq	0.02% 0.01%

### Leak Test Results

The subject source was tested for radioactivity leakage and surface contamination as described by ANSI Standard N43.6-1997. Any leakage and/or contamination detected did not exceed  $5 \times 10^{-3}$  microcuries.

Date: 7.14.08

### **Physical Properties** (for reference only)

Total Dimensions:	17.9" (45.5 cm) x 25.4" (64.5 cm)
Active Dimensions:	16.4" (41.7 cm) x 23.9" (60.7 cm)

### Physical Properties as Tested for Uniformity (for reference only)

Total Field of View:	15.0!! (39.6 err.) = 22.0!! (60.7 err.)
Useful Field of View	15.2 $(58.0 \text{ cm}) \ge 25.9 (60.7 \text{ cm})$
esciul i leiu or view.	13.7" (34.8 cm) x 21.5" (54.6 cm)

### **Source Decay Characteristics**<sup>1</sup>

Cobalt-57 half-life is 271.7 days

Photon Energies (keV)		Abundance (%)
$\gamma_1$	14.4	9.5
$\gamma_2$	122	85.6
	136	10.6
<b>y</b> 9	692	0.16

(1) "A Handbook of Radioactivity Measurement Procedures", NRCP Report No. 58, Second Edition, (February 1985).

Most gamma cameras tend to resolve the  $\gamma_2$  and  $\gamma_3$  as a single non-Gaussian peak for window alignment during quality control procedures. The medium and high-energy gamma emissions from the trace Co-56 and Co-58 (0.5 MeV — 3.0 MeV) impurities decrease relative to the Co-57 with an effective half-life of 105 days.

## RadQual

### **Decay Reference Chart for Cobalt-57 Flood Source**

### Source Serial No. BM01103681

Source Reference Date

10-Aug-08

Labeled Activity

MBq

mCi

10

370

Date	mCi*	MBq*
10-Aug-08	10.00	370.0
17-Aug-08	9.82	363.5
24-Aug-08	9.65	357.0
31-Aug-08	9.48	350.7
7-Sep-08	9.31	344.5
14-Sep-08	9.15	338.4
21-Sep-08	8.98	332.4
28-Sep-08	8.83	326.5
5-Oct-08	8.67	320.8
12-Oct-08	8.52	315.1
19-Oct-08	8.36	309.5
26-Oct-08	8.22	304.0
2-Nov-08	8.07	298.6
9-Nov-08	7.93	293.4
16-Nov-08	7.79	288.2
23-Nov-08	7.65	283.1
30-Nov-08	7.52	278.1
7-Dec-08	7.38	273.1
14-Dec-08	7.25	268.3
21-Dec-08	7.12	263.6
28-Dec-08	7.00	258.9
4-Jan-09	6.87	254.3
11-Jan-09	6.75	249.8
18-Jan-09	6.63	245.4
25-Jan-09	6.51	241.0
1-Feb-09	6.40	236.8
8-Feb-09	6.29	232.6
15-Feb-09	6.18	228.5
22-Feb-09	6.07	224.4
1-Mar-09	5.96	220.5
8-Mar-09	5.85	216.6

Date	mCi*	MBq*
15-Mar-09	5.75	212.7
22-Mar-09	5.65	209.0
29-Mar-09	5.55	205.3
5-Apr-09	5.45	201.6
12-Apr-09	5.35	198.1
19-Apr-09	5.26	194.6
26-Apr-09	5.17	191.1
3-May-09	5.07	187.7
10-May-09	4.98	184.4
17-May-09	4.90	181.2
24-May-09	4.81	177.9
31-May-09	4.72	174.8
7-Jun-09	4.64	171.7
14-Jun-09	4.56	168.7
21-Jun-09	4.48	165.7
28-Jun-09	4.40	162.7
5-Jul-09	4.32	159.9
12-Jul-09	4.24	157.0
19-Jul-09	4.17	154.3
26-Jul-09	4.10	151.5
2-Aug-09	4.02	148.8
9-Aug-09	3.95	146.2
16-Aug-09	3.88	143.6
23-Aug-09	3.81	141.1
30-Aug-09	3.75	138.6
6-Sep-09	3.68	136.1
13-Sep-09	3.61	133.7
20-Sep-09	3.55	131.4
27-Sep-09	3.49	129.0
4-Oct-09	3.43	126.8
11-Oct-09	3.37	124.5

Date	mCi*	MBq*
18-Oct-09	3.31	122.3
25-Oct-09	3.25	120.1
22-Nov-09	3.02	111.9
29-Nov-09	2.97	109.9
6-Dec-09	2.92	107.9
13-Dec-09	2.87	106.0
20-Dec-09	2.81	104.2
27-Dec-09	2.77	102.3
3-Jan-10	2.72	100.5
10-Jan-10	2.67	98.7
17-Jan-10	2.62	97.0
24-Jan-10	2.57	95.3
31-Jan-10	2.53	93.6
7-Feb-10	2.48	91.9
14-Feb-10	2.44	90.3
21-Feb-10	2.40	88.7
28-Feb-10	2.35	87.1
7-Mar-10	2.31	85.6
14-Mar-10	2.27	84.1
21-Mar-10	2.23	82.6
28-Mar-10	2.19	81.1
4-Apr-10	2.15	79.7
11-Apr-10	2.12	78.3
18-Apr-10	2.08	76.9
25-Apr-10	2.04	75.5
2-May-10	2.01	74.2
9-May-10	1.97	72.9
16-May-10	1.93	71.6

\* mCi and MBq values are reference values only, this is not a calibrated source

### **CERTIFICATE OF CALIBRATION**

### MODEL MED3400 GAMMA REFERENCE STANDARD

Radionuclide:	Cs-137	Activity:	37.66 kBq (1.018 μCi)
Serial Number:	17417	Reference Date	: 1200 PST April 1, 2002
Half Life <sup>(1)</sup> :	30.0 ± 0.2 ye	ears	
PRINCIPAL EMISSIONS <sup>(1)</sup>			4)
Type		Energy (keV)	Intensity (%)
gamma		661.66	85.21
SOURCE DESCRIPTION	¢		
Active Diameter:	1 mm	Nature of Active Deposit:	Cesium Chloride in resin matrix
Overall Diameter:	12.7 mm	Position of Active Bead:	8 mm from the end of rod
Overall Length:	102 mm		
METHOD OF CALIBRATIC	ON		
The source was calibrated a activity of the standard w intercomparisons with the N traceable to the National Inst	on a high purity g vas determined lational Institute c stitute of Standar	ermanium detector against a C using an efficiency establish of Standards and Technology. ds and Technology.	s-137 standard at 661.66 keV. The ned and verified through ongoing This standard is indirectly (implicitly)
North American Scientific, conducted by the National Ir	Inc. actively par Institute of Standa	ticipates in the Radioactivity I rds and Technology in coopera	Measurements Assurance Program tion with the Nuclear Energy Institute.
TOTAL UNCERTAINTY (9	99% Confidence L	_evel)	
Systematic	uncertainty	<b>,</b>	3.22%
Random ur	ncertainty		1.06%
Total unce	rtainty (quadratic	sum)	± 3.39%
Jeff W Calibration	A A agner Laboratory		March 28, 2002 Date
REFERENCES			
(1) Table of Radioactive Isc	topes, 7th edition, 19	986.	
• LEAK TEST CERTIFICATION ON REVERSE •			
North American Scientific, Inc	. 7435 Greenbust	n Ave., North Hollywood, CA 91605	(818) 734-8600 Fax (818) 734-5200

TECHNICAL DATA SHEET			
MODEL NO.: MED3503 Spot Marker			
<u>QUANTITY:</u> 1			
S/N OR <u>NUCLIDE LOT NO. ACTIVITY</u> Co-57 29603 2.0 MBq (55 uCi)	UNCERTAINTYREFERENCE DATEHALF-LIFE±15%November 1, 2002271.77 days		
NATURE OF ACTIVE DEPOSIT:	Evaporated CoCl <sub>2</sub>		
BACKING:	n/a		
COVER:	n/a		
ACTIVE DIAMETER/AREA:	3 mm active diameter		
OVERALL DIAMETER/DIMENSIONS:	25.4 mm diameter x 6.4 mm thick		
• LEAK TEST CERTIFICATION ON REVERSE •			
North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 734-8600 Fax (818) 734-5200			

### **CERTIFICATE OF CALIBRATION**

### MODEL MED3550 GAMMA REFERENCE STANDARD

Radionuclide:		
Serial Number:		
Half Life <sup>(1)</sup> :		

28792 30.0 ± 0.2 years

Cs-137

Activity:

Reference Date:

7.683 MBq (207.7 μCi) 1200 PDT October 1, 2002

### PRINCIPAL EMISSIONS<sup>(1)</sup>

<u>Type</u> gamma

Energy (keV) 661.660 Intensity (%)

85.21

### SOURCE DESCRIPTION

The activity is uniformly distributed throughout approximately 10 milliliters of epoxy resin and cover ed with 10 milliliters of inactive epoxy resin in a 30 milliliter polyethylene vial with an epoxy sealed cap.

### METHOD OF CALIBRATION

The standard was calibrated by direct comparison to 10 milliliters of standardized solution traceable to the National Institute of Standards and Technology, in an identical geometry, using a pressurized ion chamber. Therefore, the activity value provided is equivalent to 10 milliliters of standardized solution. This standard is indirectly (implicitly) traceable to the National Institute of Standards and Technology.

North American Scientific, Inc. actively participates in the Radioactivity Measurements Assurance Program conducted by the National Institute of Standards and Technology in cooperation with the Nuclear Energy Institute.

TOTAL UNCERTAINTY (99% Confidence Level) <u>± 5.00%</u>

Seyed Miri Calibration Laboratory

September 24, 2002 Date

REFERENCES

(1) <u>Table of Radioactive Isotopes</u>, 7th edition, 1986.

### • LEAK TEST CERTIFICATION ON REVERSE •

North American Scientific, Inc. 7435 Greenbush Ave., North Hollywood, CA 91605 (818) 734-8600 Fax (818) 734-5200

### SHIPPING PAPER FOR DANGEROUS GOODS

#### **CARRIER: Cardinal Health ORIGIN:** Bill No .: SHIPPER: **CONSIGNEE NAME AND ADDRESS: Eastlantic Vineland Diagnostic Cardinal Health** 1470 S. Main Rd. 650 Elmwood Ave. Vineland, NJ 08360 Sharon Hill, PA 19079 856-794-9674 610-461-7070 Proper Shipping Name Hazard Class UN No. UN 2915 Radioactive Material, Type A Package 7 NATURE AND QUANTITY OF DANGEROUS GOODS NO. PIECES WITHIN BOX RADIONUCLIDE CHEMICAL FORM/PHYSICAL STATE ACTIVITY - GBg ACTIVITY - mCi Co-57 2 solid epoxy 0.285 7.702) ( Cs-137 0.006 ( 3 solid epoxy 0.175) 0.000 0.000 0.000

RADIOACTIVE LABEL CATEGORY: White-I Yellow-II Transport Index: Emergency Spill-Leak-Fire-Exposure or Accident Contact: CHEMTREC 800-424-9300

Total Activity Contained in one (1) box:

This is to certify that the above named materials are properly classified, described, packaged, marked and labeled, and are in proper condition for transportation according to applicable regulations for the Department of Transportation.

Certified By:	Courier Signature:	Consignee Signature:
Euly O Donnell &	A pan Doterro	
Date: Time:	Date: Time:	Date: Time:
	10/1708 17.21	

Run:

Route:

Box:

0.000 0.000 0.000

0.291

7.877

O'Don DAILY SURVEY AND WEEKLY WIPE AREAS Ludlum Im AREA I MONDAY I TUESDAY I WEDNES THURSDAY | FRIDAY | WEEKLY WIPES (DPIN 11-11-08 ? 5-137 mi BACKGROUND CPms 410 00 1) NUC MED TABLE 2) NUC MED FLOOR BLGD= B. 2m 3) INJECTION AREA where = 0.2 nr/ 4) HOT LAB FLOOR 5) HOT LAB COUNTER 6) BEHIND L BLOCK 7) WELL ODWITCH DOSE Calibrator TECHNOLOGIST\_\_ RSO

uelin O'Don DAILY SURVEY AND WEEKLY WIPE AREAS L ELOW MONDAY I TUESDAY I WEDNESDAY AREA THURSDAY I FRIDAY EEKLY WIPES ( DPI 11-11-0 BACKGROUND CPm's  $0 \mathbf{\alpha}$ 1) NUC MED TABLE 2) NUC MED FLOOR 3) INJECTION AREA and = 0.2 mul 4) HOT LAB FLOOR 5) HOT LAB COUNTER 6) BEHIND L BLOCK 7) WELL ODUNTED DOSE. Culibrator TECHNOLOGIST RSO\_

oom Close QU Euclip O'Donne rminAYE 200 DAILY SURVEY AND WEEKLY WIPE AREAS CENSES Lidlum AREA I MONDAY I TUESDAY I WEDNESDAY I THURSDAY I FR WIPES ( DPIN 11.12.08 1-1-918-100 BACKGROUND CPms .02 1) NUC MED TABLE .02 2) NUC MED FLOOR .02 3) INJECTION AREA .02 4) HOT LAB FLOOR .02 i., ., 5) HOT LAB COUNTER .02 <u>37.3</u> 6) BEHIND L BLOCK 123.6 DOSE 7) WELL OOWNEED .02 Calibrator 2:03 E0:1 TECHNOLOGIST RSO.



This is to acknowledge the receipt of your letter/application dated

10 | 23 | 08, and to inform you that the initial processing which includes an administrative review has been performed.

There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number**  $\_\_\_ 43000$ . When calling to inquire about this action, please refer to this control number. You may call us on (610) 337-5398, or 337-5260.

NRC FORM 532 (RI) (6-96) Sincerely, Licensing Assistance Team Leader