

Cardinal Health  
Nuclear Pharmacy Services  
650 Elmwood Ave.  
Sharon Hill, PA 19079  
www.cardinal.com



**Cardinal Health**

11/12/2008

29-30895-01  
03036530

Eastlantic Vineland  
1470 S. Main Rd.  
Vineland, NJ 08360

Nuclear Medicine Department

This letter confirms the return for disposal of the following sources:

10 mCi, Co-57, Flood  
Serial #: BM01103681, Calibration Date: 08/10/08

55 uCi, Co-57, Spot Marker  
Serial #: 29603, Calibration Date: 11/02

10 uCi, Cs-137, Spot Marker  
Serial #: 1125-7-20, Calibration Date: 08/01/05

200 uCi, Cs-137, Vial  
Serial #: 28792, Calibration Date: 10/01/02

1 uCi, Cs-137, Rod  
Serial #: 17417, Calibration Date: 04/01/02

The sources were returned to Eckert & Ziegler Company- Isotope Products Laboratories for ultimate disposal.

Wipe Test Result:

Counter Efficiency:	95.74%
Bkg cpm:	118
Net cpm:	75
Net µCi:	0*

\*Passed with < 0.005 uCi removable contamination.

Thank you for choosing Cardinal Health for your radiopharmaceutical needs.

Sincerely,

Lynn Potenza  
Radiation Safety Officer

P-7

RECEIVED  
REGION I  
2008 NOV 17 PM 12:41

143000  
N°SS/RGNI MATERIALS-002