

MS 16
Q-2

Servicing Imaging Systems International, Inc.
52 Watchung Avenue
U. Montclair, New Jersey 07043
(201) 783-3638
FAX No. (201) 783-8479

License No. 29-30214-01
Docket No. 030-33828
Control No. 123675

January 6, 1997

Attention: Mr. John D. Kinneman, Chief
Nuclear Materials Safety Branch 2
Division of Nuclear Materials Safety
UNITED STATES
NUCLEAR REGULATORY COMMISSION
Region I
475 Allendale Road
King of Prussia, PA 19406-1415

RE: APPLICATION FOR MATERIAL LICENSE AMENDMENT DATED SEPTEMBER
12, 1996.
YOUR LETTER OF DECEMBER 23, 1996..

Dear Sir:

Please receive the requested information as follows:

1. Terrance Lee, M.D., is an authorized user under NRC Materials License No. 29303-22-01.
2. Each day of operation, we will check our dose calibrators for constancy, using 2 dedicated check sources, i.e. Co-57 and Cs-137, at the setting we are using for Tc-99m. We will make and maintain records as required by the regulations (10 CFR 35.50).
3. We will make quarterly inventories of our sealed sources and maintain records as required by the regulations (10 CFR 35.59).
4. The suite of our facility in West-Orange is 301-304, with the main entrance at 304. The access to the Nuclear Cardiology is through the back door at 301.

Sincerely yours,


Dr. Aharon Ben-Haim

OFFICIAL RECORD COPY

ML 10

123675
JAN - 8 1997