

**U.S. NUCLEAR REGULATORY COMMISSION**

Date: 12/8/08

**TELEPHONE CONVERSATION RECORD**

Time: 8:20am

Mail Control 142828 License No(s). 52-17776-02 Docket No(s). 03020896  
 or Report No(s).

Name of Licensee: GeoCim, Inc.

Name of Participant(s): Rolando Davila, RSO Jenny Johanen, HP

Telephone No. 787-792-2626 610-337-5071

Subject: Voidance of license amendment request letter dated 9/18/08 (re:  
 (NOTE: This will be used as the calibration of portable gauges services to others limited to calibration  
 Documents Title in ADAMS) and leak testing.)

Summary: You may do the calibration, but Condition 13E. limits leaktesting to persons specifically authorized. And you are not authorized. Condition 19 does not allow you to any cleaning, maintenance or repair of gauges. Licensee understands he is limited to calibrations of the gauges only.

Action Required: NRC explained to licensee what he may do. Control NO. 142828 is VOIDED by this telephone converstion record.

Document Availability:  Publicly Available  Non-Publicly Available  
 Non-Sensitive  Non-Sensitive Copyright  Sensitive  Sensitive Copyright  
 Immediate Release  Normal Release  Delay Release Date

Prepared &amp; SUNSI Review Completed By: Johansen / RA / Date: 12/08/2008