

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

*Indiana Nuclear Associates, LLC
Indianapolis, IN*

REPORT NUMBER(S) *2008-001*

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351

3. DOCKET NUMBER(S)

030-29663

4. LICENSEE NUMBER(S)

13-24834-01

5. DATE(S) OF INSPECTION

Dec. 7, 2008

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title

Printed Name

Signature

Date

LICENSEE'S REPRESENTATIVE

NRC INSPECTOR

Deborah A. Piskura

Deborah A. Piskura

12/07/2008

SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION

1. LICENSEE Indiana Heart Associates, P.C. REPORT NUMBER(S) 2008-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-29663	4. LICENSE NUMBER(S) 13-24834-01	5. DATE(S) OF INSPECTION December 9, 2008	
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01 – 03.08		

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02201	2. PRIORITY 5	3. LICENSEE CONTACT Kiran R. Kareti, M.D., RSO	4. TELEPHONE NUMBER 317-355-1234
-----------------------------	------------------	---	-------------------------------------

<input checked="" type="checkbox"/>	Main Office Inspection	Next Inspection Date: December 2013
<input checked="" type="checkbox"/>	Field Office 8075 N. Shadeland Ave, Indpls, IN & 1402 E. County Line Rd, Greenwood, IN	
<input type="checkbox"/>	Temporary Job Site Inspection	

PROGRAM SCOPE

This licensee was a large cardiac clinic network authorized to use licensed material permitted by Section 35.200. The licensee employed five full-time technologists who performed approximately 500+ nuclear cardiology procedures per month. The licensee received unit doses from a licensed radiopharmacy. The licensee retained a physics consultant who audited the radiation safety program on a quarterly basis.

This inspection consisted of interviews with licensee personnel, a review of select records, tour of the nuclear medicine department, and independent measurements. The inspection included observations of security of byproduct material, receipt and surveys of byproduct material packages, use of personnel monitoring, dose calibrator QA checks, and area surveys. The inspector observed licensee personnel prepare, assay and inject two cardiac stress testing dosages.