

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:

Open MRI, LLC
d/b/a Advanced Diagnostic Imaging
1120 Professional Boulevard
Evansville, IN 47714
REPORT NUMBER(S)

2. NRC/REGIONAL OFFICE

U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4351

3. DOCKET NUMBER(S)
030-34799

4. LICENSEE NUMBER(S)
13-32112-01

5. DATE(S) OF INSPECTION
November 19, 2008

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:


- ☒ 1. Based on the inspection findings, no violations were identified.
- ☐ 2. Previous violation(s) closed.
- ☐ 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- ☐ 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Robert P. Hays		11/19/08

SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION

1. LICENSEE

Open MRI, LLC

REPORT

NUMBER(S) 2008-001

2. NRC/REGIONAL OFFICE

Region III

2443 Warrenville Road, Suite 210

Lisle, IL 60532

3. DOCKET NUMBER(S)

03034799

4. LICENSE NUMBER(S)

13-32112-01

5. DATE(S) OF INSPECTION

November 19, 2008

6. INSPECTION PROCEDURES USED

87130 (10/24/02)

7. INSPECTION FOCUS AREAS

03.01-03.07

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S)

02200

2. PRIORITY

3

3. LICENSEE CONTACT

Carol Collum, CNMT

4. TELEPHONE NUMBER

812-471-7086



Main Office Inspection

Next Inspection Date: November 2011



Field Office

Temporary Job Site
Inspection

PROGRAM SCOPE

The licensee was a medical clinic located in Evansville, Indiana, with authorization by the license to use any byproduct material as needed permitted by 10 CFR 35.100, 35.200, excluding generators, and 35.300, not to exceed 1 curie, excluding iodine-131 for thyroid carcinoma therapy. One FT nuclear medicine technologist (NMT) routinely performs an average of 12-15 administrations per week, with the majority of tests being bone scans. No I-131 dosages greater than 20 millicuries are administered and average 1-2 procedures per year. Licensed material is received as unit doses from an Evansville, IN, nuclear pharmacy. No change in RSO or NMT since previous inspection. Area surveys of the storage, injection, hot lab, and scanning areas did not reveal any unusual or elevated readings.

Performance Observations

During the inspection, the licensee's NMT demonstrated/discussed: (1) survey meter use and calibration; (2) package receiving and check-in procedures; (3) unit dose prep and safe use; (4) package wipe test counting; (5) dosimetry; (6) dose calibrator tests; (7) waste handling; (8) security of licensed material; (9) sealed source inventory; (10) radiation safety program audits; (11) iodine-131 procedures and written directives; and (12) contamination surveys with no contamination events since the previous inspection.