



RIVERSIDE

REGIONAL MEDICAL CENTER ²⁰⁰⁸ DEC 22 PM 2: 03

RECEIVED
REGION I

December 18, 2008
License No.: 45-09001-01
Center Docket No. 03003330

2008003

K-8

**U.S. Nuclear Regulatory Commission
ATTN: Document Control Desk
Washington, D.C. 20555**

RE: Reply to a Notice of Violation

(1) The reason for the violation

Due to an oversight while the RSO was not on site, material that had been in storage was transferred to a disposal vendor without a current leak test.

(2) The corrective steps that have been taken and the results achieved

We have added the following statement adjacent to the inventory listing of each source that would require leak test if it was in service

"This source must pass a leak test prior to use or transfer to another person."

(3) The corrective steps that will be taken to avoid further violations

We will add the following statement adjacent to the inventory listing of each source that would require leak test if it was in service

"This source must pass a leak test prior to use or transfer to another person."

(4) The date when full compliance will be achieved.

We have completed the above change on Dec 17, 2008.

Thank you for your assistance. If you have any questions, you may e-mail at me Harold.prussia@rivhs.com or call at 757-594-2757.

Harold Prussia

CERTIFICATION

THE UNDERSIGNED OFFICIAL EXECUTING THIS CERTIFICATION UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS RESPOSNE ARE BINDING AND CERTIFIES THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

CERTIFYING OFFICER NAME AND TITLE	SIGNATURE	DATE
Diana LoVecchio Vice President of Operations and Support Services		12/19/08

**Cc: Regional Administrator
U.S. NRC Region I
475 Allendale Road
King of Prussia, PA 19406-1415**