



# RIVERSIDE

## REGIONAL MEDICAL CENTER

December 18, 2008

License No.: 45-09001-01

Center Docket No. 03003330

**U.S. Nuclear Regulatory Commission**

**ATTN: Document Control Desk**

**Washington, D.C. 20555**

**RE: Reply to a Notice of Violation**

**(1) The reason for the violation**

Due to an oversight while the RSO was not on site, material that had been in storage was transferred to a disposal vendor without a current leak test.

**(2) The corrective steps that have been taken and the results achieved**

We have added the following statement adjacent to the inventory listing of each source that would require leak test if it was in service

"This source must pass a leak test prior to use or transfer to another person."

**(3) The corrective steps that will be taken to avoid further violations**

We will add the following statement adjacent to the inventory listing of each source that would require leak test if it was in service

"This source must pass a leak test prior to use or transfer to another person."

**(4) The date when full compliance will be achieved.**

We have completed the above change on Dec 17, 2008.

Thank you for your assistance. If you have any questions, you may e-mail at me [Harold.prussia@rivhs.com](mailto:Harold.prussia@rivhs.com) or call at 757-594-2757.

Harold Prussia

**CERTIFICATION**

THE UNDERSIGNED OFFICIAL EXECUTING THIS CERTIFICATION UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS RESPOSNE ARE BINDING AND CERTIFIES THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

CERTIFYING OFFICER NAME AND TITLE	SIGNATURE	DATE
Diana LoVecchio Vice President of Operations and Support Services		12/19/08

**Cc: Regional Administrator**  
**U.S. NRC Region I**  
**475 Allendale Road**  
**King of Prussia, PA 19406-1415**

*Handwritten initials/signature*  
JEO  
2641