



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION IV
612 EAST LAMAR BLVD, SUITE 400
ARLINGTON, TEXAS 76011-4125

January 20, 2009

Bartlett Regional Hospital
ATTN: William Allen Richey, M.D.
Radiation Safety Officer
3260 Hospital Drive
Juneau, AK 99801

SUBJECT: LICENSE AMENDMENT AND NOTIFICATION PER 10 CFR 35.14(b)(4)

Please find enclosed Amendment No. 29 to NRC License No. 50-17686-01, **acknowledging receipt of your survey results for the decommissioned treadmill room or "stress lab."** **Based on the results of the survey data, there are no further actions required and the decommissioned treadmill room is suitable for unrestricted use as authorized under Item 18.G. of your license.** An environmental assessment for this action is not required, since this action is categorically excluded under 10 CFR 51.22(c)(14)(iv). You should review the enclosed document carefully and be sure that you understand all conditions. If you have any questions regarding your NRC license, please contact me at 817-276-6552.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the NRC's document system (ADAMS). ADAMS is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>.

Thank you for your cooperation.

Sincerely,

A handwritten signature in cursive script, reading "Rachel S. Browder", is written over a horizontal line.

Rachel S. Browder, Health Physicist
Nuclear Materials Safety Branch B

Docket: 030-13158
License: 50-17686-01
Control: 471985

Enclosure: As stated

MATERIALS LICENSE

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	In accordance with letter dated September 29, 2008
1. Bartlett Regional Hospital	3. License number 50-17686-01 is amended in its entirety to read as follows:
2. 3260 Hospital Drive Juneau, Alaska 99801	4. Expiration date September 30, 2014
	5. Docket No. 030-13158 Reference No.

- | | | |
|---|----------------------------------|--|
| 6. Byproduct, source, and/or special nuclear material | 7. Chemical and/or physical form | 8. Maximum amount that licensee may possess at any one time under this license |
| A. Any byproduct material permitted by 10 CFR 35.100 | A. Any | A. As needed |
| B. Any byproduct material permitted by 10 CFR 35.200 | B. Any | B. As needed |

9. Authorized use:
- A. Any uptake, dilution and excretion study permitted by 10 CFR 35.100.
- B. Any imaging and localization study permitted by 10 CFR 35.200.

CONDITIONS

10. Licensed material may be used or stored only at the licensee's facilities located at 3260 Hospital Drive, Juneau, Alaska.
11. The Radiation Safety Officer for this license is William Allen Richey, M.D.
12. Licensed material is only authorized for use by, or under the supervision of:
- A. Individuals permitted to work as an authorized user in accordance with 10 CFR 35.13 and 35.14.
- B. The following individuals are authorized users for the material and medical uses indicated:

Authorized Users

Gordon T. Blair, Jr., M.D.
William Allen Richey, M.D.
Theresa J. Shanley, M.D.
Steven T. Strickler, D.O.

Material and Use

35.100; 35.200
35.100; 35.200
35.100; 35.200
35.100; 35.200

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
50-17686-01Docket or Reference Number
030-13158

Amendment No. 29

13. For sealed sources not associated with 10 CFR Part 35 use, the following conditions apply:
- A. Sealed sources shall be tested for leakage and/or contamination at intervals not to exceed the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State.
 - B. Notwithstanding Paragraph A of this Condition, sealed sources designed to primarily emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed 3 months.
 - C. In the absence of a certificate from a transferor indicating that a leak test has been made within the intervals specified in the certificate of registration issued by the U.S. Nuclear Regulatory Commission under 10 CFR 32.210 or under equivalent regulations of an Agreement State, prior to the transfer, a sealed source received from another person shall not be put into use until tested and the test results received.
 - D. Sealed sources need not be tested if they contain only hydrogen-3; or they contain only a radioactive gas, or the half-life of the isotope is 30 days or less; or they contain not more than 100 microcuries of beta- and/or gamma-emitting material or not more than 10 microcuries of alpha-emitting material.
 - E. Sealed sources need not be tested if they are in storage and are not being used; however, when they are removed from storage for use or transferred to another person and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.
 - F. The leak test shall be capable of detecting the presence of 0.005 microcurie (185 becquerels) of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie (185 becquerels) or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(c)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. The report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region IV, 612 East Lamar Blvd., Suite 400, Arlington, Texas 76011-4125, ATTN: Director, Division of Nuclear Materials Safety. The report shall specify the source involved, the test results, and corrective action taken.
 - G. Tests for leakage and/or contamination, including leak test sample collection and analysis, shall be performed by the licensee or by other persons specifically licensed by the U.S. Nuclear Regulatory Commission or an Agreement State to perform such services.
 - H. Records of leak test results shall be kept in units of microcuries and shall be maintained for 3 years.
14. Sealed sources containing licensed material shall not be opened or sources removed from source holders by the licensee.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
50-17686-01Docket or Reference Number
030-13158

Amendment No. 29

15. The licensee shall conduct a physical inventory every 6 months, or at other intervals approved by the U.S. Nuclear Regulatory Commission, to account for all sources and/or devices received and possessed under the license. Records of inventories shall be maintained for 3 years from the date of each inventory and shall include the radionuclides, quantities, manufacturer's name and model numbers, and the date of the inventory.
16. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing financial assurance for decommissioning.
17. The licensee is authorized to transport licensed material in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
18. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. This license condition applies only to those procedures that are required to be submitted in accordance with the regulations. Additionally, this license condition does not limit the licensee's ability to make changes to the radiation protection program as provided for in 10 CFR 35.26. The U.S. Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated July 12, 2004 (ML042120579)
 - B. Letter dated April 26, 2006 (ML061960369)
 - C. Letter dated March 12, 2007 (ML070880452)
 - D. Letter dated March 27, 2007 (ML071060189)
 - E. Letter dated April 9, 2007 (ML071470469)
 - F. Letter dated April 16, 2008 (ML081440585)
 - G. Letter dated September 29, 2008 (ML083400127)

FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date: January 20, 2009By: *Rachel S. Browder*

Rachel S. Browder, Health Physicist
Nuclear Materials Safety Branch B
Region IV
Arlington, Texas 76011-4125