



REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

January 26, 2009

Docket No. 03036139

License No. 29-30780-01

Sanjiv Girish Faldu, M.D.
Owner/Radiation Safety Officer
555 Iron Bridge Road, Suite 100
Freehold, NJ 08857

SUBJECT: NRC INSPECTION REPORT NO. 03036139/2009001, FREEHOLD, NEW JERSEY SITE

Dear Dr. Faldu:

On January 12 and 16, 2009, Steven Courtemanche of this office conducted a safety inspection at the above address of activities authorized by the above listed NRC license. The inspection was an examination of your licensed activities as they relate to radiation safety and to compliance with the Commission's regulations and the license conditions. The inspection consisted of observations by the inspector, interviews with personnel, and a selective examination of representative records. Additional information provided in your correspondence dated January 15, 2009, and the telephone conversation on January 16, 2009, between you and Mr. Courtemanche of this office were also examined as part of the inspection. The findings of the inspection were discussed with you at the conclusion of the inspection by telephone on January 16, 2009.

Within the scope of this inspection, no violations were identified. During the inspection, the inspector confirmed the corrective and preventive actions taken by the licensee as documented in the licensee's letter dated January 8, 2004. The item concerning the failure to perform leak tests of sealed sources at the required six month interval is considered closed. Two other issues regarding the collection of personnel dosimetry and the performance of dose calibrator linearity tests will be examined at the next inspection of your licensed program. Though not considered as items of noncompliance, it is our understanding that the licensee is implementing procedures enhancing the collection and mailing of personnel dosimeters and ensuring that the dose calibrator linearity checks are performed as needed. We have no further questions concerning the above matters at this time.

Current NRC regulations are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Regulations, Guidance, and Communications**. The current NRC Enforcement Policy is included on the NRC's website at www.nrc.gov; select **About NRC; How We Regulate; Enforcement**; then **Enforcement Policy**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 7:00 a.m. to 6:30 p.m. EST, Monday through Friday (except Federal holidays).

S. Faldu

2

No reply to this letter is required. Please contact Mr. Courtemanche at 610-337-5075 if you have any questions regarding this matter.

Sincerely,

Original signed by James P. Dwyer

James P. Dwyer, Chief
Commercial and R&D Branch
Division of Nuclear Materials Safety

cc:
State of New Jersey

S. Faldu

2

No reply to this letter is required. Please contact Mr. Courtemanche at 610-337-5075 if you have any questions regarding this matter.

Sincerely,

Original signed by James P. Dwyer

James P. Dwyer, Chief
Commercial and R&D Branch
Division of Nuclear Materials Safety

cc:
State of New Jersey

DOCUMENT NAME: G:\WordDocs\Current\Insp Letter\L29-30780-01.2009001.doc

SUNSI Review Complete: SCourtemanche

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	N	DNMS/RI		DNMS/RI		
NAME	SCourtemanche/src*		JDwyer/JPD				
DATE	01/21/2009		01/26/09				

*see prior concurrence

OFFICIAL RECORD COPY

ML090260162