

## CONVERSATION RECORD

|TIME |DATE

1-28-2009

 VISIT  CONFERENCE  TELEPHONE INCOMING OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT

Cricket Buch, Team Leader  
of Radiology

ORGANIZATION (OFFICE, DEPT. ETC.)

SSM St. Joseph Health Center - Wentzville  
500 Medical Drive  
Wentzville, MO 63358

TELEPHONE NO.

636-327-1144

FAX

-1143

## SUBJECT

Control No. 317853- Re - Amendment request by letter received November 14, 2008 (dated April 10, 2008), regarding request to terminate their NRC license, License No. 24-24859-01.

## SUMMARY

In a telephone conversation with Ms. Buch on this date, we discussed the following items and the need to provide additional information as noted below:

1. The referenced document states that a thorough close out survey was performed and that the results of the survey and wipe tests were enclosed with the letter. The survey results were not received with the letter so please provide a copy for our review. This survey should include a diagram of the room(s)/areas where wipes and measurements were taken and identification of the approximate locations of the survey points of reference and the date of the survey. Also, please identify the instruments used for the surveys and the latest calibration results as of the survey date.
2. The letter referenced above states that the last nuclear medicine exam was performed on March 31, 2008. Please indicate which radionuclides were typically used and for what diagnostic procedure(s). Please show dates of the last receipt of each radionuclide, including the indicated activity at time of receipt, and its use and or disposition.
3. The letter referenced above, also states that the radioactive sealed sources (small calibration/reference) have been transferred to SSM St. Joseph Health Center in St. Charles, MO., which has License No. 24-15159-01. Please identify, by manufacturer, model and/or serial number (if available), isotope and activity, each sealed source transferred to them, and records confirming both the transfer of the sealed sources by this licensee and receipt (including date of receipt) of the sources by the recipient licensee. Also, please include a leak test record (current as of the transfer date) for each of the transferred sources having an activity sufficient to require leak testing, confirming their non-leaking status. (Please note that sealed sources need not be tested if they contain only hydrogen-3, or they contain only a radioactive gas, or the half-life of the isotope is 30 days or less, or they contain not more than 100 microcuries of beta- and/or gamma emitting material or not more than 10 micorcuries of alpha-emitting material.)
4. Also, please confirm (if it is in fact correct) that the licensee has never possessed or used any leaking sealed sources, based on routine leak test data.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this conversation record will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

ACTION REQUIRED


Please provide written response by February 10, 2009, to the above items, as additional information to Control No. 317853 and send it to me by facsimile (630-515-1078). No follow-up hard copy is needed. If any questions, please call me at 630-829-9829. Also, please call me right after you have faxed the response.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Loren J. Hueter



1-28-2009

ACTION TAKEN

SIGNATURE

TITLE

DATE