



RECEIVED

FEB 27 2009

HALEY & ALDRICH

North East Technical Services, Inc.
75 Aileron Ct., Suite 4
Westminster, MD 21157
Ph: 410.751.5090 Fax: 410.751.5091

Haley & Aldrich, Inc. - McLean
7926 Jones Branch Dr., Suite 870
McLean, VA 22102
ATTN: Jen Macgreger

Shipping Address: 7926 Jones Branch Dr., Suite 870
McLean, VA 22102

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	5001EZ	Gauge S/N	H2489	Leak Test Date	2/13/2009
Source	Reading in microCuries				
5224GO	0.0000611				
NJ02645	0.00000				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: D. de Sina

Date: FEB 20 2009