



REMOVED

NOV 17 2008

North East Technical Services, Inc.  
75 Aileron Ct., Suite 4  
Westminster, MD 21157  
Ph: 410.751.5090 Fax: 410.751.5091

Haley & Aldrich, Inc. - McLean  
7926 Jones Branch Dr., Suite 870  
McLean, VA 22102  
ATTN: Sandra Henning

Shipping Address: 7926 Jones Branch Dr., Suite 870  
McLean, VA 22102

## LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

Gauge Model	5001EZ	Gauge S/N	H3694	Leak Test Date	11/7/2008
Source	Reading in microCuries				
0432CX	0.00005848				
NJ04026	0.00000				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by:                     Doug de Sims                    

Date:                     NOV 10 2008