

RJ TOKARZ MEDICAL IMAGING RADIATION SAFETY CORP
44 STELTON ROAD, SUITE 115
PISCATAWAY, NJ 08854

732-424-0909

732-424-3715

*Ms 16
J-6*

March 12, 2009

TO: Dennis Lawyer

Subject: Control 143312 response

*29-31295-01
03037600*

This is my response to our discussion on March 6, 2009 regarding the license amendment for Alan Kalischer, M.D.

The cobalt-57 sheet source will **be** stored in a leaded container when on board the van. This will be done to **reduce** the exposure rate outside the van to less than 2 mR/h. There will be confirmation with an ambient exposure rate survey performed with a gm survey meter.

Attachment 2 is the floorplan for the location Care Station which is the client referenced in this application. The address of the Care Station is

The Care Station

90 Route **22 West**

Springfield, NJ 07081

143312

We will use a dedicated check source to perform an operational check on the survey meter before **each** use at each address of use.

There is no sanitary sewer in the van.

A **sample** of our agreement letter with the client is enclosed as Attachment 1.

Please call me with any questions regarding this matter.

Sincerely,

A handwritten signature in black ink, appearing to read "Robert J. Tokarz". The signature is fluid and cursive, with the first name being the most prominent.

Robert J. Tokarz
Radiation Safety Officer

CONTROL 143312: ATTACHMENT 1

MOBILE MEDICAL SERVICES

LETTER OF AGREEMENT

Between the Client (Name and address) and the Licensee (Alan Kalischer, M.D.,
Fanwood-Westfield Cardiology, 313 South Avenue, Fanwood, NJ 07023)

DATE:

This is an agreement between the Client and Licensee regarding Mobile Medical Services requiring the use of licensed byproduct material at the Client's building.

The responsibilities of the Client are as follows:

- Provide adequate space for use of **licensed** byproduct **material**.
- b. Allow Licensee to oversee use of licensed **byproduct** material.
- c. Allow Licensee authority to implement radiation safety practices and identify and correct radiation safety problems..

The responsibilities of the Licensee are as follows:

- a. Bring byproduct material and instrumentation into the Client's building.
- b. Administer the byproduct material.
- c. Survey the Client's facility to ensure compliance with 10 CFR Part 20 **before** leaving the client's facility.
- d. Quality control of the instrumentation before **first-patient** use.
- e. Stop unsafe operations.

This agreement will be applicable for the entire **period** of time over which the service is to be provided.

Signed

For the Client:

For the Licensee:

Alan Kalischer, M.D.