



MARYVIEW MEDICAL CENTER
Bon Secours Hampton Roads Health System

Maryview Medical Center
Maryview Behavioral Medicine Center
Martha Davis Cancer Center
Bon Secours Home Care
Bon Secours Heart Institute
Bon Secours Health Center at Harbour View

March 17, 2009

U.S. NRC Region I Office
Division of Nuclear Materials Safety
475 Allendale Road
King of Prussia, PA 19406-1415

K-3

ATTN: Ms. Michelle Simmons

RE: Docket No.: 03003347 License No.: 45-10831-02 Mail Control No.: 143224

Dear Ms. Simmons:

I am submitting the following confirmations as per our telephone conversation this morning.

We confirm that the RSO will receive vendor training prior to initiating patient treatments. This training will be provided by either the vendor at source installation or by the AMP.

We confirm that the enclosure is equipped with an independent radiation monitoring system having a radiation detector located inside the enclosure and a remote read-out located at the adjacent control console.

Again, thank you for your assistance in the amendment request.

Very truly yours,

Leland R. Kirkland, Ph.D.
Oncology Physicist

143224

NMSS/RCM1 143224