

**From:** Elizabeth Ullrich  
**Sent:** Monday, March 23, 2009 2:18 PM  
**To:** michelle.nowaczyk@radnet.com  
**Subject:** 2/9/09 letter to amend NRC License 07-16529-01

Papastavros' Associates Medical Imaging, LLC  
License No. 07-16529-01  
Docket No. 030-11379  
Control No. 143444

Subject: Request for additional information to your amendment letter dated February 9, 2009

Ms. Nowaczyk,

The February 9, 2009, letter requests Germanium-68 sources to calibrate the scanner detectors, 500 Mbq per source and 200 MBq total (appr 13.5 millicuries per source and 112 mCi total). Please let me know if you believe that

a - these are calibration sources that meet the requirements of 10 CFR 35.65, and do not need to be specifically listed in the license; OR

b. - they are sealed sources for diagnostic medical uses under 35.500; if so, state the total amount of material to be possessed as Item 6/7/8.D. on your license; OR

c. - they are sealed sources which do NOT meet the 35.65 requirements or the 35.500 requirements, and must be listed specifically on your license; if so, provide the manufacturer name and the model number for the sources to be possessed.

Please note that you may not reply to this request by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269). If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

thanks,  
Betsy

Betsy Ullrich, MS, CHP  
Senior Health Physicist  
Commercial and R&D Branch  
Division of Nuclear Material Safety  
Region I  
Nuclear Regulatory Commission  
(610) 337-5040  
[Elizabeth.Ullrich@nrc.gov](mailto:Elizabeth.Ullrich@nrc.gov)

---

**E-mail Properties**

Mail Envelope Properties ( )

Subject: 2/9/09 letter to amend NRC License 07-16529-01  
Sent Date: 3/23/2009 2:03:48 PM  
Received Date: 3/23/2009 2:17:00 PM  
From: Elizabeth Ullrich

Created By: Elizabeth.Ullrich@nrc.gov

Recipients:  
michelle.nowaczyk@radnet.com (michelle.nowaczyk@radnet.com)  
Tracking Status: None

Post Office:

Files	Size	Date & Time
MESSAGE	6377	3/23/2009

Options

Expiration Date:  
Priority: olImportanceNormal  
ReplyRequested: False  
Return Notification: False

Sensitivity: olNormal  
Recipients received: