

DEC 18 1958

Marvin M. Mann, Assist. Director for Compliance
Division of Inspection

Lester R. Rogers, Chief, Radiation Safety Branch
Division of Licensing and Regulation

SYLVANIA-CORNING NUCLEAR CORPORATION - BYPRODUCT MATERIAL LICENSE NO.
31-2374-1, DOCKET NO. 40-682, AND SOURCE MATERIAL LICENSE NO. C-4029.

SYMBOL: DLR:RFB

The report of the July 10, 1958, inspection of the subject company's operations at Hicksville and Bayside, New York, indicated noncompliance on two items for an Ohmart gauge covered under license No. 31-2374-1. It is noted that this gauge had been disposed of prior to the visit by inspection. The related noncompliance items included (1) not conducting leak tests and (2) improper labeling of the gauge.

The device was used only one week and in storage the remainder of the time that it was possessed by the licensee. We note the inspector did not see the improper label. For these reasons, we are calling the leak test requirement to his attention but we are not citing the licensee on these two items.

A136

OFFICE ▶	DLR:RSB					
SURNAME ▶	<i>[Signature]</i> RFB:TW:LRR					
DATE ▶	12/15/58					