

Seaman Nuclear Corporation

7315 South First Street Oak Creek, WI 53154 USA

Tel 414-762-5100 Fax 414-762-5106

Leak Test Certificate

January 13, 2009

A leak test has been performed on meter, serial number **21088**, a model C-300, containing the radionuclide Ra 226.

Owned By: CTA Livingston Inc.
P.O. Box 1153
Livingston, MT 59047

Date Sample Collected: 1/5/09
Collected By: P. Frady
Date Sample Analyzed: 1/13/09
Analyzed By: T. Lindemann


Most regulatory agencies consider a source to be leaking if a leak test reveals the presence of more than 0.005 microcurie of removable contamination.

Analysis found contamination of less than 0.005 microcurie.

Analysis authorized by Wisconsin license 079-1257-01.

Seaman Nuclear Corporation

LEAK TEST DUE: 7/5/09



Scott C. Seaman
Radiation Safety Officer

S/N: 21088
Leak Tested: 1/5/09
Due: 7/5/09
 Seaman Nuclear Corporation
Oak Creek, WI (414) 762-5100