

71-0943

Best Theratronics

413 March Road
Ottawa, Ontario
Canada K2K 0E4
Tel: 613-591-2100

March 31, 2009

Mr. David Pstrak
Branch Chief
Division of Spent Fuel Storage and Transportation
United States Nuclear Regulatory Commission

Dear Mr. Pstrak,

Further to our conference call on March 24, 2009 regarding our QA Program registration with the USNRC, Best Theratronics conducted an internal review to verify that the process for packaging and transportation of radioactive material was a seamless transition from MDS Nordion to Best Theratronics.

We have reviewed all of the applicable shipments and verified that formal procedures were followed with the following quality records on file.

1. Best Theratronics maintains a Unit History File (UHF) for every device with the following records:
 - Sealed Source Certificates which include the Test for Freedom from Leakage and Freedom from Surface Contamination.
 - Radiation survey results including a detailed survey on the device after source loading and a radiation survey of the transport package prior to shipment.
2. Devices decommissioned or moved within the US were tested for Freedom from Surface Contamination prior to shipment. The records of the test results are filed in the corresponding device UHF.
3. All packages were prepared for shipment in accordance with formal procedures. All personnel performing this work are trained on these procedures and in the Transport of Dangerous Goods. (TDG). Records of training are on file.
4. All transport packages were inspected prior to shipment according to container maintenance procedures.
5. All field work in the US was performed by trained personnel in accordance with our NRC Materials License.

As a result of our investigation, we conclude that during the time period from May 1, 2008 to the present, Best Theratronics continued to meet the spirit of the 10 CFR 71.

If you have any further questions please do not hesitate to contact me at any time.

Best Regards,



Mike de van der Schueren
Director, Compliance
613-591-2100 ext. 2488 tel
613-218-8990 cell
mdvds@theratronics.ca

NM5501
NM5524