

**Delaware State  
University**

# Memo

p-7

To: Thomas Thompson

From: Todd Campbell 

CC:

Date: 4/15/2009

Re: Mail Control No. 143599.

07-11871-06

03033438

---

In response to email sent April 14, 2009, regarding Mail Control No. 143599.

1. The sealed check source is Oxford model 261 C32, 5 uCu Cs-137.
2. Waste management: I am not sure if you received the last page of the application or if you want clarification of the Model Waste procedures. I am sending a copy of the last page as submitted in the application. We dispose of our waste by decay-in-storage with small quantities being disposed of in a sanitary sewer.

**Leak tests will be performed at the intervals approved by the NRC or an Agreement State and specified in the SSD Registration Certificate. Leak tests will be performed by an organization authorized by RC or an Agreement State to provide leak testing services to other licensees or using a leak test kit supplied by an organization authorized by NRC or an Agreement State to provide leak test kits to other licensees and according to the sealed source or plated foil manufacturer's (distributor's) and kit supplier's instruction. As an alternative, we will implement the model leak test program published in Appendix R to NUREG-1556, Vol. 7, "Program-Specific Guidance About Academic, Research, and Development, and Other Licenses of Limited Scope," dated December 1999.**

#### **11. Waste Management**

**We will use the model waste procedures published in Appendix T to NUREG-1556, Vol. 7, "Program-Specific Guidance About Academic, Research, and Development, and Other Licenses of Limited Scope," dated December 1999.**