
CONVERSATION RECORD|TIME |DATE
4-14-2009

VISIT CONFERENCE TELEPHONE

INCOMING
 OUTGOING

NAME OF PERSON(S) CONTACTED OR IN CONTACT
Joshua A. Gottschall, RSO**ORGANIZATION (OFFICE, DEPT. ETC.)**
Wilcox Associates, Inc.
Cadillac, MI 49601**TELEPHONE NO.**
231-775-7755
FAX -3135

SUBJECT

Control No. 317900- Re request by letter dated February 2, 2009, to delete a former gauge storage location (37867 Interchange Drive, Farmington Hills, Michigan) and replace it with nearby new one (37871 Interchange Drive, Farmington Hills, Michigan) for License No. 21-23314-01 and request to adopt the more recent version of NUREG-1556 Vol. 1, Rev. 1 issued in November 2001, and the submission of Items 7 through 11 of Appendix B of NUREG-1556 Vol. 1, Rev. 1., as an attachment to letter dated April 10, 2009 (submitted by facsimile on April 13, 2009).

SUMMARY

In a telephone conversation with Mr. Gottschall on this date, I discussed my review/evaluation of Items 7 through 11 of Appendix B of NUREG-1556 Vol. 1, Rev. 1. and requested that he address the following items:

1. For item 7, confirm that he, Joshua Gottschall, is the Radiation Safety Officer for the licensed program.
2. For item 10, "Material Receipt and Accountability", confirm that physical inventories will be conducted at intervals not to exceed 6 months, to account for all sealed sources and devices received and possessed under the license.
3. Regarding omissions of instructions (in both operating and emergency procedures) for gauge usage with an unshielded source extended more than 3 feet below the surface, Mr. Gotschall said that the licensee never uses sources at that depth and would confirm that statement in writing.
4. Regarding operating procedures, confirm that instructions will be added to block and brace the gauge to prevent movement during transport.
5. Regarding operating procedures, confirm that instructions will be added such that after making changes affecting the gauge storage area, the licensee will reevaluate compliance with public dose limits and ensure proper security of gauges.
6. Regarding Emergency Procedures, confirm that in addition to immediately securing the area around the gauge, people will be kept at least 15 feet away.
7. Regarding Emergency Procedures, confirm that if heavy equipment is involved, "the operator", as well as the equipment will be detained until it is determined there is no contamination involved.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this conversation record will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

ACTION REQUIRED


Licensee to provide requested response as additional information to Control No. 317573 and **submit it by facsimile (630-515-1078)**. Please **call me (630-829-9829) when you have sent the response**. No follow-up hard copy is needed.

NAME OF PERSON DOCUMENTING CONVERSATION

SIGNATURE

DATE

Loren J. Hueter



4-14-2009

ACTION TAKEN

SIGNATURE

TITLE

DATE