


Main Office
175 Sherman Ave.
New Haven Connecticut 06511 (203) 562 - 5115

Spring Gbn Medical Center
2200 Whitney Ave.
Hamden Connecticut 06518 (203) 248 - 9600



J-6
06-2849-d
03031265

Fax

To: Dennis Sawyer From: Dr. Ardesia/Bernice
 Fax: 6010.337.5269 Pages: _____
 Phone: _____ Date: _____
 Re: Leak Test CC: _____

- Urgent For Review Please Comment Please Reply Please Recycle

● **Comments:** This transmittal is intended only for the use of the individual or entity to which it is addressed and may contain information that is privileged and confidential. If the reader of this transmittal is not the intended recipient or the employee or agent responsible for delivering transmittal to the intended recipient, you are hereby notified that any dissemination, distribution or copying of this communication is strictly prohibited. If you have received this communication in error, please notify us immediately by telephone and return the original transmittal to us.

Hamden Fax: (203) 248 - 9661

New Haven Fax (203) 7844160

Message:

any questions Bernice Buck
can be reached at 203.234.9437

**Cardiothoracic and Vascular Group
Department of Nuclear Medicine - Hamden Office**

**Wipe Test of Sealed Sources (using Victoreen Wipe Test Counter)
(BKG 193, Calibration 112.2)**

and

Area Contamination Testing and Surveys

(using a Bicron Geiger Counter with a BKG of 0.01 mR/hr)

All Wipes and Surveys performed by Bernice Buck, RT(N), CNMT

Date	Radioisotope	Type of Source	Serial Number	Room	Survey mR/hr	(P/F) Pass/Fail
4/13/09	Co-57	Vial	1194-2611			P
4/13/09	Co-57	Vial	1194-2613			P
4/13/09	Co-57	Vial	1366-9-2			P
4/13/09	Co-57	Button	N/A			P
4/13/09	Co-57	Button	N/A			P
4/13/09	Cs-137	Vial	561-21-11			P
4/13/09	Cs-137	Vial	561-47			P
4/13/09	Co-57	Sheet	1202-002			P
4/13/09	Co-57	Sheet	1202-001			P
4/13/09				Hot Lab - Lead Castle	0.01	P
4/13/09				Hot Lab - Lead Shield	0.01	P
4/13/09				Hot Lab - Counter Top	0.01	P
4/13/09				Hot Lab - Radioactive Waste Boxes	0.01	P
4/13/09				Stress Lab	0.01	P
4/13/09				Stress Lab - Floor	0.01	P
4/13/09				Treadmill & EKG Machine	0.01	P
4/13/09				Stress Table	0.01	P
4/13/09				Stress Lab-Dressing Room 1	0.01	P
4/13/09				Stress Lab-Dressing Room 2	0.01	P
4/13/09				Stress Lab Prep Room & Surrounding Floors	0.01	P
4/13/09				Patient Bathroom	0.01	P
4/13/09				Employee Bathroom	0.01	P
4/13/09				Exam Room 1	0.01	P
4/13/09				Exam Room 2	0.01	P
4/13/09				Exam Room 3	0.01	P
4/13/09				Ultrasound Lab	0.01	P
4/13/09				Office# 1	0.01	P
4/13/09				Office #2	0.01	P
4/13/09				Medical Assistant Work Area	0.01	
4/13/09				Front Office	0.01	P
4/13/09				Kitchen/Break Room	0.01	P
4/13/09				Patient Waiting Room	0.01	P

**Cardiothoracic and Vascular Group
Department of Nuclear Medicine - Hamden Office**

**Wipe Test of Sealed Sources (using Victoreen Wipe Test Counter) (BKG 193,
Calibration 112.2)**

and

**Area Contamination Testing and Surveys
(using a Bicon Geiger Counter with a BKG of 0.01 mR/hr)**

All Wipes and Surveys performed by Bernice Buck, RT(N), CNMT

Date	Radioisotope	Type of Source	Serial Number	Calibration Date / Activity	(P/F) Pass/Fail
4/13/09	Co-57	Vial	1194-2611	8/06 5 mCi	P
4/13/09	Co-57	Vial,	1194-2613	8/06 5 mCi	P
4/13/09	Co-57	Vial	1366-9-2	3/1/09 5.5 mCi	P
4/13/09	Co-57	Button	N/A	1 uCi	P
4/13/09	Co-57	Button	N/A	1 uCi	P
4/13/09	Cs-137	Vial	561-21-11	7/01/99 0.220 mCi	P
4/13/09	Cs-137	Vial	561-47	4/15/97 0.247 mCi	P
4/13/09	Co-57	Sheet	1202-002	9/06 15 mCi	P
4/13/09	Co-57	Sheet	1202-001	9/06 15 mCi	P

9 Sealed Sources were picked up by the Cardinal Health delivery person (Mark) on Monday April 13,2009 at 11:25am.

Survey of the Outgoing Sources (mR/hr)

Source	BKG	1 Meter	Surface
2 – Co57 Sheet Sources in Lead Lined Cases	0.01	0.02	0.2
Cardinal Health Case with 3 Co-57 vials, 2 Cs-137 vials and 2 Cs 137 buttons	0.01	0.05	1.0

Cardiothoracic and Vascular Group
Department of Nuclear Medicine - Hamden Office

Wipe Test of Sealed Sources (using Victoreen Wipe Test Counter) (BKG 193, Calibration 112.2)

and

Area Contamination Testing and Surveys

(using a Bicon Geiger Counter with a BKG of 0.01 mR/hr)

All Wipes and Surveys performed by Bernice Buck, RT(N), CNMT

Date	Room	Survey mR/hr	(P/F) Pass/Fail
4/13/09	Hot Lab - Lead Castle	0.01	P
4/13/09	Hot Lab - Lead Shield	0.01	P
4/13/09	Hot Lab - Counter Top	0.01	P
4/13/09	Hot Cab - Radioactive Waste Boxes	0.01	P
4/13/09	Stress Lab	0.01	P
4/13/09	Stress Lab - Floor	0.01	P
4/13/09	Treadmill & EKG Machine	0.01	P
4/13/09	Stress Table	0.01	P
4/13/09	Stress Lab-Dressing Room 1	0.01	P
4/13/09	Stress Lab-Dressing Room 2	0.01	P
4/13/09	Stress Lab Prep Room & Surrounding Floors	0.01	P
4/13/09	Patient Bathroom	0.01	P
4/13/09	Employee Bathroom	0.01	P
4/13/09	Exam Room 1	0.01	P
4/13/09	Exam Room 2	0.01	P
4/13/09	Exam Room 3	0.01	P
4/13/09	Ultrasound Lab	0.01	P
4/13/09	Office # 1	0.01	P
4/13/09	Office #2	0.01	P
4/13/09	Medical Assistant Work Area	0.01	P
4/13/09	Front Office	0.01	P
4/13/09	Kitchen/Break Room	0.01	P
4/13/09	Patient Waiting Room	0.01	P