Project:	07-3080.04 Schofield Barracks	Work Order:	7907	
LAB ID:	SC&A	Analysis Type:	Gamma Spectrometry	

 Reviewer:
 Stacey Sedano

 Date:
 12/27/2007

Sample	Matrix	Collection Date	Date Received	Preparation Date	Hold Times Met? (Y, N< or N/A)	Analysis Date	Hold Times Met? (Y, N< or N/A)
SB-CZ-SS-2029-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2030-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2031-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2032-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2033-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2034-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2035-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2036-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/30/07	N/A
SB-CZ-SS-2037-BA	SOIL	9/13/07	9/25/07	9/25/07	N/A	10/31/07	N/A

Validation Item	Acceptable	Not Acceptable	Not Applicable
	(YES)	(NO)	(N/A)
Sample Chain of Custody Review		. ,	, ,
Are there printed names and signatures present in the Relinquished By and Received By Blocks?		X (1)	
Does the COC date match the Relinquished By date?	X		
Is the Received By date consistent with sample custody transfer (Relinquished By)?	X		
Have all the samples listed on the Chain of Custody have been analyzed? (Verify this by checking that the Memo and/or case narratives are consistent			
with the COC.)	X		.,
Were the sample(s) preserved appropriately?			Х
Are all the samples included in the analytical report are listed correctly on the Chain of Custody?	Х		
Are the analytes reported consistent with the project requirements? (See Attached Sheet)  Comments:	X		
Sample Receipt Checklist Review			
Did the laboratory complete the Sample Receiving Checklist?	X		
Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).	X X		
Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).	X X		
Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).	X X		
Sample Receipt Checklist Review  Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).  Comments:	X X		
Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).	X X		
Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).	X X		
Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).  Comments:	XXX		
Did the laboratory complete the Sample Receiving Checklist?  Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).  Comments:  Case Narrative/Analytical Report  Does the Case Narrative report submitted by the laboratory indicate any	X X		

Validation Item	Acceptable	Not Acceptable	Not Applicable
Are results that are flagged by laboratory necessary and complete, and are understandable comments provided?	Y	•	
Are the reporting units are correct and consistent? (pCi/g)	X		
Laboratory Quality Control Sample Review  Did the laboratory properly complete all required laboratory quality control			
samples at required frequencies?  LCS - 1 per matrix and one per batch or 1/20 samples whichever is more frequent  Matrix Spike – 1 per matrix and one per batch or 1/20 samples whichever is more frequent  Duplicates - 1 per matrix and one per batch or 1/20 samples whichever is more frequent  Blanks - 1 per matrix and one per batch or 1/20 samples whichever is more frequent	X (2)		
Are the laboratory quality control sample results acceptable (solids)? LCS - 30% -69% estimated (J); >130% estimated (J); <30% unusable (R) Matrix Spike – 20% -70% estimated (J); >130% estimated (J); <20% unusable (R) Duplicates - Duplicates – Normalized Absolute Difference (NAD)>1.96 estimated (J)	X(2,3,4)		
Comments:			
<ol> <li>Matrix Spikes are not necessary for Gamma Spec.</li> <li>LCS and NAD results are all acceptable.</li> <li>Potential Blank contamination for samples with Bi-214 activity less the blank contamination.</li> </ol>	an 0.30 pCi/g. I	No samples wer	e flagged for

Validation Item	Acceptable	Not Acceptable	Not Applicable
Other Evaluation Factors		-	
If a result has an uncertainty greater than the result, is the uncertainty is less than the required detection limit?	Х		
Are the sample hold times acceptable? (Six months or less for all Rad except <sup>3</sup> H, which is three months or less )			Х
Are total propagated uncertainty (TPU) values provided for all results?	X		
Are the aliquot sizes appropriate – (1 g minimum for dry solids)	X		
Are soil sample results reported on a dry-weight basis? (See Case Narrative)	X		
Were required detection limits achieved? (see "QA/QC Plan" Table 2.2)	X(6)		

were required detection limits achieved? (see "QA/QC Plan" Table 2.2)	λ(6)	
Comments:		
Commonto.		
B : 11 11 12 10 E 01/		
5. Detection limit of 0.5 pCi/g was not met for samples for Th-	·234 listed below:	
SB-CZ-SS-2034-BA		
OD-OZ-30-2034-DA		