Project:	07-3080.04	Work Order:	7924

LAB ID: SC&A Analysis Type: Gamma Spec

Reviewer: Bachir Badaoui

Date: 12/14/2007

Sample	Matrix	Collection Date	Date Received	Preparation Date	Hold Times Met? (Y, N< or N/A)	Analysis Date	Hold Times Met? (Y, N< or N/A)
SB-CZ-SS-2169-SD	Soil	9/24/2007	9/27/2007	10/3/2007	N/A	11/6/2007	N/A
SB-CZ-SS-2170-SD	Soil	9/24/2007	9/27/2007	10/3/2007	N/A	11/6/2007	N/A
SB-CZ-SS-2171-SD	Soil	9/24/2007	9/27/2007	10/3/2007	N/A	11/6/2007	N/A
SB-CZ-SS-2172-SD	Soil	9/24/2007	9/27/2007	10/3/2007	N/A	11/6/2007	N/A

Validation Item	Acceptable (YES)	Not Acceptable (NO)	Not Applicable (N/A)
Sample Chain of Custody Review			
Are there printed names and signatures present in the Relinquished By and Received By Blocks?		X(1)	
Does the COC date match the Relinquished By date?	Х		
Is the Received By date consistent with sample custody transfer (Relinquished By)?	Х		
Have all the samples listed on the Chain of Custody have been analyzed? (Verify this by checking that the Memo and/or case narratives are consistent with the COC Gamma Isotopes	Х		
Were the sample(s) preserved appropriately?			X
Are all the samples included in the analytical report are listed correctly on the Chain of Custody?	Х		
Are the analytes reported consistent with the project requirements? (See Attached Sheet)	Х		
Sample Receint Checklist Review			
Sample Receipt Checklist Review Did the laboratory complete the Sample Receiving Checklist?	T v		
Did the laboratory complete the Sample Receiving Checklist? Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?).	X X		
Did the laboratory complete the Sample Receiving Checklist?	XXX		
Did the laboratory complete the Sample Receiving Checklist? Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?). Comments: Case Narrative/Analytical Report	X X		
Did the laboratory complete the Sample Receiving Checklist? Are all receipt inspection items marked "Yes"? (If "No" are they acceptable?). Comments:	X		

Validation Item	Acceptable	Not Acceptable	Not Applicable
Are results that are flagged by laboratory necessary and complete, and are	Х	71000 ptuble	Тършешы
understandable comments provided? Are the reporting units are correct and consistent? (pCi/g)	X		
Comments:			
Laboratory Quality Control Sample Review	1		
Did the laboratory properly complete all required laboratory quality control samples at required frequencies? LCS - 1 per matrix and one per batch or 1/20 samples whichever is more frequent Matrix Spike – 1 per matrix and one per batch or 1/20 samples whichever is more frequent Duplicates - 1 per matrix and one per batch or 1/20 samples whichever	X(2)		
is more frequent Blanks - 1 per matrix and one per batch or 1/20 samples whichever is more requent			
Are the laboratory quality control sample results acceptable (solids)? LCS - 30% -69% estimated (J); >130% estimated (J); <30% unusable (R) Matrix Spike – 20% -70% estimated (J); >130% estimated (J); <20% unusable (R) Duplicates - Duplicates – Normalized Absolute Difference (NAD)>1.96 estimated (J)	X(3-4)		
Comments:			
 Matrix Spikes are not necessary for Gamma Spec. LCS and NAD results are all acceptable. The gamma LCS, duplicate analyses, and tracer recoveries indicate be qualified despite the lack of a matrix spike. 	acceptable data	quality, so the	data will not

Validation Item	Acceptable	Not Acceptable	Not Applicable
Other Evaluation Factors			
If a result has an uncertainty greater than the result, is the uncertainty is less than the required detection limit?	Х		
Are the sample hold times acceptable? (Six months or less for all Rad except ³ H, which is three months or less)			х
Are total propagated uncertainty (TPU) values provided for all results?	X		
Are the aliquot sizes appropriate – (1 g minimum for dry solids)	X		
Are soil sample results reported on a dry-weight basis? (See Case Narrative)	X		
Were required detection limits achieved? (see "QA/QC Plan" Table 2.2)	X		