



COLUMBUS REGIONAL HOSPITAL

FAX COVER LETTER

TO: Patricia Pelke **PHONE:** _____
_____ **FAX:** 630-515-1078

FROM: Kent Johnson **PHONE:** 812-376-5998
Columbus Regional Hospital **FAX:** 812-376-5988

DATE: 4/28/09

TOTAL NUMBER OF PAGES: 4 (Including Cover Letter)

CONFIDENTIALITY NOTICE

The documents accompanying this FAX transmission contain confidential information and are the property of the sender. The information contained in the documents is privileged and is intended only for the use of the individual(s) or entity(ies) whose name appears above. If you are not the intended recipient, be advised that any unauthorized disclosure, copying, distribution, or the taking of any action in reliance on the contents of this telecopied information is strictly prohibited. If you have received this FAX transmission in error, please notify us immediately by telephone at the number listed below to arrange for the return of the forwarded documents to us. Thank you.

If you do not receive all the pages clearly, please call. Thank you.

COMMENTS:

I spoke with Loren Hueter last week and he advised that I fax this info to you
for assignment to one of the NRC reviewers. Colleen Casey has taken care
of our amendments recently so she is familiar with our situation. We are trying
to release a mobile Nuclear Medicine coach back to the vendor and are asking

Your assistance on the best way to expedite our amendment request. Thank You!

Kent Johnson, Director of Radiology
Columbus Regional Hospital
Columbus, INdiana

2400 EAST 17TH STREET
COLUMBUS, INDIANA 47201
TELEPHONE 812-379-4441



COLUMBUS REGIONAL HOSPITAL

April 28, 2009

Colleen Casey
U.S. NRC Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352

RE: Columbus Regional Hospital (NRC Lic. #13-01631-05) Amendment request to decommission our "fixed" mobile coach and release it for unrestricted use.

Dear Ms. Casey:

As per our letter to you on Nov. 21, 2008, Columbus Regional Hospital (NRC Lic. #13-01631-05) has included decommissioning surveys of our "fixed" mobile Nuclear Medicine coach (performed after last dose and patient on April 27), which indicate no residual radioactive material attributable to licensed activities. All Nuclear material from the coach has been transferred to our authorized "new" Nuclear Medicine Department. At this time we request that this coach be deleted from our license as an area of use. We will await USNRC authorization in order to release this area for public use.

Should you have any questions or comments feel free to contact our consultant, Mr. Scott Houchin at 216-663-7000.

Sincerely,

Kent Johnson
Director of Radiology

Enclosures

April 28, 2009

Colleen Casey
U.S. NRC Region III
2443 Warrenville Road
Suite 210
Lisle, Illinois 60532-4352

RE: Columbus Regional Hospital (NRC Lic. #13-01631-05) Amendment request to decommission our "fixed" mobile coach and release it for unrestricted use.

Dear Ms. Casey:

As per our letter to you on Nov. 21, 2008, Columbus Regional Hospital (NRC Lic. #13-01631-05) has included decommissioning surveys of our "fixed" mobile Nuclear Medicine coach (performed after last dose and patient on April 27), which indicate no residual radioactive material attributable to licensed activities. All Nuclear material from the coach has been transferred to our authorized "new" Nuclear Medicine Department. At this time we request that this coach be deleted from our license as an area of use. We will await USNRC authorization in order to release this area for public use.

Should you have any questions or comments feel free to contact our consultant, Mr. Scott Houchin at 216-663-7000.

Sincerely,

Kent Johnson
Director of Radiology

Enclosures

DAILY AREA SURVEY
COLUMBUS REGIONAL HOSPITAL NUCLEAR TRAILER

MONTH APRIL YEAR 2009

	HOT LAB (mR/hr)	DESK (mR/hr)	INJECTION AREA (mR/hr)	CAMERA (mR/hr)	INITIALS	Meter
1	.02	.02	.02	.02	SLZ	520F
2	.03	.02	.02	.02	SLZ	14C
3	.02	.02	.02	.02	SLZ	14C
4	Closed					
5	Closed					
6	.02	.02	.02	.02	SLZ	14C
7	.02	.02	.02	.02	SLZ	14C
8	.02	.02	.02	.02	SLZ	520F
9	.02	.02	.02	.02	SLZ	14C
10	.02	.02	.02	.02	SLZ	14C
11	Closed					
12	Closed					
13	.02	.02	.02	.02	SLZ	14C
14	0.02	0.02	0.02	0.02	SLZ	14C
15	.02	.02	.02	.02	SLZ	520F
16	.02	.02	.02	.02	SLZ	14C
17	.02	.02	.02	.02	SLZ	14C
18	Closed					
19	Closed					
20	.02	.02	.02	.02	SLZ	14C
21	.02	.02	.02	.02	SLZ	14C
22	.02	.02	.02	.02	SLZ	520F
23	.02	.02	.02	.02	SLZ	14C
24	.02	.02	.02	.02	SLZ	14C
25	.02	.02	.02	.02	SLZ	14C
26	Closed					
27	.02	.02	.02	.02	SLZ	14C
28	.02	.02	.02	.02	SLZ	14C
29						
30						
31						

Post closing Survey = Background
4-28-09
Last day 4-27-09

RADIATION SAFETY OFFICER *[Signature]*

Meter 1: Ludlum 14C s/n 144158 Meter 2: Eberline 520 S/N 2834

WEEKLY WIPES
COLUMBUS REGIONAL HOSPITAL NUCLEAR TRAILER

MONTH APRIL

YEAR 2009

4-27-09

DATE	BKG (CPM)	HOT LAB (CPM)	DESK (CPM)	INJECTION AREA (CPM)	CAMERA (CPM)	NET CPM FOR ALL)	INITIALS	NET DPM for ALL
4/8/09	504	500	500	500	500	(0)	SKZ	0
4/10/09	493	507	507	507	507	(15)	SKZ	15
4/17/09	443	472	472	472	472	472 (29)	SKZ	31
4/23/09	490	503	503	503	503	503 (13)	SKZ	14
4/24/09	517	517	517	517	517	517 (0)	SKZ	0
4/27/09	492	492	492	492	492	492 (0)	SKZ	0

ACTION LEVELS:

RESTRICTED: 2000dpm/100cm²

UNRESTRICTED: 2000dpm/100cm²

Rad Safety Officer WJ Parrott Ph.D.

4-27-09 (After last dose delivered and last patient done in "coach")

Instrument: Ludlum 1000 scaler
S/N 154260

91.3% efficiency (Co57) on 10/17/08

DPM = CPM x 1.095