



**PPG Industries, Inc.**

P. O. Box 191 New Martinsville, WV 26155 USA 304-455-2200

James D. Rock  
Works Manager  
Natrium Plant  
Chemicals Group

*Br3*

May 15, 2009

**Certified Mail Return Receipt Requested**

U.S. Nuclear Regulatory Commission  
Division of Nuclear Materials Safety  
Region I  
475 Allendale Road  
King of Prussia, Pennsylvania 19406

RECEIVED  
REGION I  
2009 MAY 18 PM 12:28

Attention: Licensing Assistance Team

*03006658*

**Subject: Amendments to License No. 47-01542-01, PPG Industries, Inc., Natrium Facility**

Please find attached amendments requested to U.S. Nuclear Regulatory Commission Materials License No. 47-01542-01 for the PPG Industries, Inc., Natrium Facility. The amendments are due to a change of personnel. Erika A. Baldauff has replaced Leisa C. Stiller as the Radiation Safety Officer. Additionally, attached is a copy of the radiation training certificate for Erika A. Baldauff.

If you have any questions regarding this information, please call Erika Baldauff at (304) 455-2200, extension 3573.

Sincerely,

James D. Rock  
Works Manager

Attachments:

*143740*

NMSS/RGN1 MATERIALS-002

May 15, 2009



**PPG Industries, Inc., Natrium Facility**

**AMENDMENTS TO U.S. NUCLEAR REGULATORY COMMISSION  
MATERIALS LICENSE NO. 47-01542-01**

**Section 11.**

Licensed material specified in items 6A and 6B shall be used by, or under the supervision of, R.J. Feldmeier, Theresa A. Davies-Newman or ~~Liesa C. Stiller~~ Erika A. Baldauff. Licensed material specified in items 6C and 6 D shall be used by or under the supervision of James E. Martin, PH.D., CHP, Chul Lee, M.S., Thomas E. Pruss, or ~~Liesa C. Stiller~~ Erika A. Baldauff.

**Section 12.**

The Radiation Safety Officer for this license is ~~Liesa C. Stiller~~ Erika A. Baldauff.

**Section 15.A.**

Installation, initial radiation survey, relocation, removal from service of devices containing sealed sources shall be performed by only R.J. Feldmeier, Theresa A. Davies-Newman or ~~Liesa C. Stiller~~ Erika A. Baldauff in accordance with the manufacturer's instructions, or by persons specifically licensed by the Commission or an Agreement State to perform such services.

**Section 16.**

Prior to initial use and immediately after installation, relocation, dismantling, alignment, or any other activity involving the source or removal of the shielding, the licensee shall assure that a radiological survey is performed to determine radiation levels in accessible areas around, above, and below the gauge with the shutter open. This survey shall be performed only by ~~Liesa C. Stiller~~ Erika A. Baldauff or by persons authorized to perform such services by the Commission or an Agreement State. A record of the results of this survey shall be maintained for the duration of the license.

OHMART/VEGA

# Erika Ann Baldauff

PPG Industries, Inc. New Martinsville, WV

has successfully completed the

## Ohmart/VEGA Radiation Safety Course

presented


February 23, 2009 - February 27, 2009

at

Ohmart/VEGA

### Subject matter covered:

- Duties of the RSO
- Basic atomic theory
- Measurement and monitoring techniques
- Exposure calculations
- Biological effects of radiation
- Rules and Regulations
- Leak test, shutter check
- Installation, relocation, and removal procedures
- Hands on lab work
- Proper disposal practices
- Emergency procedures
- DOT shipping



Kyle Harness  
Radiation Safety Officer



Candy Brock  
Training Manager

RADAR

GUIDED RADAR

ULTRASONIC

NUCLEAR

CAPACITANCE

VIBRATION

PRESSURE

This is to acknowledge the receipt of your letter/application dated

5/15/2009, and to inform you that the initial processing which includes an administrative review has been performed.

AMEND, 47-0154200  
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 143740.  
When calling to inquire about this action, please refer to this control number.  
You may call us on (610) 337-5398, or 337-5260.

NRC FORM 532 (RI)  
(6-96)

Sincerely,  
Licensing Assistance Team Leader