



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

MAY 20 2009

James C. Clouse, D.O.
Radiation Safety Officer
Golden Valley Memorial Hospital
1600 North Second Street
Clinton, MO 64735

Dear Dr. Clouse:

This concerns your letter dated March 12, 2009, informing the Commission of your medical use of naturally occurring and accelerator-produced radioactive material (NARM).

Since the medical use of unsealed NARM is now covered under 10 CFR 35.200 and the medical use of sealed NARM calibration/reference sources is authorized under 10 CFR 35.65, you are currently in compliance with NRC requirements and an amendment to your NRC license is not required.

If you have any questions concerning this letter, please contact me at (630) 829-9842.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

Sincerely,

A handwritten signature in cursive script, appearing to read "Toye L. Simmons".

Toye L. Simmons
Materials Licensing Branch

License No. 24-16597-01
Docket No. 030-11317