



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

MAY 19 2009

Edward Wroblewski, M.A.
Radiation Safety Officer
St. Vincent Hospital & Health Care Center
2001 W. 86th Street
Indianapolis, IN 46240

Dear Mr. Wroblewski:

This concerns your letter dated February 27, 2009, informing the Commission of your medical use of naturally occurring and accelerator-produced radioactive material (NARM).

Since the medical use of unsealed NARM is now covered under 10 CFR 35.200 and the medical use of sealed NARM calibration/reference sources is authorized under 10 CFR 35.65, you are currently in compliance with NRC requirements, therefore, an amendment to your NRC license is not required.

If you have any questions concerning this letter, please contact me at (630) 829-9842.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

Sincerely,

A handwritten signature in cursive script, appearing to read "Toye L. Simmons".

Toye L. Simmons
Materials Licensing Branch

License No. 13-00133-02
Docket No. 030-01579