

**From:** Donna Marie Grace [DonnaMarie.Grace@RADNET.COM]  
**Sent:** Wednesday, June 03, 2009 5:37 PM  
**To:** Elizabeth Ullrich  
**Subject:** RE: question for Papastavros license

I will be sending a letter shortly in response to your email. Thank you for bringing this to my attention.

Donna Grace

***Donna M. Grace R.T., (R)(M)***  
***Imaging Center Manager***  
***Papastavros' Associates Medical Imaging, L.L.C.***  
***1701 Augustine Cut-off***  
***Wilmington, DE 19803***  
***office: (302) 652-3016 extension 3032***  
***cell: (302) 465-4271***  
***fax: (302) 571-6270***  
***email: [DonnaMarie.Grace@Radnet.com](mailto:DonnaMarie.Grace@Radnet.com)***

---

**From:** Elizabeth Ullrich [mailto:Elizabeth.Ullrich@nrc.gov]  
**Sent:** Tuesday, May 26, 2009 2:47 PM  
**To:** Donna Marie Grace  
**Subject:** question for Papastavros license

Ms. Grace:

Below is the email I sent to Michelle 2 months ago. Please let me know if you receive this email; you will have to respond by regular mail, as stated at the end of the email message below.

Betsy Ullrich

\*\*\*\*\*

**From:** Elizabeth Ullrich  
**Sent:** Monday, March 23, 2009 2:18 PM  
**To:** michelle.nowaczyk@radnet.com  
**Subject:** 2/9/09 letter to amend NRC License 07-16529-01

Papastavros' Associates Medical Imaging, LLC  
License No. 07-16529-01  
Docket No. 030-11379  
Control No. 143444

Subject: Request for additional information to your amendment letter dated February 9, 2009

Ms. Nowaczyk,

The February 9, 2009, letter requests Germanium-68 sources to calibrate the scanner detectors, 500 Mbq per source and 200 MBq total (appr 13.5 millicuries per source and 112 mCi total). Please let me know if you believe that

a - these are calibration sources that meet the requirements of 10 CFR 35.65, and do not need to be specifically listed in the license; OR

b. - they are sealed sources for diagnostic medical uses under 35.500; if so, state the total amount of material to be possessed as Item 6/7/8.D. on your license; OR

c. - they are sealed sources which do NOT meet the 35.65 requirements or the 35.500 requirements, and must be listed specifically on your license; if so, provide the manufacturer name and the model number for the sources to be possessed.

Please note that you may not reply to this request by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269). If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

thanks,

Betsy

Betsy Ullrich, MS, CHP

Senior Health Physicist

Commercial and R&D Branch

Division of Nuclear Material Safety

Region I

Nuclear Regulatory Commission

(610) 337-5040

[Elizabeth.Ullrich@nrc.gov](mailto:Elizabeth.Ullrich@nrc.gov)

---

### E-mail Properties

Mail Envelope Properties (C326929E7DA4314DA339281C8A199AFF078310)

Subject: RE: question for Papastavros license

Sent Date: 6/3/2009 5:37:23 PM

Received Date: 6/3/2009 5:37:23 PM

From: Donna Marie Grace

Created By: DonnaMarie.Grace@RADNET.COM

Recipients:

Elizabeth.Ullrich@nrc.gov (Elizabeth Ullrich)

Tracking Status: None

Post Office:

MDBAL01EX001.radnet.rdlx.pvt

Files	Size	Date & Time
MESSAGE	21642	6/3/2009

Options

Expiration Date:

Priority: o!ImportanceNormal

ReplyRequested: False  
Return Notification: False

Sensitivity: olNormal  
Recipients received: