

WIN

TO: JENNY JOHANSEN

610-337-5393

P-2

FROM: ABC TECH INC

PAGE: 1 OF 2

29-31056-01

03036957

143749

REF: ABC TECH INC



MICROTEC SERVICES, INC.

110 Charles Street ♦ Pasadena, TX 77506
Tel (713) 475-2274 Fax (713) 475-2362

Leak Testing
Calibrations
Repair
New Instruments

***** **Certificate of Leak Test** *****

Customer: ATLANTIC ENGINEERING LABORATORIES, INC.
RAHWAY, NJ

This is to certify that the leak test on the indicated source(s) shown below has been analyzed on the specified date and the results shown accurately represent the level of removable contamination.

***** **Source Information** *****

Serial Number: 47-12481
Isotope: Cs-137/Am-241 Be
Activity: 8.0/40 mCi.
Remarks: GAUGE 17057
Manufacturer / Model: TROXLER 3411-B
Date of Test: 22-Apr-09

***** **Leak Test Analysis Data** *****

Gross Counts: 209
Background: 209
Counting Efficiency: 52.8%
Contamination (DPM): 0
Removable Contamination: Less Than .0001 Microcuries
Less Than 5 Bq.

Date Analyzed: 28-Apr-09
Analyzed by: Q. Stokley

Comments:

(NOTE: Removable contamination equal to or greater than .005 microcuries (11,000 dpm) constitutes a leaking source. Results shown above are in accordance with 10 CFR and Texas TRCR regulations for sealed sources and plated sources.)