

Johnson & Johnson
PHARMACEUTICAL RESEARCH
& DEVELOPMENT, L.L.C.

1000 U.S. Highway 202, P.O. Box 300
Raritan NJ 08869

MS6
J-6

June 16, 2009

RECEIVED
JUN 18 PM 12:03

Dennis Lawyer
Licensing Assistance Team
Division of Nuclear Materials Safety
U.S. Nuclear Regulatory Commission, Region I
475 Allendale Road
King of Prussia, PA 19406-1415

SUBJECT: JOHNSON & JOHNSON PHARMACEUTICAL RESEARCH & DEVELOPMENT, L.L.C., REQUEST FOR ADDITIONAL INFORMATION CONCERNING APPLICATION FOR A LICENSE AMENDMENT TO LICENSE NO. 29-02608-03, CONTROL NO. 143601.

03010814

Dear Dennis Lawyer:

This letter is in response to your email dated June 11, 2009, requesting additional information pertaining to the Johnson & Johnson Pharmaceutical Research & Development, L.L.C. (JJPRD) NRC Material License (29-02608-03) amendment request submitted on May 29, 2009.

Please find below the original questions and the additional information requested:

1. In the Report, "Final Status Survey for Johnson & Johnson Pharmaceutical Research & Development" in Enclosure 4, the Static Point Measurement results did not appear to specify the measurement units. This was the case for each static point measurement information listed for Room 011, 013/014, 015 and 016. Please specify the unit of measurement for these readings.

The units of measurement were counts per minute (CPM).

2. In the Report, "Final Status Survey for Johnson & Johnson Pharmaceutical Research and Development" in Enclosure 3, Table 8 did not include the information for Ludlum 2221 which appears to have been used for static measurements in each of the rooms. Please provide the information for the Ludlum 2221 in Table 8 or state that the Ludlum 2221 was not used for measurements.

143601

NMSS/RGN1 MATERIALS-002

The Ludlum 2221 meter was not used for static, only scans. This is also specified in each of the rooms' meter information and results sheet.

If you have any questions or concerns, please feel free to contact me at (908) 704-4930.

Sincerely,

A handwritten signature in black ink, appearing to read "Jim Kwiatkowski". The signature is fluid and cursive, with the first name "Jim" written in a larger, more prominent script than the last name "Kwiatkowski".

James Kwiatkowski
Manager – Environmental, Health & Safety
Johnson & Johnson Pharmaceutical
Research & Development, L.L.C.