



# PAPASTAVROS' ASSOCIATES

MEDICAL IMAGING L.L.C.

---

June 19, 2009

Q-5  
MS-16

Ms. Betsy Ullrich, MS, CHP  
Senior Health Physicist  
Commercial and R&D Branch  
Division of Nuclear Material Safety  
Region I  
Nuclear Regulatory Commission  
475 Allendale Road  
King of Prussia, PA 19406

2009 JUN 26 AM 10:48

RECEIVED  
REGION I

Re: Papastavros' Associates Medical Imaging, LLC  
License No. 07-16529-01  
Docket No. 030-11379  
Control No. 143444

Dear Ms. Ullrich,

We provide the following information in response to your letter of 02/09/09. The Ge-68 sources used in the GE Discovery LS PET-CT unit, which is planned for installation at our Neuro Science site, qualify as calibration, transmission, and reference sources as described in 10 CFR part 35.65 and do not need to be specifically listed in the license.

We look forward to receiving the requested amendment to our radioactive materials license. Please contact me if I can provide additional information.

Sincerely,

Garth A. Koniver, M.D.  
Radiation Safety Officer

143444