

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: <i>Nuclear Diagnosis, Inc MC. Vernon, MO / Co. Charge, MO / 3rd Street, MO</i>		2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region III 2443 Warrenville Road Suite 210 Lisle, Illinois 60532-4351	
3. DOCKET NUMBER(S) <i>030-12966</i>		4. LICENSEE NUMBER(S) <i>24-17561-01</i>	
REPORT NUMBER(S) <i>2009-001</i>		5. DATE(S) OF INSPECTION <i>6/9/09</i>	

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

_____ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.
(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title		Signature		Date
LICENSEE'S REPRESENTATIVE	Printed Name			
NRC INSPECTOR	<i>Michael M LaFrance</i>	<i>[Signature]</i>		<i>6/18/09</i>

[Handwritten initials]

SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION

1. LICENSEE Nuclear Diagnosis, Inc. REPORT NUMBER(S) 2009-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road, Suite 210 Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-12966	4. LICENSE NUMBER(S) 24-17561-01	5. DATE(S) OF INSPECTION 6/9/09	
6. INSPECTION PROCEDURES USED 87130		7. INSPECTION FOCUS AREAS 03.01-03.07	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 2200	2. PRIORITY 3	3. LICENSEE CONTACT Lynn Carlton – RSO	4. TELEPHONE NUMBER 417-890-0020
----------------------------	------------------	---	-------------------------------------

<input type="checkbox"/>	Main Office Inspection	Next Inspection Date: June 2012
<input checked="" type="checkbox"/>	Field Office	Mt. Vernon, MO; Carthage, MO; and Branson, MO
<input type="checkbox"/>	Temporary Job Site Inspection	

PROGRAM SCOPE

The licensee is authorized to use licensed material permitted by 10 CFR 35.100, 200 and 300 at 11 sites in southwest Missouri area. However, the only site authorized for 35.300 activities is at the Branson facility and they had not performed those activities since the last inspection. The three sites visited (Skaggs Community Health Center; Branson, MO – McCune-Brooks Regional Hospital; Carthage, MO – and Missouri Rehabilitation Center; Mt. Vernon, MO) each performed 2-10 diagnostic studies per day, depending on patient availability. The inspector identified that the McCune-Brooks Hospital at 627 W Centennial; Carthage, MO, had not been used by the licensee in the last 12 months. The inspector informed the licensee they would be required to implement decommissioning activities at the site within 12 months to comply with NRC regulations.

Observations and Findings

The inspector observed the licensee administer licensed material and the technician had all radiation safety equipment available and implemented proper radiation safety practices. The inspector interviewed licensee staff and determined that they were aware of radiation safety practices and were implementing the radiation safety program adequately. The inspector reviewed selected documents which included: dosimetry records, radiation safety committee meeting minutes, dose administration records, dose calibrator records and radiological surveys; no abnormal issues were identified. The inspector performed independent radiological surveys and did not identify an abnormal radiation or contamination levels.