



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, ILLINOIS 60532-4352

JUN 25 2009

Bruce Canfield, C.N.M.T.
Radiation Safety Officer
Dearborn County Hospital
600 Wilson Creek Road
Lawrenceburg, IN 47025

Dear Mr. Canfield:

Enclosed is Amendment No. 12 amending your NRC Material License No. 13-17327-01 in accordance with your request. Please note that the changes made to your license are printed in **bold** font.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

Please note, based upon the close-out survey results in the letter dated April 1, 2009, and facsimile received June 18, 2009, we have authorized the release of the "old hot" laboratory at 600 Wilson Creek, Road, Lawrenceburg, Indiana, for unrestricted use.

Since the medical use of NARM (naturally-occurring and accelerator-produced radioactive materials) such as rubidium-82 generators, is now covered under 10 CFR 35.200 and the medical use of sealed NARM calibration/reference sources is authorized under 10 CFR 35.65, you are currently in compliance with NRC requirements and an amendment to your NRC license is not required.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter and its enclosure will be available electronically for public inspection in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html> (the Public Electronic Reading Room).

Sincerely,

A handwritten signature in black ink, appearing to read "William P. Reichhold".

William P. Reichhold
Materials Licensing Branch

License No. 13-17327-01
Docket No. 030-12564

Enclosure: Amendment No. 12